

National Bureau of Standards **Aprificate**

Standard Reference Material 4250B

Radioactivity Standard

Radionuclide

Cesium-134

Source identification

4250B-

Source description

Liquid in NBS borosilicate-glass ampoule $(1)^*$

Solution composition

Approximately 63 micrograms of cesium chloride per gram, of 0.1

molar hydrochloric acid (2)

Mass

grams

Radioactivity concentration

 $1.742 \times 10^6 \text{ Bg g}^{-1}$

Reference time

1200 EST March 23, 1982

Random uncertainty

0.02 percent (3)

Systematic uncertainty

1.14 percent (4)

Total uncertainty (Random plus systematic) 1.16 percent

Photon-emitting impurities

None observed (5)

Half life

 $2.065 \pm 0.001 \text{ years}$ (6)

Measuring instrument

NBS pressurized " 4π " γ ionization chamber calibrated by liquid scintillation coincidence

efficiency-extrapolation technique

This Standard Reference Material was prepared in the Center for Radiation Research, Nuclear Radiation Division, Radioactivity Group, Dale D. Hoppes, Group Leader.

Washington, D.C. 20234 April, 1982

George A. Uriano, Chief Office of Standard Reference Materials

FOOTNOTES

(1) Approximately five milliliters of solution. Ampoule specifications:

| body diameter | $16.5 \pm 0.5 \text{ mm}$ |
|----------------------|----------------------------|
| wall thickness | $0.60 \pm 0.04 \text{ mm}$ |
| barium content | less than 2.5 percent |
| lead oxide content | less than 0.02 percent |
| other heavy elements | trace quantities |

- (2) Solution density 1.000 \pm 0.002 g/ml at 22.6°C.
- (3) Half the 99-percent confidence interval of the mean (2.7 times the standard error computed from 42 ionization-chamber measurements).
- (4) Estimated uncertainty limits due to calibration of the pressurized " 4π " ionization chamber, which is the linear sum of estimated uncertainty limits due to:

| a) | half the 99-percent confidence interval of the | | |
|----|--|------|---------|
| , | mean of five coincidence measurements | 0.06 | percent |
| b) | dead time | 0.09 | percent |
| c) | resolving time | 0.15 | percent |
| d) | background | 0.05 | percent |
| e) | gravimetric measurements | 0.10 | percent |
| f) | half the 99-percent confidence interval of the mean of the ionization-chamber measurements | 0.11 | percent |
| g) | radium-226 reference sources ratios | 0.31 | percent |
| h) | efficiency extrapolation | 0.06 | percent |
| i) | modification of curve-fitting formula | 0.20 | percent |
| j) | spurious pulses | 0.01 | percent |

- (5) Limits of detection as a percentage of the gamma-ray-emission rate of the 605-keV gamma rays emitted in the decay of cesium-134 are
 - 0.1 percent between 90 and 790 keV
 - 0.01 percent between 807 and 1900 keV

provided that the impurity photons are separated in energy by five keV or more from photons emitted in the decay of cesium-134.

(6) NBS-measured half-life value. NBS Special Publication No. 626, January 1982, p. 93.