

National Bureau of Standards Certificate

Standard Reference Material 4260

Photon-Emission-Rate Standard Sample

Iron-55

This Standard Reference Material consists of iron-55 quantitatively deposited as the chloride, on the lacquered surface of a 0.01-cm thick, stainless steel disk, 1.7 cm in diameter. The deposit is covered with another layer of lacquer approximately 0.002 g/cm² thick, and the stainless steel disk is cemented to a monel disk 2.54 cm in diameter.

Using a detector located in a direction normal to the plane of the source mount, the number of manganese K-x-rays emitted per second per steradian, on March 25, 1974, was

* $\pm 3.0\%*$.

This Standard Reference Material was standardized by comparison of its emission rate with that of a working standard which had been calibrated by means of high-pressure proportional counting.

The uncertainty in the emission rate, 3.0 percent, is the linear sum of 1.5 percent, which is the limit of the random error associated with the emission-rate comparison at the 99-percent confidence limit ($2.9 S_m$, where S_m is the standard error computed from 18 measurements), and the total error, 1.5 percent, in the working standard.

An error of the order of 0.1 percent can be introduced if the detector is located at an angle 20 degrees away from the normal to the source mount, due to absorption by the lacquer.

The material from which this Standard Reference Material was prepared was examined for photon-emitting impurities using a pure germanium detector and none was found.

A half life obtained from the literature of 2.7 years^[1] is recommended.


This standard was prepared and calibrated in the Center for Radiation Research, Radioactivity Section, W. B. Mann, Chief.

Washington, D.C. 20234
August, 1974

J. Paul Cali, Chief
Office of Standard Reference Materials

Information on the Use of SRM 4260
Iron-55 Point Sources

Decay Scheme



Energies and Abundance [2,3]

Transition, Energy (keV)	Abundance
$K_{\alpha 2,}$ 5.888	0.0762
$K_{\alpha 1,}$ 5.899	0.1506
$K_{\beta 1,3,}$ 6.490	0.0303
$K_{\beta 5,}$ 6.535	--

Additional Information

The calibration of this standard is given in terms of N_0 , the number of manganese K-x-rays per second per steradian emitted in the direction perpendicular to the surface of the source. For most counting geometries, the detector efficiency, ϵ , can be calculated from the observed count rate, N , by means of the following formula:

$$\epsilon = \frac{N}{4\pi N_0}$$

It is recommended that the source be placed so that the outside edges of the detector subtend angles, relative to the perpendicular from the surface of the source, of less than 20° . This precaution will insure that errors, caused by increased absorption of K-x-rays emitted away from the perpendicular, by the film of lacquer, are less than 0.1%.

For detectors of larger solid angle, such as 2π , corrections can be calculated using the total scatter cross section for the lacquer, $10 \text{ cm}^2/\text{g}$.

[1] M. J. Martin and P. H. Blichert-Toft, Nuclear Data Tables, 8, Numbers 1-2, Section A, October 1970.

[2,3] Calculated from data in the following references:

2. P. V. Rao, M. H. Chess, and B. Crasemann, Phys. Rev., A5 997 (1972).
3. G. C. Nelson, B. G. Saunders and S. I. Salem, Atomic Data, 1, No. 4, November, 1970.