National Bureau of Standards Ernest Ambler, Director

National Bureau of Standards

Certificate

Standard Reference Material 4416L-F

Radioactivity Standard

Radionuclide

Ga11ium-67

Source identification

4416L-F-

Source description

Liquid in NBS borosilicate-

glass ampoule (1)*

Solution composition

Approximately 110 micrograms of GaCl₃ per gram of 2 molar hydro-chloric acid (2)

Mass

Radioactivity concentration

 $3.922 \times 10^6 \text{ Bq g}^{-1}$

Reference time

1500 EST April 23, 1985

Overall uncertainty

0.80 percent (3)

Photon-emitting impurities

None observed (4)

Half life

 $3.261 \pm 0.001 \text{ days}$ (5)

Measuring instrument

NBS pressurized "4π"γ ionization chamber calibrated by $4\pi(e,x)-\gamma$ coincidence efficiency-extrapo-

lation technique

This Standard Reference Material was prepared in the Center for Radiation Research, Nuclear Radiation Division, Radioactivity Group, Dale D. Hoppes, Group Leader.

Gaithersburg, MD 20899 May 1985

Stanley D. Rasberry, Chief Office of Standard Reference Materials

*Notes on back

NOTES

(1) Approximately five milliliters of solution. Ampoule specifications:

body diameter $16.5 \pm 0.5 \text{ mm}$ wall thickness $0.60 \pm 0.04 \text{ mm}$ barium content less than 2.5 percent lead oxide content other heavy elements trace quantities

- (2) Solution density 1.032 ± 0.002 g/mL at 22.6 °C.
- (3) The overall uncertainty was formed by taking three times the quadratic combination of standard deviations of the mean, or approximations thereof, for the following:

a.)	18 coincidence measurements	0.08 percent
ъ)	18 ionization-chamber measurements	0.01 percent
c)	dead time	0.08 percent
d)	resolving time	0.10 percent
e)	backgrounds	0.11 percent
f)	gravimetric measurements	0.05 percent
g)	original ionization-chamber measurements	0.02 percent
h)	efficiency extrapolation	0.10 percent
i)	half life	0.10 percent
j)	solution composition and density	0.05 percent
k)	radium 100 to radium 20 reference	
	sources ratio	0.03 percent
1)	radium reference source positioning	0.10 percent

- (4) Limits of detection as a percentage of the gamma-ray-emission rate of the 93-keV gamma rays emitted in the decay of gallium-67 are
 - 0.1 percent between 90 and 388 keV
 - 0.01 percent between 398 and 1900 keV,

provided that the impurity photons are separated in energy by five keV or more from photons emitted in the decay of gallium-67.

(5) NCRP Report No. 58, 2nd edition (1985), p. 390.