U. S. Department of Commerce Frederick B. Dent Secretary

mal Bureau of Standards

National Bureau of Standards

Certificate

Standard Reference Materials

Magnetic Gram Susceptibility

763-Aluminum 764-Platinum 765-Palladium 766-Manganese Fluoride

These Standard Reference Materials are issued primarily for the calibration of instruments used to measure magnetic susceptibility.

Magnetic gram susceptibility as a function of temperature a/

Temperature	763 (Al)	764 (Pt)	765 (Pd)	766 <mark>b</mark> / (MnF ₂)
K		$10^6 \chi_g$, cm ³	3 g ⁻¹	2
77.7	.696			-
295	.605	.993	5.28	124.0
296		.992	5.27	123.7
297	.604	.991	5.26	123.3
298		.991	5.25	123.0
299	.603	.990	5.24	122.6

a/Magnetic susceptibility values are estimated to have an uncertainty less than ± 0.5 percent.

The overall direction and coordination of technical measurements at NBS leading to certification were performed by G. A. Candela and R. E. Mundy.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. E. Michaelis.

Washington, D. C. 20234 April 5, 1973 J. Paul Cali, Chief Office of Standard Reference Materials

b/Susceptibility values have not been corrected for demagnetization or anisotropic effects.

Preparation and Handling:

The materials for these SRM's were all of the highest possible purity obtainable. Each SRM was thoroughly cleaned and etched prior to packing. The etchants used were the following:

SRM	Etchant		
Al	HF/H ₂ 0	1:1	
MnF ₂	HF/H ₂ 0	1:10	
Pd and Pt	HC1/HNO3	1:3, hot	

Finally, the materials were thoroughly rinsed with distilled $\rm H_2O_{\hbox{\scriptsize -}}$

Caution:

These SRM's should be handled only with non-ferrous instruments such as brass or plastic forceps and contact with any ferromagnetic materials must be avoided. Should any cutting or grinding of the samples be required or should any surface contamination be suspected, it is recommended that the samples be etched as described above.

Procedure:

The expression, f = 1/2k grad B^2 , relates the measurable quantities, force-per-unit volume (f) and magnetic field (B), to the field independent isotropic magnetic susceptibility (k). The component of force in a given x direction on a piece of homogeneous matter of volume V is then given by:

$$F_{x} = \frac{(k - k_{M})}{2\mu_{O}} \int_{V} \frac{dB^{2}}{dx} dV$$

1/G. A. Candela and R. E. Mundy, "Absolute magnetic susceptibilities by the Gouy and the Thorpe-Senftle methods," IRE Trans. Instr. I-2, 106 (1962).

 $k_{\mbox{M}}$ is the susceptibility of the medium. In the Gouy method, a cylinder of matter of constant cross section (a), and uniform density is orientated with its axis in the x direction. The force in the x direction is given by:

$$F_{x} = \frac{a(k - k_{M})}{2\mu_{O}} \int_{x_{O}}^{x_{1}} \frac{dB^{2}}{dx} dx = \frac{a(k - k_{M})}{2\mu_{O}} (B_{1}^{2} - B_{0}^{2})$$

where B_1 and B_0 are the fields at x_1 and x_0 , respectively. When $k >> k_M$, the gram susceptibility is then equal to:

$$\chi_{g} = \frac{2 F_{x} \mu_{o}}{M_{s} / \ell (B_{1}^{2} - B_{0}^{2})}$$

where ℓ is the length of the cylindrical sample and ${\rm M}_{\rm S}$ is the mass of the sample.

The samples were all measured in an atmosphere of dry nitrogen gas. B_1 was measured by a proton resonance magnetometer and B_0 was measured by a calibrated Hall effect gaussmeter. The magnetic fields used were $1000 \leq B_1 \leq 7500$ gauss and in all the experiments $B_0^2 < 0.005$ B_1^2 . The sample lengths were measured by direct comparison to gage blocks. The temperature was measured by a calibrated copper constantan thermocouple.

SUPPLEMENTARY INFORMATION

Although not certified, the volume susceptibilities of these SRM's are given at 297 K.

SRM	Type	Volume Susceptibility 106 k
763	Aluminum	1.63 ^a
764	Platinum	21.2
765	Palladium	63.1
766	Manganese Fluori	ide 484

^al.87 at 77.7 K.