

INSTALLING AND PROGRAMMING THE RADIO SYSTEMS ON AIRNET STATIONS

Purpose This Meteorology and Air Quality Group (MAQ) procedure describes the assembly and programming of the radio-frequency systems installed on selected AIRNET stations.

Scope This procedure applies to the individuals assigned to assemble, install, and program the radio-frequency systems on selected AIRNET stations.

In this procedure This procedure addresses the following major topics:

Topic	See Page
General Information About This Procedure	2
Who Requires Training to This Procedure?	2
Installing an RF unit	3
Programming the CR10X and RF units	5
Records Resulting from This Procedure	8

Signatures

Prepared by: _____ Jake Martinez, MAQ	Date: _____
Approved by: _____ Craig Eberhart, Air Monitoring Project Leader	Date: _____
Approved by: _____ Terry Morgan, QA Officer	Date: _____
Work authorized by: _____ Dave Fuehne, Acting MAQ Group Leader	Date: _____

06/16/05

CONTROLLED DOCUMENT

This copy is uncontrolled if no red stamp is present on printed copies. Users are responsible for ensuring they work to the latest approved revision.

General information about this procedure

Attachments This procedure has the following attachments:

Number	Attachment Title	No. of pages
1	Hazard Control Plan	2
2	Example Program for CR10X Unit	2

History of revision

This table lists the revision history and effective dates of this procedure.

Revision	Date	Description Of Changes
0	11/4/99	New document.
1	6/16/04	Added safety information about safe distance from power lines.
2		Quick-change revision to change HCP to HR.

Who requires training to this procedure?

The following personnel require training before implementing this procedure:

- Technicians assigned to install the RF systems
-

Training method

The training method for this procedure is **on-the-job** training by a previously-trained individual and is documented in accordance with the procedure for training (MAQ-024).

Prerequisites

In addition to training to this procedure, the following training is also required prior to performing this procedure:

- MAQ-011, "Logbook Use and Control"
 - Cardiopulmonary Resuscitation (CPR)
-

References

The following documents are referenced in this procedure:

- MAQ-011, "Logbook Use and Control"
- MAQ-024, "Personnel Training"
- Operator's Manual for Campbell Scientific CR10X
- Operator's Manual for Campbell Scientific PS12 power supply
- Operator's Manual for Campbell Scientific RF95 RF Modem
- "Radiotelemetry Network Instruction Manual"

Installing an RF unit

System description

The radio frequency (RF) system used on the AIRNET stations is manufactured by Campbell Scientific, Inc. The base antenna is located on the top of the Cave and is connected to a model RF-95 radio base station unit which is in turn connected to a COM port on the base computer. Stations equipped for RF monitoring have a data collection unit, a radio, and antenna. A station may communicate directly with the base station if the signal strength allows it, or it may communicate via another station that acts as a repeater.

Two stations at TA-49 are connected to the base station via a phone line. These stations use the same internal components except for a DC-112 phone modem instead of a radio unit. Because no new phone-based communications are planned, the installation of a phone system is not described in this procedure but is similar to that for the RF units.

Where RF units are used

The RF transmitters are installed on AIRNET stations that are considered critical, especially the "FFCA" stations that must meet completeness and runtime criteria. Other stations may have the RF systems installed as directed by the Environmental Measurements Project Leader.

Equipment and parts needed

To install a system, collect the following parts and equipment:

- Campbell Scientific weathertight box with mounting hardware
 - CR10X data recorder
 - PS 12 power supply unit with battery
 - RF95 RF modem
 - Johnson Design radio Model DL3420 or DL3422
 - Omni-directional or directional antenna, with mounting hardware
 - Transformer (110V to 18V, 1.11A)
 - Relay for indicating power loss, mounting socket
 - Mounting brackets, screws, and hardware for attaching box to station legs
-

Warning

Be aware of power lines in the vicinity of the installation site. Do not allow any metal object, especially antenna poles, to come within 13 feet of a power line.

Installing an RF unit, continued

Steps to install an RF system To install an RF system in an AIRNET station, perform the following steps:

Step	Action
1	Mount the equipment on the board inside the box.
2	Remove the lid on the RF modem and set the DIP switches to the ID number of the station. The ID number is the same as the AIRNET station number. Refer to the table in the "Radiotelemetry Network Instruction Manual" for the settings.
3	Install the battery into the PS12 box. Plug into INT terminal. Leave the power switch off.
4	Connect the two power leads from the PS12 to the + and – terminals in the CR10X.
5	Hook up transformer (110V to 18V, 1.11A) to charging terminals on PS12. Run cord out bottom of housing to inside of AIRNET station and plug the transformer into the outlet inside the station.
6	Connect the power leads from the radio to the power terminals inside the PS12 box.
7	Install the omni-directional or the unidirectional antenna on the side of the station. Run the antenna cable through the hole in the bottom of the Campbell box and connect the end to the radio unit.
8	Connect the communication cable (tape style) from the CR10X to the modem.
9	Connect the radio communication cable (tape style) from the modem to the RF95 radio.
10	Turn on the power switch on the PS12 and observe the red light on the modem: it should blink twice and go out. This indicates everything is connected properly.
11	Put the lids back on the PS12 unit and the modem.
12	Use the "goop" supplied with the Campbell box to seal the opening in the bottom of the box (around the transformer power cord and the antenna lead).
13	Close up the box and return to the Cave to initiate radio communications and program the unit.

Programming the CR10X and RF units

Programming the station After installing a RF system, it needs to be programmed. This is done with the computer program on the base computer and then transmitted to the station via radio.

Steps to set up primary file To create the primary file for a new AIRNET station, perform the following steps:

Step	Action
1	Run the program "PC208W" on the base computer.
2	Click the Program button, then File, New .
3	Copy another program for another station into this file. See the typical example program in Attachment 2. Ensure the station ID is changed throughout as necessary. Give the file a name that is composed of the station ID.
4	Compile the program.
5	Click the Setup button then select the Hardware tab. Ensure the following options are set: <ul style="list-style-type: none"> • <u>Dialed Using RF95 Path</u>: set to "U" if it's possible to establish direct, high quality, communication with the station from the base. If the communication quality is poor, enter "A." If communication is still not possible, set a repeater by entering "A61 60F" (this sets station 61 as a repeater to station 60).
6	Click the Data collection tab and ensure the following option boxes are checked: <ul style="list-style-type: none"> • Data logged since last call • Append to end of file. • ASCII, comma separated Turn on check box "Collect Final Storage Area 1" Enter "0" in "Arrays to collect on 1 st call"
7	Click on Schedule tab and ensure the following parameters are entered: <ul style="list-style-type: none"> • Calling interval: 01 00:00:00 • Next time to call: set to next available time interval so each station is called sequentially at 4-minute intervals. • Primary retry interval: 02:00 • Tries: 2 Turn on check box "Schedule on"
8	Click on Save Edits button before closing.

Programming the CR10X and RF units, continued

Steps to set up shadow file To create the “shadow file” for a new AIRNET station, perform the following steps:

Step	Action
1	Perform steps 1 – 4 above to create a file, but name it with the station ID and the suffix “_s”.
2	Click the Setup button then select the Hardware tab. Ensure the following options are set: <ul style="list-style-type: none"> • <u>Dialed Using RF95 Path</u>: enter “1” • <u>Security code</u>: enter “0” • <u>Call back ID number</u>: enter the station number
3	Click the Data collection tab and ensure the following option boxes are checked: <ul style="list-style-type: none"> • Most recently logged arrays • Append to end of file. • ASCII, comma separated Turn on check box “Collect Final Storage Area 1” Enter “24” in “Arrays to collect on 1 st call”
4	Click on Schedule tab and ensure the following parameters are entered: <ul style="list-style-type: none"> • After call do: enter “cd_notif” <u>Turn off</u> check box “Schedule on” (All other parameters have no affect.)
5	Click on Save Edits button before closing.

Connect to the new station After creating the files in the steps above, connect to the station by clicking the button **Connect**. Select the station number from the list and click **Connect**. If the program indicates a connection is established, click the **Send** button to send the program to the station.

Checking call back files If the station has called back because of power loss, low battery voltage, or other problem, the program will create a “shadow file” (with “_s” suffix in file name) for that call. Click the **View** button to look for the shadow files of the most recent dates and open them to see why the station has called.

Programming the CR10X and RF units, continued

Record work performed Record in the AIRNET field logbook the date of the installation of the RF system, the station number, notes on any problems encountered, and any other work performed. Follow the requirements in MAQ-011 when making logbook entries.

Records resulting from this procedure

Records

The following records are generated as a result of this procedure and will be submitted according to MAQ-011:

- Entries in the logbook to document the work performed

HAZARD REVIEW FOR INSTALLING AND PROGRAMMING THE RADIO SYSTEMS ON AIRNET STATIONS

Work tasks/Steps	Hazards, Concerns, and Potential accidents; Likelihood/ Severity	Controls, Preventive Measures (e.g., safety equipment, administrative controls, etc.)	Hazard Level from IMP 300-00-00 Hazard Grading Matrix
Install battery in radio system according to steps in this procedure.	Electrical shock. battery in unit is a small 12 volt battery, and standard small transformer is plugged into standard 110V outlet improbable / moderate = low	Battery in unit is small 12 volt, so potential for shock is minimal. Steps call for connecting battery only after all other connections made. Transformer is plugged into GFI outlet and no exposed connectors are involved.	Low
Use tools (screw drivers, wrenches, knives, pliers & wire cutters, etc.) as needed to install or mount equipment	Pinched fingers or other minor injuries from use of hand tools. occasional / negligible = minimal	Use common sense and use proper tool for the job.	Low
Install antennas on AIRNET stations according to steps in this procedure.	Electrocution from contact with overhead electrical lines. catastrophic / remote = low	Maintain at least 13 feet distance between the closest end of a metal antenna and a power line (from "General Requirements for Electrical Work Practices", 2000 edition).	Low

Wastes or residual materials resulting from process

None.

Emergency actions to take in event of control failure

For all injuries, provide first aid and see that injured person is taken to Occupational Medicine (only if immediate medical attention is not required) or the nearest hospital. Notify supervisor and group office as soon as possible.

EXAMPLE PROGRAM FOR CR10X UNIT

```

;{CR10X}
;
*Table 1 Program
 01: 10   Execution Interval (seconds)

1: Batt Voltage (P10)
 1: 1     Loc [ batvolt ]

2: If Flag/Port (P91)
 1: 28   Do if Flag 8 is Low
 2: 30   Then Do

3: Do (P86)
 1: 11   Set Flag 1 High

4: Do (P86)
 1: 18   Set Flag 8 High

5: End (P95)

6: If Flag/Port (P91)
 1: 41   Do if Port 1 is High
 2: 21   Set Flag 1 Low

7: IF (X<=>F) (P89)
 1: 6    X Loc [ _____ ]
 2: 4    <
 3: 12.5 F
 4: 21   Set Flag 1 Low

8: Initiate Telecommunications (P97)
 1: 2    RF Modem/9600 Baud
 2: 1    Disabled when User Flag 1 is High
 3: 180  Seconds Call Time Limit
 4: 600  Seconds Before Fast Retry
 5: 3    Fast Retries
 6: 0    Minutes before Slow Retry
 7: 5    Failures Loc [ _____ ]
 8: 091  Data Logger ID

9: Extended Parameters (P63)
 1: 2    Option
 2: 5    Option
 3: 4    Option
```

4: 70 Option
5: 13 Option
6: 0 Option
7: 0 Option
8: 0 Option

10: If time is (P92)

1: 0 Minutes (Seconds --) into a
2: 60 Interval (same units as above)
3: 10 Set Output Flag High (Flag 0)

11: Set Active Storage Area (P80)

1: 1 Final Storage Area 1
2: 091 Array ID

12: Real Time (P77)

1: 1120 (Same as 1220) Y,D,Hr/Mn

13: Average (P71)

1: 1 Reps
2: 1 Loc [batvolt]

*Table 2 Program

02: 0.0000 Execution Interval (seconds)

*Table 3 Subroutines

End Program