

Seventy-Second Meeting of the  
**Obstetrics and Gynecology Devices Panel**

Tuesday, August 29, 2006  
Gaithersburg Hilton, Gaithersburg, MD

**Mirabel Medical Systems, Inc. T-Scan™ 2000 ED (P050003)**  
**Discussion Questions**

1. The primary effectiveness endpoint of the pivotal study used estimates of prevalence, sensitivity, and specificity to calculate the probability that a woman who is T-Scan positive has cancer relative to a randomly selected woman from the general target population. The success criterion was set at a level of 2 or greater. Using a sensitivity of 26.4%, a specificity of 94.7%, and a cancer prevalence of 0.15%, a “relative probability” of 4.95 (95% CI 3.16; 7.14) was obtained for the overall study population, thereby meeting the pre-defined endpoint for success. Sensitivity and specificity varied, however, for the different subgroups including the intended use population and the exact prevalence of disease in this age group is uncertain.
  - a. Please discuss the clinical significance of the primary effectiveness measure and result obtained for the overall study population.
  - b. Please discuss the clinical significance of the effectiveness outcomes by covariates and for subgroups including the intended use population where the values for sensitivity and specificity were different from those used to calculate the overall study population result.
  
2. The T-Scan device is intended to be used in women aged 30-39 who are negative for both Clinical Breast Exam (CBE) and Family History. Due to the relatively low prevalence of disease in the intended population, FDA agreed to enriching the sensitivity arm with subjects outside these criteria to assist with patient accrual. In order to identify a cohort of biopsy positive women (i.e. the cohort needed to calculate sensitivity), the sensitivity arm included women who were undergoing a biopsy due to an earlier positive screening. In addition, the sample included women age 40-45 to ensure a sufficient sample size. Of the evaluable cancer patients used to determine device sensitivity (excluding post-menopausal women), 58% (50/87) were over age 39, 81% (70/87) had an abnormal CBE, and 16% (13/81) had a positive family history (first degree relative). Only 5% (4/87) of cancer-positive subjects in the sensitivity arm were consistent with the indications for use statement.

Please discuss whether you believe the degree of enrichment of the sensitivity arm affects the interpretation of the final results of the study and their applicability to the intended target population and if so, how.

3. The pivotal study was conducted at multiple sites in the United States and Israel. As shown in the table below, there were differences in sensitivity results between subjects from Israel and the United States.

	<b>Excluding Post-Menopausal Women in Se Arm</b>	<b>Including Post-Menopausal Women in Se Arm</b>
<b>Sensitivity in U.S.</b>	<b>11.5% (3/26)</b>	<b>10.3% (3/29)</b>
<b>Sensitivity in Israel</b>	<b>32.8% (20/61)</b>	<b>32.3% (21/65)</b>

In addition, specificity (within the Specificity Arm) was 95.5% in U.S. subjects and 92.7% in Israeli subjects. Please discuss any differences in patient characteristics as well as the differences in sensitivity and specificity and whether you believe the results from the two countries are poolable.

4. Eleven percent (11%) of patients (65/597) were excluded from the sensitivity arm due to technical difficulties with the device. Of the 37 cancers excluded from the sensitivity arm, 51% (19) were eliminated due to technical difficulties. All 19 of these cancer cases excluded due to technical difficulties were from U.S. sites. In contrast, only 0.7% (14/1946) of the patients were excluded for technical difficulties from the specificity arm.

Please discuss whether these losses due to technical difficulties introduce significant bias into the study.

5. No device-related adverse events were reported during the course of the study. However, according to the sponsor's calculations, for every T-Scan positive patient who has cancer, additional mammograms will be conducted on 135 normal subjects. If the prevalence of disease in women 30-39 is lower than the sponsor's estimate of 0.15%, this number of additional mammograms may be significantly higher.

Please discuss whether you believe there are any potential risks associated with these additional mammography exams in women age 30-39, taking into account that for any given woman, the T-Scan is intended to be used on a yearly basis.

6. The sponsor has proposed the following indication for use (IFU) for its device:

*The T-Scan 2000 ED is indicated for use as a complement to clinical breast examination (CBE) in asymptomatic women who are 30 to 39 years of age with a negative clinical breast exam and a negative family history for breast cancer. The device detects electrical impedance changes in breast tissue that are associated with an increased risk of breast cancer. A positive T-Scan™ result provides physicians with additional information to guide a recommendation regarding further breast examination, e.g., mammography or ultrasound. The T-Scan evaluates women's risk of breast cancer at the time of the exam (current risk) and not lifetime risk.*

Please comment on whether data provided in the PMA and discussed today provide a reasonable assurance of effectiveness and safety to support this proposed indication for use. If not, are there any simple modifications to this indication which the data clearly support?

7. Taking into account your responses to the previous questions, please discuss the overall risk/benefit profile for the T-Scan device for the intended patient population.
  
8. Please comment on the draft labeling provided by the sponsor.