

Appendix C5

The Monocytoid Cell Line Mono Mac 6 (MM6)/IL-6 *In Vitro* Pyrogen Test

[This Page Intentionally Left Blank]

ICCVAM Final Recommended Protocol for Future Studies Using the Monocytoid Cell Line Mono Mac 6 (MM6)/Interleukin (IL)-6 *In Vitro* Pyrogen Test

PREFACE

This protocol is for the detection of Gram-negative endotoxin, a pyrogen, in parenteral drugs, as indicated by the release of IL-6 from the monocytoid cell line Mono Mac 6 (MM6). This protocol is based on information obtained from 1) the European Centre for the Validation of Alternative Methods (ECVAM)¹, MM6/IL-6 Background Review Document (BRD) presented in Appendix A of the Interagency Coordinating Committee on the Validation of Alternative Methods (ICCVAM) BRD (available at http://iccvam.niehs.nih.gov/methods/pyrogen/pyr_brd.htm), and 2) information provided to the National Toxicology Program (NTP) Interagency Center for the Evaluation of Alternative Toxicological Methods (NICEATM) by Dr. Thomas Hartung, Head of ECVAM. The ICCVAM BRD includes the ECVAM Standard Operating Procedures (SOPs) for the MM6/IL-6 test (could be referred to as Monocyte Activation Test), which are based on the methodology published by Taktak et al. (1991). A table of comparison between the ICCVAM recommended protocol and the ECVAM SOPs is provided in **Table 1**.

Users should contact the relevant regulatory authority for guidance when using this ICCVAM recommended protocol to demonstrate product specific validation, and any deviations from this protocol should be accompanied by scientifically justified rationale. Future studies using the MM6/IL-6 pyrogen test may include further characterization of the usefulness or limitations of the assay for regulatory decision-making. Users should be aware that this protocol might be revised based on additional optimization and/or validation studies. ICCVAM recommends that test method users routinely consult the ICCVAM/NICEATM website (<http://iccvam.niehs.nih.gov>) to ensure that the most current protocol is used.

¹ECVAM is a unit of the Institute for Health and Consumer Protection at the European Commission's Joint Research Centre.

[This Page Intentionally Left Blank]

Table 1 Comparison of ICCVAM Recommended Protocol with the ECVAM SOPs for the MM6/IL-6 Pyrogen Test

Protocol Component	ICCVAM Protocol	ECVAM SOP ¹	ECVAM Validation SOP ¹
Test Substance	Test neat or in serial dilutions that produce no interference, not to exceed the MVD	Test neat or at minimal dilution that produces no interference	Test at MVD
Decision Criteria for Interference	Mean OD ² of PPC is 50% to 200% of 1.0 EU/mL EC	Mean OD of PPC is 50% to 200% of 1.0 EU/mL EC	Mean OD of PPC is 50% to 200% of 1.0 EU/mL EC
Incubation Plate for ELISA (The number of samples or controls in quadruplicate)	NSC (1)	NSC (1)	NSC (1)
	EC (5)	EC (5)	EC (5)
	TS (14)	TS (14)	TS (2) x EC (5) spikes = 10 TS
	PPC ³ (0)	PPC (0)	PPC (2) = 2 TS
	NPC ³ (0)	NPC (0)	NPC (2) = 2 TS
	PC ⁴ (0)	PC (0)	PC (1) = 1 TS
	NC ⁴ (0)	NC (0)	NC (1) = 1 TS
ELISA Plate	Includes seven point IL-6 SC and blank in duplicate	Includes seven point IL-6 SC and blank in duplicate	Not included
Assay Acceptability Criteria	Quadratic function of IL-6 SC $r \geq 0.95^5$	Not included	Not included
	Mean OD of NSC ≤ 0.15	Not included	Not included
	EC SC produces OD values that ascend in a sigmoidal concentration response	Endotoxin concentration (0.5 IU/mL) > background (defined as the mean +2SD (n-1))	Mean OD of each EC > Mean OD of next lower EC concentration (minimum of 4 data points needed for valid SC)
	Not included	Not included	PC = $\pm 20\%$ of the theoretical value
	Not included	Not included	OD NC < 0.200
	Not included	Not included	OD PC > LOQ ⁶
	Outliers rejected using Dixon's test	Outliers rejected using Dixon's test	Outliers rejected using Dixon's test
Decision Criteria for Pyrogenicity	Endotoxin concentration TS > ELC ⁷ TS	Endotoxin concentration TS > ELC TS	OD TS > OD 0.5 EU/mL EC

Abbreviations: EC = Endotoxin control; ELC = Endotoxin limit concentration; ELISA = Enzyme-linked immunosorbent assay; EU = Endotoxin units; IL-6 = Interleukin-6; IU = International units; LOQ = Limit of quantification; MM6 = Mono Mac 6; MVD = Maximum valid dilution; NC = Negative control; NPC = Negative product control; NSC = Negative saline control; OD = Optical density; PC = Positive control;

PPC = Positive product control; SC = Standard curve; SD = Standard deviation; SOP = Standard operating procedure; TS = Test substance

¹ECVAM MM6/IL-6 SOP and ECVAM MM6/IL-6 Validation SOP are presented in Appendix A of the ICCVAM BRD (available at http://iccvam.niehs.nih.gov/methods/pyrogen/pyr_brd.htm).

²Mean OD values are corrected (i.e., reference filter reading, if applicable, and NSC are subtracted).

³In the ICCVAM MM6/IL-6 protocol, PPC and NPC are assessed in the interference test described in **Section 4.3**, which is performed prior to the ELISA. In the ECVAM SOPs, PPC and NPC were only included in the ECVAM validation study.

⁴PC and NC were only included in the ECVAM validation study. PC is 50 pg/mL endotoxin in saline. NC is 0.9% saline.

⁵Correlation coefficient (r), an estimate of the correlation of x and y values in a series of n measurements.

⁶LOQ is the mean OD of the NSC + 10x the SD of the mean OD for the NSC.

⁷Where unknown, the ELC is calculated (See **Section 12.2**).

1.0 PURPOSE AND APPLICABILITY

The purpose of this protocol is to describe the procedures used to evaluate the presence of Gram-negative endotoxin, a pyrogen, in parenteral drugs. The presence of Gram-negative endotoxin is detected by its ability to induce the release of IL-6 from Mono Mac 6 (MM6) cells, a human cell line derived from a patient with acute monocytic leukemia (Zeigler-Heitbrock et al. 1988). The concentration of IL-6 released by incubation of MM6 cells with a test substance or controls (i.e., positive and negative) is quantified using an enzyme-linked immunosorbent assay (ELISA) that includes monoclonal or polyclonal antibodies specific for IL-6. The amount of pyrogen present is determined by comparing the values of endotoxin equivalents produced by MM6 cells exposed to the test substance to those exposed to an internationally harmonized Reference Standard Endotoxin (RSE)¹ or an equivalent standard expressed in Endotoxin Units (EU)/mL. A test substance is considered pyrogenic if the endotoxin concentration of the test substance exceeds the Endotoxin Limit Concentration (ELC) for the test substance.

The relevance and reliability of this test method to detect non-endotoxin pyrogens have not been demonstrated in a formal validation study, although data are available in the literature to suggest that this assay has the potential to serve this purpose.

2.0 SAFETY AND OPERATING PROCEDURES

All procedures should be performed following standard laboratory precautions, including the use of laboratory coats, eye protection, and gloves. If necessary, additional precautions required for specific chemicals will be identified in the Material Safety Data Sheet (MSDS).

The stop solution used in the ELISA kit is acidic and corrosive and should be handled with the proper personal protective devices. If this reagent comes into contact with skin or eyes, wash thoroughly with water. Seek medical attention, if necessary.

Tetramethylbenzidine (TMB) solution contains a hydrogen peroxide substrate and 3, 3', 5, 5'-TMB. This reagent is a strong oxidizing agent and a suspected mutagen. Appropriate personal protection should be used to prevent bodily contact.

Bacterial endotoxin is a toxic agent (i.e., can induce sepsis, shock, vascular damage, antigenic response) and should be handled with care. Skin cuts should be covered and appropriate personal protective devices should be worn. In case of contact with endotoxin, immediately flush eyes or skin with water for at least 15 minutes (min). If inhaled, remove the affected individual from the area and provide oxygen and/or artificial respiration as needed. Skin absorption, ingestion, or inhalation may produce fever, headache, and hypotension.

¹RSEs are internationally harmonized reference standards (e.g., WHO-lipopolysaccharide [LPS] 94/580 *Escherichia coli* [*E. coli*] O113:H10:K-; United States Pharmacopeia [USP] RSE *E. coli* LPS Lot G3E069; USP RSE *E. coli* Lot G; FDA *E. coli* Lot EC6). Equivalent endotoxins include commercially available *E. coli*-derived LPS Control Standard Endotoxin (CSE) or other *E. coli* LPS preparations that have been calibrated with an appropriate RSE.

3.0 MATERIALS, EQUIPMENT AND SUPPLIES

3.1 Source of Cells

The MM6 cell line is a human monocytic cell line originally described by Professor H.W.L. Ziegler-Heitbrock at the Institute for Immunology, University of Munich, Germany (Ziegler-Heitbrock et al. 1988). The MM6 cell line may be purchased from the German Collection of Microorganisms and Cell Cultures (DSMZ, <http://www.dsmz.de>) by individuals working at non-profit organizations. Prior to transaction, a legal agreement must be reached with Professor H.W.L. Ziegler-Heitbrock stating that the cells will be used for research purposes only. Any contract research organization or pharmaceutical company wanting to obtain the MM6 cell line must contact Professor H.W.L. Ziegler-Heitbrock to negotiate a fee for provision and a royalty payment per batch of product tested. Contact information for Professor H.W.L. Ziegler-Heitbrock is as follows: Professor Dr. H.W.L. Ziegler-Heitbrock, University of Leicester, Dept. of Microbiology, University Road, Leicester LE1 9HN, United Kingdom, e-mail: ziehei@gmx.de.

MM6 cells should be maintained according to the instructions provided by the DSMZ and Professor Dr. H.W.L. Ziegler-Heitbrock, which should stipulate the permissible limit to the passage number for these cells.

3.2 Equipment and Supplies

For all steps in the protocol, excluding the ELISA procedure, the materials that will be in close contact with samples (e.g., pipet tips, containers, solutions) should be sterile and pyrogen-free.

3.2.1 Utilization of MM6 cells

3.2.1.1 Equipment

- Centrifuge
- Hood; Bio-safety, laminar flow (recommended)
- Incubator; cell culture ($37\pm 1^{\circ}\text{C} + 5\% \text{CO}_2$)
- Inverted Microscope
- pH meter
- Pipetter; multichannel (8- or 12-channel)
- Pipettors; single-channel adjustable (20, 200, and 1000 μL)
- Repeating pipetter
- Vortex mixer
- Water bath

3.2.1.2 *Consumables*

- Centrifuge tubes; polystyrene (15 and 50 mL)
- Combitips; repeating pipetter (1.0 and 2.5 mL)

- Cryotubes; screw-cap (2 mL)
- Filters; sterile, 0.22 µm
- Flasks; tissue culture
- Phosphate buffered saline (PBS); sterile
- Pipettes; sterile
- Plates; microtiter, 96-well, polystyrene, tissue culture
- Pyrogen-free saline (PFS)
- Reaction tubes; polystyrene (1.5 mL)
- RPMI-1640 cell culture medium supplemented as described in **Section 4.3** to yield either RPMI-C or RPMI-M
- Tips; pipetter, sterile, pyrogen-free (20 and 200 µL)
- Tubes; polystyrene

3.2.2 ELISA

3.2.2.1 *Equipment*

- Microplate mixer
- Microplate reader (450 nm with an optional reference filter in the range of 540-590 nm)²
- Microplate washer (optional)
- Multichannel pipetter

3.2.2.2 *Consumables*

- Container; storage, plastic
- Deionized water; nonsterile
- Plates; microtiter, 96-well, polystyrene
- Pyrogen-free water (PFW)
- Reservoirs; fluid
- Tips; pipetter, sterile and nonsterile
- Tubes; polystyrene (12 mL)

3.2.2.3 *ELISA Kit*

An ELISA that measures IL-6 release is used. A variety of IL-6 ELISA kits are commercially available and the IL-6 ELISA procedure outlined in this protocol is intended to serve as an example for using an ELISA kit. The IL-6 ELISA should be calibrated using an IL-6

² The TMB chromagen is measured at OD₄₅₀. However, the use of an IL-1β ELISA kit with a chromagen other than TMB is acceptable. The ELISA should be measured at a wavelength appropriate for the specific chromagen used.

international reference standard (e.g., World Health Organization [WHO] 89/548) prior to use. The IL-6 cytokine assay kits do not provide the RSE or endotoxin equivalent; therefore, this reagent must be purchased separately. Results obtained using these products are subject to the assay acceptability and decision criteria described in **Sections 8.0** and **9.0**. IL-6 ELISA kit components may include the following:

- ELISA plates coated with anti-human IL-6 capture antibody; monoclonal or polyclonal
- Buffered wash solution
- Dilution buffer
- Enzyme-labeled detection antibody
- Human IL-6 reference standard
- PFS
- Stop solution
- TMB³/substrate solution

3.3 Chemicals

- Endotoxin (e.g., WHO-lipopolysaccharide [LPS] 94/580 *Escherichia coli* [*E. coli*] O113:H10:K-; United States Pharmacopeia [USP] RSE *E. coli* LPS Lot G3E069; USP RSE *E. coli* Lot G; U.S. Food and Drug Administration [FDA] *E. coli* Lot EC6)

3.4 Solutions

- RPMI-1640 cell culture medium; supplemented as described in **Section 4.3**

4.0 ASSAY PREPARATION

All test substances, endotoxin, and endotoxin-spiked solutions should be stored as specified in the manufacturer's instructions. The preparation of MM6 cells for use in the assay is outlined in **Section 6.1**.

4.1 Endotoxin Standard Curve

An internationally harmonized RSE or equivalent is used to generate the endotoxin standard curve. The use of any other *E. coli* LPS requires calibration against a RSE using the MM6/IL-6 pyrogen test. A standard endotoxin curve consisting of a Negative Saline Control (NSC) and five RSE concentrations (0.125, 0.25, 0.50, 1.0, and 2.0 EU/mL) are included in the incubation step (refer to **Table 4-1**) and then transferred to the ELISA plate. To prepare the endotoxin standard curve, first obtain a 2000 EU/mL stock solution by addition of PFW to the lyophilized content of the stock vial by following the instructions provided by the manufacturer (e.g., 5 mL of PFW is added to a vial containing 10,000 EU). To reconstitute the endotoxin, the stock vial should be vortexed vigorously for at least 30 min or sonicated in

³The use of an IL-6 ELISA kit with a chromagen other than TMB is acceptable.

a bath sonicator for at least 5 min. Subsequent dilutions should be vortexed vigorously immediately prior to use. The stock solution is stable for not more 14 days when stored at 2 to 8°C or for up to 6 months when kept in a -20°C freezer. An endotoxin standard curve is prepared as described in **Table 4-1** by making serial dilutions of the stock solution in PFS with vigorous vortexing at each dilution step. Dilutions should not be stored, because dilute endotoxin solutions are not as stable as concentrated solutions due to loss of activity by adsorption, in the absence of supporting data to the contrary.

Table 4-1 Preparation of Endotoxin Standard Curve

Stock Endotoxin EU/mL ¹	µL of Stock Endotoxin	µL of PFS	Endotoxin Concentration EU/mL
2000 ^{2,3}	40	3960	20 ⁴
20	100	900	2.0
2.0	500	500	1.0
1.0	500	500	0.50
0.50	500	500	0.25
0.25	500	500	0.125
0	0	1000	0

Abbreviations: EU = Endotoxin units; PFS = Pyrogen-free saline

Each stock tube should be resonicated and vortexed vigorously before the subsequent dilution.

¹To reconstitute the endotoxin, the stock vial should be vortexed vigorously for at least 30 min or sonicated in a bath sonicator for at least 5 min. Subsequent dilutions should be vortexed vigorously immediately prior to use.

²A 2000 EU/mL stock solution of endotoxin is prepared according to the manufacturer's instructions.

³The stock solution is stable for not more 14 days when stored at 2 to 8°C or for up to 6 months when kept in a -20°C freezer.

⁴This concentration is not used in the assay.

4.2 Cell Culture Medium

MM6 cells are maintained in RPMI containing 10% FBS, denoted as RPMI-M. For use in the ELISA procedure, the concentration of FBS is reduced to 2% and referred to as RPMI-C. Each medium is prepared and stored as described by the manufacturer.

4.2.1 RPMI-M

- Bovine insulin; 0.23 IU/mL
- FBS; heat-inactivated at 55±1°C (50 mL or a 10% final concentration)
- HEPES buffer; 20 mM
- L-Glutamine; 2 mM
- MEM non-essential amino acids; 0.1 mM
- Oxaloacetic acid; 1 mM
- Penicillin/streptomycin (10,000 IU/mL penicillin, 10 mg/mL streptomycin)
- RPMI-1640 medium (500 mL)
- Sodium pyruvate; 1 mM

4.2.2 Starting a Culture of MM6 Cells

To initiate a culture of MM6 cells, remove a vial of the primary stock from liquid nitrogen. Thaw the vial on ice. Gently mix and transfer the cells to a 50 mL centrifuge tube and add 10 mL of RPMI-M. Centrifuge at 100 x g for 5 min at room temperature (RT). Remove the supernatant and resuspend the cells in ice-cold RPMI-M. Centrifuge at 100 x g for 5 min at RT. Remove the supernatant and resuspend the MM6 cells in 2 mL of RPMI-M. Add 8 mL of RPMI-M to a tissue culture flask and transfer the cell suspension to the flask. Cells should be examined microscopically to ensure that the cells are not clumped together. Place the flasks in a cell culture incubator and maintain the cells at $37\pm 1^{\circ}\text{C} + 5\% \text{CO}_2$.

4.2.3 Propagation of MM6 Cells

Remove the cell culture flask from the incubator and examine the cells under a microscope to determine that the morphology of the cells is consistent with the appearance of MM6 cells that previously yielded acceptable results. Centrifuge at 100 x g for 8 min at RT. Remove the supernatant, resuspend the cell pellet in 4 mL of RPMI-M, and gently pipet up and down to mix. It is advisable that cell number and cell viability be determined using appropriate methods (e.g., hemocytometer and vital dye or flow cytometer and fluorescent marker). The percentage of cell viability should exceed 80% for further propagation. The results of these examinations should be included in the study report. Transfer the cells (2×10^5 cells/mL) to new tissue culture flasks and add RPMI-M. Place the flasks in a cell culture incubator and maintain the cells at $37\pm 1^{\circ}\text{C} + 5\% \text{CO}_2$.

4.2.4 Preparation of a MM6 Cell Bank

To initiate a bank of MM6 cells, centrifuge the cell culture(s) at 100 x g for 8 min at 2 to 8°C . Remove the supernatant and resuspend the cells in FBS at 2 to 8°C . It is advisable to determine cell number and cell viability as outlined in **Section 4.2.3** and adjust the cell concentration to 4×10^6 cells/mL and store on ice for 10 min. Add an equal volume of ice-cold FBS containing 10% dimethyl sulfoxide (DMSO) drop-wise to the cell suspension (final concentration is 2×10^6 cells/mL with 5% DMSO). Transfer the cell suspension to sterile, pyrogen-free cryotubes (1 mL/tube). Place the tubes in a well-insulated polystyrene box and store in a -80°C freezer for greater than 48 hours (hr) and then transfer to a liquid nitrogen container.

4.3 **Interference Test**

For every test substance lot, interference testing must be performed to check for interference between the test substance and the cell system and/or ELISA. The purpose of the interference test is to determine whether the test substance (or specific lot of test substance) has an effect on cytokine release.

4.3.1 Interference with the Cell System

All test substances must be labeled as pyrogen-free (i.e., endotoxin levels at an acceptable level prior to release by the manufacturer) to ensure that exogenous levels of endotoxin do not affect the experimental outcome. Liquid test substances should be diluted in PFS. Solid test substances should be prepared as solutions in PFS or, if insoluble in saline, dissolved in DMSO and then diluted up to 0.5% (v/v) with PFS, provided that this concentration of DMSO does not interfere with the assay. To ensure a valid test, a test substance cannot be

diluted beyond its Maximum Valid Dilution (MVD) (refer to **Section 12.3**). The calculation of the MVD is dependent on the ELC for a test substance. The ELC can be calculated by dividing the threshold human pyrogenic dose by the maximum recommended human dose in a single hour period (see **Section 12.2**) (USP 2007; FDA 1987). Furthermore, test substances should not be tested at concentrations that are cytotoxic to MM6 cells.

4.3.1.1 *Reference Endotoxin for Spiking Test Substances*

The WHO-LPS 94/580 [*E. coli* O113:H10:K-] or equivalent internationally harmonized RSE is recommended for preparation of the endotoxin-spike solution and the endotoxin standard curve (see **Section 4.1**).

4.3.1.2 *Spiking Test Substances with Endotoxin*

Non-spiked and endotoxin-spiked test substances are prepared in quadruplicate and an *in vitro* pyrogen test is performed. A fixed concentration of the RSE (i.e., 1.0 EU/mL or a concentration equal to or near the middle of the endotoxin standard curve) is added to the undiluted test substance (or in serial two-fold dilutions, not to exceed the MVD). An illustrative example of endotoxin spiking solutions is shown in **Table 4-2**. For non-spiked solutions, 150 µL of RPMI-C and 50 µL of the test substance (i.e., equivalent to the negative product control [NPC]) are added to a well. Endotoxin-spiked solutions are prepared by adding 100 µL of RPMI-C, 50 µL of the test substance, and 50 µL of an endotoxin-spike solution (1.0 EU/mL) (i.e., equivalent to the positive product control [PPC]). Finally, MM6 cells (50 µL) are added to each well and the wells are mixed and incubated as outlined in **Section 6.1.3, Steps 6-7**. An ELISA is then performed as outlined in **Section 6.2**, without the IL-6 standard curve.

Table 4-2 Preparation of Endotoxin-Spiked and Non-Spiked Solutions for Determination of Test Substance Interference

Sample Addition	Spiked	Non-spiked
	µL/well ¹	
RPMI-C	100	150
Endotoxin-spike solution ²	50	0
Test substance (neat and each serial dilution)	50	50
MM6 cells ³	50	50
Total ⁴	250	250

Abbreviations: MM6 cells = Mono Mac 6

¹n=4 replicates each

²Endotoxin concentration is 1.0 EU/mL in RPMI-C.

³MM6 cells are resuspended in RPMI-C (2.5 x 10⁶ cells/mL).

⁴A total volume of 250 µL per well is used for the incubation.

The optical density (OD) values of the endotoxin-spiked and non-spiked test substances are calibrated against the endotoxin calibration curve. The resulting EU value of the non-spiked test substance is subtracted from the corresponding EU value of the endotoxin-spiked test substance at each dilution. The spike recovery for each sample dilution is calculated as a percentage by setting the theoretical value (i.e., endotoxin-spike concentration of 1.0 EU/mL) at 100%. For example, consider the following interference test results in Table 4-3:

Table 4-3 Example of Interference Data Used to Determine Sample Dilution

Sample Dilution	% Recovery of Endotoxin Control
None	25
1:2	49
1:4	90
1:8	110

If a spike recovery between 50% and 200% is obtained, then no interference of the test substance with either the cell system or the ELISA is demonstrated (i.e., the test substance does not increase or decrease the concentration of IL-6 relative to the endotoxin spike). The lowest dilution (i.e., highest concentration) of a test substance that yields an endotoxin-spike recovery between 50% and 200% is determined. The test substance is then diluted in serial two-fold dilutions beginning at this dilution, not to exceed the MVD, for use in the assay. Based on the results illustrated in **Table 4-3**, the initial dilution of the test substance to be used in the *in vitro* pyrogen test would be 1:4 (i.e., the lowest dilution between 50% and 200% of the 1.0 EU/mL EC).

4.3.2 Interference at the MVD

If the data obtained from the experiment in **Section 4.2.1** suggests the presence of interference at the MVD, then consideration should be given for using another validated pyrogen test method.

5.0 CONTROLS

5.1 Benchmark Controls

Benchmark controls may be used to demonstrate that the test method is functioning properly, or to evaluate the relative pyrogenic potential of chemicals (e.g., parenteral pharmaceuticals, medical device eluates) of a specific class or a specific range of responses, or for evaluating the relative pyrogenic potential of a test substance. Appropriate benchmark controls should have the following properties:

- consistent and reliable source(s) for the chemicals (e.g., parenteral pharmaceuticals, medical device eluates)
- structural and functional similarities to the class of substance being tested
- known physical/chemical characteristics
- supporting data on known effects in animal models
- known potency in the range of response

5.2 Endotoxin Control

The EC (i.e., MM6 cells incubated with an internationally harmonized RSE) serves as the positive control in each experiment. The results should be compared to historical values to insure that it provides a known level of cytokine release relative to the NSC.

5.3 Negative Saline Control

The NSC (i.e., MM6 cells incubated with PFS instead of the test substance) is included in each experiment in order to detect nonspecific changes in the test system, as well as to provide a baseline for the assay endpoints.

5.4 Solvent Control

Solvent controls are recommended to demonstrate that the solvent is not interfering with the test system when solvents other than PFS are used to dissolve test substances.

6.0 EXPERIMENTAL DESIGN

6.1 Incubation with Test Samples and Measurement of IL-6 Release

6.1.1 Preincubation of MM6 Cells

To perform an ELISA on the following day, obtain an MM6 cell suspension (30 to 50 mL) from propagation flasks and centrifuge at 100 x g for 8 min at RT. Remove the supernatant, resuspend the cell pellet in 2 mL of RPMI-C and gently pipet up and down to mix. It is advisable to determine cell number and cell viability as outlined in **Section 4.2.3**. The percentage of viable MM6 cells should exceed 80% to be suitable for use in the test. The results of these examinations should be included in the study report. Transfer the cells (4×10^5 cells/mL) to new tissue culture flasks and add RPMI-C. Place the flasks in a cell culture incubator and maintain the cells at $37 \pm 1^\circ\text{C} + 5\% \text{CO}_2$ for 16 to 24 hr. In general, the preincubation of 2.0×10^7 cells in 50 mL RPMI-C will provide enough cells for one 96-well assay plate

6.1.2 Preparation of MM6 Cells for the Incubation Assay

Prepare the MM6 cells just prior to addition to the incubation plate (**Section 6.1.3, Step 5**). Centrifuge 30 to 50 ml of cell suspension at 100 x g for 8 min at RT. Pour off the supernatant and resuspend the cells in approximately 2 ml of RPMI-C. It is advisable that cell number and cell viability be determined as outlined in **Section 4.2.3**. The percentage of viable MM6 cells should exceed 80% to be suitable for use in the test. The results of these examinations should be included in the study report. Dilute the cells with RPMI-C to a volume that gives a concentration of 2.5×10^6 cells/ml.

6.1.3 Incubation Plate

Test substances should be vortexed vigorously for at least 30 min or sonicated in a bath sonicator for at least 5 min prior to use in the assay. Test substances should be prepared in serial two-fold dilutions beginning at a level of dilution that did not show interference with the test system (see **Section 4.2**) in as many subsequent dilutions that are necessary to be within the linear range of the endotoxin standard curve, not to exceed the MVD. Each incubation plate can accommodate an endotoxin standard curve, a NSC, and 14 test substances (see **Table 6-1**).

Table 6-1 Overview of Incubation Plate Preparation in the MM6/IL-6 Pyrogen Test

Number of Wells	Sample	RPMI-C	EC	Test Sample	MM6 ¹	Mix the samples; incubate for 16 to 24 hr at 37±1°C in a humidified atmosphere with 5% CO ₂ .	Mix the samples; immediately transfer to an ELISA plate ⁵ and run ELISA.
		µL					
20 ²	EC	100	50	0	100		
4	NSC	100	0	0 ³	100		
56 ⁴	Test samples (1-14)	100	0	50	100		

Abbreviations: EC = Endotoxin control; IL-6 = Interleukin-6; NSC = Negative saline control; MM6 = Mono Mac 6

¹MM6 cell concentration is 2.5 x 10⁶ cells/mL.

²Five EC concentrations (0.125, 0.25, 0.50, 1.0, and 2.0 EU/mL) in quadruplicate

³50 µl of PFS is added instead of the test sample.

⁴14 test samples (n=4 each) per plate

⁵An IL-6 standard curve is prepared in Columns 11 and 12 on the ELISA plate (see **Table 6-3**). Therefore, 80 wells are available for test samples and controls on the incubation plate.

6.1.4 Incubation Assay for IL-6 Release

MM6 cells are prepared in a microtiter plate using a laminar flow hood (refer to **Section 6.1.2**). All consumables and solutions must be sterile and pyrogen-free. Each plate should be labeled appropriately with a permanent marker. An overview of the incubation plate preparation is shown in **Table 6-1**. The incubation procedure is outlined below:

Step 1. Refer to the suggested incubation plate template presented in **Table 6-2**.

Step 2. Using a pipetter, transfer 100 µL of RPMI-C into each well.

Step 3. Transfer 50 µL of test sample or 50 µL of PFS for the NSC into the appropriate wells as indicated in the template.

Step 4. Transfer 50 µL of the EC (standard curve) in quadruplicate into the appropriate wells according to the template.

Step 5. Transfer 100 µL of a well-mixed MM6 cell suspension into each well.

Step 6. Place the covered plate in a tissue culture incubator for 16 to 24 hr at 37±1°C in a humidified atmosphere containing 5% CO₂.

Step 7. Remove 150 µL of the supernatant from each well, without disrupting the cells, and transfer to the IL-6 ELISA plate.

Table 6-2 Incubation Plate - Sample and Control Template

	1	2	3	4	5	6	7	8	9	10	11	12
A	EC ¹ 2.0	EC 2.0	EC 2.0	EC 2.0	TS3	TS3	TS3	TS3	TS11	TS11	Void ³	Void
B	EC 1.0	EC 1.0	EC 1.0	EC 1.0	TS4	TS4	TS4	TS4	TS11	TS11	Void	Void
C	EC 0.50	EC 0.50	EC 0.50	EC 0.50	TS5	TS5	TS5	TS5	TS12	TS12	Void	Void
D	EC 0.25	EC 0.25	EC 0.25	EC 0.25	TS6	TS6	TS6	TS6	TS12	TS12	Void	Void
E	EC 0.125	EC 0.125	EC 0.125	EC 0.125	TS7	TS7	TS7	TS7	TS13	TS13	Void	Void
F	NSC	NSC	NSC	NSC	TS8	TS8	TS8	TS8	TS13	TS13	Void	Void
G	TS1 ²	TS1	TS1	TS1	TS9	TS9	TS9	TS9	TS14	TS14	Void	Void
H	TS2	TS2	TS2	TS2	TS10	TS10	TS10	TS10	TS14	TS14	Void	Void

Abbreviations: EC = Endotoxin control; NSC = Negative saline control; TS = Test substance

¹EC value (e.g., EC 2.0) represents the endotoxin concentration in EU/mL.

²TS number (e.g., TS 1) represents an arbitrary sequence for individual test substances.

³Columns 11 and 12 are reserved for the IL-6 standard curve on the ELISA plate (see **Table 6-3**).

6.2 ELISA to Measure IL-6 Release

6.2.1 IL-6 Standard Curve

An IL-6 standard supplied with the ELISA kit is used. IL-6 standards are typically supplied in lyophilized form and should be reconstituted according to the manufacturer's instructions. The stock solution should be diluted in RPMI-C to the following concentrations: 0, 62.5, 125, 250, 500, 1000, 2000, and 4000 pg/mL in volumes of at least 500 µL. Each well on the ELISA plate will receive 50 µL of an IL-6 blank or standard.

6.2.2 ELISA

The manufacturer's instructions provided with the ELISA kit should be followed and a typical experimental design is outlined below. The ELISA should be carried out at RT and therefore all components must be at RT prior to use. Frozen specimens should not be thawed by heating them in a water bath. A suggested ELISA plate template is shown in **Table 6-3**, which includes a five-point EC standard curve, an eight-point IL-6 standard curve (0 to 4000 pg/mL), and available wells for up to 14 test substances and a NSC each in quadruplicate. The EC standard curve, the NSC, and the test sample supernatants are transferred directly from the incubation plate. The IL-6 standard curve is prepared as described in **Section 6.2.1**. An overview of the ELISA plate preparation is shown in **Table 6-4**.

Step 1. After pipetting up and down very carefully three times (avoid detachment of the adherent MM6 cells) to mix the supernatant, transfer 50 μ L from each well of the Incubation Plate (A1-10; H1-10) to the ELISA plate.

Step 2. Add 50 μ L of each IL-6 standard (0 to 4000 pg/mL) into the respective wells on the ELISA plate.

Step 3. Add 200 μ L of the enzyme-labeled detection antibody (neat as supplied, or diluted, if necessary) to each of the wells.

Step 4. Cover the microtiter plate(s) with adhesive film and incubate for 2 to 3 hr at RT.

Step 5. Decant and wash each well three times with 300 μ L Buffered Wash Solution and then rinse three times with deionized water. Place the plates upside down and tap to remove water.

Step 6. Add 200 μ L of TMB/Substrate Solution to each well and incubate at RT in the dark for 15 min. If necessary, decrease the incubation time.

Step 7. Add 50 μ L of Stop Solution to each well.

Step 8. Tap the plate gently after the addition of Stop Solution to aid in mixing.

Step 9. Read the OD₄₅₀ within 15 min of adding the Stop Solution. Measurement with a reference wavelength of 540 to 590 nm is recommended⁴.

⁴The TMB chromagen is measured at OD₄₅₀. However, the use of an IL-1 β ELISA kit with a chromagen other than TMB is acceptable. The ELISA should be measured at a wavelength appropriate for the specific chromagen used.

Table 6-3 ELISA Plate - Sample and Control Template

	1	2	3	4	5	6	7	8	9	10	11	12
A	EC ¹ 2.0	EC 2.0	EC 2.0	EC 2.0	TS3	TS3	TS3	TS3	TS11	TS11	IL-6 ³ 0	IL-6 0
B	EC 1.0	EC 1.0	EC 1.0	EC 1.0	TS4	TS4	TS4	TS4	TS11	TS11	IL-6 62.5	IL-6 62.5
C	EC 0.50	EC 0.50	EC 0.50	EC 0.50	TS5	TS5	TS5	TS5	TS12	TS12	IL-6 125	IL-6 125
D	EC 0.25	EC 0.25	EC 0.25	EC 0.25	TS6	TS6	TS6	TS6	TS12	TS12	IL-6 250	IL-6 250
E	EC 0.125	EC 0.125	EC 0.125	EC 0.125	TS7	TS7	TS7	TS7	TS13	TS13	IL-6 500	IL-6 500
F	NSC	NSC	NSC	NSC	TS8	TS8	TS8	TS8	TS13	TS13	IL-6 1000	IL-6 1000
G	TS1 ²	TS1	TS1	TS1	TS9	TS9	TS9	TS9	TS14	TS14	IL-6 2000	IL-6 2000
H	TS2	TS2	TS2	TS2	TS10	TS10	TS10	TS10	TS14	TS14	IL-6 4000	IL-6 4000

Abbreviations: EC = Endotoxin control; NSC = Negative saline control; TS = Test substance

¹EC value (e.g., EC 2.0) represents the endotoxin concentration in EU/mL.

²TS number (e.g., TS1) represents an arbitrary sequence for individual test substances.

³IL-6 values in columns 11 and 12 are in pg/mL.

Table 6-4 Overview of ELISA Procedure

Material transfer from Incubation Plate (µL)	IL-6 standard (0 to 4000 pg/mL) (µL)	Enzyme-labeled Antibody (µL)	Cover the Incubation Plate and incubate for 2 to 3 hr at RT.	Decant and wash each well three times with 300 µL Buffered Wash Solution and three times with deionized water.	TMB/Substrate Solution (µL)	Incubate for less than 15 min at RT in dark.	Stop Solution (µL)	Read each well at OD ₄₅₀ with a 540 to 590 nm reference filter.
50	50	200			200		50	

Abbreviations: OD₄₅₀ = Optical density at 450 nm; RT = Room temperature

7.0 EVALUATION OF TEST RESULTS

7.1 OD Measurements

The OD of each well is obtained by reading the samples in a standard microplate spectrophotometer (i.e., plate reader) using a visible light wavelength of 450 nm (OD_{450}) with a reference filter of 540 to 590 nm (recommended)⁵. OD values are used to determine assay acceptability and in the decision criteria for pyrogen detection (see **Sections 8.0** and **9.0**).

8.0 CRITERIA FOR AN ACCEPTABLE TEST

An EC (five-point standard curve) and a NSC should be included in each experiment. An IL-6 standard curve should be included in each ELISA as shown in the template presented in **Table 6-3**. An assay is considered acceptable only if the following minimum criteria are met:

- The quadratic function of the IL-6 standard curve produces an $r \geq 0.95$ ⁶ and the OD of the blank control is below 0.15.
- The endotoxin standard curve produces OD values that ascend in a sigmoidal concentration response.

An outlying observation may be excluded if the aberrant response is identified using acceptable statistical methodology (e.g., Dixon's test [Dixon 1950; Barnett and Lewis 1994] or Grubbs' test [Barnett and Lewis 1994; Grubbs 1969; Iglewicz and Houghlin 1993]).

9.0 DATA INTERPRETATION/DECISION CRITERIA

9.1 Decision Criteria for Pyrogen Detection

A test substance is considered pyrogenic when the endotoxin concentration of the test substance exceeds the ELC for the test sample. The ELC can be calculated as shown in **Section 12.2**.

10.0 STUDY REPORT

The test report should include the following information:

Test Substances and Control Substances

- Name of test substance
- Purity and composition of the substance or preparation
- Physicochemical properties (e.g., physical state, water solubility)
- Quality assurance data
- Treatment of the test/control substances prior to testing (e.g., vortexing, sonication, warming, resuspension solvent)

⁵The TMB chromagen is measured at OD_{450} . However, the use of an IL-1 β ELISA kit with a chromagen other than TMB is acceptable. The ELISA should be measured at a wavelength appropriate for the specific chromagen used.

⁶Correlation coefficient (r), an estimate of the correlation of x and y values in a series of n measurements.

Justification of the In Vitro Test Method and Protocol Used

Test Method Integrity

- The procedure used to ensure the integrity (i.e., accuracy and reliability) of the test method over time
- If the test method employs proprietary components, documentation on the procedure used to ensure their integrity from “lot-to-lot” and over time
- The procedures that the user may employ to verify the integrity of the proprietary components

Criteria for an Acceptable Test

- Acceptable concurrent positive control ranges based on historical data
- Acceptable negative control data

Test Conditions

- Cell system used
- Calibration information for the spectrophotometer used to read the ELISA
- Details of test procedure used
- Description of any modifications of the test procedure
- Reference to historical data of the model
- Description of evaluation criteria used

Results

- Tabulation of data from individual test samples

Description of Other Effects Observed

Discussion of the Results

Conclusion

A Quality Assurance Statement for Good Laboratory Practice (GLP)-Compliant Studies

- This statement should indicate all inspections made during the study and the dates any results were reported to the Study Director. This statement should also confirm that the final report reflects the raw data.

If GLP-compliant studies are performed, then additional reporting requirements provided in the relevant guidelines (e.g., OECD 1998; EPA 2003a, 2003b; FDA 2003) should be followed.

11.0 REFERENCES

- Barnett V, Lewis T. 1994. Outliers in Statistical Data. In: Wiley Series in Probability and Mathematical Statistics. Applied Probability and Statistics. 3rd Ed. New York: John Wiley & Sons.
- Dixon WJ. 1950. Analysis of extreme values. Annals of Mathematical Statistics 21:488-506.

EPA. 2003a. Good Laboratory Practice Standards. Toxic Substances Control Act. 40 CFR 792.

EPA. 2003b. Good Laboratory Practice Standards. Federal Insecticide, Fungicide, and Rodenticide Act. 40 CFR 160.

FDA. 1987. Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-product Endotoxin Test for Human and Animal Parenteral Drugs, Biological Products, and Medical Devices. Rockville, MD:U.S. Department of Health and Human Services (DHHS), Food and Drug Administration (FDA).

FDA. 2003. Good Laboratory Practices for Nonclinical Laboratory Studies. 21 CFR 58.

Grubbs FE. 1969. Procedures for detecting outlying observations in samples. *Technometrics* 11(1):1-21.

Iglewicz B, Houghlin DC. 1993. How to detect and handle outliers. In: *ASQC Basic Reference in Quality Control*. Vol. 14. Milwaukee, WI: ASQ Quality Press.

Organization for Economic Cooperation and Development (OECD). 1998. OECD Series on Principle of Good Laboratory Practice and Compliance Monitoring. No. 1. OECD Principles of Good Laboratory Practice (as revised in 1997). Organisation for Economic Co-operation and Development (OECD), ENV/MC/CHEM (98)17. Paris: OECD.

Taktak YS, Selkirk S, Bristow AF, Carpenter A, Ball C, Rafferty B, Poole S. 1991. Assay of pyrogens by interleukin-6 release from monocytic cell lines. *J Pharm Pharmacol* 43:578-582.

USP. 2007. The U.S. Pharmacopeia. USP30 NF25<85>. Ed. The U.S. Pharmacopeial Convention. Rockville, MD:The U.S. Pharmacopeial Convention.

12.0 TERMINOLOGY AND FORMULA

12.1 Assay Sensitivity (λ)¹

The variable λ is defined as the labeled sensitivity (in EU/mL) of the LAL Reagent in endpoint assays (e.g., the BET gel-clot technique). For kinetic BET assays, λ is the lowest point used in the endotoxin standard curve.

12.2 Endotoxin Limit Concentration (ELC)^{1,2}

The ELC for parenteral drugs is expressed in Endotoxin Units (EU) per volume (mL) or weight (mg). The ELC is equal to K/M , where:

K is the threshold human pyrogenic dose of endotoxin (EU) per body weight (kg). K is equal to 5.0 EU/kg for intravenous administration. For intrathecal administration, K is equal to 0.2 EU/kg (see also **Section 12.5**).

M is the rabbit test dose or the maximum recommended human dose of product (mL or mg) per body weight (kg) in a single hour period (see also **Section 12.8**).

For example, if a non-intrathecal product were used at an hourly dose of 10 mL per patient, then the ELC would be 0.50 EU/mL.

12.3 Maximum Valid Dilution (MVD)^{1,2}

The MVD is the maximum allowable dilution of a test substance at which the endotoxin limit can be determined. The calculation of the MVD is dependent on the ELC for a test substance. When the ELC is known, the MVD is¹:

$$\text{MVD} = (\text{ELC} \times \text{Product Potency [PP]})/\lambda$$

As an example, for Cyclophosphamide Injection, the ELC is 0.17 EU/mg, PP is 20 mg/mL, and the assay sensitivity is 0.065 EU/mL. The calculated MVD would be 1:52.3 or 1:52. The test substance can be diluted no more than 1:52 prior to testing.

If the ELC is not known, the MVD is¹:

$$\text{MVD} = \text{PP}/\text{Minimum Valid Concentration (MVC)}$$

$$\text{where, MVC} = (\lambda \times M)/K$$

where, M is the maximum human dose

As an example, for Cyclophosphamide Injection, the PP is 20 mg/mL, M is 30 mg/kg, and assay sensitivity is 0.065 EU/mL. The calculated MVC is 0.390 mg/mL and the MVD is 1:51.2 or 1:51. The test substance can be diluted no more than 1:51 in the assay prior to testing.

12.4 Negative Product Control (NPC)

For interference testing, the NPC is a test sample to which pyrogen-free saline (PFS) is added. The NPC is the baseline for determination of cytokine release relative to the endotoxin-spiked PPC.

¹From FDA (1987)

²From USP (2007)

12.5 Parenteral Threshold Pyrogen Dose (K)^{1,2}

The value K is defined as the threshold human pyrogenic dose of endotoxin (EU) per body weight (kg). K is equal to 5.0 EU/kg for parenteral drugs except those administered intrathecally; 0.2 EU/kg for intrathecal drugs.

12.6 Positive Product Control (PPC)

For interference testing, the PPC is a test substance spiked with the control standard endotoxin (i.e., 0.5 EU/mL or an amount of endotoxin equal to that which produces ½ the maximal increase in optical density (OD) from the endotoxin standard curve) to insure that the test system is capable of endotoxin detection in the product as diluted in the assay.

12.7 Product Potency (PP)^{1,2}

The test sample concentration expressed as mg/mL or mL/mL.

12.8 Rabbit Pyrogen Test (RPT) Dose or Maximum Human Dose (M)^{1,2}

The variable M is equal to the rabbit test dose or the maximum recommended human dose of product per kg of body weight in a single hour period. M is expressed in mg/kg or mL/kg and varies with the test substance. For radiopharmaceuticals, M equals the rabbit dose or maximum human dose/kg at the product expiration date or time. Use 70 kg as the weight of the average human when calculating the maximum human dose per kg. If the pediatric dose/kg is higher than the adult dose, then it shall be the dose used in the formula.