



May 1973

U.S. ATOMIC ENERGY COMMISSION

REGULATORY GUIDE

DIRECTORATE OF REGULATORY STANDARDS

REGULATORY GUIDE 8.6

STANDARD TEST PROCEDURE FOR GEIGER-MÜLLER COUNTERS

A. INTRODUCTION

Section 20.201 of 10 CFR Part 20, "Standards for Protection Against Radiation," requires that licensees make such surveys as are necessary to comply with the regulations in 10 CFR Part 20, including appropriate measurements of levels of radiation. Standard test procedures are necessary to ascertain that the instruments used for measuring radiation are operating properly. This guide provides an acceptable procedure for testing the operating characteristics of Geiger-Müller counters prior to making calibrations and measurements required by § 20.201.

B. DISCUSSION

The Radiation Detectors Committee of the Institute of Electrical and Electronics Engineers' (IEEE) Nuclear Science Group along with the Electron Tubes Standards Committee of the IEEE Electron Devices Group

produced a standard test procedure for Geiger-Müller counters which was approved as IEEE No. 309 by the IEEE Standards Committee on September 11, 1969. This procedure was subsequently approved as American National Standard N42.3-1969¹ by the American National Standards Institute (ANSI) on December 23, 1969.

C. REGULATORY POSITION

The standard test procedures for Geiger-Müller counters specified in ANSI N42.3-1969 (IEEE No. 309), "Test Procedure for Geiger-Müller Counters," are suitable for use in determining whether or not the operating characteristics of Geiger-Müller counters are acceptable.

¹Copies may be obtained from the Institute of Electrical and Electronic Engineers, 345 East 47th Street, New York, N.Y. 10017.

USAEC REGULATORY GUIDES

Regulatory Guides are issued to describe and make available to the public methods acceptable to the AEC Regulatory staff of implementing specific parts of the Commission's regulations, to delineate techniques used by the staff in evaluating specific problems or postulated accidents, or to provide guidance to applicants. Regulatory Guides are not substitutes for regulations and compliance with them is not required. Methods and solutions different from those set out in the guides will be acceptable if they provide a basis for the findings requisite to the issuance or continuance of a permit or license by the Commission.

Published guides will be revised periodically, as appropriate, to accommodate comments and to reflect new information or experience.

Copies of published guides may be obtained by request indicating the divisions desired to the U.S. Atomic Energy Commission, Washington, D.C. 20545. Attention: Director of Regulatory Standards. Comments and suggestions for improvements in these guides are encouraged and should be sent to the Secretary of the Commission, U.S. Atomic Energy Commission, Washington, D.C. 20545. Attention: Chief, Public Proceedings Staff.

The guides are issued in the following ten broad divisions:

1. Power Reactors
 2. Research and Test Reactors
 3. Fuels and Materials Facilities
 4. Environmental and Siting
 5. Materials and Plant Protection
 6. Products
 7. Transportation
 8. Occupational Health
 9. Antitrust Review
 10. General
-