

510(K) SUMMARY
PREPARED FEBRUARY 12, 2004

MANUFACTURER SUBMITTING 510(K) NOTIFICATION:

Mediware Dallas Service Center
Mediware Information Systems, Inc.
4545 Fuller Dr., Ste. 320
Irving, TX 75038

CONTACT PERSON:

Marsha Senter
Director, Regulatory Affairs/Quality Assurance
4545 Fuller Dr., Ste. 320
Irving, TX 75038
972.536.2955
972.536.2990 (Fax)

DEVICE NAME:

Proprietary Name: HCLL®, Release 2.6
Common Name: Software, Blood Bank, Stand Alone
Classification Name: Unclassified

PREDICATE DEVICES:

Cerner PathNet Blood Bank Transfusion, version 306
SoftBank® II, version 21.1
LifeTrak®/Lab, version 2.03

DEVICE DESCRIPTION:

The HCLL System is a stand-alone blood bank software device that aids in the management of single, multi-site, or centralized transfusion services.

INTENDED USE:

HCLL addresses all phases of transfusion services activities, and helps transfusion service personnel:

- Maintain a reliable patient database
- Manually and electronically register, admit and transfer patients
- Manually and electronically receive patient orders
- Manipulate and label products

510(k) SUMMARY
PREPARED FEBRUARY 12, 2004

- Print ABC Codabar and ISBT 128 blood labels
- Perform electronic crossmatch
- Issue blood products under normal and emergency conditions
- Receive inventory
- Ship specimens, reagents, derivatives and blood products
- Test patient and units specimens
- Interface to blood bank testing instrument

COMPARISON OF TECHNOLOGICAL CHARACTERISTICS TO PREDICATE DEVICE:

The technological characteristics are similar in that the systems are fully scalable to enterprise level and that the RDBMS, operating system and hardware are substantially equivalent to the predicate devices.

SAFETY AND EFFECTIVENESS DATA:

The HCLL System was developed using an established procedure for software development.

The assessment of the nonclinical (alpha) testing is that the HCLL System design input requirements had been met. The design validation process demonstrated that the HCLL System meets the requirements for the intended use.

Clinical/user site (beta) testing was conducted to validate the HCLL software. Testing was conducted at three different clinical facilities. The results of the beta testing demonstrated that the HCLL software met the required specifications and functioned as expected.

CONCLUSION:

The conclusions drawn for the nonclinical and clinical tests demonstrate HCLL, Release 2.6, is substantially equivalent to the predicate device(s) when utilized within its intended use.