

## ↓ eterans Health Administration Warning Syste Published by VA Central Office

April 28, 2004

Item:	Implantable Cardioverter Defibrillator (ICD), models <u>Micro-Jewell II 7223Cx</u> and <u>GEM DR 7271</u> manufactured by Medtronic, Inc. and implanted in <u>1997</u> <u>and 1998</u> . Of 6268 units originally implanted, an estimated 1800 remain in use worldwide.
Specific Incident:	During a cardiac event requiring cardioversion or defibrillation, internal capacitors may take longer to charge and can cause a delay or non-delivery of appropriate shock therapy.
Action:	<ol> <li>Immediately check patient records in CPRS to identify patients with the ICD implants described in this notification.</li> </ol>
	<ol> <li>Contact your local Medtronic sales representative (or Medtronic Technical Services, listed in the Contact section of this notification if you cannot reach the local representative) for assistance with determining whether the identified implant requires follow-up action as identified by Medtronic.</li> </ol>
	<ol> <li>If follow-up action is necessary, see the attached letter from Medtronic describing required actions.</li> </ol>
Source:	Manufacturer and FDA
Contact:	Medtronic Technical Services: 1-800-723-4636
	Edmund Keung, MD Director, VA Western Pacemaker Surveillance Center and National ICD Surveillance Center: (415) 750-2077
	Ronald Jones Pacemaker Engineer, VA Eastern Pacemaker Surveillance Center: (202) 745-8504

Letter from Medtronic (PDF Format)