

K071148

DEC 22 2008

510(k) Summary
Prepared August 27, 2008
Revised December 18, 2008

Sponsor: Siemens Medical Solutions USA, Inc.,
Ultrasound Division
1230 Shorebird Way
P.O. Box 7393
Mountain View, California 94039-7393

Contact Person: Sheila W. Pickering Ph.D.
Telephone: (650) 943 7187
Fax: (650) 943 7053

Submission Date: April 18, 2008

Device Name: Acuson S2000 ABVS Ultrasound System

Common Name: Diagnostic Ultrasound System with Accessories

Classification:
Regulatory Class: II
Review Category: Tier II
Classification Panel: Radiology

Ultrasonic Pulsed Doppler Imaging System	FR # 892.1550	Product Code 90-IYN
Ultrasonic Pulsed Echo Imaging System	FR # 892.1560	Product Code 90-IYO
Diagnostic Ultrasound Transducer	FR # 892.1570	Product Code 90-ITX

A. Legally Marketed Predicate Devices

The Acuson S2000 ABVS Ultrasound system is substantially equivalent to the Acuson S2000 ultrasound system (K072786) and the U-Systems ultrasound system.

B. Device Description:

The Acuson S2000 ABVS Ultrasound System has been designed to meet the following product safety standards:

- UL 60601-1, Safety Requirements for Medical Equipment
- IEC 60601-2-37 Diagnostic Ultrasound Safety Standards
- CSA C22.2 No. 601-1, Safety Requirements for Medical Equipment
- **AIUM/NEMA UD-3, 1998 Standard for Real Time Display of Thermal and Mechanical Acoustic Output Indices on Diagnostic Ultrasound Equipment**
- **AIUM/NEMA UD-2, 1998 Acoustic Output Measurement Standard for Diagnostic Ultrasound**
- 93/42/EEC Medical Devices Directive
- **Safety and EMC Requirements for Medical Equipment**
 - EN/IEC 60601-1
 - EN/IEC 60601-1-1
 - EN/IEC 60601-1-2
- IEC 1157 Declaration of Acoustic Power
- ISO 10993-1 Biocompatibility

C. Intended Use

The Modified S2000, the S2000 ABVS Ultrasound System, is intended for the following applications: General Radiology, Abdominal, Fetal, Small Parts, Transcranial, OB/GYN, Cardiac, Pelvic, Neonatal/Adult Cephalic, Pediatric, Urology, Vascular, Musculoskeletal, Superficial Musculoskeletal, and Peripheral Vascular applications. The system supports the transducers listed in the Ultrasound Indications for Use tables, including the 14L5BV for B-mode imaging of a patient's breast using an optional automatic scanning function. The device is not intended to be used as a replacement for screening mammography.

The system also provides for the measurement of anatomical structures and analysis as provided in the original S2000 Ultrasound System, which provides information that is used by medical health care professionals for clinical diagnosis purposes.

D. Substantial Equivalence

The submission device is substantially equivalent to the predicate devices with regard to both intended use and technological characteristics.

E. Performance Data

The S2000 modifications are verified and validated according to the company's design control process.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Sheila W. Pickering, Ph.D.
Senior Director of Regulatory Affairs
Siemens Medical Solutions USA, Inc.
Ultrasound Division
1230 Shorebird Way
MOUNTAIN VIEW WAY CA 94039-7393

DEC 22 2008

Re: K081148
Trade/Device Name: Acuson S2000 ABVS Ultrasound System
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: II
Product Code: IYN, IYO, and ITX
Dated: November 21, 2008
Received: November 26, 2008

Dear Dr. Pickering:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the Acuson S2000 ABVS Ultrasound System, as described in your premarket notification:

Transducer Model Number

CW2 Probe
CW5 Probe
EC9-4 Curved Array
9L4 Linear Array
14L5 Multi-D Array
4P1 Phased Array
6C2 Curved Array
4C1 Curved Array
4V1 Phased Array

10V4 Phased Array
14L5 SP Linear Array
7CF2 Curved Array Mechanical 3D
9EVF4 Curved Array
V5Ms Multiplane TEE
18L6HD Linear Array
8V3 Phased Array

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>

If you have any questions regarding the content of this letter, please contact Lauren Hefner at (240) 276-3666.

Sincerely yours,



Joyce M. Whang, Ph.D.
Acting Director, Division of Reproductive,
Abdominal, and Radiological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure(s)

Indications for Use

510(k) Number (if known): K081148

Device Name: Acuson S2000 ABVS Ultrasound System

Indications For Use:


The Modified S2000, the S2000 ABVS Ultrasound System, is intended for the following applications: General Radiology, Abdominal, Fetal, Small Parts, Transcranial, OB/GYN, Cardiac, Pelvic, Neonatal/Adult Cephalic, Pediatric, Urology, Vascular, Musculoskeletal, Superficial Musculoskeletal, and Peripheral Vascular applications. The system supports the transducers listed in the Ultrasound Indications for Use tables, including the 14L5BV for B-mode imaging of a patient's breast using an optional automatic scanning function. The device is not intended to be used as a replacement for screening mammography.

The system also provides for the measurement of anatomical structures and analysis as provided in the original S2000 Ultrasound System, which provides information that is used by medical health care professionals for clinical diagnosis purposes.

Prescription Use X AND/OR
(Part 21 CFR 801 Subpart D)

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)



(Division Sign-Off) _____ Concurrence of CDRH, Office of Device Evaluation (ODE)
Division of Reproductive, Abdominal and
Radiological Devices
510(k) Number K08 1148

Prescription Use (Per 21 CFR 801.10)
Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

Device Name:

CW2 Probe for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal					P					
Abdominal					P					
Intraoperative (Note 9)					P					
Intraoperative Neurological										
Pediatric					P					
Small Organ (Note 1)					P					
Neonatal Cephalic					P					
Adult Cephalic					P					
Cardiac					P					
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel					P					
Laparoscopic										
Musculo-skeletal Conventional					P					
Musculo-skeletal Superficial					P					
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063803; E = added under Appendix E

Additional Comments:

Note 1 For example: breast, testes, thyroid, penis, prostate, etc.

Note 9 For example: vascular, abdominal

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and Radiological Devices

510(k) Number

K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

K081148

Device Name:

CW5 Probe for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal					P					
Abdominal					P					
Intraoperative (Note 9)					P					
Intraoperative Neurological										
Pediatric					P					
Small Organ (Note 1)					P					
Neonatal Cephalic					P					
Adult Cephalic					P					
Cardiac					P					
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel					P					
Laparoscopic										
Musculo-skeletal Conventional					P					
Musculo-skeletal Superficial					P					
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063803; E = added under Appendix E

Additional Comments:

Note 1 For example: breast, testes, thyroid, penis, prostate, etc.

Note 9 For example: vascular, abdominal

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and Radiological Devices

510(k) Number

K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known): K081148

Device Name:

EC9-4 Curved Array Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Abdominal		P	P	P		P	P		BMDC	Note 2,3,4,5,6,,7,8,10, 11
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ (Note 1)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Neonatal Cephalic		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Transvaginal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										


N = new indication; P = previously cleared by FDA K# 063803; E = added under Appendix E

Additional Comments:

Note 1 For example: breast, testes, thyroid, penis, prostate, etc.

Note 2 Ensemble tissue harmonic imaging

Note 3 SieClear multi-view spatial compounding



 (Division Sign-Off)
 Division of Reproductive, Abdominal and
 Radiological Devices
 510(k) Number K081148

- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear multi-view spatial compounding
- Note 14 eSie™ Touch elasticity imaging/FTI

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)



(Division Sign-Off)

Division of Reproductive, Abdominal and
Radiological Devices

510(k) Number

K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known): **K081148**

Device Name:

9LA Linear Array Transducer for use with ACUSON S2000 ABYS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Small Organ (Note 1)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Neonatal Cephalic		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Laparoscopic										
Musculo-skeletal Conventional		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Musculo-skeletal Superficial		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11, 14
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085; E = added under Appendix E

Additional Comments:


- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology

- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear multi-view spatial compounding
- Note 14 eSie™ Touch elasticity imaging/FTI

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)



(Division Sign-Off)

Division of Reproductive, Abdominal and
Radiological Devices

510(k) Number

K081148

Ultrasound Indications for Use Form

510 (k) Number (if known): **K081148**

Device Name:

14L5 Multi-D Array Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal										
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ (Note 1)		N								Note 2,3,4,5,7,8,10, 11, 14
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085; E = added under Appendix E

Additional Comments:

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear multi-view spatial compounding

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)



(Division Sign-Off)

Division of Reproductive, Abdominal and
Radiological Devices

510(k) Number K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known): **K081148**

Device Name:

4P1 Phased Array Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Abdominal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic										
Adult Cephalic		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Cardiac		P	P	P	P	P	P		BMDC	Note 2,3,4,5,6,7,8,10
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063803; E = added under Appendix E


Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)



(Division Sign-Off)

Division of Reproductive, Abdominal and
Radiological Devices

510(k) Number

K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known): K081148

Device Name:

6C2 Curved Array Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Abdominal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Small Organ										
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085; E = added under Appendix E

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear multi-view spatial compounding

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)
 Concurrence of CDRH, Office of Device Evaluation (ODE) Prescription Use (Per 21 CFR 801.109)

Diagnostic Ultrasound Indications for Use Form

[Signature]
 (Division Sign-Off)

Division of Reproductive, Abdominal and
 Radiological Devices

510(k) Number

K081148

510 (k) Number (if known):

Device Name:

K081148

4C1 Curved Array Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Abdominal		P	P	P		P	P		BMDC	Note 2,3,4,5,6,7,8, 10, 11,
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085; E = added under Appendix E

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear multi-view spatial compounding

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

[Signature]
(Division Sign-Off)

Division of Reproductive, Abdominal and Radiological Devices

510(k) Number

K081148

Device Name:

4V1 Phased Array Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:


Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10
Abdominal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085; E = added under Appendix E

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear multi-view spatial compounding

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)
 Concurrence of CDRH, Office of Device Evaluation (ODE) Prescription Use (Per 21 CFR 801.109)


 (Division Sign-Off)
 Division of Reproductive, Abdominal and Radiological Devices
 510(k) Number K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known): K081148

Device Name:

10V4 Phased Array Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:


Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Abdominal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Small Organ										
Neonatal Cephalic		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Adult Cephalic										
Cardiac		P	P	P	P	P	P		BMDC	Note 3,4
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085; E = added under Appendix E

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)
 Concurrence of CDRH, Office of Device Evaluation (ODE) Prescription Use (Per 21 CFR 801.109)



 (Division Sign-Off)
 Division of Reproductive, Abdominal and
 Radiological Devices
 510(k) Number K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

K081148

Device Name:

**14L5 SP Linear Array Transducer for use with ACUSON S2000 ABVS
Ultrasound System**

Indications For Use:

Diagnostic imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal										
Abdominal										
Intraoperative (Note 9)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10
Intraoperative Neurological		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11
Pediatric										
Small Organ (Note 1)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Transesophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P		P	P		BMDC	Note 2,3,4,5,6, 7,8,10, 11,14
Laparoscopic										
Musculo-skeletal Conventional		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085; E = added under Appendix E

Additional Comments:

- Note 1 For example: breast, testes, thyroid, penis, prostate, etc.
- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 9 For example: vascular, abdominal
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear multi-view spatial compounding

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)



(Division Sign-Off)

Division of Reproductive, Abdominal and
Radiological Devices

510(k) Number

K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

K081148

Device Name:

7CF2 Curved array mechanical 3D transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10,11,13
Abdominal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10,11,13
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										


N = new indication; P = previously cleared by FDA K# 063803; E = added under Appendix E

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology
- Note 11 Advanced Sieclear multi-view spatial compounding
- Note 13 STIC

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)
Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)



(Division Sign-Off)

Division of Reproductive, Abdominal and
Radiological Devices

510(k) Number K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known): **K081148**

Device Name:

9EVF4 Curved Array Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10,11
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10,11
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10,11
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063803; E = added under Appendix E

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)



(Division Sign-Off)

Division of Reproductive, Abdominal and
Radiological Devices

510(k) Number K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known):

K081148

Device Name:

V5Ms Multiplane TEE Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal										
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ										
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal		P	P	P	P	P	P		BMDC	
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063803; E = added under Appendix E

Additional Comments: n/a

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use (Per 21 CFR 801.109)

[Handwritten Signature]

(Division Sign-Off)
 Division of Reproductive, Abdominal and Radiological Devices
 510(k) Number *K081148*

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known): **K081148**

Device Name:

18L6HD Linear Array Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
Ophthalmic										
Fetal										
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric										
Small Organ (Note 1)		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Neonatal Cephalic										
Adult Cephalic										
Cardiac										
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Laparoscopic										
Musculo-skeletal Conventional		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Musculo-skeletal Superficial		P	P	P		P	P		BMDC	Note 2,3,4,5,7,8,10, 11,14
Other (specify)										

N = new indication; P = previously cleared by FDA K# 063085; E = added under Appendix E

Additional Comments:

Note 1 For example: breast, testes, thyroid, penis, prostate, etc.

Note 2 Ensemble tissue harmonic imaging

Note 3 SieClear multi-view spatial compounding

Note 4 Tissue Equalization Technology

Note 5 3-Scape real-time 3D imaging


Note 7 B&W SieScape panoramic imaging

Note 8 Power SieScape panoramic imaging

Note 10 Clarify VE vascular enhancement technology
Note 11 Advanced Sieclear multi-view spatial compounding
Note 14 eSie™ Touch elasticity imaging/FTI

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE) Prescription Use (Per 21 CFR
801.109)



(Division Sign-Off)

Division of Reproductive, Abdominal and
Radiological Devices

510(k) Number

K081148

Diagnostic Ultrasound Indications for Use Form

510 (k) Number (if known): *K081148*

Device Name:

8V3 Phased Array Transducer for use with ACUSON S2000 ABVS Ultrasound System

Intended Use:

Ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation									
	A	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Color Velocity Imaging	Combined (Specify)	Other (Specify)
<i>Ophthalmic</i>										
Fetal		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Abdominal										
Intraoperative Abdominal										
Intraoperative Neurological										
Pediatric		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Small Organ										
Neonatal Cephalic		P	P	P	P	P	P		BMDC	Note 2,3,4,5,7,8,10
Adult Cephalic										
Cardiac		P	P	P	P	P	P		BMDC	Note 3,4,6
Trans-esophageal										
Transrectal										
Transvaginal										
Transurethral										
Intravascular										
Peripheral vessel										
Laparoscopic										
Musculo-skeletal Conventional										
Musculo-skeletal Superficial										
Other (specify) Neonatal Cardiac		P	P	P	P	P	P		BMDC	Note 3,4,6

N = new indication; P = previously cleared by FDA K# 063085; E = added under Appendix E

Additional Comments:

- Note 2 Ensemble tissue harmonic imaging
- Note 3 SieClear multi-view spatial compounding
- Note 4 Tissue Equalization Technology
- Note 5 3-Scape real-time 3D imaging
- Note 6 Cadence contrast agent imaging

- Note 7 B&W SieScape panoramic imaging
- Note 8 Power SieScape panoramic imaging
- Note 10 Clarify VE vascular enhancement technology

