

DEC 03 2008

SECTION 5
510(k) SUMMARY

510(k) SUMMARY

1. Submitter:

Boston Scientific Corporation
100 Boston Scientific Way
Marlborough, MA 01752
Telephone: 508-683-4454
Fax: 508-683-5939

Contact: Marybeth Gamber
Regulatory Affairs Manager
Date Prepared: June 18, 2008

2. Device:

Trade Name: 2-Port Through-The-PEG (TTP) Jejunostomy Tube Kit
Classification Name: Tube, Gastrointestinal (and Accessories)
Regulation Number: 876.5980
Product Code: KNT
Classification: Class II

3. Predicate Device:

Through-The-PEG (TTP) Jejunostomy Tube Kit K072476
Manufactured by Boston Scientific, Inc.

4. Device Description:

The 2-Port TTP Jejunostomy Tube consists of a two-port device designed to be placed through a Boston Scientific gastrostomy tube to provide enteral access for decompression and delivery of nutrition and/or medication. The 2-Port TTP Jejunostomy Tube is available with a pigtail tip and in two lengths, 80cm and 105cm. It may be placed by either a pullwire (pull) or guidewire (push) technique. The proposed device is available within a kit which contains the following: a stiffening cannula with female-to-female luer fitting, a guidewire, lubricating jelly, gauze, a female-to-female luer adapter, port caps, and a retention ring.

5. Intended Use:

The 2-Port TTP Jejunostomy Tube Kit is intended to provide enteral access for decompression and delivery of nutrition and/or medication.

6. Technological Characteristics:

The proposed 2-Port TTP Jejunostomy Tube Kit is similar in design, materials, and manufacturing processes to the predicate 3-Port TTP Jejunostomy Tube Kit (K072476).

7. Performance Data:

In-vitro testing has been performed and all components, subassemblies, and/or full devices met the required specifications for the completed tests.

8. Conclusion:

Boston Scientific Corporation has demonstrated that the proposed 2-Port Through-The-PEG (TTP) Jejunostomy Tube Kit is substantially equivalent to Boston Scientific Corporation's currently marketed 3-Port Through-The-PEG (TTP) Jejunostomy Tube Kit (K072476).

000012



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Ms. Marybeth Gamber
Manager, Regulatory Affairs
Boston Scientific Corporation
100 Boston Scientific Way
Endoscopy Division, M11
MARLBOROUGH MA 01752

DEC 03 2008

Re: K081739

Trade/Device Name: EndoVive Two-Port Through The PEG (TTP) Jejunal Feeding Tube
(J-Tube) Kit

Regulation Number: 21 CFR 876.5980

Regulation Name: Gastrointestinal tube and accessories

Regulatory Class: II

Product Code: KNT

Dated: November 25, 2008

Received: November 26, 2008

Dear Ms. Gamber:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. However, you are responsible to determine that the medical devices you use as components in the kit have either been determined as substantially equivalent under the premarket notification process (Section 510(k) of the act), or were legally on the market prior to May 28, 1976, the enactment date of the Medical Device Amendments. Please note: If you purchase your device components in bulk (i.e., unfinished) and further process (e.g., sterilize) you must submit a new 510(k) before including these components in your kit. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, and labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on the labeling regulation, please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll free number (800) 638-2041 or (240) 276-3150, or at its Internet address <http://www.fda.gov/cdrh.dsma/dsmamain.html>.

Sincerely yours,



Joyce M. Whang, Ph.D.

Acting Director, Division of Reproductive,
Abdominal, and Radiological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

