



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
WASHINGTON, D.C. 20555-0001

April 7, 2006

SECRETARY

COMMISSION VOTING RECORD

DECISION ITEM: SECY-06-0055

TITLE: REPORT TO CONGRESS ON ABNORMAL  
OCCURRENCES: FISCAL YEAR 2005

The Commission (with all Commissioners agreeing) approved the subject paper as recorded in the Staff Requirements Memorandum (SRM) of April 7, 2006.

This Record contains a summary of voting on this matter together with the individual vote sheets, views and comments of the Commission.

A handwritten signature in black ink, appearing to read "Annette Vietti-Cook".

Annette L. Vietti-Cook  
Secretary of the Commission

Attachments:

1. Voting Summary
2. Commissioner Vote Sheets

cc: Chairman Diaz  
Commissioner McGaffigan  
Commissioner Merrifield  
Commissioner Jaczko  
Commissioner Lyons  
OGC  
EDO  
PDR

SECY NOTE: THIS VOTING RECORD WILL BE RELEASED TO THE PUBLIC 5 DAYS  
AFTER DISPATCH OF THE REPORT TO CONGRESS.

VOTING SUMMARY - SECY-06-0055

RECORDED VOTES

	APRVD	DISAPRVD	ABSTAIN	NOT PARTICIP	COMMENTS	DATE
CHRM. DIAZ	X				X	3/21/06
COMR. McGAFFIGAN	X				X	3/22/06
COMR. MERRIFIELD	X				X	3/27/06
COMR. JACZKO	X					3/31/06
COMR. LYONS	X					3/20/06

COMMENT RESOLUTION

In their vote sheets, all Commissioners approved the staff's recommendation and some provided additional comments. Subsequently, the comments of the Commission were incorporated into the guidance to staff as reflected in the SRM issued on April 7, 2006.

NOTATION VOTE  
RESPONSE SHEET

TO: Annette Vietti-Cook, Secretary  
FROM: CHAIRMAN DIAZ  
SUBJECT: **SECY-06-0055 - REPORT TO CONGRESS ON  
ABNORMAL OCCURRENCES: FISCAL YEAR  
2005**

Approved  *[Signature]* Disapproved \_\_\_\_\_ Abstain \_\_\_\_\_  
Not Participating \_\_\_\_\_

COMMENTS:

Approved with one minor edit as attached.

*[Signature]*  
\_\_\_\_\_  
SIGNATURE  
  
3.21.06  
\_\_\_\_\_  
DATE

Entered on "STARS" Yes  No \_\_\_\_\_

## AGREEMENT STATE LICENSEES

During this reporting period, six events at Agreement State-licensed facilities were significant enough to be reported as AOs based on the criteria in Appendix A to this report.

AS 05-01      Iridium-192 Brachytherapy Seed Medical Event at LDS Hospital in Salt Lake City, Utah

Criterion IV, "For Medical Licensees," of Appendix A to this report states, in part, that a medical event that results in a dose that is (1) equal to or greater than 1 Gy (100 rads) to a major portion of the bone marrow, to the lens of the eye, or the gonads, or (2) equal to or greater than 10 Gy (1,000 rads) to any other organ; and represents a prescribed dose or dosage that is delivered to the wrong treatment site, will be considered for reporting as an AO.

Date and Place — October 26, 2004; LDS Hospital; Salt Lake City, Utah

Nature and Probable Consequences — A patient received 27.56 Gy (2,756 rads) instead of the prescribed 5 Gy (500 rads) during a high dose-rate (HDR) treatment for larynx cancer. The event involved an iridium-192 (Ir-192) source with an activity of 244.2 GBq (6.6 Ci). The error was caused by the use of the diameter instead of the radius of a circular tool to mark the treatment site in a computer software program. As a result, the area treated was 2 centimeters (cm) away from the intended treatment site. The error was discovered before the third fraction. The prescribing physician stopped the treatment until dosimetry information was completed. The licensee notified the patient and the patient's referring physician of the event. The licensee determined that the impact of the additional dose is probable acute radiation effects and possible late or chronic toxicities. X

Cause(s) — This event was caused by human error. The incorrect size button corresponding to the circle tool was used, which caused the diameter instead of the radius to be used in the dosing plan. This caused the incorrect dose to be administered to the incorrect location.

### Actions Taken to Prevent Recurrence

Licensee — The licensee suggested that the software manufacturer print the word "RADIUS" on the "size" button located adjacent to the circle tool. To date, the manufacturer has not responded to this issue. The licensee will measure the distance on the brachytherapy device's hard copy output with a ruler to confirm that the distance is entered correctly. The licensee also modified the HDR dose check program so that, in addition to confirming the doses to coordinates entered into the device's input, user specified point coordinates may be manually entered into the check program and compared to what is calculated.

State Agency — The Utah Division of Radiation Control investigated the event on November 3, 2004 and approved the corrective actions that the licensee implemented to prevent the recurrence.

NOTATION VOTE

RESPONSE SHEET

TO: Annette Vietti-Cook, Secretary  
FROM: COMMISSIONER MCGAFFIGAN  
SUBJECT: **SECY-06-0055 - REPORT TO CONGRESS ON  
ABNORMAL OCCURRENCES: FISCAL YEAR  
2005**

Approved  Disapproved  Abstain

Not Participating

COMMENTS: *Minor edits are attached.*

*Edward J. McGaffigan Jr.*  
\_\_\_\_\_  
SIGNATURE

*March 22, 2006*  
\_\_\_\_\_  
DATE

Entered on "STARS" Yes  No



**ACRONYMS and ABBREVIATIONS**

AEA	Atomic Energy Act		
AO	abnormal occurrence		
Bq	becquerel		
CFR	<i>Code of Federal Regulations</i>		
Ci	curie		
cm	centimeter	<i>cGy</i>	<i>centigray</i>
Cs-137	cesium-137		
FR	<i>Federal Register</i>		
FY	Fiscal Year		
GBq	gigabecquerel		
Gy	gray		
HDR	high dose-rate		
I-123	iodine-123		
I-125	iodine-125		
I-131	iodine-131		
Ir-192	iridium-192		
in	inch		
MBq	megabecquerel		
$\mu$ Ci	microcurie		
mCi	millicurie		
NRC	U.S. Nuclear Regulatory Commission		
SNM	special nuclear material		
Sr-90	strontium-90		
Sv	sievert		
Tc-99m	technetium-99 metastable		
TEDE	total effective dose equivalent		
Y-90	yttrium-90		

05-03 Medical Event at St. Joseph Regional Medical Center in South Bend, Indiana

Criterion IV, "For Medical Licensees," of Appendix A to this report states, in part, that a medical event that results in a dose that is (1) equal to or greater than 1 Gy (100 rads) to a major portion of the bone marrow, to the lens of the eye, or to the gonads or (2) equal to or greater than 10 Gy (1,000 rads) to any other organ; and represents a prescribed dose or dosage that is delivered to the wrong treatment site will be considered for reporting as an AO.

Date and Place — Between January 26 and March 22, 2004 (reported March 25, 2005 due to a misinterpretation of reporting requirements by the licensee), South Bend, Indiana

Nature and Probable Consequences — The licensee reported in March and April 2005, that between January 26 and March 22, 2004, three patients received unintended radiation doses to the skin of their thighs from cesium-137 brachytherapy sources. The vaginal applicator used for the treatments was loaded with incorrectly sized cesium-137 sources, which ~~had the ability to migrate~~ <sup>migrated</sup> from the intended treatment position through the placement spring when the patient moved to a more up-right position. As a result of the sources moving, the patient's inner thighs received unintended doses of radiation. Approximately two weeks after treatment, the patients developed skin lesions on their inner thighs. The licensee determined that these patients received unintended doses to a small area of the skin on the upper thigh of approximately 2000, 1500, and 2000 cGy (rad), respectively. Based on clinical observations, the licensee determined that all patients received the respective prescribed doses to the intended treatment areas. The referring physician and patients were notified of the event. The licensee referred the patients to other institutions and care providers for specialized followup wound care to treat the recurring skin ulcerations. The NRC retained a medical consultant during the inspection associated with the event. The long-term health effects on the patients, as a result of the unintended doses, is unknown.

Cause(s) — The causes of these events were improper source selection, inadequate manufacturer instructions, inadequate management oversight, and inadequate procedures.

Actions Taken to Prevent Recurrence — Corrective actions taken by the licensee involved modifying the applicator by using different hardware to hold the sources in place, revising their procedures, and retraining the staff on the new procedures.

This event is closed for the purpose of this report.



NOTATION VOTE  
RESPONSE SHEET

TO: Annette Vietti-Cook, Secretary  
FROM: COMMISSIONER MERRIFIELD  
SUBJECT: **SECY-06-0055 - REPORT TO CONGRESS ON  
ABNORMAL OCCURRENCES: FISCAL YEAR  
2005**

Approved  Disapproved  Abstain   
Not Participating

COMMENTS:

*See attached edits.*


  
\_\_\_\_\_  
SIGNATURE

*3/27/06*  
\_\_\_\_\_  
DATE

Entered on "STARS" Yes  No

Comments from Commissioner Merrifield on SECY-06-0055:

I approve the staff recommendation to submit the abnormal occurrences report to Congress. I compliment the staff for their efforts in preparing this report. This is an annual report submitted to Congress. In past years, I provided significant comments on previous versions of this report because I did not believe the staff effort properly presented a Commission level report to Congress. This year, it is obvious that staff made a considerable effort to properly integrate past Commission guidance on previous reports. Keep up the good work.



3/27/06

NOTATION VOTE

RESPONSE SHEET

TO: Annette Vietti-Cook, Secretary  
FROM: COMMISSIONER JACZKO  
SUBJECT: **SECY-06-0055 - REPORT TO CONGRESS ON  
ABNORMAL OCCURRENCES: FISCAL YEAR  
2005**

Approved  Disapproved  Abstain

Not Participating

COMMENTS:

  
\_\_\_\_\_  
SIGNATURE

3/31/06  
\_\_\_\_\_  
DATE

Entered on "STARS" Yes  No


NOTATION VOTE  
RESPONSE SHEET

TO: Annette Vietti-Cook, Secretary  
FROM: COMMISSIONER LYONS  
SUBJECT: **SECY-06-0055 - REPORT TO CONGRESS ON  
ABNORMAL OCCURRENCES: FISCAL YEAR  
2005**

Approved  Disapproved \_\_\_\_\_ Abstain \_\_\_\_\_

Not Participating \_\_\_\_\_

COMMENTS:

  
\_\_\_\_\_  
SIGNATURE  
  
3/20/06  
\_\_\_\_\_  
DATE

Entered on "STARS" Yes  No \_\_\_\_\_