



National Kidney Disease Education Program

Laboratory Working Group

Recommendations for Improving Serum Creatinine Measurement: An Update

Gary L. Myers, PhD



U.S. Department of Health
and Human Services



National Institutes of Health



National Institute of Diabetes
and Digestive and Kidney Diseases

NKDEP Lab Working Group

Key Recommendations for IVD Manufacturers & Clinical Laboratories

Interim recommendations until a revised GFR estimating equation is available

- **Implement the MDRD equation now for creatinine methods that are not re-calibrated to be traceable to IDMS**
- **For re-calibrated creatinine methods:**
 - ▶ **Add 0.1 mg/dL (8.8 $\mu\text{mol/L}$) to the creatinine value in the MDRD equation to be similar to the bias that existed in the MDRD study lab**

NKDEP Lab Working Group

Key Recommendations for IVD Manufacturers & Clinical Laboratories

Long-term recommendations

- **IVD manufacturers and clinical labs should coordinate introduction of re-calibration to coincide with a revised GFR estimating equation based on creatinine values traceable to IDMS**
 - **NKDEP will have a revised equation in 2005**
- **Clinical laboratories should report estimated GFR as >60 ml/min/1.73m² when values are above 60**
- **Report serum creatinine values as mg/dL to two decimal places. Values reported as $\mu\text{mol/L}$ should be reported as the nearest whole number.**

NKDEP Lab Working Group

Key Recommendations

- **IVD manufacturers should target optimal creatinine method performance at 1.0 mg/dL (88 $\mu\text{mol/L}$), and ensure comparable trueness and precision throughout the AMR**
- **Precision at lower creatinine concentrations needs to be improved to allow acceptable GFR_{est} at values $>60 \text{ mL/min/1.73m}^2$, and for pediatric populations**
- **After re-calibration to IDMS, a realistic total error goal for creatinine is a maximum 10% increase in the relative error of the estimated GFR**
 - **Typical values: bias $<5\%$, and CV $<8\%$, at creatinine $\geq 1.0 \text{ mg/dL}$ (88 $\mu\text{mol/L}$)**
- **IVD manufacturers must address analytical non-specificity in current routine methods**

NKDEP Lab Working Group

Key Recommendations for NKDEP in Collaboration with other Professional Organizations

- ▶ **Identify the impact on clinical decision criteria that may result from re-calibration of serum creatinine to be traceable to IDMS**
- ▶ **Develop a replacement for the MDRD eqn. that uses serum creatinine measurement traceable to IDMS**
- ▶ **Coordinate introduction of method traceability to IDMS with the appropriate GFR estimating equation**
- ▶ **Develop guidelines to implement appropriate GFR estimating equations for re-calibrated creatinine and to communicate the resultant changes in clinical interpretation of serum creatinine**

NKDEP Lab Working Group

Key Recommendations for NKDEP in Collaboration with other Professional Organizations

- **Communicate the clinical issues associated with re-calibrated serum creatinine**
 - ▶ **Reference interval change**
 - ▶ **Creatinine clearance values and reference interval change**
 - ▶ **Pharmacy impact on drug dose adjustment**
 - ▶ **Pediatric GFR estimating equations**
- **Coordinate with PT/EQAS providers to ensure appropriate grading adjustments are made during transition to IDMS traceability**
- **Establish a small group of reference labs that can perform high throughput reference measurement procedures**
- **Implement educational programs on the proper use of the MDRD equation to assess CKD risk.**

NKDEP Lab Working Group

Key Recommendations for National Metrology Institutes, reference laboratories, and members of JCTLM

- **Provide tools to assist IVD manufacturers to reduce analytical bias, since many routine methods can meet or exceed the imprecision goal of <8%**
 - ▶ **By the end of 2005, develop commutable reference materials for serum creatinine (NIST SRM 967 is expected to fulfill this need when available)**
 - ▶ **By the end of 2005, make available LC-IDMS reference method. Additional reference laboratories will be needed to meet the anticipated demand**

NKDEP Lab Working Group

Key Recommendations for PT and EQAS Providers

- **Introduce a regularly recurring proficiency program that uses commutable serum materials with target values traceable to IDMS reference measurement procedures**
 - ▶ **Permits an ongoing assessment of routine method performance and the evaluation of accuracy transfer processes used by manufacturers**
 - ▶ **The CAP Creatinine Accuracy Calibration Verification/Linearity Survey, LN24, has these attributes**