PUBLIC SUBMISSION

As of: September 02, 2008 Received date: Not specified

Status: Pending_Post Tracking No. 806eccb5

Comments Due: September 08, 2008

Submission Type: Web

Docket: MSHA-2008-0008

Safety Standards Regarding the Recommendations of the Technical Study Panel on the Utilization of Belt Air and the Composition and Fire Retardant Properties of Belt Materials in Underground Coal Mining

Comment On: MSHA-2008-0008-0001

Safety Standards Regarding the Recommendations of the Technical Study Panel on the Utilization of Belt Air and the Composition and Fire Retardant Properties of Belt Materials in Underground Coal Mining

Document: MSHA-2008-0008-DRAFT-0005

Comment from Bruce Levinson, The Center for Regulatory Effectiveness

Submitter Information

Name: Bruce Levinson

Address:

1601 Connecticut Avenue, N.W.

Suite 500

Washington, DC, 20009 **Email:** levinson@thecre.com

Phone: 202-265-2383

Organization: The Center for Regulatory Effectiveness

General Comment

In response to MSHA????????s request at the Birmingham public hearing on August 28, attached please find a copy of test standard recommended by CRE in our testimony for measuring toxic gases emitted by heated conveyor belts, Boeing Specification Support Standard (BSS) 7239, ????????Test Method For Toxic Gas Generation By Materials On Combustion.????????

Attachments

MSHA-2008-0008-DRAFT-0005.1: Comment from Bruce Levinson, The Center for Regulatory Effectiveness

AB59-COMM-3

This standard covers the test methods for determining the toxic gas generating characteristics of aircraft materials, using the National Bureau of Standards (NBS) Smoke Density Chamber for sample combustion. Gases specifically covered are carbon monoxide (CO), hydrogen cyanide (BCN), nitrogen oxides (NO_X = NO + NO₂), and sulfur dioxide (SO₂). These procedures may be used for other toxic gases if so specified. This test may be conducted simultaneously with the NBS smoke generation test described in Smoke Generation by Naterials on Combustion (BSS 7238).

1.1 PDRPOSE

The purpose of this specification is to determine the levels of several toxic gases (in ppm) released by a sample during combustion under specified thermal exposure

1.2 CLASSIFICATION

2

3

Not applicable to this specification.

APPLICABLE DOCUMENTS

The current issue of the following references shall be a port of this standard to the extent herein indicated:

AMIHCO-MBS Smoke Density Chamber, Cat. 4-5800B, Instruction Manual 941-B

MBS Technical Note 708, "Interlaboratory Evaluation of Smoke Density Chamber"

Dragerwerk Ag Lubeck, "Detector Tube Handbook," Second Edition, Available from Suppliers of Drager Equipment

BSS 7238 Test Hethod for Smoke Generation by Materials on Combustion

BSS 7242 Determination of the Concentration of Cyanide, Chloride, and Fluoride Ions in Solutions from Combustion

CONTENTS

Not applicable to this specification.

DEFINITIONS

Not applicable to this specification.

TEST SPECIMEN REQUIREMENTS

5.1 Specimen control shall be as specified in BSS 7238 Smoke Generation by Materials on Combustion. A minimum of two specimens shall be tested to measure the evolution of any specific toxicant. If the observed concentration of the toxicant exceeds 50 percent of the maximum permitted value, for either specimen, a third specimen shall be tested.

equipment/apparatus

6.1 COMMUNICATION CRAMMER

The NBS Smoke Chamber shall be used for sample combustion.

(4) Rev A (5) Rev A (6) Rev A (7) Rev A (2) Rev A (3) Rev A (4) Rev A (9) Rev A (10) Rev A (11) Rev A (8) Rev A ACTIVE PAGES: ORIGINAL ISSUE: REV.: BSS "A" 1-18-68 10-13-78 TEST METEOD FOR TOXIC GAS GENERATION BY MATERIALS ON COMBUSTION E. H. Thompson BY. Page 1 of 11 ENG DE TELL BOEING -SPECIFICATION SUPPORT STANDARD I'M Lettent


FSCM NO. 81205

6.2 GAS ANALYSIS EQUIPMENT

2.36

Certain of the following items will be required, depending upon the specific gases to be measured.

- Dräger Multigas Detector, Hodel 21/32 (Bandpump). Available from Safety and Supply Co., 5510 E. Marginal Way South, Seattle, MA 98168.
- b. Colorimetric Gas Detector Tubes, as appropriate for the toxicant to be determined. Available from Mational Drager Co., F.O. Box 120, Pittsburgh, PA 15230.
- c. MSA Sand Pump. Available from Hine Safety Appliances (MSA) Co., P.O. Box 426, Pittsburgh, PA 15288.
- d. Sydrogen Chloride Detector Tube, Part No. 91636. Available from Mine Safety Appliances Co.
- Digital Timer. A digital timer is recommended for determining the times to initiate and terminate sampling unless automatic timers are used.
- f. Sampling System. A pumping system capable of providing sample flow rates equivalent to those produced by Drager and MSA hand pumps is recommended in place of hand held pumps when large numbers of tests are to be done. A system producing a flow rate of 6.7 x 10-6s3/s (400 ml/min.) is required when scrubber sampling is used. Automatic timing controls to begin and end sampling are desirable. Two suitable systems are shown in Figure 1.



ORIGINAL ISSUE: 10-13-78

REVISED: "A" 1-18-86

6.2 GAS ANALYSIS EQUIPMENT (Continued)


States and the state of the sta

- g. Gas Scrubbers. Impinger type scrubbers (bubblers) shown in Figure 2 are suitable for absorbing toxic gases into 1.00 x 10-5m³ (10.0 al) of absorbing solutions at flow rates of up to 6.7 x 10-6m³/s (400 ml/min). They may be fabricated by any custom glass blower. Other scrubber designs can be used with the approval of The Boeing Company Quality Control, provided data is furnished demonstrating that a scrubbing efficiency of greater than 95 percent is obtainable at the flow rates employed in this test with the toxicant concentrations expected.
- b. Bigh Precision pH-meter. The Orion Model 801 or equivalent is recommended, to provide the accuracy and precision needed when using specific ion electrodes.
- i. Electrodes.
 - (1) Fluoride Specific Ion electrode, Orion Hodel 94-09A or equivalent. Required for determination of BF.
 - (2) Chloride Specific Ion Electrode, Orion Hodel 94-17A or equivalent. Required for determination of BCI at concentrations above 500 ppm.
 - (3) Cyanide Specific Ion Electrode, Orion Model 94-06 or equivalent.
 Recommended as an alternate to brager tubes for determination of BCN.
 - (4) Double Junction Reference Electrode, Orion Nodel 94-02 or equivalent. Required for use with above specific ion electrodes.
- j. Continuous Toxic Gas Monitors. Continuous monitoring equipment for any of the gases covered in this specification may be substituted for these analytical methods with the approval of Boeing, provided that information is presented to demonstrate that:
 - (1) The equipment is able to measure the toxic gas concentration in the smoke chamber within \pm 20 percent.
 - (2) The total quantity of smoke chamber contents withdrawn by all sampling devices during the total duration of the test does not exceed 10 percent of the total volume of the smoke chamber.

BSS

7239 Page 3

574



ORIGINAL ISSUE: 10-13-78

REVISED: -3- 1-18-88

PROCEDURE

7.1 CALIBRATION

F . .

» TV T 🕶 •

7

- a. Prior to each day's work the Dräger pump model 21/31 shall be tested as specified in the Dräger Detector Tube Handbook.
- b. Calibration of the colorimetric gas detection tubes is not required. The tubes shall be stored as described in the manufacturer's literature and used before the expiration date printed on the package.
- c. Calibration of any gas sampling system shall be accomplished monthly, and more frequently if any variation in operation is suspected. The calibration procedure shall demonstrate that the system will ampirate the proper volume of gas within + 5 percent through the detection devices (colorimetric tubes or scrubbers) when operated for the time period and at the flow settings used during test. This may be accomplished either using a soap bubble flowmeter (available from Kin-Tech Laborstories, Inc., Texas City, Texas) or by drawing water into a gas burette attached to the sampling port. A record shall be maintained of these settings, and the causes of any major changes in these values shall be ascertained and corrected.

7.2 TEST PROCEDURES

Approved gas analysis methods for this specification are listed in Table I. Specific gas detection tubes are suitable for the determination of CD, HCN, SO2, and MO_N (MO+ MO2) in the concentration ranges normally produced in the HBS smoke chamber. A convenient alternative method for CO uses the continuous NDIR monitor listed. Gas detection tubes can be used for low concentrations of BT and RCL: however, the higher concentrations of these gases produced by burning some materials will require gas scrubbing followed by specific ion electrode analysis, as described in BSS 7242, Determination of the Concentration of Cyanide, Chloride, and Fluoride Ions in Solutions from Combustion. Substitution of other analytical techniques is permissible subject to the approval of Boeing Quality Control when data is provided demonstrating that such techniques will produce equivalent results. These tests can be done simultaneously with the smoke generation tests (BSS 7238) without affecting the smoke measurements.

BSS

7239 Page 5

TEST PROCEDURES (Continued)

TABLE I APPROVED GAS ANALYSIS KETSODS

TOXIC GAS	MAXIMUM MEASURABLE CONCENTRATION PPM	ARALYTICAL METEOD	EQUIPMENT REQUIRED	сожентя
co	12,000	Colorimetric Tubes	Pump, Tubes	Higher ranges available
	5,000	Instrumental, Hondispersive Infrared	Beckman Model 864 or equivalent	Furnishes continuous record
87	15	Colorimetric Tubes	Yump, Tubes	Cannot increase range by decreasing number of pump strokes
The control of the co	2,500	Scrubbing, Specific Ion Electrode	Precision pR Meter, Fluoride Electrode	See 855 7242
BCL	100	Colorimetric Tubes	Yump, Tubes	
}	500	MSA Tube	MSA Pump, Tubes	
	10,000	Scrubbing, Specific Ion Electrode	Precision pR-Neter, Chloride Electrode	See BSS 7242
нох	500	Colorimetric Tubes	Pump, Tubes	Use Mitrous Fumes tubes only
so ₂	3,000	Colorimetric Tubes	Pumps, Tubem	
ROL	150	Colorimetric Tubes	Pump- Tubes	
	2,500	Scrubbing, Specific Ion Electrode	Precision pH-Heter. Cyanide Electrode	See BSS 7242
	EP BCL HO _X SO ₂	TOXIC CONCENTRATION CONCENTRATION FPM CO 12,000 5,000 EF 15 2,500 ECL 100 500 10,000 NO _X 500 SO ₂ 2,000	TOXIC CONCENTRATION ANALYTICAL METHOD CO 12,000 Colorimetric Tubes 5,000 Instrumental, Nondispersive Infrared EF 15 Colorimetric Tubes 2,500 Scrubbing, Specific Ion Electrode HCL 100 Colorimetric Tubes 10,000 Scrubbing, Specific Ion Electrode 10,000 Scrubbing, Specific Ion Electrode 10,000 Colorimetric Tubes HOx 500 Colorimetric Tubes S02 2,000 Colorimetric Tubes	TOXIC CONCENTRATION ANALYTICAL REQUIRED CO 12,000 Colorimetric Pump, Tubes 5,000 Instrumental, Beckman Model 864 or equivalent Infrared IF 15 Colorimetric Pump, Tubes 2,500 Scrubbing, Specific Ion Electrode ECL 100 Colorimetric Pump, Tubes Tubes MSA Tube MSA Pump, Tubes 10,000 Scrubbing, Specific Ion Electrode 10,000 Scrubbing, Colorimetric Pump, Tubes MSA Pump, Tubes 10,000 Scrubbing, Specific Ion Electrode 10,000 Colorimetric Chloride Electrode HOx S00 Colorimetric Pump, Tubes Tubes S02 2,000 Colorimetric Pump, Tubes

7.2.1

7.2

SAMPLING TIME

Unless otherwise specified, initiate sampling 240 seconds (4 min.) after beginning the smoke test.

TEST SEQUENCE 7.2.2

(a) Chamber Conditioning

The chamber shall be preconditioned by burning at least four specimens evolving HF immediately before measuring HF evolution from a test material. Either specimens of the test material itself or conditioning specimens with a 3 mil rediar coating may be employed. This treatment decreases the rate at which HF is adsorbed by the chamber walls from the chamber atmosphere. Heasurement of HF evolution shall immediately follow this conditioning, and reconditioning is necessary whenever an interruption in testing occurs (test of a non-HF evolving material, lunch or overnight break, equipment malfunction, etc.)

BSS 7239 Page 6

X-2000 41V. 5/8

7.2.2 TEST SEQUENCE (Continued)

7.7

(b) Colorizatric Gas Detection Tubes

A list of available colorimetric gas detection tubes and of their measuring ranges is given in Table II. Select a tube and number of strokes (or sample volume), so that an on-scale stain will be produced by the toxic gas concentration expected.

TABLE II AVAILABLE GAS DETECTOR TUBES FOR VARIOUS TOXICANTS

2.

TOXIC GAS	COLORIMETRIC GAS DETECTOR TUBE	NO. OF STROKES N	Measuring Range PPH	COMMENTS
co	10/b	10 5 2	10 to 300 20 to 600 50 to 1500	
	0.1 percent	i	100 to 3000	
ef	1.5/b	20	1.5 to 15	Cannot extend range by decreasing number of strokes.
BC1	1/a	10 5 2 1	1 to 10 2 to 20 5 to 50 10 to 100	
		5 ·	0.100 0 to 500	Requires 40 sec for 1 stroke
¥O _x	0.5/a	5	0.5 to 10	Tubes specific for RO2 alone are also available.
	2/a	10 5	2 to 50 5 to 100	
	20/a	2	20 to 500	
502	1/a	10 5 2	1 to 20 2 to 40 5 to 100	NO2 produces negative interference.
	20/a	10 5 2 1	20 to 200 40 to 400 100 to 1000 200 to 2000	NO ₂ produces negative interference. B ₂ S produces positive interference.
BCN	2/a	5 2 1	2 to 30 5 to 75 10 to 150	

- (1) Break off the ends of the detector tube. Attach it as shown in Figure 3 to a gas sampling port of the MBS chamber. A total of four detector tubes and/or bubblers may be installed for a single smoke run.
- (2) Attach a handpump (Dräger or MSA) or the gas sampling system to the external opening of the MSS chamber gas sampling port.
- (3) At the designated time during the smoke test, open the valve in the chamber sampling port and either activate the hand pump for the proper number of strokes or draw the proper volume of gas through the tube using the gas sampling system, at the proper flow rate for the specific tube. BSS 7239 Page 7

ORIGINAL ISSUE: 10-13-78

REVISED: "A" 1-18-55


7.2.2 TEST SEQUENCE (Continued)

- (4) After completing the smoke test and purging the chamber, note the stain (color change) produced. Discard any result if flow blockage due to soot deposition has occurred.
- (5) Record the following:
 - Sample Identification
 - Smoke Test Run Humber
 - Test Conditions (flux, flaming or smoldering, etc.)
 - Time of Initiating Sample
 - Duration of Sampling
 - Tube Used (Toxicant, Part No.)
 - LX (ppm) length of observed stain where x is specific toxicant
 - n_O = standard number of strokes for tube
 - n = actual number of strokes used

25

- vo = standard value for tube used
 - $= n_0 1.00 \times 10^{-4} \text{ m}^3 \text{ (100 ml)}$
- \sim v = volume aspirated \sim n 1.00 x 10^{-4} m³
- f = flowmeter reading during sampling, if flow system used
- cx (ppm) = measured concentration of toxicant x in smoke chamber

BSS 7239



ORIGINAL ISSUE: 10-13-78

REVISED: 4 1-18-88

7.2.3 GAS SCRUBBERS

- a. Place 1.00 x 10-5 m^3 (10.0 m^2) of 0.1 M sodium hydroxide solution in a clean, dry bubbler. Avoid wetting the gas inlet tube above the liquid level.
- b. Frior to the test, attach the bubbler to a gas sampling port of the NBS chamber as shown in Figure 3.
- c. At the designated time during the test, open the valve in the sampling line. Adjust the flow rate to 6.7 x $10^{-6} m^3/s$ (400 ml/min) and continue until 10^{-3} m³ (1 liter), unless otherwise specified, has been aspirated.
- d. After completing the smoke test and purging the chamber, rinse the gas inlet tube by using a rubber bulb to draw absorbing solution up into the tube several times being cautious not to let the solution reach the bulb. Transfer the bubbler contents to a screw capped plastic bottle for subsequent analysis.
- e. Record the following:

Sample Identification

Test Conditions (flux, flaming or smoldering)

Run Mumber

Sampling Port Location

Time of Initiating Sampling

puration of Sampling

Playmeter Setting During Sampling

- f. As soon as possible, preferably on the day of testing, and in no case later than 72 hours following completion of a test series, analyze the solutions using the analytical methods described in BSS 7242. Record the following data:
 - C_{m}^{X} = Measured molar concentration of analyte in the bubbler solution.
 - C^{\times} (ppm) = C^{\times}_{m} Y = Concentration of toxicant in parts per million in MBS chamber atmosphere,

where

- r . Specific toxicant measured
- F = Analytical Factor = 2.5 x 105 for a 10-3mJ (1 liter) gas sample at 30 degrees C and 105 Pa (1 atm) pressure, scrubbed by 1.00 x 10-5mJ (10.0 ml) of absorbing solution

BSS 7239

7239 Page 10

7.3 CALCULATIONS

The second of the second of the second

For each toxicant measured, calculate the average value of ${\tt C}$ (ppm) and the standard deviation as follows:

$$\frac{c_{x}^{bbm}}{c_{x}^{bbm}} = \frac{c_{x}^{bbm}}{\sum c_{x}^{bbm}}$$

$$S^{x} = \sqrt{\frac{n \sum (C_{x}^{x})^{2} - (\sum C_{ppm}^{x})^{2}}{n (n-1)}}$$

where:

CM = Average value of CM

\(\sum_{ppm} = Sum of measured C_{ppm} valves

 $\sum \left(C_{ppm}^{x}\right)^{2}$ = Sum of squared C_{ppm}^{x} valves

n = Number of individual values of Cappa (Note this is not equivalent to the use of symbol n in Section 7.2.2b.(5))

SX = Measurement standard deviation for toxicants

REPORTING

The test report for this specification may be combined with the smoke density test report of BSS 7238. The following data on each specimen shall be reported unless otherwise specified.

- Complete specimen identification, i.e. material composition and construction, alpha-numeric identifier, thickness, and weight.
- b. Test conditions, e.g. radiant flux, flaming or smoldering, etc.
- c. The analytical method used in measuring C_{ppm}^{Ξ}
- d. The values of $C^{\mathbf{x}}$ (pps) at 240 seconds (4 min), unless otherwise specified, for each toxicant tested.
- e. The values of CX (ppm) and of SX for each toxicant measured.
- f. Comparison of the $\frac{C_{ppm}^{X}}{C_{ppm}}$ values with the requirements as called out in the applicable specification.

BSS

7239 Page 11