

August 2001

Environmental Technology Verification Report


TSI INCORPORATED
MODEL 3320 AERODYNAMIC PARTICLE
SIZER (APS™)

Prepared by



Battelle

Under a cooperative agreement with



ETV ✓ ETV ✓ ETV ✓

THE ENVIRONMENTAL TECHNOLOGY VERIFICATION
PROGRAM



ETV Joint Verification Statement

TECHNOLOGY TYPE: Continuous Ambient Fine Particle Monitor

APPLICATION: MEASURING FINE PARTICULATE MASS IN
AMBIENT AIR

**TECHNOLOGY
NAME:** Model 3320 Aerodynamic Particle Sizer (APSSM)

COMPANY: TSI Incorporated

ADDRESS: 500 Technology Court **PHONE:** 800- 677-2708
St. Paul, MN 55164 **FAX:** 650-490-3860

WEB SITE: <http://www.tsi.com>

E-MAIL: particle@tsi.com

The U.S. Environmental Protection Agency (EPA) has created the Environmental Technology Verification (ETV) Program to facilitate the deployment of innovative or improved environmental technologies through performance verification and dissemination of information. The goal of the ETV Program is to further environmental protection by substantially accelerating the acceptance and use of improved and cost-effective technologies. ETV seeks to achieve this goal by providing high-quality, peer-reviewed data on technology performance to those involved in the design, distribution, financing, permitting, purchase, and use of environmental technologies.

ETV works in partnership with recognized standards and testing organizations; with stakeholder groups that consist of buyers, vendor organizations, and permittees; and with the full participation of individual technology developers. The program evaluates the performance of innovative technologies by developing test plans that are responsive to the needs of stakeholders, conducting field or laboratory tests (as appropriate), collecting and analyzing data, and preparing peer-reviewed reports. All evaluations are conducted in accordance with rigorous quality assurance protocols to ensure that data of known and adequate quality are generated and that the results are defensible.

The Advanced Monitoring Systems (AMS) Center, one of six technology centers under ETV, is operated by Battelle in cooperation with EPA's National Exposure Research Laboratory. The AMS Center has recently evaluated the performance of continuous monitors used to measure fine particulate mass and species in ambient air. This verification statement provides a summary of the test results for the TSI Incorporated Model 3320 Aerodynamic Particle Sizer (APS).

VERIFICATION TEST DESCRIPTION

The objective of this verification test is to provide quantitative performance data on continuous fine particle monitors under a range of realistic operating conditions. To meet this objective, field testing was conducted in two phases in geographically distinct regions of the United States during different seasons of the year. The first phase of field testing was conducted at the ambient air monitoring station on the Department of Energy's National Energy Technology Laboratory campus in Pittsburgh, PA, from August 1 to September 1, 2000. The second phase of testing was performed at the California Air Resources Board's ambient air monitoring station in Fresno, CA, from December 18, 2000, to January 17, 2001. Specific performance characteristics verified in this test include inter-unit precision, agreement with and correlation to time-integrated reference methods, effect of meteorological conditions, influence of precursor gases, and short-term monitoring capabilities. The Model 3320 APS reports measurement results in terms of PM_{2.5} mass and, therefore, was compared with the federal reference method (FRM) for PM_{2.5} mass determination. Additionally, comparisons with a variety of supplemental measurements were made to establish specific performance characteristics.

Quality assurance (QA) oversight of verification testing was provided by Battelle and EPA. Battelle QA staff conducted a data quality audit of 10% of the test data, and performance evaluation audits were conducted on the FRM samplers used in the verification test. Battelle QA staff conducted an internal technical systems audit for Phase I and Phase II. EPA QA staff conducted an external technical systems audit during Phase II.

TECHNOLOGY DESCRIPTION

The Model 3320 APS is a general-purpose particle spectrometer that measures aerodynamic diameter and light-scattering intensity. It provides high-resolution, real-time aerodynamic size measurements in the range of 0.5 to 20 μm . It also measures light-scattering intensity in the equivalent optical size range of 0.37 to 20 μm . A patented, double-crest optical system detects the occurrence of particle coincidence and minimizes the effects of poor signals near the lower detection threshold. The Model 3320 APS accelerates aerosol sample flow through an accelerating orifice. The aerodynamic size of a particle determines its rate of acceleration, with larger particles accelerating more slowly as a result of increased inertia. As particles exit the nozzle, they cross through two partially overlapping laser beams in the detection area. Light is scattered as each particle crosses through the overlapping beams. An elliptical mirror, placed at 90 degrees to the laser beam axis, collects the light and focuses it onto an avalanche photodetector. The avalanche photodetector then converts the light pulses into electrical pulses. Two partially overlapping laser beams allow each particle to generate a single two-crested signal. Peak-to-peak time-of-flight is measured with 4-nanosecond resolution for aerodynamic sizing. The amplitude of the signal is logged for light-scattering intensity. Small particles that may have only one detectable crest are binned separately. Particles with more than two crests, indicative of coincidence, also are binned separately, but are not used to build aerodynamic-size or light-scattering distributions. The Model 3320 APS also features a control knob, built-in display, microprocessor-controlled volumetric flow control, barometric pressure correction, and separate pumps for sheath and total flows. It includes the Aerosol Instrument Manager[®] software, a 32-bit, Windows-based platform with advanced data management capabilities. The Model 3320 APS aerosol inlet has a 3/4-in. outside diameter; and the sensor is 15 in. (38 cm) long, 12 in. (30 cm) wide, and 7 in. (18 cm) high. It weighs 22 pounds. In this verification test, no size selective inlet was used. Average particle size distributions were determined every 15 minutes. The distributions were converted to PM_{2.5} mass concentrations using vendor-supplied software. This software makes assumptions about particle density, particle volume, etc., to estimate mass from the size distributions. (Note: This model is being discontinued in September 2001, at which point a new model will be introduced. The results of this verification test are not applicable to the new model.)

VERIFICATION OF PERFORMANCE

Inter-Unit Precision: During Phase I, regression analysis showed $r^2 = 0.983$ for the 15-minute data and $r^2 = 0.972$ for the 24-hour averages from the duplicate monitors. The slopes of the regression lines were 0.670 (0.005) and 0.627 (0.081), respectively, for the 15-minute data and 24-hour averages, indicating a substantial bias

between the two monitors. The intercept of the regression line was 0.33 (0.07) for the 15-minute data and 0.80 (0.96) for the 24-hour data. The calculated coefficient of variation (CV) for the 15-minute data was 25.2%; and, for the 24-hour averages, the CV was 25.6%. Much of these CV values may be attributed to the bias between the monitors rather than to random differences in the readings. During Phase II, regression analysis showed r^2 values of 0.973 and 0.998, respectively, for 15-minute and 24-hour average data. The slopes of the regression lines were 1.234 (0.010) and 1.280 (0.040), respectively, for the 15-minute data and 24-hour averages, indicating a significant bias between the two monitors. An intercept of 1.23 (0.34) $\mu\text{g}/\text{m}^3$ was observed for the 15-minute data and an intercept of -9.2 (13.6) $\mu\text{g}/\text{m}^3$ for the 24-hour averages. The calculated CV for the 15-minute data was 21.5%; and, for the 24-hour averages, the CV was 18.8%.

Comparability/Predictability: During Phase I, comparisons of the 24-hour averages with $\text{PM}_{2.5}$ FRM results showed r^2 values of 0.100 and 0.093, respectively for Monitor 1 and Monitor 2. The slopes of the regression lines were 0.204 (0.433) and 0.118 (0.301), respectively, for Monitor 1 and Monitor 2. During Phase II, the regression results show r^2 values of 0.803 and 0.762 for Monitor 1 and Monitor 2, respectively, with slopes of the regression lines for Monitor 1 and Monitor 2 of 0.578 (0.121) and 0.555 (0.195), respectively. These results from both phases indicate a substantial negative bias of the Model 3320 APS monitors relative to the FRM.

Meteorological Effects: Multivariable model analysis of the 24-hour average data during Phase I showed no conclusive influence of meteorology on the readings of the two monitors relative to the FRM. Multivariable analysis of the 24-hour average data during Phase II ascribed to wind speed a statistically significant influence on the readings of one of the monitors relative to the FRM values at a 90% confidence level. However, under typical conditions during Phase II, the effect of this parameter on the instrument readings was only about 0.2% relative to the simple linear regression against FRM results.

Influence of Precursor Gases: Multivariable analysis of the 24-hour average data during Phase I showed no conclusive influence of the ambient precursor gases on the readings of the two monitors relative to the FRM values at the 90% confidence level. Multivariable analysis of the 24-hour average data during Phase II showed that none of the ambient precursor gases measured had a statistically significant influence on one of the Model 3320 APS monitors, but that both carbon monoxide and nitric oxide had an effect on the other monitor. The combined effects of these gases on the readings of Monitor 1 under typical conditions during Phase II were less than 2% compared to the simple linear regression against FRM results.

Short-Term Monitoring: In addition to 24-hour FRM samples, short-term sampling was performed on a five-sample-per-day basis in Phase II. The Model 3320 APS results were averaged for each of the sampling periods and compared with the gravimetric results. Linear regression of these data showed slopes of 0.489 and 0.556, respectively, for Monitor 1 and Monitor 2, consistent with the negative bias seen relative to the 24-hour FRM results. For the individual sampling periods, both correlation and slope relative to the FRM were best during periods with the lowest $\text{PM}_{2.5}$ concentrations.

Other Parameters: The two monitors required little maintenance during either phase of testing, with the exception of recalibration of the sample flow rates of one monitor during Phase I. Substantial difficulties associated with data collection resulted in the loss of a considerable amount of data during the two phases of testing. During Phase I, delays in instrument installation and losses from the data system resulted in capture of only about 10 days of data from one monitor and 12 days from the other. During Phase II, data completeness was approximately 90% for one monitor and 50% for the other.

August 2001

Environmental Technology Verification Report

ETV Advanced Monitoring Systems Center

TSI INCORPORATED MODEL 3320 AERODYNAMIC PARTICLE SIZER (APS™)

by

Kenneth Cowen
Thomas Kelly
Basil Coutant
Karen Riggs

Battelle
Columbus, Ohio 43201

Notice

The U.S. Environmental Protection Agency (EPA), through its Office of Research and Development, has financially supported and collaborated in the extramural program described here. This document has been peer reviewed by the Agency and recommended for public release. Mention of trade names or commercial products does not constitute endorsement or recommendation by the EPA for use.

Foreword

The U.S. EPA is charged by Congress with protecting the nation's air, water, and land resources. Under a mandate of national environmental laws, the Agency strives to formulate and implement actions leading to a compatible balance between human activities and the ability of natural systems to support and nurture life. To meet this mandate, the EPA's Office of Research and Development provides data and science support that can be used to solve environmental problems and to build the scientific knowledge base needed to manage our ecological resources wisely, to understand how pollutants affect our health, and to prevent or reduce environmental risks.

The Environmental Technology Verification (ETV) Program has been established by the EPA to verify the performance characteristics of innovative environmental technology across all media and to report this objective information to permittees, buyers, and users of the technology, thus substantially accelerating the entrance of new environmental technologies into the marketplace. Verification organizations oversee and report verification activities based on testing and quality assurance protocols developed with input from major stakeholders and customer groups associated with the technology area. ETV consists of six technology centers. Information about each of these centers can be found on the Internet at <http://www.epa.gov/etv/>.

Effective verifications of monitoring technologies are needed to assess environmental quality and to supply cost and performance data to select the most appropriate technology for that assessment. In 1997, through a competitive cooperative agreement, Battelle was awarded EPA funding and support to plan, coordinate, and conduct such verification tests for "Advanced Monitoring Systems for Air, Water, and Soil" and report the results to the community at large. Information concerning this specific environmental technology area can be found on the Internet at http://www.epa.gov/etv/07/07_main.htm.

Acknowledgments

The authors wish to acknowledge the support of all those who helped plan and conduct the verification test, analyze the data, and prepare this report. In particular we would like to thank the staff at the Department of Energy's National Energy Technology Laboratory, including Richard Anderson, Don Martello, and Curt White, for their assistance in conducting Phase I of the verification test reported here. We would like to thank the California Air Resources Board for its assistance in conducting Phase II of verification testing. We would like to acknowledge the efforts of ETV stakeholders for their assistance in planning this verification test and for reviewing the test/QA plan and the verification reports. Specifically, we would like to acknowledge Judith Chow of Desert Research Institute, Jeff Cook of the California Air Resources Board, Tim Hanley of EPA, and Rudy Eden of the South Coast Air Quality Management District. We also would like to thank Tim Hanley of EPA for the loan of a BGI FRM sampler for Phase II.

Contents

Notice	ii
Foreword	iii
Acknowledgments	iv
List of Abbreviations	x
1. Background	1
2. Technology Description	2
3. Test Design and Procedures	3
3.1 Introduction	3
3.2 Test Design	3
3.3 Reference Method and Supplemental Measurements	4
3.3.1 PM _{2.5} Mass	4
3.3.2 Supplemental Measurements	5
3.4 Data Comparisons	6
3.5 Site Layout/Instrument Installation	7
3.5.1 Phase I	7
3.5.2 Phase II	8
4. Quality Assurance/Quality Control	9
4.1 Data Review and Validation	9
4.2 Deviations from the Test/QA Plan	9
4.3 Calibration and Parameter Checks of Reference Sampler	9
4.3.1 Flow Rate Calibration and Verification	10
4.3.2 Temperature Sensor Calibration and Verification	10
4.3.3 Pressure Sensor Calibration and Verification	11
4.3.4 Leak Checks	11
4.4 Collocated Sampling	12
4.4.1 Phase I—Pittsburgh	12
4.4.2 Phase II—Fresno	12
4.4.3 Summary	12

4.5	Field Blanks	14
4.5.1	Phase I—Pittsburgh	14
4.5.2	Phase II—Fresno	14
4.6	Data Collection	14
4.6.1	Reference Measurements	14
4.6.2	Model 3320 APS Monitors	15
4.7	Assessments and Audits	15
4.7.1	Technical Systems Audit	15
4.7.2	Performance Evaluation Audit	16
4.7.3	Audit of Data Quality	16
5.	Statistical Methods	17
5.1	Inter-Unit Precision	17
5.2	Comparability/Predictability	17
5.3	Meteorological Effects/Precursor Gas Influence	18
5.4	Short-Term Monitoring Capabilities	18
6.	Test Results	20
6.1	Phase I—Pittsburgh (August 1 - September 1, 2000)	20
6.1.1	Inter-Unit Precision	21
6.1.2	Comparability/Predictability	24
6.1.3	Meteorological Effects	26
6.1.4	Influence of Precursor Gases	26
6.2	Phase II—Fresno (December 18, 2000 - January 17, 2001)	26
6.2.1	Inter-Unit Precision	27
6.2.2	Comparability/Predictability	30
6.2.3	Meteorological Effects	32
6.2.4	Influence of Precursor Gases	33
6.2.5	Short-Term Monitoring	33
6.3	Instrument Reliability/Ease of Use	36
6.4	Shelter/Power Requirements	36
6.5	Instrument Cost	37

7. Performance Summary	38
7.1 Phase I—Pittsburgh (August 1 - September 1, 2000)	38
7.2 Phase II—Fresno (December 18, 2000 - January 17, 2001)	39
7.3 Other Parameters	39
8. References	39

Figures

Figure 2-1. TSI Model 3320 Aerodynamic Particle Sizer	2
Figure 3-1. Site Layout During Phase I of Verification Testing	7
Figure 3-2. Site Layout During Phase II of Verification Testing	8
Figure 4-1. Comparison of Collocated PM _{2.5} FRM Samplers During Phase I of Verification Testing	13
Figure 4-2. Comparison of Collocated PM _{2.5} FRM Samplers During Phase II of Verification Testing	13
Figure 6-1a. 15-Minute PM _{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase I of Verification Testing	22
Figure 6-1b. Correlation Plot of the 15-Minute PM _{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase I of Verification Testing	22
Figure 6-2a. 24-Hour PM _{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase I of Verification Testing	23
Figure 6-2b. Correlation Plot of 24-Hour PM _{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase I of Verification Testing	23
Figure 6-3a. Daily PM _{2.5} FRM Concentrations and the 24-Hour PM _{2.5} Averages from Duplicate Model 3320 APS Monitors During Phase I of Verification Testing	24
Figure 6-3b. Correlation Plot of 24-Hour Averages from Duplicate Model 3320 APS Monitors and the PM _{2.5} FRM Results During Phase I of Verification Testing	25
Figure 6-4a. 15-Minute PM _{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase II of Verification Testing	28

Figure 6-4b.	Correlation Plot of 15-Minute PM _{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase II of Verification Testing	28
Figure 6-5a.	24-Hour PM _{2.5} Concentrations from Duplicate Model 2230 APS Monitors During Phase II of Verification Testing	29
Figure 6-5b.	Correlation Plot of 24-Hour PM _{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase II of Verification Testing	30
Figure 6-6a.	Midnight-to-Midnight Averages from Duplicate Model 3320 APS Monitors and the PM _{2.5} FRM Results During Phase II of Verification Testing	31
Figure 6-6b.	Correlation Plot of 24-Hour Averages from Duplicate Model 3320 APS Monitors and the PM _{2.5} FRM During Phase II of Verification Testing	31
Figure 6-7.	Correlation Plot of the Time-Weighted Average for the Short-Term Samples and the PM _{2.5} FRM	34
Figure 6-8.	Correlation Plot of Short-Term Monitoring Results and the Corresponding Averages from the Duplicate Model 3320 APS Monitors During Phase II of Verification Testing	35

Tables

Table 6-1.	Summary of Daily Values for the Measured Meteorological Parameters During Phase I of Verification Testing	20
Table 6-2.	Summary of Daily Values for the Measured Precursor Gas Concentrations During Phase I of Verification Testing	20
Table 6-3.	Linear Regression and Coefficient of Variation Results for 15-Minute and 24-Hour Average PM _{2.5} Concentrations from Duplicate Model 3320 APS Monitors for Phase I	21
Table 6-4.	Comparability of the Model 3320 APS Monitors with the PM _{2.5} FRM in Phase I	25
Table 6-5.	Summary of Daily Values for the Measured Meteorological Parameters During Phase II of Verification Testing	26
Table 6-6.	Summary of Daily Values for the Measured Precursor Gas Concentrations During Phase II of Verification Testing	27

Table 6-7.	Linear Regression and Coefficient of Variation Results for 15-Minute and 24-Hour Average PM _{2.5} Concentrations from Duplicate Model 3320 APS Monitors for Phase II	29
Table 6-8.	Comparability of the Model 3320 APS Monitors with the PM _{2.5} FRM for Phase II	32
Table 6-9.	Summary of Short-Term PM _{2.5} Levels During Phase II of Verification Testing ..	34
Table 6-10.	Regression Analysis Results for the Short-Term Monitoring	36

List of Abbreviations

ADQ	audit of data quality
AMS	Advanced Monitoring Systems
APS	Aerodynamic Particle Sizer
¹⁴ C	carbon 14
CARB	California Air Resources Board
cm	centimeter
CO	carbon monoxide
CV	coefficient of variation
DOE	U.S. Department of Energy
DPI	digital pressure indicator
DRI	Desert Research Institute
EPA	U.S. Environmental Protection Agency
ETV	Environmental Technology Verification
FRM	federal reference method
H ₂ S	hydrogen sulfide
IMPROVE	Interagency Monitoring for Protection of Visual Environments
in.	inch
L/min	liters per minute
m	meters
mg	milligram
mm	millimeter
NETL	National Energy Technology Laboratory
NIST	National Institute of Standards and Technology
NO	nitric oxide
NO ₂	nitrogen dioxide
NO _x	nitrogen oxides
O ₃	ozone
ppb	parts per billion
QA/QC	quality assurance/quality control
QMP	Quality Management Plan
R&P	Rupprecht & Patashnick
SFS	sequential filter sampler
SLAMS	state and local air monitoring stations
TOR	thermal optical reflectance
TSA	technical systems audit
µg	microgram
WINS	well impactor ninety six

Chapter 1

Background

The U.S. Environmental Protection Agency (EPA) has created the Environmental Technology Verification (ETV) Program to facilitate the deployment of innovative environmental technologies through performance verification and dissemination of information. The goal of the ETV Program is to further environmental protection by substantially accelerating the acceptance and use of improved and cost-effective technologies. ETV seeks to achieve this goal by providing high-quality, peer-reviewed data on technology performance to those involved in designing, distributing, permitting, purchasing, and using environmental technologies.

ETV works in partnership with recognized testing organizations; with stakeholder groups consisting of regulators, buyers, and vendor organizations; and with the full participation of individual technology developers. The program evaluates the performance of innovative technologies by developing test plans that are responsive to the needs of stakeholders, conducting field or laboratory tests (as appropriate), collecting and analyzing data, and preparing peer-reviewed reports. All evaluations are conducted in accordance with rigorous quality assurance protocols to ensure that data of known and adequate quality are generated and that the results are defensible.

The EPA's National Exposure Research Laboratory and its verification organization partner, Battelle, operate the Advanced Monitoring Systems (AMS) Center under ETV. The AMS Center recently evaluated the performance of fine particle monitors for use in continuous monitoring of fine particulate matter in ambient air. This verification report presents the procedures and results of the verification test for the TSI Model 3320 Aerodynamic Particle Sizer (APS™).

Chapter 2

Technology Description

The following description of the Model 3320 APS is based on information provided by the vendor. (Note: This model is being discontinued in September 2001, at which point a new model will be introduced. The results of this verification test are not applicable to the new model.)

The Model 3320 APS is a general-purpose particle spectrometer that measures aerodynamic diameter and light-scattering intensity. It provides high-resolution, real-time aerodynamic size measurements in the range of 0.5 to 20 μm . It also measures light-scattering intensity in the equivalent optical size range of 0.37 to 20 μm . A patented, double-crest optical system detects the occurrence of particle coincidence and minimizes the effects of poor signals near the lower detection threshold. Average particle size distributions were determined every 15 minutes. The distributions were converted to $\text{PM}_{2.5}$ mass concentrations using vendor-supplied software. This software makes assumptions about particle density, particle volume, etc., to estimate mass from the size distributions.

The Model 3320 APS accelerates aerosol sample flow through an accelerating orifice. The aerodynamic size of a particle determines its rate of acceleration, with larger particles accelerating more slowly as a result of increased inertia. As particles exit the nozzle, they cross through two partially overlapping laser beams in the detection area. Light is scattered as each particle crosses through the overlapping beams. An elliptical mirror, placed at 90 degrees to the laser beam axis, collects the light and focuses it onto an avalanche photodetector. The avalanche photodetector then converts the light pulses into electrical pulses. Two partially overlapping laser beams allow each particle to generate a single two-crested signal. Peak-to-peak time-of-flight is measured with 4-nanosecond resolution for aerodynamic sizing. The amplitude of the signal is logged for light-scattering intensity. Small particles that may have only one detectable crest are binned separately. Particles with more than two crests, indicative of coincidence, also are binned




Figure 2-1. TSI Model 3320 Aerodynamic Particle Sizer

separately, but are not used to build aerodynamic-size or light-scattering distributions. The Model 3320 APS also features a control knob, built-in display, microprocessor-controlled volumetric flow control, barometric pressure correction, and separate pumps for sheath and total flows. It includes the Aerosol Instrument Manager[®] software, a 32-bit, Windows-based platform with advanced data management capabilities. The Model 3320 APS aerosol inlet has a 3/4-in. outside diameter; and the sensor is 15 in. (38 cm) long, 12 in. (30 cm) wide, and 7 in. (18 cm) high. It weighs 22 pounds.

Chapter 3 Test Design and Procedures

3.1 Introduction

The objective of this verification test is to provide quantitative performance data on continuous fine particle monitors under a range of realistic operating conditions. To meet this objective, field testing was conducted in two phases in geographically distinct regions of the United States during different seasons of the year. Performing the test in different locations and in different seasons allowed sampling of widely different particulate matter concentrations and chemical composition. At each site, testing was conducted for one month during the season in which local $PM_{2.5}$ levels were expected to be highest. The verification test was conducted according to the procedures specified in the *Test/QA Plan for Verification of Ambient Fine Particle Monitors*.⁽¹⁾

The first phase of field testing was conducted at the ambient air monitoring station on the Department of Energy's (DOE's) National Energy Technology Laboratory (NETL) campus in Pittsburgh, PA. Sampling during this phase of testing was conducted from August 1 to September 1, 2000. The second phase of testing was performed at the California Air Resources Board's (CARB's) Air Monitoring Station in Fresno, CA. This site is also host to one of the EPA's $PM_{2.5}$ Supersites being managed by Desert Research Institute (DRI). This phase of testing was conducted from December 18, 2000, to January 17, 2001.

3.2 Test Design

Specific performance characteristics verified in this test include

- Inter-unit precision
- Agreement with and correlation to time-integrated reference methods
- Effect of meteorological conditions
- Influence of precursor gases
- Short-term monitoring capabilities.

To assess inter-unit precision, duplicate Model 3320 APS monitors were tested in side-by-side operation during each phase of testing. During Phase I, the units tested were Serial Numbers 1150 and 1158. During Phase II, the units tested were Serial Numbers 1158 and 1165. Collocation of the Model 3320 APS monitors with reference systems for time-integrated sampling of fine particulate mass and chemical speciation provided the basis for assessing the

degree of agreement and/or correlation between the continuous and reference methods. Each test site was equipped with continuous monitors to record meteorological conditions and the concentration of key precursor gases (ozone, nitrogen oxides, sulfur dioxide, etc.). The data from the meteorological and gas monitors were used to assess the influence of these parameters on the performance of the fine particle monitors being tested. Reference method sampling periods of 3, 5, and 8 hours were used in Phase II of this test to establish the short-term monitoring capabilities of the continuous monitors being tested. Statistical calculations, as described in Chapter 5, were used to establish each of these performance characteristics.

Additionally, other performance characteristics of the technologies being verified, such as reliability, maintenance requirements, and ease of use, were assessed. Instrumental features that may be of interest to potential users (e.g., power and shelter requirements, and overall cost) are also reported.

3.3 Reference Method and Supplemental Measurements

Since no appropriate absolute standards for fine particulate matter exist, the reference methods for this test were well established, time-integrated methods for determining particulate matter mass or chemical composition. It is recognized that comparing real-time measurements with time-integrated measurements does not fully explore the capabilities of the real-time monitors. However, in the absence of accepted standards for real-time fine particulate matter measurements, the use of time-integrated standard methods that are widely accepted was necessary for performance verification purposes. It should be noted that there are necessary differences between continuous and time-integrated, filter-based techniques. For example, in time-integrated sampling, particulate matter collected on a filter may remain there for up to 24 hours, whereas continuous monitors generally retain the particulate sample for one hour or less. Thus, the potential for sampling artifacts differs. Also, in the case of particle mass measurements, the mass of particulate matter is determined after equilibration at constant temperature and humidity, conditions that are almost certain to differ from those during sampling by a continuous monitor.

The Model 3320 measures particle size distributions. These distributions were converted to $PM_{2.5}$ mass concentration values and, therefore, were compared with the federal reference method (FRM) for $PM_{2.5}$ mass determination.⁽²⁾ Additionally, comparisons with a variety of supplemental measurements were made to establish specific performance characteristics. Descriptions of the reference method and supplemental measurements used during the verification test are given below.

3.3.1 $PM_{2.5}$ Mass

The primary comparisons of the Model 3320 APS readings were made relative to the FRM for $PM_{2.5}$ mass determination, i.e., the 24-hour time-averaged procedure detailed in 40 CFR Part 50.⁽²⁾ This method involves manual sampling using any of a number of designated commercially available filter samplers, followed by gravimetric analysis of the collected sample. In this method, a size-selective inlet is used to sample only that fraction of aerosol of interest (i.e., $< 2.5 \mu m$ aerodynamic diameter). The air sample is drawn into the sampler at a fixed rate

(16.7 L/min) over 24 hours, and the aerosol is collected on a Teflon filter for gravimetric analysis. After equilibration of the sample and filter in a temperature- and humidity-controlled environment, the sample is weighed on a microbalance. The particulate sample weight is determined by subtracting the weight of the filter alone, determined prior to sampling after similar equilibration. Protocols for sample collection, handling, and analysis are prescribed by the EPA⁽²⁾ and were followed for this verification test.

Filter samples for the PM_{2.5} FRM were collected daily during each phase of the testing using a BGI FRM Sampler (RFPS-0498-116), and the PM_{2.5} mass was determined according to the procedures mentioned above. In Phase I, a single BGI FRM sampler (Serial Number 311) was operated daily from noon to noon to collect the FRM samples. During Phase II, two BGI FRM samplers (Serial Number 287 and Serial Number 311) were used and were operated on alternate days to facilitate a midnight-to-midnight sampling schedule.

Collocated samples were collected during each phase to establish the precision of the FRM. A discussion of the collocated sampling is presented in Section 4.4 of this report.

3.3.2 Supplemental Measurements

Various supplemental measurements were used to further establish the performance of the continuous monitors being tested. Meteorological conditions were monitored and recorded continuously throughout each phase of the verification test. These measurements included temperature, relative humidity, wind speed, direction, barometric pressure, and solar radiation. These data were provided to Battelle for Phase I by DOE/NETL and for Phase II by DRI. Likewise, the ambient concentrations of various precursor gases including ozone and nitrogen oxides also were measured continuously during the verification test and used to assess the influence of these parameters on the performance of the monitors tested. Continuous measurements of sulfur dioxide, hydrogen sulfide, nitric oxide, nitrogen dioxide, nitrogen oxides, and ozone were provided for Phase I by DOE/NETL; and continuous measurements of carbon monoxide, ozone, nitric oxide, nitrogen dioxide, and nitrogen oxides were provided for Phase II by DRI. These gases were of interest as potential chemical precursors to aerosol components, and as indicators of ambient pollutant levels.

During Phase I, samples for chemical speciation were collected using an Andersen RAAS speciation sampler configured with five sample trains (one channel at 16.7 L/min and four channels at approximately 8 L/min). The 16.7 L/min channel was operated with a Teflon filter for PM_{2.5} mass determination. Samples for carbon analysis were collected at 8 L/min on quartz filters and analyzed by the IMPROVE thermal optical reflectance method at DRI. Nitrate and sulfate samples were collected on nylon filters downstream of a magnesium-oxide-coated compound annular denuder, and analyzed by ion chromatography at Consol.

To supplement the 24-hour samples, additional samples for PM_{2.5} mass were collected at the Fresno site over shorter sampling periods (i.e., 3-, 5-, 8-hour) to assess the capabilities of the monitors being tested in indicating short-term PM_{2.5} levels. A medium-volume sequential filter sampling (SFS) system sampling at a flow rate of 113 L/min was used to collect the short-term mass and speciation samples during Phase II. The SFS was configured to take two simultaneous

samples (i.e., Teflon-membrane/drain disk/quartz-fiber and quartz-fiber/sodium-chloride-impregnated cellulose-fiber filter packs) at 20 L/min through each sampling port. Anodized aluminum nitric acid denuders were located between the inlets and the filters to remove gaseous nitric acid. The remaining 73 L/min required for the 113 L/min total inlet flow was drawn through a makeup air sampling port inside the plenum. The timer was set to take five sets of sequential samples every 24 hours. Solenoid valves, controlled by a timer, switched between sets of filters at midnight each day. A vacuum pump drew air through the paired filter packs when the valves were open. The flow rate was controlled by maintaining a constant pressure across a valve with a differential pressure regulator.

The filters were loaded at the DRI's Reno, NV, laboratory into modified Nuclepore filter holders that were plugged into quick-disconnect fittings on the SFS. One filter pack contained a 47-mm-diameter Teflon-membrane filter with quartz-fiber backup filter. A drain disc was placed between the Teflon-membrane and quartz-fiber filters to ensure a homogeneous sample deposit on the front Teflon-membrane filter and to minimize fiber transfer from one filter to the other. The Teflon-membrane filter collected particles for mass and elemental analysis. The other filter pack contained a 47-mm-diameter quartz-fiber filter with a sodium-chloride-impregnated cellulose-fiber backup filter on a separate stage. The deposit on the quartz-fiber filter was analyzed for sulfate, nitrate, and carbon. The sodium-chloride-impregnated cellulose-fiber backup filter was analyzed for nitrate to estimate losses due to volatilization of ammonium nitrate from the front filter during sampling.

This sequential filter sampler was operated from midnight to 5:00 a.m. (0000-0500), from 5:00 a.m. to 10:00 a.m. (0500-1000), from 10:00 a.m. to 1:00 p.m. (1000-1300), from 1:00 p.m. to 4:00 p.m. (1300-1600), and from 4:00 p.m. to midnight (1600-2400). These short-term sampling measurements were appropriately summed over 24 hours for comparison with the corresponding 24-hour results of the FRM reference samplers to establish the relationship between the two sets of measurements.

3.4 Data Comparisons

The primary means used to verify the performance of the Model 3320 APS monitors was comparison with the 24-hour FRM results. Additional comparisons were made with the supplemental meteorological conditions and precursor gas concentrations to assess the effects of these parameters on the response of the monitors being tested. The short-term monitoring results from Fresno in Phase II of the verification test also were used to assess the capabilities of the Model 3320 APS monitors to indicate short-term levels of ambient $PM_{2.5}$. The comparisons were based on statistical calculations as described in Section 5 of this report.

Comparisons were made independently for the data from each phase of field testing and, with the exception of the inter-unit precision calculations, the results from the duplicate monitors were analyzed and reported separately. Inter-unit precision was determined from a statistical inter-comparison of the results from the duplicate monitors.

3.5 Site Layout/Instrument Installation

In each phase of testing, the two Model 3320 APS monitors were installed in Battelle's instrument trailer, which is a converted 40-foot refrigerator semi-trailer. The Model 3320 APS monitors were placed on a counter top, below a 7.6-cm (3 in.) port through the roof of the trailer. Flexible plastic tubes were used to supply the sampled air to the duplicate monitors. These tubes were tied together inside the trailer, and a single inlet line sampled the outside air. No size selective inlet was used. Data generated by the Model 3320 APS monitors were recorded on laptop computers during each phase of testing as described in Section 4.6.2.

3.5.1 Phase I

Phase I verification testing was conducted at the DOE/NETL facility within the Bruceeton Research Center. The facility is located in the South Park area of Pittsburgh, PA, approximately 7 miles from downtown. The air monitoring station where testing was conducted is located on the top of a relatively remote hill within the facility and is impacted little by road traffic. The layout of the testing facility is illustrated schematically in Figure 3-1.

For this test, Battelle provided temporary facilities to augment the permanent facilities in use by the DOE/NETL air monitoring staff. These temporary facilities included a temporary Battelle/ETV platform (16-foot by 14-foot scaffold construction) and a Battelle instrument trailer. The Battelle trailer was positioned parallel with, and approximately 8 m (25 feet) from, the DOE/NETL instrument trailer. The Battelle/ETV platform was located between the two trailers, with the surface at a height of approximately 2 meters (6 feet).

Most of the DOE/NETL continuous monitoring equipment, including the continuous precursor gas monitors, was located inside the DOE/NETL instrument trailer. A DOE/NETL Rupprecht & Patashnick (R&P) Co. Partisol FRM sampler used to evaluate FRM precision was located outside on a DOE/NETL platform. The Model 3320 monitors were installed inside the Battelle




Figure 3-1. Site Layout During Phase I of Verification Testing (not drawn to scale)

trailer, and the BGI FRM sampler was installed on the Battelle/ETV platform. A vertical separation of approximately 2 to 3 meters and a horizontal separation of approximately 3 meters existed between the inlets of the Model 3320 monitors and the BGI FRM sampler. A 10-meter (33-foot) meteorological tower was located approximately 20 meters (65 feet) to the north of the DOE/NETL instrument trailer.

3.5.2 Phase II

Phase II of verification testing was conducted at the CARB site on First Street in Fresno. This site is located in a residential/commercial neighborhood about three miles north of the center of Fresno. The two BGI FRM samplers and a 3-meter (10-foot) meteorological tower were located on the roof of the two-story building housing the CARB office. Continuous precursor gas monitors were located inside the CARB office space and sampled through a port in the roof of the building. The two BGI FRM samplers were located on the southernmost edge of the rooftop to be as close as possible to the instrument trailer. The Battelle trailer used during Phase I of this verification test also was used during Phase II. For Phase II, the Battelle trailer was located in the parking lot adjacent to the building in which the CARB site is located. The trailer was positioned approximately 25 meters (80 feet) to the south of the building, as shown in Figure 3-2. The Model 3320 monitors were located in the Battelle trailer and installed in the same fashion as in Phase I of the verification test. A difference in elevation of approximately 6 meters (20 feet) existed between the top of the trailer and the roof of the building housing the CARB site and between the inlets of the Model 3320 monitors and the BGI FRM samplers. In addition to the two BGI FRM samplers used to collect the reference samples, an R&P Partisol FRM sampler was operated on the rooftop by CARB. This sampler was positioned approximately 25 meters (80 feet) to the northeast of the BGI FRM samplers and was used to measure the precision of the FRM reference values. The sequential filter sampler used to collect the short-term samples was located near the R&P FRM sampler.




Figure 3-2. Site Layout During Phase II of Verification Testing (not drawn to scale)

Chapter 4

Quality Assurance/Quality Control

4.1 Data Review and Validation

Test data were reviewed and approved according to the AMS Center quality management plan (QMP)⁽³⁾ and the test/QA plan,⁽¹⁾ and Battelle's one-over-one policy. The Verification Test Coordinator or the Verification Testing Leader or designee reviewed the raw data, laboratory notebook entries, and data sheets that were generated each day and approved them by initialing and dating the records.

Data from the Model 3320 APS monitors were reviewed by the Verification Test Coordinator before being used in statistical calculations. Daily PM_{2.5} concentration averages calculated from the continuous Model 3320 APS data were considered valid if the percent data recovery for the 24-hour sampling period (i.e., noon to noon for Phase I, or midnight to midnight for Phase II) was 75% or greater.

4.2 Deviations from the Test/QA Plan

The following deviations from the test/QA plan were documented and approved by the AMS Center Manager. Neither of these deviations had any deleterious effect on the verification data.

- Calibration checks of the temperature and pressure sensors were not performed within one week of the start of Phase II. Subsequent checks of these sensors indicated proper calibration.
- The distance between the reference samplers and the monitors being tested was increased to approximately 25 meters to accommodate changes in the overall site layout for Phase II.

In addition, although not formally a deviation from the test/QA plan, we note that the relative humidity of the continuing weighing room used by Consol in Phase I occasionally deviated from the specified limits. The impact of this occurrence was minimal, as noted in Section 4.4.1.

4.3 Calibration and Parameter Checks of Reference Sampler

The BGI FRM samplers provided by Battelle for this verification test were calibrated using National Institute of Standards and Technology (NIST)-traceable flow meters and temperature and pressure sensors. The calibration and verification of these samplers are described below.

4.3.1 Flow Rate Calibration and Verification

Prior to Phase I of the verification test, a three-point calibration of the sampler flow rate was performed on June 22, 2000. Flows were measured at three set points (16.7 L/min, and approximately +10% and -10% of 16.7 L/min) using a dry gas meter (American Meter Company, Battelle asset number LN 275010, calibrated January 21, 2000). If necessary, the flows were adjusted manually until agreement with the dry gas meter fell within $\pm 2\%$ of the sampler's indicated flow reading.

The on-site operators checked the flow rate of the BGI FRM sampler both before and after Phase I of the verification test using an Andersen Instruments Inc. dry gas meter (identification number 103652, calibrated March 30, 2000). The flow rate was checked prior to testing on both July 19, 2000, and July 30, 2000. In both cases, the measured flow rate was verified to be within 4% of the flow rate indicated by the sampler. After testing, the flow rate was again checked on September 11, 2000, using the same Andersen dry gas meter. In this case, the flow rate did not fall within the 4% acceptance limit. This failure is probably linked to the failure of the ambient temperature thermocouple, on September 7, 2000, after completion of the Phase I sampling (see Section 4.3.2).

Prior to Phase II of the verification test, single point calibration checks of the duplicate BGI FRM samplers were performed at 16.7 L/min on December 15, 2000. These flow rate checks were performed using a BGI DeltaCal calibrator (BGI Inc., serial number 0027, calibrated October 24, 2000), and the measured flow rates were within 4% of the indicated flow on each sampler. Weekly flow rate checks also were performed throughout Phase II using the DeltaCal flow meter. In each case, the measured flow rates were within $\pm 4\%$ of the indicated reading of the BGI FRM and within $\pm 5\%$ of the nominal 16.7 L/min setpoint.

Calibration of the flow rate for the SFS sampler used during Phase II was maintained by DRI through daily flow checks with a calibrated rotameter, and through independent performance evaluation audits conducted by Parson's Engineering. No additional flow verification was performed for this test.

4.3.2 Temperature Sensor Calibration and Verification

Both the ambient temperature sensor and the filter temperature sensor of the BGI FRM sampler were checked at three temperatures (approximately 5, 22, and 45°C) on June 20, 2000. The sensor readings were compared with those from an NIST-traceable Fluke Model 52 thermocouple gauge (Battelle asset number LN 570068, calibrated October 15, 1999). Agreement between the sampler temperature sensors and the calibrated thermocouple was within $\pm 2^\circ\text{C}$ at each temperature.

The temperature sensors also were checked at the DOE/NETL site both before and after Phase I of the verification test by the on-site operators. Prior to testing, the sensors were checked on July 19, 2000, and July 30, 2000, against the readings from a mercury thermometer (Ever Ready, serial number 6419, calibrated October 29, 1999). For these checks, agreement between the sensors and the thermometer was within $\pm 2^\circ\text{C}$. After the verification period, the ambient

temperature sensor suffered a malfunction on September 7. The filter temperature sensor was checked on September 11, 2000, and showed agreement with the mercury thermometer within $\pm 2^{\circ}\text{C}$. The sensor was replaced, after completing Phase I, with a new factory-calibrated sensor provided by BGI.

The temperature sensors for the two BGI FRM samplers were checked on January 16, 2001, against readings from a Fluke Model 52 thermocouple gauge (Battelle asset number LN 570077, calibrated October 26, 2000). For each BGI FRM, both the ambient and filter temperature sensor readings agreed with the thermocouple readings within $\pm 2^{\circ}\text{C}$.

4.3.3 Pressure Sensor Calibration and Verification

Before Phase I, the barometric pressure sensor in the BGI FRM sampler was calibrated against an NIST-traceable Taylor Model 2250M barometer (Battelle asset number LN 163610, calibrated January 12, 2000) and an NIST-traceable convectron gauge (Granville-Phillips Co., Battelle asset number LN 298084, calibrated August 25, 1999) on June 17 and 18, 2000. The sensor was calibrated at ambient pressure and under a reduced pressure (approximately 100 mm mercury below ambient).

Checks of the pressure sensor were performed at the DOE/NETL site both before and after Phase I of the verification test. The pressure sensor was checked on July 19, 2000, and July 30, 2000, using an NIST-traceable Taylor Model 2250M barometer (Battelle asset number LN 163609, calibrated January 12, 2000). On September 11, 2000, the pressure sensor of the BGI FRM sampler was again checked against the same barometer, but did not agree within the acceptance criterion of 5 mm mercury. This failure is possibly associated with the failure of the ambient temperature sensor on September 7, 2000.

The ambient pressure sensor for both BGI FRM samplers used in Phase II was checked against the pressure readings of a BGI DeltaCal on January 16, 2001. Agreement between the BGI FRM pressure readings and those of the DeltaCal was within 5 mm mercury for both samplers.

4.3.4 Leak Checks

Leak checks of the BGI FRM sampler were performed every fourth day during Phase I of the verification test. These leak checks were conducted immediately following the cleaning of the WINS impactor and were performed according to the procedures in the operator's manual for the BGI FRM sampler. All leak checks passed the acceptance criteria provided in the operator's manual.

Leak checks of the BGI FRM and SFS samplers were performed daily during Phase II of the verification test. These leak checks were conducted during set-up for each 24-hour sampling period. All leak checks passed before the sampler set-up was completed.

4.4 Collocated Sampling

4.4.1 Phase I—Pittsburgh

To establish the precision of the PM_{2.5} FRM, the BGI FRM sampler was collocated with an R&P FRM sampler for Phase I, including a period of two weeks prior to and one week after Phase I of the verification test. During the sampling periods before and after Phase I, the BGI and R&P FRM samplers were located on the same platform and within 4 meters of one another. During the Phase I testing period, these samplers were separated by a distance of approximately 25 meters. The samples from the BGI FRM sampler were collected and analyzed by Consol, and the samples from the R&P FRM sampler were collected and analyzed by on-site Mining Safety and Health Administration staff.

Figure 4-1 shows the results of the collocated FRM sampling conducted for Phase I. These data were analyzed by linear regression; and the calculated slope, intercept, and r^2 values are 0.939 (0.033), 1.28 (0.66) $\mu\text{g}/\text{m}^3$, and 0.957, respectively, where the values in parentheses are 95% confidence intervals (CIs). Despite completely independent operations (i.e., separate sampling staff and weighing facilities), these data show very good agreement between the BGI FRM and the R&P FRM samplers. The data also indicate that, although the humidity in the conditioning/ weighing room at Consol was not always within the specified FRM limits, the influence of the elevated humidity was not severe.

4.4.2 Phase II—Fresno

During Phase II of testing, duplicate BGI FRM samplers (SN 287 and SN 311) were used to collect the 24-hour FRM reference samples. These samplers were operated one at a time on alternate days to facilitate midnight-to-midnight sampling. Likewise, an R&P Partisol sampler was used by CARB to collect 24-hour FRM samples. The R&P FRM sampler was located approximately 25 meters from the BGI FRM samplers. The same on-site operators performed the sampling for the two FRM samplers; however, DRI performed the gravimetric analyses for the BGI FRM samplers and CARB performed the analyses for the R&P FRM sampler.

Figure 4-2 shows the results for the collocated FRM sampling conducted for Phase II. Only 12 days of collocated sampling were available from the Fresno site. The linear regression of these data shows a slope of 1.096 (0.047), intercept of -1.0 (2.1) $\mu\text{g}/\text{m}^3$, and r^2 value of 0.982, where the numbers in parentheses indicate the 95% CIs.

4.4.3 Summary

The results from the collocated FRMs in both Pittsburgh and Fresno show agreement that is consistent with the goals for measurement uncertainty of PM_{2.5} methods run at state and local air monitoring stations (SLAMS). These goals are identified in Appendix A to 40 CFR Part 58, Section 3.5⁽⁴⁾ which states: “The goal for acceptable measurement uncertainty has been defined as 10 percent coefficient of variation (CV) for total precision and $\pm 10\%$ for total bias.” Since the collocated FRMs in both Pittsburgh and Fresno were operated by independent organizations, a




Figure 4-1. Comparison of Collocated $\text{PM}_{2.5}$ FRM Samplers During Phase I of Verification Testing




Figure 4-2. Comparison of Collocated $\text{PM}_{2.5}$ FRM Samplers for Phase II of Verification Testing

comparison to the SLAMS data quality objectives for $PM_{2.5}$ is an appropriate way to assess whether the measurement systems were producing data of acceptable quality. In both Pittsburgh and Fresno, the results of the collocated sampling meet the data quality objectives for the total bias. In Fresno, the collocated sampling results show a CV of 6.3%, which meets the data quality objectives for precision. In Pittsburgh, the calculated CV was 10.5%. However, this value is driven largely by scatter in the low concentration regime. When a single data pair is removed, the CV becomes 9.1%, which meets the data quality objectives for total precision. (It should be noted, as well, that the Fresno collocated results consist of only 12 data points.) Thus, the collocated FRM results from Pittsburgh and Fresno show that the reference measurements were suitable for verifying the performance of continuous fine particle monitors.

4.5 Field Blanks

4.5.1 Phase I—Pittsburgh

During Phase I, at least 10% of the collected reference samples were field blanks. The observed filter mass difference of the field blanks ranged from $-7 \mu\text{g}$ to $16 \mu\text{g}$, and the corresponding $PM_{2.5}$ concentrations (which were determined using an assumed sample volume of 24 m^3) were all less than $0.7 \mu\text{g}/\text{m}^3$, averaging $1.5 \mu\text{g}/\text{m}^3$. FRM results for Phase I were not blank corrected.

4.5.2 Phase II—Fresno

During Phase II, at least 10% of the collected reference samples (both the BGI FRM samplers and the DRI sequential filter sampler) were field blanks. The results were added to a database containing historical field blank data. On average, these blanks showed mass differences of $2 \mu\text{g}$, with a standard deviation of $8 \mu\text{g}$. Assuming a sample volume of 24 m^3 (i.e., FRM value), the average blank accounts for approximately $0.1 \mu\text{g}/\text{m}^3$. Assuming a sample volume of 3.6 m^3 (i.e., three-hour short-term sample from sequential filter sampler), the average blank accounts for approximately $0.6 \mu\text{g}/\text{m}^3$. These blank values were negligible, even for the short-term sampling periods, in comparison with the $PM_{2.5}$ mass levels that were present during the Phase II testing (see Section 6.2). FRM results for Phase II were blank corrected using the data available from the historical database.

4.6 Data Collection

4.6.1 Reference Measurements

During Phase I, daily records of the sampling activities for the BGI FRM sampler were recorded on individual data sheets by the on-site operators, and summary data from the BGI FRM sampler were downloaded daily using portable data logging modules. Information recorded on the data sheets included identification of the sampling media (i.e., filter ID numbers) and the start and stop times for the sampling periods. Summary data from the sampler included the parameters listed above, in addition to the sampling duration, volume sampled, and average temperature and pressure readings.

During Phase II, summary data from the BGI FRM samplers were logged daily on sampling sheets by the on-site operators. These data included sample identification, start times for the sampling period, sampling duration, volume sampled, and average temperature and pressure readings.

4.6.2 Model 3320 APS Monitors

Data from each of the Model 3320 APS monitors were recorded every 15 minutes on an on-site laptop computer. These data were converted from particle concentration to mass using vendor-supplied software that incorporated assumptions about particle density, morphology, etc. The converted data were saved as text files and imported into Excel for subsequent analysis. Copies of the data were stored by the Verification Test Coordinator on a floppy disk, as well as on a computer hard drive.

4.7 Assessments and Audits

4.7.1 Technical Systems Audit

Phase I—Pittsburgh

The technical systems audit (TSA) ensures that the verification tests are conducted according to the test/QA plan⁽¹⁾ and that all activities associated with the tests are in compliance with the ETV pilot QMP.⁽³⁾ The Battelle Quality Manager conducted an internal TSA on August 3, 2000, at the Pittsburgh test site. All findings noted during this TSA were documented and submitted to the Verification Test Coordinator for correction. The corrections were documented by the Verification Test Coordinator and reviewed by Battelle's Quality Manager, Verification Testing Leader, and AMS Center Manager. None of the findings adversely affected the quality or outcome of this phase of the verification test. All corrective actions were completed to the satisfaction of the Battelle Quality Manager. The records concerning this TSA are permanently stored with the Battelle Quality Manager.

Phase II—Fresno

An internal TSA was conducted by the Battelle Quality Manager on January 9, 2001, at the Fresno test site. An external TSA was also conducted concurrently by EPA quality staff, Ms. Elizabeth Betz and Ms. Elizabeth Hunike. All findings noted during these TSAs were documented and submitted to the Verification Test Coordinator for corrective action. None of the findings adversely affected the quality or outcome of this phase of the verification test for the BAM 1020. All corrective actions were completed to the satisfaction of the Battelle Quality Manager and the EPA.

4.7.2 Performance Evaluation Audit

Phase I—Pittsburgh

The reference sampler provided by Battelle for this verification test was audited during Phase I to ensure that it was operating properly. During Phase I of the verification test, the flow rate of the BGI FRM sampler was audited on August 28, using a dry gas meter (American Meter Company, Battelle asset number LN 275010, calibrated April 17, 2000). The measured flow rate was within the $\pm 4\%$ acceptance criterion with respect to the internal flow meter and within the $\pm 5\%$ acceptance criterion with respect to the nominal flow rate.

Both temperature sensors in the BGI FRM sampler were checked on August 28, using a Fluke 52 thermocouple (Battelle asset number LN 570068, calibrated October 15, 1999). Agreement between each sensor and the thermocouple was within the $\pm 2^\circ\text{C}$ acceptance criterion.

Phase II—Fresno

A performance evaluation audit was conducted to ensure that the two BGI FRM samplers used during Phase II of testing were operating properly. The flow rates of the samplers were audited on January 16 and 17, 2001, using a dry gas meter (Schlumberger, Serial Number 103620, calibrated July 6, 2000). For each sampler, the measured flow rate was within the $\pm 4\%$ acceptance criterion with respect to the internal flow meter and within the $\pm 5\%$ acceptance criterion with respect to the nominal flow rate.

The temperature readings for the two samplers were checked with a mercury thermometer (Fisher Scientific, Serial Number 7116). Agreement between each sensor and the thermocouple was within the $\pm 2^\circ\text{C}$ acceptance criterion.

The pressure sensors for the two samplers were checked against a Druck digital pressure indicator (DPI) (Serial Number 6016/00-2, calibrated June 28, 2000). Agreement between each sensor and the DPI was within the acceptance criterion of ± 5 mm mercury.

4.7.3 Audit of Data Quality

Battelle's Quality Manager ensured that an audit of data quality (ADQ) of at least 10% of the verification data acquired during the verification test was completed. The ADQ traced the data from initial acquisition, through reduction and statistical comparisons, to final reporting. Reporting of findings followed the procedures described above for the Phase I TSA. All findings were corrected to the satisfaction of the Battelle Quality Manager, and none of the findings adversely affected the quality of the verification test for the Model 3320 APS monitors.

Chapter 5 Statistical Methods

Performance verification is based, in part, on statistical comparisons of continuous monitoring data with results from the reference methods. A summary of the statistical calculations that have been made is given below.

5.1 Inter-Unit Precision

The inter-unit precision of the Model 3320 APS monitors was determined based on procedures described in Section 5.5.2 of EPA 40 CFR 58, Appendix A, which contains guidance for precision assessments of collocated non-FRM samplers. Simultaneous measurements from the duplicate Model 3320 APS monitors were paired, and the behavior of their differences was used to assess precision. For both the 15-minute readings and the 24-hour PM_{2.5} measurements, the coefficient of variation (CV) is reported. The CV is defined as the standard deviation of the differences divided by the mean of the measurements and expresses the variability in the differences as a percentage of the mean. As suggested by the EPA guidance, only measurements above the limit of detection were used in precision calculations. Inter-unit precision was assessed separately for each phase of the verification test.

5.2 Comparability/Predictability

The comparability between the Model 3320 APS results and the PM_{2.5} FRM was assessed, since the Model 3320 APS yields measurements with the same units of measure as the PM_{2.5} FRM. The relationship between the two was assessed from a linear regression of the data using the PM_{2.5} FRM results as the independent variable and the Model 3320 APS monitor results as the dependent variable as follows:

$$C_i = \mu + \beta \times R_i + \varepsilon_i \quad (1)$$

where R_i is the i^{th} 24-hour FRM PM_{2.5} measurement; C_i is the average of the 15-minute Model 3320 APS measurements over the same 24-hour time period as the i^{th} reference measurement; μ and β are the intercept and slope parameters, respectively; and ε_i is error unexplained by the model. The average of the hourly Model 3320 APS measurements is used because this is the quantity that is most comparable to the reference sampler measurements.

Comparability is expressed in terms of bias between the Model 3320 APS monitor and the PM_{2.5} FRM and the degree of correlation (i.e., r^2) between the two. Bias was assessed based on the slope and intercept of the linear regression of the data from the PM_{2.5} FRM and the Model 3320 APS monitor. In the absence of bias, the regression equation would be $C_i = R_i + \varepsilon_i$ (slope = 1, intercept = 0), indicating that the 24-hour average of hourly Model 3320 APS measurements is simply the PM_{2.5} FRM measurement plus random error. A value of r^2 close to 1 implies that the amount of random error is small; that is, the variability in the hourly measurements is almost entirely explained by the variability in the PM_{2.5} FRM measurements.

Quantities reported include r^2 , intercept, and slope, with estimates of the 95% CI for the intercept and slope. Comparability to the FRM was determined independently for each of the two duplicate Model 3320 APS monitors being tested and was assessed separately for each phase of the verification test.

5.3 Meteorological Effects/Precursor Gas Influence

The influence of meteorological conditions on the correlation between the Model 3320 APS monitors and the PM_{2.5} FRM reference results was evaluated, by using meteorological data such as temperature and humidity as parameters in multivariable analyses of the reference/monitor comparison data. The same evaluation was done with ambient precursor pollutant concentrations as the model parameters. The model used is as follows:

$$C_i = \mu + \beta \times R_i + \sum \gamma_j \times X_{ji} + \varepsilon_i \quad (2)$$

where X_{ji} is the meteorological or precursor gas measurement for the i^{th} 24-hour time period, γ_j is the associated slope parameter, and other notation is as in Equation 1. Comparability results are reported again after these variables are adjusted for in the model. Additionally, estimates of γ_j are provided. Meteorological effects and precursor gas interferences were assessed independently for each of the two duplicate Model 3320 APS monitors tested and were assessed separately for each phase of the verification test.

Note that the multivariable model ascribes variance unaccounted for by linear regression against the FRM to the meteorological or precursor gas parameters. The model treats all candidate parameters equally. The model discards the least significant parameter and is rerun until all remaining variables have the required significance (i.e., predictive power). The results of the model should not be taken to imply a cause-and-effect relationship. It is even possible that the parameters identified as significant for one unit of a monitoring technology may differ from those identified for the duplicate unit of that technology, due to differences in the two data sets.

5.4 Short-Term Monitoring Capabilities

This assessment was based on linear regression analysis of results from the Model 3320 APS monitors and the short-term (3-, 5-, and 8-hour) sampling results from the two BGI FRM samplers generated in Phase II only. The analysis was conducted, and the results are reported, in

a fashion identical to that for the comparability results for the 24-hour samples described in Section 5.2.

These comparisons were made only after establishing the relationship between the short-term sampling results and the corresponding 24-hour FRM results. The comparison between the two sets of reference measurements was made by linear regression using the weighted sum of the results from the short-term sampling as the dependent variable and the 24-hour FRM results as the independent variable in the regression analysis. The comparison used Equation 1, replacing the average of 15-minute measures with the average of short-term sampler measurements. Independent assessments were made for each of the duplicate Model 3320 APS monitors, and the data from each phase of testing were analyzed separately.

Chapter 6 Test Results

6.1 Phase I—Pittsburgh (August 1 - September 1, 2000)

Samples were collected daily between August 1 and September 1, 2000, using a PM_{2.5} FRM sampler. During this period, the daily PM_{2.5} concentration as measured by the BGI FRM sampler ranged from 6.1 µg/m³ to 36.2 µg/m³, with an average daily concentration of 18.4 µg/m³. Typically, the PM_{2.5} composition was dominated by sulfate and carbon species. On average, the measured sulfate concentration, determined by ion chromatography, accounted for approximately 47% of the daily PM_{2.5} mass. Total carbon, as measured by the IMPROVE thermal optical reflectance (TOR) method, accounted for approximately 38% of the PM_{2.5} mass, with elemental carbon contributing approximately 22% and organic carbon contributing approximately 77% of the total carbon. Additionally, nitrate contributed about 8.3% of the daily PM_{2.5} concentration.

Table 6-1 summarizes the meteorological conditions during Phase I, and Table 6-2 summarizes the observed concentrations of the measured precursor gases during this period.

Table 6-1. Summary of Daily Values for the Measured Meteorological Parameters During Phase I of Verification Testing

	Wind Speed (mph)	Vertical Wind Speed (mph)	Wind Direction (degrees)	Air Temp. @ 10 m (C)	Air Temp. @ 2 m (C)	RH (%)	Solar Radiation (W/m ²)	Press. (mbar)	Total Precip. (in.)
Average	3.35	0.09	196	20.0	16.6	89.4	162.8	979.7	0.0014
Max.	6.45	0.29	298	24.1	22.5	95.8	246.1	986.7	0.0297
Min	1.88	-0.03	106	14.6	12.1	80.2	47.9	974.5	0.0000

Table 6-2. Summary of Daily Values for the Measured Precursor Gas Concentrations During Phase I of Verification Testing

	SO ₂ (ppb)	H ₂ S (ppb)	NO (ppb)	NO ₂ (ppb)	NO _x (ppb)	O ₃ (ppb)
Average	6.9	1.5	3.1	10.1	13.0	24
Max	12.8	2.9	10.4	17.4	27.4	51
Min	2.7	-0.6	0.14	5.3	5.3	5

6.1.1 Inter-Unit Precision

Particle concentration measurements were made every 15 minutes by the duplicate Model 3320 APS monitors during Phase I of the verification test. These 15-minute particle concentration readings were converted to mass concentrations using software supplied by the vendor. Figure 6-1a shows the calculated fine particulate mass data from the two Model 3320 APS monitors for Phase I of the verification test. Figure 6-1b is a scatter plot of these same data points. The duplicate monitors were installed on August 8 (approximately one week after the start of the verification test), at which point one of the monitors was found to have a flow rate problem. This monitor was removed by the vendor and was returned for repair. After correction of the problem, the monitor was reinstalled on August 21. Data from the other Model 3320 APS were not available from August 10 to 21, as a result of problems with the data collection system that developed after a power outage at the test site on August 10 and that were not detected until the second monitor was reinstalled. For the period when data are available from both of the Model 3320 APS monitors, the duplicate monitors showed similar temporal patterns in PM_{2.5} concentrations, but displayed an offset between their concentration values.

For comparison with the PM_{2.5} FRM reference measurements, the 15-minute data were averaged from noon to noon for each day to correspond with the 24-hour sampling periods used in Phase I of the verification test. In Figure 6-2a the noon-to-noon averages for Phase I of the verification test are presented for the two Model 3320 APS monitors. A correlation plot of these data is shown in Figure 6-2b.

These data were analyzed by linear regression, and the results of this analysis are presented in Table 6-3. The CV values for these data were also determined according to Section 5.1, and the calculated CV is shown in Table 6-3. The regression analysis of the 15-minute data shows a coefficient of determination $r^2 = 0.983$. These results also indicate a substantial bias between the two monitors, with Monitor 1 generally reading about 50% higher than Monitor 2 [slope = 0.670 (0.005)]. The calculated intercept of the correlation plot is 0.33 (0.07) $\mu\text{g}/\text{m}^3$, which is significantly different from zero at the 95% confidence level.

Table 6-3. Linear Regression and Coefficient of Variation Results for 15-Minute and 24-Hour Average PM_{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase I

Parameter	15-Minute Data	24-Hour Average Data
Slope (95% CI)	0.670 (0.005)	0.627 (0.081)
Intercept ($\mu\text{g}/\text{m}^3$) (95% CI)	0.33 (0.07)	0.80 (0.96)
r^2	0.983	0.972
CV	25.2%	25.6%




Figure 6-1a. 15-Minute PM_{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase I of Verification Testing




Figure 6-1b. Correlation Plot of the 15-Minute PM_{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase I of Verification Testing




Figure 6-2a. 24-Hour PM_{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase I of Verification Testing




Figure 6-2b. Correlation Plot of 24-Hour PM_{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase I of Verification Testing

The regression results of the 24-hour average concentrations indicate an r^2 of 0.972, a slope of 0.627 (0.081), and an intercept of 0.80 (0.96) $\mu\text{g}/\text{m}^3$, which is not statistically different from zero at 95% confidence. The slope of the regression line is statistically different from unity at 95% confidence indicating a bias between the duplicate monitors. The calculated CV for the 24-hour averages is 25.6%, which is close to the CV of the 15-minute data of 25.2%. Much of these CV values may be attributable to the observed bias between the duplicate monitors, rather than to random differences between the two.

6.1.2 Comparability/Predictability

In Figure 6-3a, the noon-to-noon averages of the Model 3320 APS measurements are shown, along with the $\text{PM}_{2.5}$ FRM measurements for Phase I of the verification test. These $\text{PM}_{2.5}$ concentrations were analyzed by linear regression according to Section 5.2 to establish the comparability of each of the Model 3320 APS monitors with the $\text{PM}_{2.5}$ FRM sampler. The resulting comparisons are plotted in Figure 6-3b, and the calculated slope, intercept, and r^2 value of the regression analyses are presented in Table 6-4 for each monitor.




Figure 6-3a. Daily $\text{PM}_{2.5}$ FRM Concentrations and the 24-Hour $\text{PM}_{2.5}$ Averages from Duplicate Model 3320 APS Monitors During Phase I of Verification Testing




Figure 6-3b. Correlation Plot of the 24-Hour Averages from Duplicate Model 3320 APS Monitors and the PM_{2.5} FRM Results During Phase I of Verification Testing

Table 6-4. Comparability of the Model 3320 APS Monitors with the PM_{2.5} FRM in Phase I

Regression Parameter	Monitor 1	Monitor 2
Slope (95% CI)	0.204 (0.433)	0.118 (0.301)
Intercept (μg/m ³) (95% CI)	7.4 (10.1)	5.0 (7.2)
r ²	0.100	0.093

Regression analysis of the 24-hour averages with the PM_{2.5} FRM measurements show r² values of 0.100 and 0.093, respectively, for Monitor 1 and Monitor 2. For both Monitor 1 and Monitor 2, the regression results show a substantial negative bias relative to the FRM. For Monitor 1, the slope of the regression line is 0.204 (0.433), and for Monitor 2 the slope is 0.118 (0.301), where the numbers in parentheses are the respective 95% confidence intervals. Thus, the slopes are substantially less than 1.0 and not significantly different from zero. Since the Model 3320 APS calculates PM_{2.5} mass from particle concentrations derived from individual particle counts, it is likely that errors in assumptions about particle density, particle shape, etc., in the calculations

may account for some of the observed bias. Neither intercept was significantly different from zero at the 95% confidence level.

6.1.3 Meteorological Effects

A multivariable model, as described in Section 5.3, was used to determine if variability in the readings of the Model 3320 APS monitors could be accounted for by meteorological conditions. Based on the limited data available for this analysis, there was no conclusive evidence of meteorological influence on the readings of the two monitors.

6.1.4 Influence of Precursor Gases

As described in Section 5.3, multivariable analysis was performed to determine if precursor gases had an influence on the readings of the Model 3320 APS monitors. Based on the limited data available for this analysis, there was no conclusive evidence of precursor gas influence on the readings of the two monitors.

6.2 Phase II—Fresno (December 18, 2000 - January 17, 2001)

During Phase II, daily 24-hour PM_{2.5} concentrations averaged 74 µg/m³ and ranged from 4.9 µg/m³ to 146 µg/m³. A strong diurnal pattern was observed in the PM_{2.5} concentration, with the peak levels occurring near midnight. Particle composition was dominated by nitrate and carbon. On average, the overall PM_{2.5} concentration comprised 22% nitrate and 40% total carbon. Sulfate accounted for only about 2% of the daily PM_{2.5} mass.

Table 6-5 summarizes the meteorological conditions during Phase II, and Table 6-6 summarizes the observed concentrations of the measured precursor gases during this period.

Table 6-5. Summary of Daily Values for the Measured Meteorological Parameters During Phase II of Verification Testing

	Wind Speed (m/sec)	Wind Direction (Degrees)	Standard Deviation of Wind Direction (Degrees)	Air Temp. (C)	RH (%)	Solar Radiation (W/m ²)	Press. (mm Hg)
Average	1.43	186	34.2	8.3	75.4	88.2	756.2
Max	4.18	260	48.8	12.8	92.0	123.5	761.7
Min	0.91	116	21.3	4.6	51.6	17.1	747.3

Table 6-6. Summary of Daily Values for the Measured Precursor Gas Concentrations During Phase II of Verification Testing

	CO (ppm)	O ₃ (ppb)	NO (ppb)	NO ₂ (ppb)	NO _x (ppb)
Average	1.9	13	61.8	32.6	94.4
Max	3.3	28	119.9	50.3	170.2
Min	0.4	6	4.1	14.8	18.9

6.2.1 Inter-Unit Precision

As in Phase I, particle concentration readings were recorded every 15 minutes by the duplicate Model 3320 APS monitors. These particle concentrations were subsequently converted to PM_{2.5} mass concentration values by the APS software. The 15-minute mass concentrations from the two Model 3320 APS monitors for Phase II of the verification test are shown in Figure 6-4a. In Figure 6-4b, these data are plotted against one another to illustrate the correlation between the two monitors. Gaps in the data are periods during which no data were available from one or both Model 3320 APS monitors because of problems with the laptop computer used for data logging.

For comparison with the PM_{2.5} FRM reference measurements, the 15-minute data were averaged from midnight to midnight for each day to correspond with the 24-hour sampling periods used in Phase II of the verification test. In Figure 6-5a, the midnight-to-midnight averages for Phase II of the verification test are presented for the Model 3320 APS monitors. A correlation plot of these data is shown in Figure 6-5b.

The results of a linear regression analysis of these data are presented in Table 6-7. The CVs for the 15-minute and the midnight-to-midnight average values were also calculated and are shown in Table 6-7.

Regression analysis of the 15-minute data from the duplicate monitors shows an r^2 of 0.973, but indicates a bias between the two monitors, with Monitor 2 typically reading higher than Monitor 1 [slope = 1.234]. The regression results also show an intercept of 1.23 (0.34) $\mu\text{g}/\text{m}^3$. The calculated CV for the 15-minute data is 21.5%. Much of the calculated CV may be attributable to the bias between the duplicate monitors rather than to random differences in the readings.

Regression analysis of the 24-hour average concentration results for the duplicate monitors shows an r^2 of 0.998. These results show a slope of the regression line that is statistically different from unity at the 95% confidence level [slope = 1.280], indicating a bias between the two monitors. The calculated intercept is not statistically significant at 95% confidence. The 24-hour results show a CV of 18.8%, comparable to that of the 15-minute data.




Figure 6-4a. 15-Minute PM_{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase II of Verification Testing




Figure 6-4b. Correlation Plot of 15-Minute PM_{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase II of Verification Testing




Figure 6-5a. 24-Hour PM_{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase II of Verification Testing




Figure 6-5b. Correlation Plot of 24-Hour PM_{2.5} Concentrations from Duplicate Model 3320 APS Monitors During Phase II of Verification Testing

Table 6-7. Linear Regression and Coefficient of Variation Results for 15-Minute and 24-Hour Average PM_{2.5} Concentrations from Duplicate Model 3320 APS Monitors for Phase II

Parameter	15-Minute Data	24-Hour Average Data
Slope (95% CI)	1.234 (0.010)	1.280 (0.040)
Intercept (µg/m ³) (95% CI)	1.23 (0.34)	0.32 (0.98)
r ²	0.973	0.998
CV	21.5%	18.8%

6.2.2 Comparability/Predictability

In Figures 6-6a and 6-6b, the midnight-to-midnight averages of the Model 3320 APS measurements are shown, along with the PM_{2.5} FRM measurements for Phase II of the verification test. From Figure 6-6a it appears that the Model 3320 APS averages followed a similar temporal pattern as the FRM data, but in general the FRM averages exceeded the Model 3320 APS values, sometimes by more than a factor of two. A correlation plot of these 24-hour data is shown in Figure 6-6b. These PM_{2.5} concentration values were analyzed by linear regression according to Section 5.2 to establish the comparability of each of the Model 3320 APS monitors with the PM_{2.5} FRM sampler. The calculated slope, intercept, and r² value of the regression analyses are presented in Table 6-8 for each monitor.

The r² values of the regression analyses were 0.803 for Monitor 1 and 0.762 for Monitor 2. Both Monitors 1 and 2 showed a substantial negative bias relative to the FRM results. The slopes of the regression lines were 0.578 (0.121) and 0.555 (0.195), respectively. The intercept of the regression line for Monitor 2 was not statistically different from zero at the 95% confidence level. For Monitor 1, the regression results show an intercept of -14.4 (10.4) µg/m³.




Figure 6-6a. Midnight-to-Midnight Averages from Duplicate Model 3320 APS Monitors and the PM_{2.5} FRM Results During Phase II of Verification Testing




Figure 6-6b. Correlation Plot of 24-Hour Averages from Duplicate Model 3320 APS Monitors and the PM_{2.5} FRM During Phase II of Verification Testing

Table 6-8. Comparability of the Model 3320 APS Monitors with the PM_{2.5} FRM in Phase II

Regression Parameter	Monitor 1	Monitor 2
Slope (95% CI)	0.578 (0.121)	0.555 (0.195)
Intercept (µg/m ³) (95% CI)	-14.4 (10.4)	-9.2 (13.6)
r ²	0.803	0.762

The vendor has suggested that because of the installation configuration, substantial particle loss may have occurred during sampling, contributing to the negative bias relative to the FRM. However, the vendor performed the installation and approved that configuration before any verification testing took place. Furthermore, errors in assumptions about particle density or other conversion factors may be more likely causes of bias between the Model 3320 APS and the FRM.

6.2.3 Meteorological Effects

As with the data from Phase I, a multivariable model was used to determine if the meteorological conditions had an influence on the readings of the Model 3320 APS monitors in Phase II. This analysis involved a backward elimination process to remove from the model those parameters showing no statistically significant influence on the results. This analysis shows no influence of meteorology on the results of Monitor 2 at the 90% confidence level. For Monitor 1, the analysis shows the following relationship:

$$\text{Monitor 1} = 0.674 * \text{FRM} + 8.99 * \text{WS} - 34.3 \mu\text{g}/\text{m}^3$$

where FRM is the measured PM_{2.5} FRM values in µg/m³, and WS is the wind speed in meters per second. The magnitude of the implied effect can be estimated in comparison to the linear regression against the FRM data above (Table 6-8). For example, using the average values for PM_{2.5} and the various meteorological parameters during Phase II (Section 6.2), the above equation would predict an average PM_{2.5} reading of 28.4 µg/m³ for Monitor 1:

$$\begin{aligned} \text{Monitor 1} &= 0.674 * 74.0 + 8.99 * 1.43 - 34.3 \mu\text{g}/\text{m}^3 \\ &= 28.43 \mu\text{g}/\text{m}^3 \end{aligned}$$

Based on the linear regression results (Table 6-8) and the average PM_{2.5} concentration during Phase II, Monitor 1 would read

$$\begin{aligned} \text{Monitor 1} &= 0.578 * 74.0 - 14.4 \mu\text{g}/\text{m}^3 \\ &= 28.37 \mu\text{g}/\text{m}^3 \end{aligned}$$

i.e., a difference of only 0.2%.

6.2.4 Influence of Precursor Gases

Multivariable analysis was also performed to establish if a relationship exists between precursor gases (carbon monoxide, nitrogen dioxide, nitric oxide, total nitrogen oxides, and ozone) and the Model 3320 APS readings. This analysis shows that none of the measured gases influenced the readings of Monitor 2 at the 90% confidence level. For Monitor 1, the following relationship was determined:

$$\text{Monitor 1} = 0.633 \cdot \text{FRM} - 0.363 \cdot \text{NO} + 9.96 \cdot \text{CO} - 14.4 \mu\text{g}/\text{m}^3$$

where the concentrations of nitric oxide and carbon monoxide are in ppb and ppm respectively.

Substituting the average values for each of these parameters from Phase II (Section 6.2) into this equation, the predicted average value from the multivariable analysis would be:

$$\begin{aligned} \text{Monitor 1} &= 0.633 \cdot 74.0 - 0.363 \cdot 61.8 + 9.96 \cdot 1.9 - 14.4 \mu\text{g}/\text{m}^3 \\ &= 28.9 \mu\text{g}/\text{m}^3 \end{aligned}$$

i.e., a difference of less than 2% relative to the average of 28.4 $\mu\text{g}/\text{m}^3$ predicted by the linear regression against FRM data alone (Table 6-8).

6.2.5 Short-Term Monitoring


During Phase II of the verification test, short-term monitoring was conducted on a five-sample-per-day basis throughout the test period. Table 6-9 presents the averages and the ranges of $\text{PM}_{2.5}$ concentrations for these sampling periods during Phase II. Figure 6-7 shows the correlation between the time-weighted sum of the short-term measurements from the sequential filter sampler and the 24-hour FRM measurements. The slope and intercept of the regression line are 0.930 (0.077) and 2.2 (6.6) $\mu\text{g}/\text{m}^3$, respectively, with an r^2 value of 0.960, where the numbers in parentheses are 95% CIs.

Table 6-9. Summary of Short-Term PM_{2.5} Levels During Phase II of Verification Testing

PM _{2.5} Concentration (µg/m ³)	Sampling Period				
	0000-0500	0500-1000	1000-1300	1300-1600	1600-2400
Average	81.0	52.2	56.8	46.7	87.7
Maximum	163.2	131.4	140.9	136.6	180.7
Minimum	3.4	7.7	4.8	2.2	7.2

Figure 6-7. Correlation Plot of the Time-Weighted Average for the Short-Term Samples and the PM_{2.5} FRM

In Figure 6-8, the averages of the Model 3320 APS readings for all the short-term monitoring periods are plotted versus the corresponding PM_{2.5} concentration values from the reference sampler. Linear regression analysis of these data was performed separately for each Model 3320 APS, and the results are presented in Table 6-9. Regression analyses were also performed separately for each of the five time periods during which the short-term samples were collected (i.e., 0000-0500, 0500-1000, 1000-1300, 1300-1600, and 1600-2400). These regression results are also presented in Table 6-9.



The short-term monitoring results indicate a substantial negative bias between the Model 3320 APS monitors and the reference measurements, consistent with the results above for 24-hour FRM data. The magnitude of this bias differs among the five short-term monitoring periods. When all sampling periods are included in the analysis, the slopes of the regression lines are 0.489 for Monitor 1 and 0.556 for Monitor 2. When the individual sampling periods are analyzed separately, the slopes of the regression lines range from 0.405 to 0.777 for Monitor 1 and from 0.533 to 0.830 for Monitor 2. For both monitors, the agreement with the reference measurements was best (i.e., the slopes were closest to 1.0) for the sampling periods with the lowest PM_{2.5} concentrations (i.e., 0500-1000, 1000-1300, and 1300-1600). The correlation between the Model 3320 APS monitors and the reference measurements also was highest for these periods. For Monitor 1, the r^2 values ranged from 0.633 to 0.899 for the individual sampling periods and was 0.643 overall. For Monitor 2, the r^2 values ranged from 0.645 to 0.886, with an overall r^2 of

0.660. (It should be noted that the reference measurements have not been corrected to account for the observed difference between the time-weighted average of the short-term samples and the FRM.).

Figure 6-8. Correlation Plot of Short-Term Monitoring Results and the Corresponding Averages from the Duplicate Model 3320 APS Monitors During Phase II of Verification Testing

Table 6-10. Regression Analysis Results for the Short-Term Monitoring

Short-Term Monitoring Period	Monitor 1			Monitor 2		
	Slope	Intercept ($\mu\text{g}/\text{m}^3$)	r^2	Slope	Intercept ($\mu\text{g}/\text{m}^3$)	r^2
						0.660
All	0.489	-3.1	0.643	0.556	-3.3	
0000-0500	0.495	-6.6	0.633	0.586	-5.6	0.645
0500-1000	0.777	-13.7	0.814	0.830	-9.4	0.724
1000-1300	0.729	-9.2	0.899	0.824	-9.9	0.820
1300-1600	0.595	-2.7	0.838	0.802	-7.0	0.886
1600-2400	0.405	-9.4	0.729	0.533	-14.5	0.742

6.3 Instrument Reliability/Ease of Use

During instrument installation in Phase I of verification testing, one of the two Model 3320 APS monitors had problems with its sample flow rate. This monitor was returned to the factory for recalibration and was installed on site for approximately the last 10 days of sampling. No maintenance was performed on the other monitor from the time of installation on August 8. However, difficulties with the data collection system resulted in the loss of substantial data (approximately 10 days) from this monitor such that only approximately 12 days of data were captured in Phase I.

During Phase II, substantial problems with the data collection system were again encountered. These difficulties resulted in the loss of several periods of data for each monitor. These problems appeared to stem from the simultaneous use of the computer's two serial ports. The problems were eventually resolved by replacing the laptop computer that was originally supplied for the data collection. Other than the computer problems, no maintenance was performed on the two Model 3320 APS monitors. Data recovery during Phase II was approximately 90% for one monitor and approximately 50% for the other.

6.4 Shelter/Power Requirements

As recommended by the vendor, the Model 3320 APS monitors were installed and operated inside an instrument trailer during each phase of testing. Both instruments and their associated data collection computer were run on a single 15 A circuit. In general, the monitors were relatively easy to install (only 2 to 3 hours for both monitors) and ran mostly unattended during the testing. Again, the primary problems encountered were with the data collection system. Vendor literature indicates an operative temperature range of 10-40°C; however, these limits were not verified during this test.

6.5 Instrument Cost

The price of the Model 3320 APS as tested is approximately \$40,000.

Chapter 7

Performance Summary

The Model 3320 APS is a semi-continuous particle monitor designed to provide indications of the ambient particulate matter concentrations at time periods as short as several seconds. Duplicate Model 3320 APS monitors were evaluated under field test conditions in two separate phases of this verification test. The duplicate monitors were operated side by side. The results from each phase of this verification test are summarized below.

7.1 Phase I—Pittsburgh (August 1 - September 1, 2001)

Inter-unit precision was assessed in Phase I for both 15-minute and 24-hour data. The duplicate Model 3320 APS monitors were highly correlated with one another for the 15-minute data and 24-hour averages. Regression analysis showed $r^2 = 0.983$ for the 15-minute data and $r^2 = 0.972$ for the 24-hour averages. The slopes of the regression lines were 0.670 (0.005) and 0.627 (0.081), respectively, for the 15-minute data and 24-hour averages, indicating a substantial bias between the two monitors. The intercept of the regression line was 0.33 (0.07) $\mu\text{g}/\text{m}^3$ for the 15-minute data and 0.80 (0.96) $\mu\text{g}/\text{m}^3$ for the 24-hour data. The calculated CV for the 15-minute data was 25.2% and, for the 24-hour averages, the CV was 25.6%. Much of these CVs may be attributed to the bias between the monitors rather than to random differences in the readings.

Comparisons of the 24-hour averages with $\text{PM}_{2.5}$ FRM results showed r^2 values of 0.100 and 0.093, respectively, for Monitor 1 and Monitor 2. The slopes of the regression lines were 0.204 and 0.118, respectively, for Monitor 1 and Monitor 2, indicating a substantial negative bias of the monitors relative to the FRM in Phase I.

Multivariable analysis of the 24-hour average data showed no conclusive influence of meteorology on the readings of the two monitors relative to the FRM values at the 90% confidence level.

Multivariable analysis of the 24-hour average data also showed no conclusive influence of the ambient precursor gases on the readings of the two monitors relative to the FRM values at the 90% confidence level.

7.2 Phase II—Fresno (December 18, 2000 - January 17, 2001)

During Phase II, inter-unit regression analysis showed r^2 values of 0.973 and 0.998, respectively, for 15-minute and 24-hour average data from the duplicate Model 3320 APS monitors. The slopes of the regression lines were 1.234 and 1.280, respectively, for the 15-minute data and 24-hour averages, indicating a significant bias between the two monitors. The calculated CV for the 15-minute data was 21.5%; and, for the 24-hour averages, the CV was 18.8%.

Comparison of the 24-hour averages with $PM_{2.5}$ FRM results showed slopes of the regression lines for Monitor 1 and Monitor 2 of 0.578 (0.121) and 0.555 (0.195), respectively, indicating a substantial negative bias between the Model 3320 APS monitors and the FRM in Phase II. The regression results show r^2 values of 0.803 and 0.762 for Monitor 1 and Monitor 2, respectively. A statistically significant intercept of -14.4 (10.4) $\mu\text{g}/\text{m}^3$ was observed for Monitor 1, and an intercept of -9.2 (13.6) was observed for Monitor 2.

Multivariable analysis of the 24-hour average data showed no influence of meteorology on the readings of one monitor at the 90% confidence level. However, the multivariable model ascribed to wind speed had a statistically significant influence on the readings of the other monitor. Under typical conditions during Phase II, the multivariable results differed by only 0.2% from the simple linear regression against FRM results.

Multivariable analysis of the 24-hour average data showed that none of the ambient precursor gases had a statistically significant influence on one of the Model 3320 APS monitors, but that both carbon monoxide and nitric oxide had an effect on the other monitor. The combined effects of these gases, under typical conditions, had an effect of less than 2% on the readings of Monitor 1 during Phase II, relative to the linear regression against FRM results.

In addition to 24-hour FRM samples, short-term sampling was performed on a five-sample-per-day basis in Phase II. The Model 3320 APS results were averaged for each of the sampling periods and compared with the gravimetric results. Linear regression of all of these data showed slopes of 0.489 and 0.556, respectively, for Monitor 1 and Monitor 2, consistent with the negative bias seen relative to the 24-hour FRM results. For the individual sampling periods, both the Model 3320 APS correlation and slope relative to the FRM were best during periods with the lowest $PM_{2.5}$ concentrations.

7.3 Other Parameters

The two monitors required little maintenance during either phase of testing, with the exception of recalibration of the sample flow rates of one monitor during Phase I. Substantial difficulties associated with data collection resulted in the loss of a considerable amount of data during the two phases of testing. Delays in instrument installation and losses from the data system resulted in capture of approximately 10 days of data from one monitor and 12 days from the other. Data recovery during Phase II was approximately 90% for one monitor and approximately 50% during Phase I for the other.

Chapter 8

References

1. *Test/QA Plan for the Verification of Ambient Fine Particle Monitors*, Battelle, Columbus, Ohio, June 2000.
2. “National Ambient Air Quality Standards for Particulate Matter; Final Rule,” U.S. Environmental Protection Agency, 40 CFR Part 50, *Federal Register*, 62 (138):38651-38701, July 18, 1997.
3. *Quality Management Plan (QMP) for the Advanced Monitoring Systems Pilot*, Version 2.0, Battelle, Columbus, Ohio, October 2000.
4. “Quality Assurance Requirements for State and Local Air Monitoring Stations (SLAMS).” Appendix A to 40 CFR Part 58, *Federal Register*, 62 (138), p.65, July 18, 1997.