

Subject/Title: ASI VDRL ANTIGEN TEST		Doc#: 6004-950 NCCLS
Effective Date: 11/01	Supersedes Revision/Date: 02/00	Revision: 11/01
Prepared by: ASI	QA Approval by:	Copy/Dept.:

FOR IN VITRO DIAGNOSTIC USE

Cat. No. 950010

- 1 **INTENDED USE:** The **ASI VDRL Antigen Test** is a qualitative and semiquantitative nontreponemal flocculation test for the detection of reagin antibodies in human serum. These materials are intended to be acquired, possessed and used only by health professionals.

- 2 **SUMMARY AND EXPLANATION:** *Treponema pallidum*, the etiological agent of syphilis, induces the production of at least two types of antibodies in human infection: anti-treponemal antibodies that can be detected by FTA-ABS antigen (1), and anti-nontreponemal antibodies (reagin) that can be detected by the VDRL test (2).

- 3 **PRINCIPLE OF THE PROCEDURE:** The **ASI VDRL Antigen Test** is a microscopic nontreponemal flocculation test to be used for the detection of reagin. The procedure is based on the VDRL antigen being combined at a correct ratio with buffered saline and then mixed with heat-inactivated serum.

- 4 **REAGENTS**
 - 4.1 ASI VDRL ANTIGEN - 0.03% Cardiolipin and 0.9% Cholesterol in absolute alcohol. Sufficient Lecithin (approximately 0.20 to 0.22%) is added to produce standard reactivity.
 - 4.2 ASI VDRL BUFFERED SALINE - Phosphate-buffered saline, pH 5.9-6.1, containing 0.05% formaldehyde as preservative.

- 5 **WARNINGS AND PRECAUTIONS**
For *In Vitro* Diagnostic Use
 - 5.1 ASI VDRL ANTIGEN is highly flammable and is irritating to eyes, respiratory system and skin. There is possible risk of irreversible effects and there is risk to an unborn child. Avoid contact with skin and eyes. Do not breathe aerosols. Wear suitable protective clothing. Keep container tightly closed. Keep away from sources of ignition. No smoking. Target organs are blood, intestines, liver, muscle and nervous tissue.
 - 5.2 Observe universal precautions in handling and disposing of the specimens utilized in this test. The CDC/NIH Health Manual "Biosafety in Microbiological and Biomedical Laboratories" describes how these materials should be handled in accordance with Good Laboratory Practice (3).
 - 5.3 Do not pipet by mouth.
 - 5.4 Do not smoke, eat, drink or apply cosmetics in areas where serum samples are handled.

5.5 Any cuts, abrasions or other skin lesions should be suitably protected.

Title: ASI VDRL ANTIGEN TEST		Doc#: 6004-950 NCCLS
Effective Date: 11/01	Supersedes Revision/Date: 02/00	Revision: 11/01

6 HANDLING AND PROCEDURAL NOTES

- 6.1 In order to obtain reliable and consistent results, the instructions in the package insert must be strictly followed. Do not modify the handling and storage conditions for reagents or samples.
- 6.2 Do not use past the expiration date indicated on the kit.
- 6.3 Do not interchange components of one kit with those of another kit.
- 6.4 Keep the VDRL ANTIGEN and VDRL BUFFERED SALINE tightly closed at all times to prevent evaporation.
- 6.5 All glassware, needles and syringes must be clean and dry before use. Rinse all equipment with water, alcohol and acetone in this specific order.
- 6.6 Do not use glass slides with concavities, wells or glass rings.

7 **STORAGE INSTRUCTIONS:** Store VDRL ANTIGEN and BUFFERED SALINE at room temperature (15-30°). The VDRL ANTIGEN should be protected from light.

8 INDICATIONS OF DETERIORATION

- 8.1 Turbidity or precipitation in controls is indicative of deterioration and the control should not be used.
- 8.2 Bacterial contamination of reagents or specimens may cause false positive results.
- 8.3 Any visible discoloration of VDRL ANTIGEN or VDRL BUFFERED SALINE may be indicative of deterioration and the reagent should not be used.

9 SPECIMEN COLLECTION AND STORAGE

- 9.1 Only serum is suitable for use in this test. Plasma is not acceptable.
- 9.2 Samples may be maintained in their original tubes at 2-8°C for up to four (4) hours. If longer storage is required, the serum must be separated from the red cells and stored at -15°C or below.
- 9.3 Frozen samples must be thawed at room temperature before use.
- 9.4 Samples should be free from bacterial contamination, hemolysis or lipemia.
- 9.5 If necessary before testing, centrifuge the specimens at a force sufficient to sediment cellular components.
- 9.6 Samples to be sent out for testing should be placed on ice packs and packaged like any other biohazardous material that could potentially transmit infection.

Title: ASI VDRL ANTIGEN TEST		Doc#: 6004-950 NCCLS
Effective Date: 11/01	Supersedes Revision/Date: 02/00	Revision: 11/01

10 MATERIALS PROVIDED

VDRL ANTIGEN	2 x 5 ml
VDRL BUFFERED SALINE	2 x 60 ml

11 ADDITIONAL MATERIALS REQUIRED

- 11.1 Mechanical rotator adjustable to 180±5 rpm and circumscribing 3/4 inch diameter, with humidity cover
- 11.2 VDRL control sera: reactive, weak reactive, nonreactive
- 11.3 Saline (0.9% NaCl solution)
- 11.4 Non-disposable calibrated 18-gauge needle without bevel
- 11.5 Non-disposable glass syringe, 1 ml or 2 ml
- 11.6 Bottles, 30 ml, round, narrow-mouth, approximately 35 mm in diameter with ground glass stoppers and flat inner bottom surfaces
- 11.7 Slides, 2 x 3" with rings approximately 14 mm in diameter. The rings can be paraffin or ceramic, but must be sufficiently high to prevent spillage during rotation.
- 11.8 Micropipettor, calibrated to deliver 50 µl
- 11.9 Pipets, glass serological: 1 ml in 1/10 increments, 5 ml in 1/10 increments, 10 ml in 1/10 increments
- 11.10 Timing device, minute and second capability
- 11.11 Microscope capable of 100x magnification

12 TEST PROCEDURE

12.1 PREPARATION OF THE VDRL ANTIGEN SUSPENSION

- 12.1.1 The VDRL ANTIGEN SUSPENSION must be prepared fresh each day.
- 12.1.2 The temperature of the VDRL BUFFERED SALINE, VDRL ANTIGEN and equipment should be between 23 and 29°C before preparing the suspension. All glassware, pipets and equipment must be dry.
- 12.1.3 With a 1 ml serological pipet, deliver 0.4 ml of VDRL BUFFERED SALINE to the bottom of a 30 ml round, glass stoppered bottle with flat inner bottom.
- 12.1.4 Pipet 0.5 ml of VDRL ANTIGEN gradually into the VDRL BUFFERED SALINE while continuously but gently rotating the bottle on a flat surface. Add the VDRL ANTIGEN drop by drop at a rate allowing approximately 6 seconds for 0.5 ml of VDRL ANTIGEN. Keep the pipet tip in the upper third of the bottle. Do not splash the BUFFERED SALINE onto the pipet. The proper rotation speed is obtained when the center of the bottle circumscribes a 2-inch (5 cm) diameter circle approximately three (3) times per second. Expel the last drop of VDRL ANTIGEN from the pipet without touching the pipet to the BUFFERED SALINE, and continue rotation of the bottle for 10 seconds.
- 12.1.5 Add 4.1 ml of VDRL BUFFERED SALINE from a 5 ml pipet.
- 12.1.6 Cap the bottle and shake it from bottom to top and back approximately 30 times in 10 seconds. The VDRL ANTIGEN SUSPENSION is ready for use and is usable for 8 hours.
- 12.1.7 Mix the VDRL ANTIGEN SUSPENSION by gently swirling it each time it is used. Do not mix the SUSPENSION by forcing it back and forth through the syringe and needle since this may cause breakdown of particles and loss of reactivity.
- 12.1.8 To achieve reliable and reproducible test results, the VDRL ANTIGEN SUSPENSION, controls and test specimens must be at 23-29°C when the tests are performed.

12.2 PREPARATION OF THE SAMPLES

- 12.2.1 Each serum sample must be heat-inactivated for 30 minutes at 56°C prior to testing.

12.2.2 If heat-inactivation occurs more than four (4) hours prior to testing, reheat the serum for an additional 10 minutes at 56°C.

Title: ASI VDRL ANTIGEN TEST		Doc#: 6004-950 NCCLS
Effective Date: 11/01	Supersedes Revision/Date: 02/00	Revision: 11/01

12.3 ASSAY PROTOCOL B QUALITATIVE

- 12.3.1 Using a micropipettor, pipet 50 µl of serum into one 14 mm test circle.
- 12.3.2 Gently resuspend the VDRL ANTIGEN SUSPENSION.
- 12.3.3 Draw a sufficient volume of VDRL ANTIGEN SUSPENSION into the needle and glass syringe assembly. Dispense several drops into the 30 ml SUSPENSION bottle to make sure the passage is clear.
- 12.3.4 Holding the VDRL ANTIGEN SUSPENSION dispensing needle and syringe in a vertical position, dispense exactly one (1) free-falling drop of ANTIGEN SUSPENSION into each circle containing serum.
- 12.3.5 Place the slide onto the mechanical rotator and cover to maintain humidity. Rotate the slide at 180±5 rpm for four (4) minutes.
- 12.3.6 Immediately after rotating the slide, remove it from the rotator and read the test microscopically, using 100x magnification. Record the results.

12.4 ASSAY PROTOCOL B SEMIQUANTITATIVE

- 12.4.1 To quantitate serum samples for determination of endpoint, dilutions can be prepared directly on the glass slide.
- 12.4.2 Using a micropipettor, dispense 50 µl of saline into the circles numbered 2-4. Do not spread.
- 12.4.3 Dispense 50 µl of serum onto circles 1 and 2 of the glass slide.
- 12.4.4 Mix the saline and the serum in circle 2 by drawing the mixture up and down in the pipet 5 or 6 times. Avoid any bubble formation.
- 12.4.5 Transfer 50 µl from circle 2 to circle 3 and mix as in step (4) above. Repeat this serial dilution procedure to circle 4 and discard 50 µl from the last circle. Circles 1 through 4 represent a dilution series as follows:

Circle:	1	2	3	4
Dilution:	1:1	1:2	1:4	1:8

- 12.4.6 Gently resuspend the VDRL ANTIGEN SUSPENSION in the 30-ml bottle, and draw sufficient volume into the syringe and needle assembly.
- 12.4.7 Holding the VDRL ANTIGEN SUSPENSION dispensing needle and syringe in vertical position, dispense several drops into the 30-ml bottle to clear the needle of air. Then add exactly one (1) free-falling drop of ANTIGEN SUSPENSION to each circle.
- 12.4.8 Place the slide onto the automatic rotator and cover to maintain humidity. Rotate the slide at 180±5 rpm for four (4) minutes.
- 12.4.9 Immediately after rotation, read the test microscopically using 100x magnification.
- 12.4.10 If the highest dilution tested (1:8) is reactive, further serial dilutions must be performed until endpoint is observed.

13 **QUALITY CONTROL:** Controls with graded reactivity should be included in each test run to confirm optimal reactivity of the ANTIGEN SUSPENSION. If control samples do not yield the expected response, the assay should be considered invalid and the assay repeated. If the repeat assay does not elicit the expected results for the control

Title: ASI VDRL ANTIGEN TEST		Doc#: 6004-950 NCCLS
Effective Date: 11/01	Supersedes Revision/Date: 02/00	Revision: 11/01

14 INTERPRETATION OF RESULTS

14.1 INTERPRETATION OF RESULTS- QUALITATIVE

Read and record the results as follows:

Medium to large clumps	Reactive (R)
Small clumps	Weakly Reactive (W)
No clumping or very slight roughness	Nonreactive (N)

14.2 INTERPRETATION OF RESULTS- SEMIQUANTITATIVE

The highest dilution that produces a reactive, not weakly reactive, result is the endpoint titer. In the following example, the titer reported would be 1:2.

TITER	1:1	1:2	1:4	1:8	1:16
RESULT	RR	R	W	N	N

15 LIMITATIONS OF THE PROCEDURE

- 15.1 Biological false positive reactions occur occasionally with the VDRL test. Such reactions sometimes occur in samples from individuals with a history of drug abuse, or with diseases such as lupus, erythematosus, malaria, vaccinia, mononucleosis, leprosy, viral pneumonia and after smallpox vaccinations.
- 15.2 Prozone reactions may occur in which reactivity with undiluted serum is inhibited. The prozone phenomenon often gives weakly reactive or "rough" nonreactive results in the qualitative test. Therefore, all specimens with these results must be quantitatively tested until endpoint is reached or until no reactivity is observed.
- 15.3 In manufacturing, the VDRL ANTIGEN and BUFFERED SALINE are tested only with serum. A modification of the serum test reagents and procedures for CSF testing appears in the "Manual of Tests for Syphilis" (2). The user is responsible for modifying the reagents and procedures and for provision of the required quality control materials according to this manual.
- 15.4 Pinta, yaws, bejel and other treponemal diseases produce positive reactions in this test (2).
- 15.5 The results of the **ASI VDRL Antigen Test** must be confirmed by a treponemal test.
- 15.6 In accord with all diagnostic methods, a final diagnosis should not be made on the result of a single test, but should be based on a correlation of test results with other clinical findings.

16 EXPECTED VALUES

The **ASI VDRL Antigen Test** is evaluated for the equivalence, in its pattern of reactivity, against the CDC Reference VDRL Antigen Suspension.

17 REFERENCES

1. Hunter EF, Deacon WE, Myer PE. 1964. *Public Health Reports*, **79**:410-412.
2. Larsen SA, Hunter EF, Kraus SJ (ed.). 1990. *Manual of Tests for Syphilis*, Public Health Service, Washington, D.C.
3. *Biosafety in Microbiological and Biomedical Laboratories*, 3rd ed. 1993. HHS Publication No. (CDC) 93-8395, Public Health Service, Washington, D.C.