Validity Testing Information Part I

Laboratory Name: Clinical Laboratory Partners

Address: 129 East Cedar Street, Newington, CT 06111

Responsible Person: Dr. Alan Wu (Printed Name)

I certify that the answers and information provided are true and correct as of this date. Any false, fictitious, or fraudulent answers or information provided may violate Federal Law and could subject me to prosecution, monetary penalties, or both (Sec 18 U.S.C. 1001; 31 U.S.C. 3801-812).

Signature, Responsible Person

Date

Ha Wa

Man Wu Printed Name, Responsible Person

Validity Testing Information Part II

Conduct an audit of all DOT regulated specimens from the date your laboratory started validity testing. Summarize your findings in an Excel spread sheet in both hard copy and electronic format. Provide the following information in a separate column of the spreadsheet/audit for each DOT regulated specimen that was reported either adulterated or substituted:

- Specimen ID number
- Laboratory Accession Number
- Date of receipt
- · Date reported
- Reported result (i.e., adulterated or substituted)
- Quantitative test result (e.g., actual creatinine concentration and specific gravity reading; actual pH reading; adulterant identity and its concentration if applicable)

Note: Retain a copy of this information to ensure that you would be able to retrieve additional data.

I certify that the answers and information provided are true and correct as of this date. Any false, fictitious, or fraudulent answers or information provided may violate Federal Law and could subject me to prosecution, monetary penalties, or both (Sec 18 U.S.C. 1001; 31 U.S.C. 3801-812).

Mul	10-10-0
Signature, Responsible Person	Date
Alan Wr	
Printed Name, Responsible Person	

samhsatnp

	T	$\neg \neg$
1D Number	Specimen	
Number	Accession	
Date	Receipt	
Reported	Date	
Results	Reported	
	Quantitative Findings	



National Laboratory Certification Program


January 24, 2001

0647
Dr. Alan H.B. Wu
Clinical Laboratory Partners, LLC
Toxicology Laboratory
129 East Cedar Street
Newington, CT 06111

Dear Dr. Wu:

The enclosed critique was developed from the inspection reports of the inspectors who conducted the fifth maintenance inspection of your laboratory under the National Laboratory Certification Program (NLCP). Based upon our review of these reports, the laboratory's performance in this inspection appeared to meet most of the Minimum Technical Criteria. However, the laboratory must take steps to correct the following issues raised:

Dr. Wu Page 2 of 3 01/24/01



The laboratory must submit, within 30 calendar days of receipt of this letter, documentation to demonstrate that corrective actions have been implemented to address the issues raised. In responding to these issues, please organize the material in your document in accordance with the sections and item numbers as listed in this correspondence. Once these issues have been successfully addressed, RTI will recommend to the Department of Health and Human Services (HHS) that the laboratory's certification be continued. The laboratory must also review the enclosed critique and take all necessary corrective actions. All corrective actions must be implemented within 30 days receipt of this correspondence and will be reviewed at the next inspection.

If you have any questions or if we can be of further assistance, please call me at (919) 541-6176 or Dr. Michael Baylor at (919) 541-7043.

Sincerely,

Susan Crumpton

NLCP Technical Analyst

Swan Cyton

Enclosure

cc: Project Files/M5

NATIONAL LABORATORY CERTIFICATION PROGRAM

Document Review and Critique

Laboratory I.D. Number: <u>0647</u> Document No. Final

Laboratory:

Clinical Laboratory Partners, LLC

Location:

Newington, CT

Document Reviewed:

[] Application Form [XX] Inspection Report #M5

Date: 9 November 2000

[] Other_

Status:

[XX] Appeared to meet most of the minimum technical criteria

[] Appeared to meet a number of the minimum technical criteria

Failed to meet a number of the minimum technical criteria

Failed to meet a significant number of the minimum technical criteria

A review of the National Laboratory Certification Program (NLCP) inspection reports has been completed. The laboratory has appeared to meet most of the minimum technical criteria required for the inspection phase of the Program.

Deficiencies identified as a result of the inspection are detailed on the following pages. The laboratory is required to correct the deficiencies before its next inspection.

The following deficiencies were identified, and appear in the same order as the corresponding questions in the Laboratory Inspection Report:

Section E. Standard Operating Procedures - Procedures Manual

Section F. Chain-of-Custody, Accessioning, and Security

Section G. Quality Control and Quality Assurance

Section H. Initial Tests

Section I. Specimen Validity Tests

Section J. Confirmatory Tests

Section K. Records Audit

\/or	Final
Ver.	rınaı

Lab ID# 0647

Section L. Certification and Reporting

Section M. Laboratory Information Management System (LIMS)

Section N. Personnel