



ORAL MEDICATION SYRINGE DISPENSERS AND PROTECTIVE CAPS

By Audrey Morrison, RN, BSc in Nursing

The following are two cases that the Food and Drug Administration received through its Medical Device Reporting (MDR) system.

Case 1. A family member was given a standard hypodermic syringe with a luer lock that had a removable tip shield to administer an oral suspension to a 5-month-old infant. Inadvertently, the family member administered the medication to the infant without removing the tip shield. The tip ejected and lodged in the infant's airway. Attempts to remove the tip shield were unsuccessful. The infant subsequently died.

Case 2. In this injury report, a single dose of liquid antibiotic was sent home with the grandmother of a patient. The medication was put into an oral medication syringe and a protective cap was put on the end of the syringe. When the grandmother administered the medication, she was unaware that the syringe was fitted with a protective cap. When she activated the syringe, the tip ejected and lodged in the infant's throat.

What went wrong?

In the first case, the healthcare practitioner gave the family member a standard hypodermic syringe with a luer lock that was not intended for administering oral medications. This type of syringe did not provide complete closure around the tip shield and allowed fluid to be drawn up into the syringe with the tip shield in place. The translucent tip shield was similar to the body of the syringe and appeared to be part of the syringe. When the medication was administered with the tip shield in place, the tip shield ejected and lodged in the baby's airway. In the second case, the proper type of syringe was used but the caregiver was not aware that she should remove the protective cap. In both cases, apparently incomplete instructions were given to the caregiver on the safe use of the syringe for oral administration of medications.

What precautions can you take?

- Be aware of the difference between hypodermic syringes for injection and oral medication syringe dispensers.
- Use only syringes labeled "FOR USE ONLY BY MOUTH" when administering liquid oral medications.
- Provide users with oral **and** written instructions for use and reinforce with practice demonstrations to insure the user understands the instructions.
- Reinforce the following with the parent or caregiver:
 - Read all instructions for use before administering oral liquid medications.
 - Remove and discard any plastic caps, covers, or tip shields before use.
 - Keep all medications and dispensing equipment out of reach of children.
- Alert caregivers to the potential for choking with tip shields and caps if accidentally ingested or lodged in the mouth or throat.

Audrey Morrison, RN, BSc in Nursing, is a nurse-consultant in CDRH's Office of Surveillance and Biometrics.

In This Issue:	
Oral Medication Syringe Dispensers and Protective Caps	1
Olympus America Issues Voluntary Recall of Suspect Video Bronchoscopes	2
Effects of Repeated Ethylene Oxide Sterilization on Synthetic Absorbable Sutures	3
FDA Safety Alert: Recall of Ob/Gyn and Surgical Products Manufactured by A & A Medical/Rocket USA/LifeQuest	5
A&A Medical, Inc., LifeQuest Medical, Inc., and Rocket USA (A&A) Recall Information	6
FDA Patient Safety News	7

OLYMPUS AMERICA ISSUES VOLUNTARY RECALL OF SUSPECT VIDEO BRONCHOSCOPES

On November 30, 2001, Olympus America, Inc., of Melville, New York initiated a voluntary recall of their bronchoscopes after a medical facility reported that bacteria growing in video bronchoscopes may have been responsible for a patient infection requiring medical treatment. A bronchoscope is a flexible tube with a small light and camera used to evaluate the airway and lungs to diagnose or rule out respiratory problems such as pneumonia and lung cancer.

Olympus issued a recall notification letter to hospitals and medical facilities about the suspect bronchoscopes on November 30, 2001, requesting they be returned to Olympus for modification.

The biopsy port on the video bronchoscope is not intended to be removed from the bronchoscope. However, the medical facility, which reported one illness

among patients treated with the bronchoscopes, discovered that the biopsy ports on the bronchoscopes became loose and could be removed by twisting. Olympus America is repairing the bronchoscopes by applying adhesive to the biopsy port and is replacing the plastic biopsy port with a port made from stainless steel.

On February 27, 2002, Olympus issued a second recall notification letter to facilities that did not respond to their first letter.

The following are the Olympus Bronchoscope Models that have been recalled:

BF-40, BF-P40, BF-IT40, BF-3C40, BF-XP40, BF-XT40, BF-240, BF-P240, BF-IT240, BF-6C240, BF-160, BF-P160, BF-IT160, BF-3C160, BF-XT160.

The bronchoscopes were distributed nationwide and to Canada, Mexico, Dominican Republic, Argentina, Brazil, Panama, Chile, Peru, Ecuador, Columbia, Venezuela, Paraguay, Uruguay, Costa Rica, and El Salvador between June 5, 1997 and December 10, 2001.

FDA continues to work with Olympus to get this problem resolved.

If you recently had a lung or airway examination with a bronchoscope and are experiencing respiratory problems, FDA recommends that you contact your physician. In addition, you should report your problem to FDA's MedWatch Reporting Program by telephone, 1-800-332-1088 or on line at <http://www.fda.gov/medwatch>.

For additional information the recall contact at Olympus is:

Laura Storms-Tyler,
Director, Regulatory Affairs and Quality Assurance
Olympus America Inc.,
Two Corporate Center Drive,
Melville, NY 11747-3157
(631) 844-5688

To contact the Consumer Staff in FDA's Center for Devices and Radiological Health call 1-888-463-6332. When prompted, press 2, press 1, press 3, press 1. To speak to a Consumer Affairs Specialist (from 8:00 AM to 4:30 PM EST), press 5. To request information after business hours, press 4 and leave a message.

Why Are We So Excited?

Because the CDRH Contacts Listing is ready for you to join!

Visit: <http://www.fda.gov/cdrh/contactslisting>





CDRH often needs to find people who have an interest in medical devices or radiation-emitting products, so we can invite them to discuss policy issues about these devices. To do that, we maintain a data base of people interested in interacting with us. Then we can get in touch with you when an issue we want to discuss comes up.

Sign up at no cost to help CDRH:

- examine device issues,
- develop device policy, and
- inform you about devices relating to your specific interests.


Choose your:

- areas of interest
- geographic region
- specialties and affiliations







CDRH finds you by your interests and sends you Email, on timely issues like:

- upcoming teleconference viewing locations,
- recalls and safety issues,
- new web sites that pertain to your particular interest area,
- collaboration with you to research a particular issue.



Food and Drug Administration
 Center for Devices and Radiological Health
 Office of Health and Industry Programs



EFFECTS OF REPEATED ETHYLENE OXIDE STERILIZATION ON SYNTHETIC ABSORBABLE SUTURES*

By Terry O. Woods, Ph.D., Stanley A. Brown, D.Eng., Katharine Merritt, Ph.D.,
Scott G. McNamee, Ph.D., and Victoria M. Hitchins, Ph.D.

Editor's note: Although the August 14, 2000 Enforcement Priorities Guidance [1] exempts "open-but-unused" single-use devices (SUDs), the effects of resterilizing SUDs raise serious concerns.

Abstract

Scientists in the Food and Drug Administration's laboratory conducted a study to determine the effect of repeated ethylene oxide (EO) sterilization on sutures that had been opened-but-not-used. Four types of commonly used synthetic absorbable sutures were subjected to 1 and 2 EO resterilization cycles. Knot tensile strength was determined for new sutures and for sutures that had been subjected to 1 and 2 EO resterilization cycles. As has been found with other types of single-use devices, no general conclusions can be made for absorbable sutures. The strengths of different types of sutures increased, decreased, or stayed the same after repeated sterilization. In addition, the inner packages of some sutures were not intact after resterilizing, possibly exposing the sutures to increased humidity. This humidity can produce degradation leading to loss of strength immediately after exposure, after additional shelf aging, and after clinical use.

Introduction

Many absorbable sutures are provided with two layers of packaging: an outer layer that maintains sterility and an inner layer that provides a moisture barrier. Sutures with opened outer packages and intact inner packages, often referred to as opened-but-unused devices [1], are commonly resterilized for reuse although labeled for single use only. Mechanical properties of the polymers used in many sutures can be affected by sterilization methods.

To investigate the possible effects of resterilizing on these single-use devices (SUD), we examined the effect of repeated ethylene oxide (EO) sterilizations on the knot tensile strength of four types of synthetic absorbable sutures. Ethylene oxide sterilization of the sutures was done by the staff at the Materials Management Center at the National Institutes of Health Clinical Center in Bethesda, Maryland.

Methods

The following table summarizes the four types** of size 1 synthetic absorbable sutures that were examined.

Name	Structure	Composition
Dexon II	coated, braided	polyglycolic acid
Vicryl	coated, braided	copolymers of glycolide & lactide
Polysorb	coated, braided	copolymers of glycolide & lactide
PDS	monofilament	polydioxanone

The sutures were sterilized 1 or 2 times with EO (90:10 sterilant mixture with 90% hydrochlorofluorocarbons and 10% EO 75-21-8) for 130 minutes at 54^N-56^N C and


50% - 65% relative humidity, followed by a 12 hour aeration period. Before each EO cycle, the outer packages were removed and the inner packages were placed into new outer packages to simulate handling of opened-but-not used sutures. All packages were carefully handled to avoid mechanically altering or damaging them.

Mean knot tensile strength was determined for the new samples sutures and after 1 or 2 EO resterilization cycles. Suture strength and knotting of the test strand was determined according to United States Pharmacopeial (USP) <881>. [2] All new samples met the USP limit on average knot pull strength given in the USP official monograph on absorbable surgical suture. [3]

Results

After resterilization using EO, the out-of-package mean knot tensile strength for different types of synthetic absorbable sutures showed a range of responses: increasing, decreasing, or staying the same for different suture types (Figure 1). However, all tested sutures, regardless of EO exposure, met USP limits on average knot pull strength.

Figure 1. Mean knot tensile strength for four types of synthetic absorbable sutures new and after resterilization by EO once or twice. USP limit is 5.08 kgf for size 1 synthetic absorbable sutures




The two-tailed Student's t-test was used for statistical analyses. The test showed no difference in the strength of Dexon II and Vicryl initially and after 1 and 2 EO resterilization cycles. The strength of the Polysorb sutures was not affected after 1 EO cycle but decreased significantly after 2 EO cycles. The PDS II sutures showed a significant increase in strength after 1 EO cycle and an additional increase in strength after 2 EO cycles that was not statistically significant when compared to the 1 EO cycle.

The EO process had readily apparent effects on the suture inner packages.

SUTURES - From Page 3

- Visual observation revealed that some inner packages were not intact after resterilizing (Figure 2).

Figure 2: Resterilized suture inner package with ruptured seal.



- The seals on a number of the inner packages were breached, showing gaps in the seals that were easily visible without magnification.
- Some resterilized inner packages (after both 1 and 2 EO cycles) were visibly puffier than new packages and appeared to have been inflated. When the puffy inner packages were compressed with the fingers before opening, some deflated, indicating that the seal had broken.
- Most of the inner packages on the resterilized sutures contained wrinkles or creases that were not present on the new packages.
- Delamination was observed between the foil pack and an external polymer layer for a number of packages.
- Behavior of the adhesive seal on some of the resterilized packages (again after both 1 and 2 EO resterilization cycles) was different than that of new packages. When the resterilized packages were opened, some tore across the foil package, while all of the new packages separated at the adhesive seal.

Discussion and Conclusions

The testing highlighted a number of concerns related to resterilizing opened-but-unused sutures.

- Since the sterilization method and expiration date do not always appear on the inner package, a reprocessor may not have this information about an individual suture package. Even when the initial sterilization method is known, a reprocessor may not know

the details of the sterilization protocol used on a given type of suture. In fact, for the suture types tested, package labeling indicated the method of sterilization for only one type and no details of the sterilization process were given.

- Seals on some inner packages were destroyed during resterilizing, exposing the absorbable sutures to humidity different than provided in the original packaging. The primary mechanism for degradation of the tested sutures is hydrolysis [4]. Loss of seal integrity might not cause an initial strength loss, as was observed in two of the tested suture types. However, for all absorbable sutures with the inner seal destroyed during repackaging, exposure to increased humidity for an extended time will cause suture degradation. This leads to a loss of strength after shelf aging and a loss of strength and possible changes in the degradation behavior after clinical use. These changes in the suture material could result in wound dehiscence or other complications.
- The change in strength of these sutures implies potential changes in the *in vivo* degradation behavior of these polymers. Sutures that become stronger after resterilizing may degrade more slowly due to an increase in the crystalline fraction of the polymer. Sutures that become weaker could degrade faster. Further laboratory testing is needed to confirm or refute these concerns.
- For some types of sutures the original expiration date is not listed on the inner suture package. If the expiration date is related to the mechanical characteristics of the suture or to the package that insures its integrity rather than its sterility, the original date may be lost. If the expiration date is related to an assurance of sterility, then a new expiration date should be applied.

The most significant conclusion to be drawn from this study is the observation that, as is true for other types of devices [5], it is not possible to make general conclusions about the

effects of reesterilized on absorbable sutures. Suture strength was not affected for some sutures; others increased, and others decreased in strength with repeated EO sterilization cycles.

References

1. U.S. Food & Drug Administration, Center for Devices & Radiological Health, Office of Compliance. *Guidance for Industry and for FDA staff: Enforcement Priorities for Single-Use Devices Reprocessed by Third Parties and Hospitals*, 8/14/2000, <http://www.fda.gov/cdrh/reuse/1168.html>.
2. *The United States Pharmacopeial Convention. USP 24-NF 19 Supplement 2*. Rockville, MD: The United States Pharmacopeial Convention, July 1, 2000; 2903-4.
3. *The United States Pharmacopeial Convention. USP 24-NF 19 Supplement 2*. Rockville, MD: The United States Pharmacopeial Convention, July 1, 2000; 2859.
4. Middleton, J.C. and Tipton, A.J., Jr. *Synthetic Biodegradable Polymers as Medical Devices*. Medical Plastics and Biomaterials. 1998; Vol.5 No. 2. March-April:30-39.
5. Brown, S.A., Merritt, K., Woods, T.O., and Hitchins, V.M., "The Effects of Use and Simulated Reuse on PTCA Balloons and Catheters," *Biomedical Instrumentation & Technology* 2001;35:312-322.

Terry O. Woods, Ph.D., Stanley A. Brown, D.Eng., Katharine Merritt, Ph.D., Scott G. McNamee, Ph.D., and Victoria M. Hitchins, Ph.D. are scientists in CDRH's Office of Science and Technology.

*Adapted from an article by the authors published in *Biomedical Instrumentation & Technology* 2001;35:391-94.

**Polysorb, United States Surgical Corporation, Norwalk, CT; Dexon II, Davis+Geck, Division of American Cyanamid Company, Wayne, NJ; Vicryl, Ethicon, Inc., Johnson & Johnson, Somerville, NJ; and PDS II, Ethicon, Inc., Johnson & Johnson, Somerville, NJ. Representative products and manufacturers are named for identification only and the list does not imply endorsement by the U.S. Department of Health and Human Services.

**FDA Safety Alert:
Recall of Ob/Gyn and Surgical Products
Manufactured by A & A Medical/Rocket USA/LifeQuest**

(You are encouraged to copy and distribute this information)

March 15, 2002

To: Hospital Administrators
Risk Managers
Director, Central Supply
Director, Department of Obstetrics and Gynecology
Director, Department of Surgery
Ambulatory Surgical Centers

I am writing to alert you to a recall of all medical devices manufactured by A & A Medical, Inc. of Alpharetta, GA because they may not have undergone sterilization even though they may be labeled as sterile or ethylene oxide processed. *As a result, these devices could cause serious and possibly life-threatening infections.*

Other Companies Affected by the Recall

A & A Medical also does business as A&A Medical/Rocket USA, and LifeQuest. The recall includes all products labeled as sterile and shipped since 1999 nationwide and internationally.

Please note that these products may be sold by firms other than those listed above. FDA is working to identify all distributors of these products. A list of distributors and the products they receive from A & A Medical is being placed on FDA's web site at <http://www.fda.gov/cdrh/recalls/recall31402.html>. We will continue to update this list until all distributors have been identified and listed. In addition, these distributors are being asked to contact all customers who received the affected products.

Products Included in the Recall

This recall includes all products manufactured under the name A & A Medical, Rocket USA, or LifeQuest that are labeled as sterile or as ethylene oxide processed. This firm manufactures many types of Ob/Gyn and surgical devices. The recall includes, but is not limited to currettes (flexible and rigid), uterine dilators, fetal blood samplers, and laparoscopy accessories. A list of known products is attached.

Recommendations

- Do not use any A & A Medical, Inc., Rocket USA, or LifeQuest products.
- Periodically consult the FDA web site for a listing of distributors of A & A Medical, Inc. products. If you have any products from these distributors, contact the distributor for further instructions. Not all of the distributors' products may be affected by this recall.

Additional Information

Individuals seeking additional information should call the company at 1-800-424-1234 or 770-343-8400 or contact FDA's Center for Devices and Radiological Health, Rockville, Maryland at 1-800-638-2041. Additional information regarding this recall can also be found on the FDA's MedWatch web site at <http://www.fda.gov/bbs/topics/NEWS/2002/NEW00799.html>.

Should you have questions regarding this letter, please contact, Jan Davis, Office of Surveillance and Biometrics (HFZ-510), 1350 Piccard Drive, Rockville, Maryland, 20850, by fax at 301-594-2968, or by e-mail at phann@cdrh.fda.gov. Additionally, a voice mail message may be left at 301-594-0650 and your call will be returned as soon as possible.

FDA SAFETY ALERT - From Page 5

All of the FDA medical device postmarket safety notifications can be found on the World Wide Web at <http://www.fda.gov/cdrh/safety.html>. Postmarket safety notifications can also be obtained through e-mail on the day they are released by subscribing to our list server. You may subscribe at <http://list.nih.gov/archives/dev-alert.html>.

Reporting Adverse Events to FDA

The Safe Medical Devices Act of 1990 (SMDA) requires hospitals and other user facilities to report deaths and serious injuries associated with the use of medical devices, including the recalled devices. We request that you follow the procedures established by your facility for such mandatory reporting.

We also encourage you to report medical device malfunctions. You can report these directly to the device manufacturer. You can also report to MedWatch, the FDA's voluntary reporting program. You may submit reports to MedWatch one of four ways: online at <http://www.accessdata.fda.gov/scripts/medwatch/>; by telephone at 1-800-FDA-1088; by FAX at 1-800-FDA-0178; or by mail to MedWatch, Food and Drug Administration, HF-2, 5600 Fishers Lane, Rockville, MD 20857.

Sincerely yours,

/s/

David W. Feigal, Jr., M.D., M.P.H.
Director,
Center for Devices and Radiological Health

List of Known Ob/Gyn and Surgical Devices Recalled by A & A Medical, Inc.

(This is a partial list as known at the time of this alert.)

Curette (flexible and rigid, all sizes)	Collection set tubing	Aspiration sets
Laminaria	IUD removal instruments	Mucus samplers
Biopsy pipettes/Endometrial sampling sets	Uterine sounds	Pratt dilator set
Ovum forceps	Tenaculum forceps	Needle extenders and guide
Fetal bladder drain	Fetal blood sampler	Harvesting Pump and accessories
Loop/ball electrodes	Laparoscopy accessories	

**A&A Medical, Inc., LifeQuest Medical, Inc., and
Rocket USA (A&A) Recall Information**
(<http://www.fda.gov/cdrh/recalls/aamedicalrecallhome.html>)

- FDA Issues Nationwide / International Alert on Recalled OB/GYN and Surgical Devices Due to Health Risk (<http://www.fda.gov/cdrh/recalls/recall31402.html>)
- Press Release: FDA Issues Nationwide/International Alert on Potentially Dangerous OB/GYN Surgical Devices (<http://www.fda.gov/bbs/topics/NEWS/2002/NEW00799.html>)
- Safety Alert: Recall of Ob/Gyn and Surgical Products Manufactured by A & A Medical/Rocket USA/LifeQuest (<http://www.fda.gov/cdrh/safety/safety031502.html>)
- Urgent Device Recall Letter sent to Distributors and Customers of Ob/Gyn and Surgical Products Manufactured by A & A Medical/Rocket USA/LifeQuest (<http://www.fda.gov/cdrh/recalls/recallletteraa.html>)
- Q & A for the Recall of A & A Medical, Inc. Devices (<http://www.fda.gov/cdrh/recalls/qaarecall.html>)
- Self Certification Form for A & A Medical, Inc. Recall (<http://www.fda.gov/cdrh/recalls/formaa.html>)
- List of Distributors for A & A Medical, Inc. (Recall) (<http://www.fda.gov/cdrh/recalls/listaa.html>)

FDA PATIENT SAFETY NEWS

By Morgan E. Warner, B.A.

FDA's Center for Devices and Radiological Health (CDRH) is testing a new program to educate healthcare professionals about patient safety issues. "FDA Patient Safety News" is a monthly, 15-minute television program airing on healthcare education networks including Health Science Television Network (HSTN), Long Term Care Network (LTCN), VA's Education Network, and GE TiP TV. The programs provide information on how to protect patients and how to use medical products safely.

The programs provide information about new device approvals, product recalls, safety alerts, and tips for improving patient safety. Some of the episodes may require immediate action by you, such as removing a recalled item from stock or following the recommendations of a Safety Alert or a "safety tip."

Visit the FDA Patient Safety News web site at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfTopic/psn/index.cfm>, to view the most current program. You will also find more information about the broadcasts plus links for more information on each episode featured in the programs.



Topics from May 2002 Broadcast

- Campaign to Encourage Patients to "Speak Up"
- Preventing Ventilator Deaths and Injuries
- FDA Approves Home Monitoring System for Pacemaker
- Alert on Implanted Neurostimulators and Diathermy
- Avoiding Patient Injuries from Circumcision Clamps
- Update on A & A Medical Recall (see page 5)
- Warning on Using Nitrous Oxide After Intraocular Gas Injection

If you have questions about the web site or any of the programs, write to PSNews@cdrh.fda.gov.

Morgan E. Warner, B.A., is a research assistant in the Center's Office of Surveillance and Biometrics, Medical Product Surveillance Network Team.

ELECTRONIC NOTIFICATION FOR THE *USER FACILITY REPORTING BULLETIN* IS NOW AVAILABLE

If you would like to be notified electronically (via e-mail) when a new issue of the *User Facility Reporting Bulletin* is released, you can sign-up for our List Service at:

<http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfCDRHNew/listman.cfm>

USER FACILITY REPORTING BULLETIN

FDA produces the *User Facility Reporting Bulletin* quarterly to assist hospitals, nursing homes, and other medical device user facilities in complying with their statutory reporting requirements under the Safe Medical Devices Act of 1990, the Medical Device Amendments of 1992, and the Food and Drug Administration Modernization Act of 1997. The *Bulletin's* contents may be freely reproduced. Comments should be sent to the Editor.

Editor: Nancy Lowe • Assistant Editor: Mary Ann Wollerton • Associate Editor: Edie Seligson

Department of Health and Human Services • Public Health Service • Food and Drug Administration
Center for Devices and Radiological Health, HFZ-230 • Rockville, MD 20857 • FAX: 301-594-0067
e-mail: nsl@cdrh.fda.gov