

# QUALITY ASSURANCE PROJECT PLAN (QAPjP) and QA Report for Pacific 2001

**Prepared by: A.L. Norman and Andrew Gillespie**  
**Date: June 25, 2001**

## Table of Contents

1.	Principal Investigator .....	3
2.	Team Members .....	3
3.	Measurement Program.....	3
4.	Measurement Species and Units .....	3
5.	Representative Size Range (if PM) .....	3
6.	Measurement Platform (surface, airborne) .....	3
7.	Measurement Sites (surface only).....	3
8.	Measurement Objective(s) .....	3
9.	Measurement Details .....	3
9.1.	Field Measurements .....	3
9.1.1.	Measurement Principle .....	3
9.1.2.	Instrumentation (Manufacturer/Model) .....	3
9.1.3.	Flow System .....	4
9.1.4.	Inlet Height Above Ground (if surface) .....	4
9.1.5.	Nominal Flow Rate.....	4
9.1.6.	Flow Measurement/Control.....	4
9.1.7.	Flow Temperature and Pressure .....	4
9.1.8.	Sampling Times/Period/Frequency.....	4
9.1.9.	Sampling Methods .....	4
9.1.10.	Filter Type/Coating Type/Reagent Type .....	4
9.1.11.	Planned Changes to Instruments or Methods During Study.....	4
9.2.	Laboratory Measurements (If Applicable) .....	4
9.2.1.	Laboratory Name and Address.....	4
9.2.2.	Analytical Method(s) .....	4
9.2.3.	Sample Extraction or Work-up.....	4
9.2.4.	Analytical Detection Limits .....	5
10.	Quality Assurance/Quality Control .....	5
10.1.	Field Quality Assurance/Quality Control .....	5
10.1.1.	Traceability .....	5
10.1.2.	Calibration .....	5
10.1.3.	Zeros and spans .....	5
10.1.4.	Blanks.....	5
10.1.5.	Field Quality Control procedures .....	5
10.1.6.	Precision determination.....	5
10.1.7.	Comparison with other measurements .....	5
10.1.8.	Inspections and Audits .....	5
10.2.	Laboratory Quality Assurance/Quality Control .....	6
10.2.1.	Traceability .....	6
10.2.2.	Calibration procedures .....	6
10.2.3.	Blanks.....	6
10.2.4.	Other lab QC .....	6
10.2.5.	Precision determination.....	6
10.2.6.	Comparison with other methods .....	6
10.2.7.	Audits .....	7
11.	Data Management and Quality Control .....	7
11.1.	Raw Data Recording.....	7

## QAPjP / QA Report

11.2.	Final Data Reporting.....	7
11.3.	Data Quality Control and Validation.....	7
11.4.	Validity Flags.....	7
11.5.	Below Method Detection Limit Values .....	7
11.6.	Derived Parameters .....	8
11.7.	Explanation of Zero or Negative Data.....	8
12.	Data Quality Objectives (Pre-Study) .....	8
12.1.	Accuracy .....	8
12.2.	Precision .....	8
12.3.	Comparability.....	8
12.4.	Representativeness .....	8
12.5.	Completeness .....	8
12.6.	Other Quality Information.....	8

**1. Principal Investigator**

Dr. A. L. Norman  
Department of Physics and Astronomy  
The University of Calgary  
Calgary, Alberta  
T2N 1N4, Canada

**2. Team Members**

Andrew Gillespie  
Isotope Sciences Laboratory  
Department of Physics and Astronomy  
The University of Calgary

**3. Measurement Program**

Hivol Particle Chemical Characterization Measurements

**4. Measurement Species and Units**

Hivol: PM 2.5 Aerosol  $\text{SO}_4^{2-}$ ,  $\mu\text{gm}^{-3}$  at 0°C and 1 atm.  
Gas: (possibly  $\text{SO}_2$ ), ppbv

**5. Representative Size Range (if PM)**

Hivol PM < 2.5  $\mu\text{m}$

**6. Measurement Platform (surface, airborne)**

Hivol: Surface - 2m above ground level on trailer

**7. Measurement Sites (surface only)**

Cassier Tunnel, Langely and Burnaby South High School.

**8. Measurement Objective(s)**

Hivol: Determine the amount of  $\text{SO}_4^{2-}$  (and  $\text{SO}_2$ ), and determine their isotopic ratios to assist with aerosol sulphate source identification and formation mechanisms prior to and after fine aerosol formation.

**9. Measurement Details**

**9.1. Field Measurements**

**9.1.1. Measurement Principle**  
Filtration

**9.1.2. Instrumentation (Manufacturer/Model)**

Sierra Miscu (1970's version no model numbers available) with a WSAG1200 PM2.5 conversion kit.

**9.1.3. Flow System**

The flow inlet contains an impactor plate to intercept particles above PM 2.5 followed by a filter pack to collect the aerosols of interest. The sampler head is connected to a vacuum motor in series with a mechanical Sierra Miscu flow control system adjusted to operate at a constant flow of 40 scfm. The exhaust from the motor is shunted downwind of the samplers through 20 ft of pipe.

**9.1.4. Inlet Height Above Ground (if surface)**

Hivol: 2m above ground level on top of platform.

**9.1.5. Nominal Flow Rate**

40 scfm (standard cubic feet per minute)

**9.1.6. Flow Measurement/Control**

Kurz model 341 calibrator, serial # 032, 15-60 scfm using mass flow measurements is used to set flow prior to sample program, after motor brushes have been changed, and at the end of the sampling program.

**9.1.7. Flow Temperature and Pressure**

25 degrees C and 1 atmosphere

**9.1.8. Sampling Times/Period/Frequency**

Continuous sampling every 24 hours.

**9.1.9. Sampling Methods**

HVS protocol.

**9.1.10. Filter Type/Coating Type/Reagent Type**

Particulate: Glass fibre.

Sulphur Dioxide:  $K_2CO_3$  impregnated cellulose nitrate filters.

**9.1.11. Planned Changes to Instruments or Methods During Study**

None.

**9.2. Laboratory Measurements (If Applicable)**

**9.2.1. Laboratory Name and Address**

Isotope Science Laboratory

University of Calgary

University Dr. NW

Calgary AB

T2N 1N4

**9.2.2. Analytical Method(s)**

IRMS analysis of  $^{34}S/^{32}S$  in  $SO_2$  and  $SO_4^{2-}$  and  $^{18}O/^{16}O$  in  $SO_4$ .

**9.2.3. Sample Extraction or Work-up**

Sonication in deionized, distilled water followed by filtration and

precipitation as BaSO<sub>4</sub>.

#### **9.2.4. Analytical Detection Limits**

10 µg S, 20 µg O.

### **10. Quality Assurance/Quality Control**

#### **10.1. Field Quality Assurance/Quality Control**

##### **10.1.1. Traceability**

Flow rate calibrations are traceable to NIST Test numbers 213-21/190522.

##### **10.1.2. Calibration**

Calibration in the field is performed using the Kurz flow calibrator and logged prior to and after the sampling program at each site. Also a recalibration is performed after motor brushes are changed.

##### **10.1.3. Zeros and spans**

N/A

##### **10.1.4. Blanks**

As per HVS protocol.

##### **10.1.5. Field Quality Control procedures**

Disposable gloves will be worn and filters will be in contact with clean tweezers only. Filters will be inspected for holes and contamination such as grass, insects, etc. which will be removed prior to analysis. Filters will be placed in clean tinfoil and placed in labelled ziplock bags. HVS and head cleaned with distilled, deionized water prior to operation and if contaminated (eg. After grass cutting is observed).

##### **10.1.6. Precision determination**

Precision determinations will be assessed prior to the field study by operating duplicate instruments simultaneously for 24 hours.

##### **10.1.7. Comparison with other measurements**

The Hivol PM2.5 sulphate concentrations for the isotope samples will be cross-checked with those from similar filtration sampling programs.

##### **10.1.8. Inspections and Audits**

Instruments and filters to be inspected daily. Anomalies such as motor noise, filter tear and other flags will be recorded in a daily log book. Flow rate calibration and inspection to be performed a) upon Hivol setup, b) when brushes are changed, c) prior to Hivol takedown.

## **10.2. Laboratory Quality Assurance/Quality Control**

### **10.2.1. Traceability**

Sulphur isotopes are measured against an internal laboratory reference SO<sub>2</sub> (Quemont) similar to, and calibrated to the isotope composition of the international standard, V-CDT. The calibration takes place daily using internal laboratory standards that bracket the isotope composition of the sample material (SW and STB) that are calibrated on a nominal biannual basis to the international reference materials NBS-127, NZ-1 and Zero.

Oxygen isotopes are measured relative to a reference gas CO and calibrated relative to the international standard V-SMOW. The calibration is performed daily using an internal laboratory isotope BaSO<sub>4</sub> standard (Fisher-Old) the ratios for which have been determined relative to V-SMOW by a disproportionation technique (Ph.D. thesis Shakur).

### **10.2.2. Calibration procedures**

Six initial standards are run at the start of each day followed by two (or more) calibration standards every ten samples. The isotope composition and amount of standard brackets the ratio and size of the samples analyzed. Standards calibrated to international standards VCDT and VSMOW.

### **10.2.3. Blanks**

Laboratory filter blanks are performed for each sample preparation batch and include the multiple sources of contamination including reagent blank, DD water blanks, glassware blanks, and filter handling contamination.

### **10.2.4. Other lab QC**

Calibration check solutions have been used to assess the quality of the analysis procedures. Duplicate sulphate solutions across a range of concentrations have been prepared and analyzed in an identical manner to samples to ensure quality control.

### **10.2.5. Precision determination**

Gravimetric precision determined by creating replicate solutions of known SO<sub>4</sub><sup>2-</sup> concentration, then precipitating SO<sub>4</sub><sup>2-</sup>, and weighing the resulting BaSO<sub>4</sub>. ±0.02 mg.

Isotope ratio precision is determined from the standard deviation of replicate measurements of samples: ± 0.5 ‰ for sulphur isotopes, ±1 ‰ for oxygen isotopes.

### **10.2.6. Comparison with other methods**

If SO<sub>2</sub> concentration samples are collected then they will be compared to SO<sub>2</sub> concentration measured by pulse-fluorescence and K<sub>2</sub>CO<sub>3</sub> impregnated filters at the same sites.

#### **10.2.7. Audits**

No independent audits of the laboratory will be performed specifically for this study. However, the laboratory is the source of several International Atomic Energy Association standards for sulphur isotopes and has undergone numerous intercomparison programs with NIST and other calibration materials.

### **11. Data Management and Quality Control**

#### **11.1. Raw Data Recording**

Once daily entry into lab-book and log files including:

Sample Name (e.g. GVRD - CasT)

Start Date (e.g. 25/08/01) time (10:45)

Stop Date (e.g. 26/08/01) time (8:50)

Flow rate: 40 scfm

Sampling time recorded by HVS timer. (x minutes)

Sampling time recorded by operator (y minutes =Start to Stop time)

#### **11.2. Final Data Reporting**

24 hour data will be recorded and reported.

#### **11.3. Data Quality Control and Validation**

All filter samples will be flagged as either valid (V) or invalid (I). Raw data will be inspected by the PI and all samples containing <24 hours of sampling will be flagged as invalid. Sample identifier tags, and times recorded on sample zip-lock storage bags to be checked against the spreadsheet information prior to sample processing.

#### **11.4. Validity Flags**

OK Valid data

TF Torn Filter

VC Visible Contamination (with descriptor)

WB Worn Brushes (Hivol sampling interrupted)

PNF Protocol Not Followed (and descriptor)

RN Raining during sampling (and descriptor: sprinkle, drizzle, shower, downpour)

#### **11.5. Below Method Detection Limit Values**

Samples below detection limits will be corrected and evaluated using a series of standards bracketing the sample size. Replicates of a laboratory reference will be used to identify the precision of the measurements.

**11.6. Derived Parameters**

Non-sea-salt sulphate isotope ratios will be calculated using sulphate to sodium ratios for sea water (0.252) and isotope mass balance calculations.

**11.7. Explanation of Zero or Negative Data**

In the event of a Hivol motor failure no sample data will be recorded. Below detection limit or blank corrections may produce zero or negative values in the data set.

**12. Data Quality Objectives (Pre-Study)**

**12.1. Accuracy**

The accuracy of the isotope measurements for particles between 0.03 µm to 2.5 µm cannot be ascertained at this time.

**12.2. Precision**

Measurement precision for replicate gravimetric determinations < ±15%. δ-value replicate measurements < ± 1.

**12.3. Comparability**

Measurements for concentration using Filtration techniques are expected to compare within ± 15% with other techniques.  
Isotope measurements are expected to compare within ± 1 %.

**12.4. Representativeness**

Samples collected on days when RH (relative humidity) < 98% are expected to be representative for the approximate size range 0.03 µm to 2.5 µm.

**12.5. Completeness**

Completeness objective = 80 % calculated as the number of 24 hour periods with valid data (including periods of instrument failure, power failure, contamination etc.).

**12.6. Other Quality Information**

---

End of Pre-Study QAPjP