AP8

RADIUM-228 IN WATER

PART A

PRINCIPLE

Radium in water is concentrated by co-precipitation with barium sulfate. The barium sulfate is metastasized to barium carbonate. The carbonate is dissolved and passed through a chromatographic column to separate radium/barium from actinium. The radium/barium fraction is eluted and barium-133 is counted by gamma spectrometry to quantify the chemical yield. Actinium is eluted and co-precipitated with cerium fluoride for beta counting.

REFERENCES

- 1. Sill C. W. (1987). Determination of Radium-226 by high-resolution alpha spectrometry.
- 2. Burnett, H. B. & Cable, P. <u>Radium-228 in natural waters via extraction chromatography</u>. 38th Annual Conference on Bioassay, Analytical, and Environmental Radiochemistry.
- 3. Eichrom Technologies, Inc., <u>Analytical Procedure RAW03, Rev. 0.5</u>, 2001.

Certification Record for

AP8

RADIUM-228 IN WATER

CHECKPOINTS

| 1. J | OB HAZARD ANALYSIS (JHA) | · · · · · · · · · · · · · · · · · · · | | • | |
|--------|--------------------------|---------------------------------------|------|---|------|
| 2. N | MSDS/HAZARDS DISCUSSED | | | | |
| | TRACER ADDED | | | | |
| | BARIUM SULFATE PRECIPITA | ATE | | | |
| | ACTINIUM SEPARATION | | | | |
| | ACTINIUM DEPOSITION | | | | |
| 8. F | INAL CALCULATIONS | | | | |
| | | | | | |
| | | | | | |
| | ANALYST'S SIGNATURE: | | | | |
| | CEDTIFIED DV | | | | |
| | CERTIFIED BY: | | 3.00 | | |
| | DATE: | | | | |
| | | | | | |
| | ANALYSIS VALUE: | | | | |
| | YO LOWD LIVE LY | | | | |
| | KNOWN VALUE: | | | | |
| | MEASURED/KNOWN: | | | | |
| | , | | - | | |
| See Ta | ısk, Batch | for the original data. | | | |
| • | | | | | |
| COMN | MENTS: | | | | |
| | | | | | |
| | | | | | |

PART B

1.0 PURPOSE AND SCOPE

This procedure provides the analytical method for determination of radium-228 in water.

2.0 REAGENTS

All chemicals are hazardous. See MSDS for specific precautions. See step 2.0 of AP8 JHA. Unless otherwise indicated, all references to water should be understood to mean reagent grade water.

Barium carrier, 30 mg/mL: Dissolve 13.3 g of BaCl₂•2H₂O in 100 mL of water. Filter the solution through a DM-450, 47mm filter (or equivalent) and dilute to 250 mL with water.

Barium chloride dihydrate, BaCl₂•2H₂O: crystalline.

Ba-133, NIST traceable standardized solution, approximately 300-600 pCi for each sample.

Cerium carrier, 1 mg/mL: AA quality Ce solution of 1000 µg/mL.

Ethanol, 95%: Dilute 475 mL ethyl alcohol to 500 mL with water.

Hydrofluoric acid, HF, concentrated, 28 M: CAUTION: Skin contact with HF causes very severe burns.

Ln resin, 2 mL pre-packed columns or cartridges, 100-150 mm particle size.

Nitric acid, HNO₃, concentrated, 16 M.

Nitric acid, HNO₃, 1 M, slowly add 31.5 mL 16 M HNO₃ to 200 mL of water. Dilute to 500 mL with water and mix.

Nitric acid, 0.35 M: Add 175 mL of 1 M HNO₃ to 500 mL of water. Prepare in a volumetric flask.

Nitric acid, 0.095 \underline{M} : Add 47.5 mL of 1 \underline{M} HNO₃ to 500 mL of water. Prepare in a volumetric flask.

Potassium carbonate, 50% (w/v) K₂CO₃: Add 500 g of K₂CO₃ in 950 mL of water. Dilute to 1 L with water and filter through a DM-450, 47 mm filter (or equivalent).

Ra-228, NIST traceable standardized solution.

Sulfuric acid, H₂SO₄, concentrated, 18 M.

3.0 APPARATUS AND MATERIALS

Analytical balance

Beakers, appropriate for sample volume

Centrifuge

Centrifuge tube, 50 mL

Column rack

DM-450 filter, 47 mm or equivalent

Extension funnels

Metricel polypropylene filter, 0.1 µm, 25 mm or equivalent

Gamma spectrometry system

Gas-flow proportional counting system

Glass filter setup

Hot plate

Polysulfone filter apparatus

Stainless steel disks

Vortex mixer

4.0 PROCEDURE

4.1 General Requirements

Before proceeding, you must be certified, as indicated in QCP1 of this manual, and Section 3 of the Quality Program (QP) Manual. See page 2 of this section for a copy of the certification record.

- 4.2 Sample Preparation
 - 4.2.1 Measure sample in a volumetric flask and transfer to an appropriate size beaker. Use water for the blank and Laboratory Control Standard (LCS). See step 4.2.1 of AP8 JHA.

NOTE: If HF was initially added as a preservative to the sample to prevent Pu hydrolysis, transfer the sample to a platinum crucible, evaporate to dryness on a hot plate, and proceed to step 4.3.3.

- 4.2.2 Add a known amount of Ba-133 tracer to all samples, blank, and LCS. See step 4.2.2 of AP8 JHA.
- 4.2.3 Add approximately 30 pCi of Ra-228 to the LCS. See step 4.2.3 of AP8 JHA.
- 4.2.4 While stirring, carefully add 10 mL of 18 M H₂SO₄. Heat sample to boiling. See step 4.2.4 of AP8 JHA.
- 4.2.5 Add 1 mL Ba carrier to all samples, blank, and LCS. See step 4.2.5 of AP8 JHA.
- 4.2.6 Filter the precipitate through a DM-450 filter. Pour the supernate in the

dilute waste acid stream. See step 4.2.6 of AP8 JHA.

- 4.2.7 Transfer the filter paper to a centrifuge tube.
- 4.2.8 Add 10 mL of water and vortex to remove the precipitate from the filter paper. Discard the filter paper. Centrifuge the precipitate at 2000 RPM for 5 minutes and discard the supernate. See step 4.2.8 of AP8 JHA.
- 4.2.9 Add 10 mL of water and vortex to rinse the precipitate. Centrifuge at 2000 RPM for 5 minutes and discard supernate. See step 4.2.9 of AP8 JHA.
- 4.2.10 Repeat step 4.2.9.
- 4.2.11 Add 20 mL 50% K₂CO₃ to the precipitate. Vortex. Heat for at least 30 minutes in boiling water, with occasional stirring, to metastasize the precipitate to BaCO₃. See step 4.2.11 of AP8 JHA.
- 4.2.12 Centrifuge the samples at 2000 RPM for 5 minutes and discard the supernate. See step 4.2.12 of AP8 JHA.
- 4.2.13 Add 10 mL of water and vortex to rinse the precipitate. Centrifuge at 2000 RPM for 5 minutes and discard the supernate. See step 4.2.13 of AP8 JHA.
- 4.2.14 Repeat step 4.2.13 two more times.
- 4.2.15 Allow the Ac-228 to grow in at least 24 hours before proceeding to step 4.2.16.

Note: Confirm that low background proportional detectors will be available. Proceed with 2 samples at a time. The detector is the limiting factor, but the samples must be counted within 2 hours.

Note: Use Ln Resin Columns or Cartridges Only

- 4.2.16 Snap the plastic bottom off each Ln column to be used and allow solution to elute into a waste beaker.
- 4.2.17 Add 10 mL 0.095 M HNO₃ to each column and collect in a waste beaker for disposal into the dilute acid waste. See step 4.2.17 of AP8 JHA.
- 4.2.18 Prepare load solution by adding 10 mL 0.095 M HNO₃ to the precipitate from step 4.2.13 and vortex to dissolve the precipitate. See step 4.2.18 of AP8 JHA.
- 4.2.19 Place labeled centrifuge tubes under each column for Ba-133 and Ra-228.
- 4.2.20 Pour the load solution into each column and collect. See step 4.2.20 of AP8 JHA.

4.2.21 Rinse each centrifuge tube with 5 mL 0.095 M HNO₃ and pour into each column and save in the same centrifuge tube as from step 4.2.19. See step 4.2.21 of AP8 JHA.

Note: Record Ac-228 Separation Time.

- 4.2.22 Rinse the column two times with 5 mL 0.095 M HNO₃ and collect in the same centrifuge tube as in Step 4.2.19. See step 4.2.22 of AP8 JHA.
- 4.2.23 Pour the solution from step 4.2.21 into a specimen cup and submit to the counting room for gamma analysis for the Ba-133 yield determination.
- 4.2.24 Place a new, labeled centrifuge tube under each column to collect the Ac-228 fraction.
- 4.2.25 Elute the Ac-228 from the column with 10 mL of 0.35 \underline{M} HNO₃. See step 4.2.25 of AP8 JHA.
- 4.2.26 Add 100 μL Ce carrier and 2 mL 28 M HF to each sample, mix, and wait 10 minutes. See step 4.2.26 of AP8 JHA.
- 4.2.27 Assemble polysulfone filtering apparatus for each sample. Wet the filter paper with 5 mL 95% ethanol.
- 4.2.28 Filter each sample onto a 0.1 μm filter paper. See step 4.2.28 of AP8 JHA.
- 4.2.29 Count the CeF₃ from step 4.2.27 immediately on a low background proportional counter long enough to achieve the desired statistics.

5.0 CALIBRATIONS

- 5.1 Low Background Alpha/Beta Counter Efficiency Calibration
 - 5.1.1 Add enough Ra-228 to achieve desired statistics into an appropriate size beaker. Carefully take the solution to dryness.
 - 5.1.2 Snap the plastic bottom off each Ln column to be used and allow solution to elute into a waste beaker.
 - 5.1.3 Add 10 mL 0.095 M HNO₃ to each column and collect in a waste beaker for disposal into the dilute acid waste. See step 4.2.21 of AP8 JHA.
 - 5.1.4 Add 10 mL 0.095 M HNO₃ to sample from step 5.1.1. Pour solution onto a Ln column and allow to drain into a waste beaker. See step 4.2.21 of AP8 JHA.
 - 5.1.5 Rinse the beaker with 5 mL 0.095 M HNO₃ and pour into each column and allow the solution to drain into a waste beaker. See step 4.2.21 of AP8 JHA.

- 5.1.6 Rinse Ln column twice more with 5 mL 0.095 M HNO₃ and pour into each column and allow the solution to drain into a waste beaker. See step 4.2.21 of AP8 JHA.
- 5.1.7 Place a labeled centrifuge tube under each Ln column and add 10 mL of 0.35 M HNO₃. See step 4.2.24 of AP8 JHA.
- 5.1.8 Go to step 4.2.25 and continue through step 4.2.28.
- 5.1.9 Calculate the Ac-228 counting efficiency using the efficiency equation in section 6.
- 5.1.10 The Laboratory Manager or designee must review and approve the Ac-228 counting efficiency.
- 5.2 Ba-133 Calibration for Yield Determination
 - 5.2.1 Add an appropriate amount of a Ba-133 standard to a specimen cup and bring the volume to 25 mL with 0.095 M HNO₃. See step 4.2.21 of AP8 JHA.
 - 5.2.2 Submit to the counting room to establish the counting efficiency of Ba-133 in the specimen cup geometry.
 - 5.2.3 The Laboratory Manager must review and approve the Ba-133 counting efficiency.

6.0 CALCULATIONS

Critical data values will be documented on standard forms maintained as critical records. The following equations define the critical data values. All data will be recorded and reduced according to these calculations. TPU has not been evaluated for this non-routine procedure.

$$Concentration = \frac{G - B}{T_2 \cdot E \cdot Y \cdot Q \cdot e^{-\lambda_{Ra}T_3} \cdot e^{-\lambda_{Ac}T_1}} \cdot \frac{\lambda_{Ac}T_2}{1 - e^{-\lambda_{Ac}T_2}} = pCi / unit$$

$$2\sigma \ Error = \frac{1.96\sqrt{G+B}}{T_2 \cdot E \cdot Y \cdot Q \cdot e^{-\lambda_{Ra}T_3} \cdot e^{-\lambda_{Ac}T_1}} \cdot \frac{\lambda_{Ac}T_2}{1 - e^{-\lambda_{Ac}T_2}} = pCi/unit$$

$$MDC = \frac{3 + 4.65\sqrt{B}}{T_2 \cdot E \cdot Y \cdot Q \cdot e^{-\lambda_{Ra}T_3} \cdot e^{-\lambda_{Ac}T_1}} \cdot \frac{\lambda_{Ac}T_2}{1 - e^{-\lambda_{Ac}T_2}} = pCi/unit$$

To calculate efficiency:

$$E = \frac{G_E - B}{E_{ACT}}$$

To calculate radiochemical yield:

$$Y = \frac{Ba - 133_{MA}}{Ba - 133_{KA}}$$

where:

 $\lambda_{Ac} = \text{Ac-}228 \text{ decay constant, } 1.8846 \times 10^{-3}, \text{min}^{-1}$ $\lambda_{Ra} = \text{Ra-}228 \text{ decay constant, } 2.2915 \times 10^{-7}, \text{min}^{-1}$

B = background counts

 $Ba-133_{MA}$ = Ba-133 measured activity by gamma count

 $Ba-133_{KA} = Ba-133$ known activity added C = concentration in pCi/unit E = counting efficiency, cpm/pCi $E_{ACT} = pCi of efficiency standard$

G = gross counts

G_E = gross counts of efficiency standard MDC = minimum detectable concentration

Q = sample quantity

TPU = total propagated uncertainty T₁ = Ac-228 decay time, min

 $T_1 = Re-220$ decay time, $T_2 = count time, min$

 $T_3 = \text{Ra-228 decay time, min, from sample collection to step } 4.2.20$

Y = chemical yield

7.0 <u>RECORDS</u>

- 7.1 Reference QP Manual for general record requirements.
- 7.2 The raw count data is saved during the weekly backup of the low background alpha/beta counter to the ORISE network disks and during the weekly backups using Tivoli software or the equivalent for data from the Canberra Apex Gamma system.
- 7.3 Hard copies of assignment and calculation sheets are maintained in the archived site file. Electronic copies of assignment and calculation sheets are saved during the daily incremental backup of the network system. The following data sheets should be completed and retained:
 - Ra-228 Analysis Assignment Form
 - Ra-228 Data Sheet
 - Ra-228 Concentration and Uncertainty Report (This report may be generated using approved Excel spreadsheets or from the database, if available.)

(AP 8, Rev 4) Ra-228 ANALYSIS ASSIGNMENT FORM

| Assigned To: | Date: | | Batch: | | |
|-----------------|---------------------------------------|---------------------------------------|----------------|--------|--|
| Task #: | LWR #: | · · · · · · · · · · · · · · · · · · · | Activity Lev*: | ···· | |
| Sample #'s: | | · | . ' | | |
| | | | | | |
| | · · · · · · · · · · · · · · · · · · · | | | | |
| | | QC REQUIRED: | | | |
| | | (0.2222) | | | |
| NK | | | | | |
| LICATE | Sample # | ···· | # Replicates: | | |
| | Ra-228 STD # | | QUANTITY: | INITIA | |
| | | | UNITS: | | |
| | | | | | |
| TRIX SPK | Sample # | | | | |
| | Ra-228 STD # | | QUANTITY: | | |
| | | | UNITS: | | |
| | Pipette # | Volume | | Weight | |
| | | | | | |
| | Ba-133 STD # | · | QUANTITY: | | |
| | | | UNITS: | | |
| | N TO | • | | | |
| CIAL INSTRUCTIO | JINO: | | | | |
| | | | , | | |
| | | | | | |
| | | - | | | |

(AP 8, Rev 4) Ra-228 DATA SHEET

| | (| | | | | | | | |
|-----------|-----|-------------|---|--|--|--|--|--|--|
| Carrier # | | | | | | | | | |
| Sample # | | | | | | | | | |
| Quantity | | | | | | | | | |
| Units | | | | | | | | | |
| | | | | | | | | | |
| Carrier # | | | | | | | | | |
| Sample # | | | | | | | | | |
| Quantity | | | | | | | | | |
| Units | 1.0 | | | | | | | | |
| | | | | | | | | | |
| Carrier # | | | | | | | | | |
| Sample # | | | | | | | | | |
| Quantity | | | · | | | | | | |
| Units | | | | | | | | | |
| | | ÷ | | | | | | | |
| Carrier # | | | | | | | | | |
| Sample # | | | : | | | | | | |
| Quantity | | | | | | | | | |
| Units | | | 1 | | | | | | |

(AP8, Rev 4) RADIUM-228 CONCENTRATION & UNCERTAINTY REPORT

OPERATOR INITIALS DATE BATCH # TASK #

| DETECTOR # | | 1, | | | | |
|--------------------------------|-----|----|---|---|-----|--|
| CARRIER TRAY# | | | | | | |
| | · . | | | | | |
| SAMPLE NO. | | | | | | |
| Ac-228 SEPARATION DATE | | | | | · · | |
| Ac-228 SEPARATION TIME | | | | | | |
| END COUNT DATE | | | | | | |
| Ac-228 DECAY TIME (min) | | · | | | | |
| LENGTH OF COUNT TIME (min) | | | | · | | |
| SAMPLE COUNTS | | | | | | |
| BACKGROUND COUNTS | | | · | | | |
| DETECTOR EFFICIENCY cpm/pCi | | | · | | | |
| SAMPLE QUANTITY (L) | | | | | | |
| Ba-133 MEAS. ACTIVITY (pCi) | | | | | | |
| Ba-133 KNOWN ACTIVITY (pCi) | | | | | | |
| Ba-133 YIELD | | | | | | |

(AP8, Rev 4) RADIUM-228 CONCENTRATION & UNCERTAINTY REPORT

| OPERATOR | DATE | BATCH # | | TASK# | | |
|------------------------|------------|---------|-------------|-------|-------|----|
| | | | | | | |
| | | | | | | |
| | · | | 1 | | | |
| SAMPLE # | | | | | , | |
| Concentration, pCi/L | | | | | | ٠. |
| 2-sigma error | | | | | | |
| 4.65-sigma MDC | | | | | | |
| | | | | | | |
| | | | | | · | |
| SAMPLE # | | | | | | |
| Concentration, pCi/L | | | | | | |
| 2-sigma error | | | | | | |
| 4.65-sigma MDC | | | | | • | |
| , | · · · | | | | - | |
| | | | | | | • |
| Ra-228 Known pCi | | | Meas/Known | | | |
| Uncertainty | | | Uncertainty | | | |
| BLANK CORRECT? | YES[] NO[] | | | | INIT | |
| LCS CORRECT? | YES[] NO[] | | | | INIT | |
| BATCH YIELD CORRECT | | | | | INIT | |
| IF NO, SPECIFY REASON: | 1 | | | | | |
| | | | | | | |
| ANALYST: | | · | | | DATE: | |
| REVIEWED BY: | | | | | DATE: | · |
| GIVEN TO: | | | · _ | | DATE: | _ |
| | | | | | | |

Laboratory Procedures Manual ORISE/IEAV