

# Radiation Control Program

## Violations Summary Report -- 2002 Violations

### Inspections Performed by X-Ray Inspectors

From 1/1/2002 Through 12/31/2002

<b>Description of Violation</b>	<b>Number of Violations</b>
OS&P not complete/not available; certificate, RPP, or NOV not available	2,946
Compliance test &/or corrective action not performed	1,646
Time-temp. chart, timer, therm. not available; Wrong safelight filter, wattage, &/or distance	1,018
Collimator indicators, PBL, light field, &/or centers alignment inadequate	918
Personnel not credentialed; operators, physician, physicist	384
Registration or certification: none, not current, expired	364
Public dose surveys not performed; Dose to public excessive	356
Annual protective device test not performed	332
Occupational dose not determined; Current year dose not determined	320
Patient exposure at skin entrance (ESE) exceeded limits	264
Output reproducibility &/or mA linearity inadequate	232
Indicators of x-ray production, multiple tubes, battery charge, AEC operation not available	230
Timer inaccurate or non-reproducible; Exposure untimed	198
Technique chart not available	194
'Notice to Employees' not posted	160
Radiation area caution signs not posted; No machine label	142
Mammo QC, physicist eval., or physician review of QC not performed; phantom tech. not patier	122
Annual fluoroscopic entrance exposure measurements not performed	102
Current personnel dose unknown; PM records not available	100
Other x-ray & other mammo	82
Mammo QC failed &/or phantom failed without ceasing mammograms	80

Technique chart does not match techniques used clinically	58
Technique factors not indicated prior to exposure: kVp, mA, mAs, time, AEC	50
Mammo anti-scatter grids not available; Compression paddle misaligned	42
Mammo patient film not provided or not retained	38
Fluoro collimation (spot film, input phosphor, Min @ max, &/or stepless) not adequate	38
Protective devices &/or holding devices not available or not used	32
Mammo image not processed w/in 24 hrs.; Batch processing not performed properly	30
Inadequate HVL	26
Incomplete technique chart	24
QA/QC records not available	22
Mammo QC corrective action not performed	18
Inappropriate PM device; PM device used improperly; Supplier not NVLAP	12
Operator collimation inadequate	12
X-Ray machine not specifically designed for mammography	10
Service company not registered &/or used by registrant	10
Patient viewing system not available or not used; Two-way aural communication not available	8
Therapy annual calibration not performed	8
Training records for assembler, calibrator, consultant not available	6
RSO quarterly audit records not available	6
Minimum source to skin distance (SSD) exceeded (radiographic & dental)	6
Fluoro EER exceeds limit	6
CT dose measurements, spot checks not performed	4
Service company servicing non-registered x-ray machine	4
Analytical x-ray use records not available; Industrial x-ray inventory not available	4
Code not used	4
Tube housing assembly not stable or hand held during radiography	4

Fluoro high level operates w/o continuous pressure &/or continuous signal	4
Fluoro minimum SSD inadequate	4
Code not used	2
Overexposure of occupationally exposed person	2
Overexposure report to Agency or individual not performed	2
Industrial x-ray interlock bypass warning not posted	2
Survey instrument calibration not performed	2
Code not used	2
Code not used	2
Exposure switch not the 'deadman' type	2
Mobile unit exposure cord does not allow operator position 6 feet from patient	2
Therapy filter interlock or indicator system inadequate	2
Code not used	2
Report Compiled on: August 09, 2005	