

# Radiation Control Program

## Violations Summary Report -- 2003 Violations

### Inspections Performed by X-Ray Inspectors From 1/1/2003 Through 12/31/2003

Description of Violation	Number of Violations
OS&P not complete/not available; certificate, RPP, or NOV not available	3,252
Compliance test &/or corrective action not performed	1,520
Time-temp. chart, timer, therm. not available; Wrong safelight filter, wattage, &/or distance	992
Collimator indicators, PBL, light field, &/or centers alignment inadequate	732
Registration or certification: none, not current, expired	426
Personnel not credentialed; operators, physician, physicist	422
Annual protective device test not performed	358
Public dose surveys not performed; Dose to public excessive	328
Occupational dose not determined; Current year dose not determined	286
Patient exposure at skin entrance (ESE) exceeded limits	278
Indicators of x-ray production, multiple tubes, battery charge, AEC operation not available	208
Output reproducibility &/or mA linearity inadequate	162
'Notice to Employees' not posted	160
Mammo QC, physicist eval., or physician review of QC not performed; phantom tech. not patient	156
Technique chart not available	152
Radiation area caution signs not posted; No machine label	138
Timer inaccurate or non-reproducible; Exposure untimed	122
Current personnel dose unknown; PM records not available	114
Technique chart does not match techniques used clinically	96
Annual fluoroscopic entrance exposure measurements not performed	74
Other x-ray & other mammo	64

Mammo QC failed &/or phantom failed without ceasing mammograms	62
Fluoro collimation (spot film, input phosphor, Min @ max, &/or stepless) not adequate	56
Mammo anti-scatter grids not available; Compression paddle misaligned	50
Technique factors not indicated prior to exposure: kVp, mA, mAs, time, AEC	40
Mammo patient film not provided or not retained	38
Mammo image not processed w/in 24 hrs.; Batch processing not performed properly	34
Protective devices &/or holding devices not available or not used	32
Survey instrument calibration not performed	28
Operator collimation inadequate	28
Patient viewing system not available or not used; Two-way aural communication not available	28
QA/QC records not available	24
Inadequate HVL	24
Mammo QC corrective action not performed	14
Inappropriate PM device; PM device used improperly; Supplier not NVLAP	14
X-Ray machine not specifically designed for mammography	12
Service company not registered &/or used by registrant	10
Deliberate exposure to radiation	6
Tube housing assembly not stable or hand held during radiography	6
Fluoro EER exceeds limit	6
Therapy annual calibration not performed	6
Training records for assembler, calibrator, consultant not available	4
Service company servicing non-registered x-ray machine	4
Analytical x-ray use records not available; Industrial x-ray inventory not available	4
Minimum source to skin distance (SSD) exceeded (radiographic & dental)	4
Diagnostic tube housing inadequate	4

Fluoro cumulative timer inadequate	4
Fluoro minimum SSD inadequate	4
Overexposure of occupationally exposed person	2
Mammo AEC not providing constant density; Mammo kVp inaccurate	2
Code not used	2
Primary or secondary barriers inadequate	2
Fluoro high level operates w/o continuous pressure &/or continuous signal	2
Fluoro primary barrier not adequate	2
Code not used	2
Report Compiled on: August 09, 2005	





