



# Certificate

## Standard Reference Material<sup>®</sup> 1001

### X-ray Film Step Tablet

This Standard Reference Material (SRM) is intended for use in the calibration of optical densitometers and similar equipment used in the photographic, graphic arts, and X-ray fields. SRM 1001 is a double emulsion X-ray film step tablet 25.4 cm (10 in) long by 3.5 cm (1.375 in) wide. The tablet has 17 steps 1.3 cm (0.5 in) wide perpendicular to the long edge of the film with certified transmission densities ranging from less than 0.200 to greater than 4.000. The serial number is located at one end of the step tablet.

**Determination of Certified Transmission Density:** The transmission density of each step of the tablet was measured using the NIST Diffuse Transmittance Densitometer [1]. The instrument and measurements conform to the conditions specified for ISO Standard Diffuse Visual Transmission Density,  $D_T(90^\circ \text{ Opal}; S_H; \leq 10^\circ; V_T)$  [2,3]. The certified transmission densities listed in Table 2 were measured on a 3 mm diameter circle at the center of each step, and apply only to that area. When measured, the side of the tablet with the serial number was in contact with the diffuser of the densitometer.

**Expiration of Certification:** The certification of SRM 1001 units with a calibration date of 13 July 2004 or later is deemed to be valid, within the uncertainties specified, until 13 December 2008, provided the SRM is stored and handled in accordance with the "Instructions for Use" and "Storage and Handling" sections of this certificate. SRM 1001 units with a calibration date earlier than 13 July 2004, retain their original two-year period of certification. However, certification will be nullified if the SRM is damaged or contaminated. Registration (see attached sheet) will facilitate notification.

**Discussion of Uncertainties:** Uncertainties were calculated according to the procedures outlined in reference 4. Measured (Type A) uncertainties were assumed to be Gaussian-distributed, and were calculated from the standard deviations. Estimated or inferred (Type B) uncertainties were also assumed to be Gaussian-distributed, and were calculated from the measurement equation. The uncertainties were combined by adding their variances in quadrature. Table 1 is a list of all identifiable sources of uncertainty. The measured standard deviations are listed in (a); the inferred standard deviations are listed in (b).

**Source<sup>1</sup>:** Film step tablets were produced by GE Inspection Technologies, AGFA division, Goose Creek, SC.

The technical direction and physical measurements leading to certification were provided by E.A. Early with assistance from C.L. Cromer, D. Dummer, T.R. O'Brian, X. Xiong, and R.D. Saunders of the NIST Optical Technology Division and P.F. Wychorski of the Eastman Kodak Co.

Albert C. Parr, Chief  
Optical Technology Division

Robert L. Watters, Jr., Chief  
Measurement Services Division

Gaithersburg, MD 20899  
Certificate Issue Date: 30 December 2005  
*See Certificate Revision History on Page 3*

---

<sup>1</sup>Certain commercial equipment, instruments, or materials are identified in this certificate in order to specify adequately the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.

The support aspects involved in the issuance of this SRM were coordinated through the NIST Measurement Services Division.

## INSTRUCTIONS FOR USE

Remove the film from its protective sleeve and place the center of a step on the diffuser of the densitometer to be calibrated with the side of the step tablet with the serial number in contact with the diffuser. Calibrate the densitometer using the transmission density of that step, and repeat this procedure with other steps of the step tablet.

**Storage and Handling:** The densities of this X-ray film step tablet may change with time. To minimize such changes, the film should be stored in a cool, dry place where it will not be exposed to light or other radiant energy or to chemical fumes, or to dust in the air. Scratches, abrasion marks, or foreign matter on the film can change the density. Fingerprints are a common source of contamination. Fingerprints on the film surfaces can be avoided by handling the step tablet only by the edges and by wearing clean cloth gloves available from photographic film dealers for this purpose. Any attempt to clean a film step tablet, other than to remove dust with a soft camel-hair brush, is not recommended as it is likely to alter the certified values.

Table 1. Tabulation of Uncertainties	
Source of Uncertainty	Uncertainty in Transmission Density
(a) Measured (Type A) Uncertainties	
Source Stability	$\pm 0.001$
Signal Noise	$\pm 0.001$
Amplifier Gain	$\ll 0.001$
Detector Linearity	$\pm 0.001$
Step Uniformity	$\pm 0.001$
(b) Inferred (Type B) Uncertainties	
Voltmeter Accuracy	$\ll 0.001$
Diffusion Coefficient	$< 0.001$
Opal Reflectance	$< 0.001$
Spectral Product	$< 0.001$
Combined Uncertainty ( $1\sigma$ )	$\pm 0.002$
Expanded Uncertainty ( $3\sigma$ )	$\pm 0.006$

## REFERENCES

- [1] Early, E.A.; O'Brian, T.R.; Saunders, R.D.; Parr, A.C.; *Standard Reference Materials: Film Step Tablet Standards of Diffuse Visual Transmission Density – SRM 1001 and SRM 1008*; Natl. Inst. Stand. Technol. Spec. Publ. 260-135, U.S. Government Printing Office: Washington, DC (1998).
- [2] ISO 5-2: 2001, *Photography – Density Measurements - Part 2: Geometric Conditions for Transmission Density*; International Organization for Standardization: Geneva, Switzerland (2001).
- [3] ISO 5-3: 1995, *Photography – Density Measurements - Part 3: Spectral Conditions*; International Organization for Standardization: Geneva, Switzerland (1995).
- [4] ISO; *Guide to the Expression of Uncertainty in Measurement*; ISBN 92-67-10188-9, 1st ed.; International Organization for Standardization: Geneva, Switzerland (1993); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at <http://physics.nist.gov/Pubs/>.

**Certificate Revision History:** 30 December 2005 (This technical revision reports an extension of the certification period); 25 October 2004 (This revision reports the clarification additional text to the "Expiration of Certification" section); 20 May 2004 (Extension of the certification period for units with a calibration date of 13 July 2004 or later.); 14 October 1997 (Original certificate date).

*Users of this SRM should ensure that the certificate in their possession is current. This can be accomplished by contacting the SRM Program at: telephone (301) 975-6776; fax (301) 926-4751; email [srminfo@nist.gov](mailto:srminfo@nist.gov); or via the Internet at <http://www.nist.gov/srm>.*

## APPENDIX

<b>Table 2. Transmission Densities of SRM 1001</b>	
<b>Serial No.: 0401001</b>	<b>Calibration Date: 13 July 2004</b>
<b>Step Number</b>	<b>Transmission Density<sup>(a)</sup></b>
1	0.154
2	0.239
3	0.488
4	0.735
5	0.981
6	1.230
7	1.468
8	1.731
9	1.983
10	2.252
11	2.506
12	2.783
13	3.030
14	3.279
15	3.523
16	3.791
17	4.187

<sup>(a)</sup> Expanded uncertainty ( $k = 3$ ) of the transmission density of each step is  $\pm 0.006$  (see Table 1).