

**ATTACHMENT 2: WESTBAY® TRANSDUCER PERFORMANCE CHECK AND MAINTENANCE FORM**

**SOP-5226-2**

Records Use Only

**Westbay® Transducer Performance Check and Maintenance Form**



**Los Alamos National Laboratory  
Water Stewardship Program**

**Westbay® Transducer Performance Check and Maintenance Form**

Transducer Serial Number: \_\_\_\_\_

psi Rating: \_\_\_\_\_

Error Tolerance (psi): \_\_\_\_\_

Meteorological Station						Calibration Check					Inspector Name	Comments (Equipment Condition, Describe Maintenance, Calibration Check)
Date	Station Name	Time of Reading (MST)	Barometric Pressure (mb)	Barometric Pressure (psia)	Add 0.026 for Calibration check at TA-64	Calibration Check Location	Time of Reading (MST)	Transducer Barometric Pressure (psia)	Acceptable Range (psia)	Calib Check Passed		

NOTE: To convert millibar (mb) pressure measurements to psia, use conversion factor: 0.01450psi/mb