



TEST REPORT

Report Number: 3143860MIN-001

Project Number: 3143860

Testing performed on the
XPO100

To
RTCA/DO-160F:2007
Section 21

Conducted and Radiated Emissions for Category M Equipment

For
Invacare Corporation

Test Performed by:
Intertek Testing Services NA, Inc.
7250 Hudson Blvd., Suite 100
Oakdale, MN 55128

Test Authorized by:
Invacare Corporation
One Invacare Way, PO Box 4028
Elyria, OH 44036

Prepared by: 
Norman Shpilsher

Date: February 14, 2008

Reviewed by: 
Uri Spector

Date: February 14, 2008

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to copy or distribute this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program. This report must not be used to claim product endorsement by A2LA, NIST nor any other agency of the U.S. Government.



2.0 TEST SUMMARY

Referring to the performance criteria and the operating mode during the tests specified in this report, the equipment complies with the requirements according to the following standards.

TEST STANDARD	TEST	RESULT
21.4	Conducted RF Emissions	Pass
21.5	Radiated RF Emissions	Pass