

Carolyn R. Hill¹, John M. Butler¹, Cynthia J. Sprecher², Robert S. McLaren², Dawn R. Rabbach², Benjamin E. Krenke², and Douglas R. Storts²



¹ U.S. National Institute of Standards and Technology, 100 Bureau Drive, Gaithersburg, MD 20899-8312 USA
² Promega Corporation, Madison, WI 53711 USA




Promega has developed several new STR kits to address recommendations of additional loci requested by the European community [1,2]. These kits include D10S1248, D2S441, D22S1045, D12S391, D1S1656, and SE33 as well as the 10 STRs and the sex-typing marker amelogenin present in the Applied Biosystems' SGM Plus™ kit. In order to evaluate the performance of new PCR primer sets compared with ones currently in use, concordance testing was performed on over 1400 samples from U.S. population groups. Comparisons were made with previous genotyping results from commonly used STR kits including PowerPlex® 16, Identifier™, and MiniFiler™ as well as in-house assays. From almost 100,000 alleles compared between the PowerPlex® ESX and ESI Systems (2 kits x 17 loci x 2 alleles/locus x 1443 samples), a total of 7 differences were observed. Additional allele differences were observed when comparisons were made to currently available kits with many of the dropouts coming from the present commercial kits or published primer sets. Sequence analysis was performed on all discordant samples to ascertain the primer binding site mutation. An additional primer was added in the final PowerPlex ESX 16 and 17 Systems to correct for allele dropout with D22S1045. Our results indicate that these new kits enable reliable STR typing with sensitive DNA detection and high powers of discrimination. With an overall concordance of greater than 99.9% to STR loci typed with currently available kits, these new kits should permit reliable extensions of DNA database and forensic casework efforts.

[1] Gill, P., Fereday, L., Morling, N., Schneider, P.M. (2006) The evolution of DNA databases--recommendations for new European loci. *Forensic Sci. Int.* 156:242-244.
 [2] Gill, P., Fereday, L., Morling, N., Schneider, P.M. (2006) Letter to editor -- New multiplexes for Europe-amendments and clarification of strategic development. *Forensic Sci Int.* 163:155-157.

Purpose of Concordance Studies

When different primer sets are utilized, there is a concern that allele dropout may occur due to primer binding site mutations that impact one set of primers but not another.



Applied Biosystems has taken the strategy of not changing primer sequences with equivalent STR loci between their kits (except with MiniFiler) and uses mobility modifiers to adjust spacing between loci in the same dye channel where needed. Promega has moved primer positions in order to optimize spacing between STR allele ranges with new kits--thus, necessitating concordance studies to check for potential allele dropout due to primer binding site mutations. Published differences between STR kits due to allele dropout have been noted on the NIST STRBase website at <http://www.cstl.nist.gov/biotech/strbase/NullAlleles.htm>.

Total Number of Samples Attempted = 1461
 1443 with complete profiles

U.S. Population Samples (663 samples)

- Previously studied with Identifier™, MiniFiler™, Yfiler™, PP16, miniSTRs, and many additional assays (>200,000 allele calls)
- 260 African Americans, 260 Caucasians, 140 Hispanics, and 3 Asians

U.S. Father/Son pairs (786 samples)

- Previously studied with Identifier, MiniFiler, Yfiler, 23plex
- ~100 fathers/100 sons for each group: African Americans, Caucasians, Hispanics, and Asians


NIST SRM 2391b PCR DNA Profiling Standard (12 samples)

- Genomic Components 1-10 (includes 9947A and 9948)
- ABI 007 and K562

Comparisons for Each Locus between Various Kits and In-House Assays


As many as five different PCR primer sets from various kits and assays have been compared to one another

Locus	Kit/Assay 1	Kit/Assay 2	Kit/Assay 3	Kit/Assay 4	Kit/Assay 5
Amelogenin	PP-ESX17	PP-ESI17	Identifier	NIST 23plex	MiniFiler
D3S1358	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
TH01	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
D21S11	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
D18S51	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
D10S1248	PP-ESX17	PP-ESI17	NIST 23plex	NIST NC01	MiniFiler
D1S1656	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
D2S1338	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
D16S539	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
D22S1045	PP-ESX17	PP-ESI17	NIST 23plex	NIST NC01	MiniFiler
vWA	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
D8S1179	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
FGA	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
D2S441	PP-ESX17	PP-ESI17	NIST 23plex	NIST NC02	MiniFiler
D12S391	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
D19S433	PP-ESX17	PP-ESI17	Identifier	PP16	MiniFiler
SE33	PP-ESX17	PP-ESI17	SE33 monoplex	PP16	MiniFiler




PCR Product Size Ranges and Dye Labels for STR Loci in New Promega Kits

Five new European loci (boxed) are placed at the upper end with ESI vs ESX systems. ESX 16 and ESX 17 (as well as ESI 16 vs ESI 17) systems only differ by the addition of SE33 (German-recommended locus)



Example Data from PowerPlex ESX 17 System using 0.5 ng DNA and 30 cycles




Materials and Methods:
 DNA template: 0.5-1 ng input DNA
 PCR (GeneAmp 9700; manufacturer recommended protocols with some half reactions)
 ABI 3130xl; 36cm array, POP-4, 10s 8kV

Amplicon Size Differences (bp) for STR Loci from Several Kits Relative to ESX 17

STR Locus	ESI 17	PP16	Identifier
D1S1656	87	--	--
D2S441	260	--	74
D2S1338	26	--	-4
D8S1179	-128	0	-78
D10S1248	200	--	--
D12S391	164	--	--
D16S539	-187	-11	-42
D18S51	-151	0	-19
D19S433	-30	0	-100
D21S11	0	0	-17
D2S1045	225	--	--
FGA	-118	55	-56
TH01	-85	0	11
vWA	0	0	28
SE33	42	--	--
Amelogenin X	0	19	19
Amelogenin Y	0	19	19

Summary of Allele Discordance Observed

Number of Discordant Results Observed between Various STR Kits/Assays



Comparisons made with ~660 samples
 Comparisons made with ~1120 samples
 Comparisons made with ~1440 samples

7 differences between ESX 17 and ESI 17 beta-test materials
 96,124 allele comparisons (1,443 samples x 17 loci x 2 alleles/locus x 2 kits)

0.0071% discordance (primarily D22S1045, which has been corrected with an additional forward primer)
 0.0031% after correction

Details Regarding the 25 Discordant Results Observed (from >100,000 allele comparisons)

Locus	ESX 17	ESI 17	Identifier	MiniFiler	PP16	NC01	23plex	Sequence Reason for Discordance
D16S539	12,12	12,13	12,13	12,13	12,13	--	--	no loci discordance
D3S1358	14,17	14,17	14,16	14,16	14,17	--	--	G→C SNP 11 bp downstream - Identifier reverse primer
D19S433	13,14	13,14	14,14	--	--	--	--	T→A SNP 8 bp downstream - Identifier reverse primer
D19S433	13,14,2	13,14,2	14,2,14,2	--	--	--	--	T→A SNP 8 bp downstream - Identifier reverse primer
D22S1045	17,17	15,17	--	--	15,17	15,17	15,17	G→T SNP 15 bp upstream - ESX forward primer (corrected)
D22S1045	17,17	15,17	--	--	15,17	15,17	15,17	G→T SNP 15 bp upstream - ESX forward primer (corrected)
D22S1045	17,17	15,17	--	--	15,17	15,17	15,17	G→T SNP 15 bp upstream - ESX forward primer (corrected)
D22S1045	15,16	15,16	--	--	15,16	15,16	15,16	G→T SNP 15 bp upstream - ESX forward primer (corrected)
D1S1656	15,3,15,3	14,15,3	--	--	--	--	--	G→T SNP 30 bp upstream - ESX forward primer
SE33	26,2,27,2	26,2,27,2	--	--	--	--	--	G→C SNP 11 bp upstream - monoplex/SEFiler forward primer
SE33	20,28,3	20,28,3	20,28,2	--	--	--	--	T→G deletion 410 bp upstream - monoplex/SEFiler forward primer
SE33	24,2,28,2	24,2,28,2	23,2,28,2	--	--	--	--	G→T SNP 134 bp upstream - monoplex/SEFiler forward primer
SE33	21,2,26,2	21,2,26,2	21,2,21,2	--	--	--	--	G→T SNP 134 bp upstream - monoplex/SEFiler forward primer
SE33	24,2,25,2	24,2,25,2	24,2,24,2	--	--	--	--	G→T SNP 134 bp upstream - monoplex/SEFiler forward primer
SE33	19,19	19,25,2	19,25,2	--	--	--	--	G→T SNP 75 bp upstream - ESX reverse primer
D16S539	9,11	9,11	9,11	9,9	9,11	--	--	T→C SNP 34 bp downstream - MiniFiler reverse primer
D16S539	11,12	11,12	11,12	12,12	11,12	--	--	G→T SNP 34 bp downstream - MiniFiler reverse primer
D16S539	9,11	9,11	9,11	9,9	9,11	--	--	T→C SNP 34 bp downstream - MiniFiler reverse primer
D16S539	11,14	11,14	11,14	14,14	11,14	--	--	T→C SNP 34 bp downstream - MiniFiler reverse primer
D16S539	9,11	9,11	9,11	9,9	9,11	--	--	T→C SNP 34 bp downstream - MiniFiler reverse primer
D16S539	11,13	11,13	11,13	13,13	11,13	--	--	T→C SNP 34 bp downstream - MiniFiler reverse primer
D16S539	11,12	11,12	11,12	12,12	11,12	--	--	T→C SNP 34 bp downstream - MiniFiler reverse primer
D16S539	9,12	9,12	9,12	9,9	9,12	--	--	T→C SNP 34 bp downstream - MiniFiler reverse primer
D16S539	11,12	11,12	11,12	12,12	11,12	--	--	T→C SNP 34 bp downstream - MiniFiler reverse primer
D18S51	13,15	13,15	15,15	13,15	13,15	--	--	G→T SNP 172 bp upstream - Identifier reverse primer

NIST Funding: Interagency Agreement 2008-DN-R-121 between the National Institute of Justice and NIST Office of Law Enforcement Standards.
 Disclaimer: Points of view of the authors and do not necessarily represent the official position or policies of the US Department of Justice. Certain commercial equipment, instruments and materials are identified in order to specify experimental procedures as completely as possible. In no case does such identification imply a recommendation or endorsement by the National Institute of Standards and Technology nor does it imply that any of the materials, instruments or equipment identified are necessarily the best available for the purpose.