


**Testing and Evaluation Protocol for Spectroscopic  
Personal Radiation Detectors (SPRDs) for Homeland  
Security**

**T&E Protocol N42.48, 2010**

**Version 1.02**

## Table of Content

1.	Scope.....	1
2.	References.....	1
3.	Compliance Level Information.....	1
4.	Test and evaluation steps .....	1
5.	Recording test results.....	2
6.	Test report .....	2
7.	Guidance for testing ANSI N42.42 data format requirements .....	2
8.	Considerations .....	3

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 1 of 94

## Testing and Evaluation Protocol Alarming Personal Radiation Detectors for use in Homeland Security

### 1. Scope

This document establishes the protocol for testing alarming personal radiation detectors based on the performance requirements established in ANSI N42.48, “American National Standard Performance Criteria for Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.”

### 2. References

This protocol shall be used in conjunction with the following documents:

[R1] ANSI/IEEE N42.48, “American National Standard Performance Criteria for Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.”

[R2] ANSI/IEEE N42.42, “Data format standard for radiation detectors used for homeland security.”

[R3] NIST Handbook 150:2006, NVLAP Procedures and General Requirements

[R4] NIST Handbook 150-23:2007 (DRAFT) NVLAP Radiation Detection Instruments


### 3. Compliance Level Information

Instrument under test might meet all the requirements listed in the ANSI/IEEE N42.48 standard. Therefore, different agencies developed documents describing the compliance levels required for particular applications of the instruments under test. Examples of such compliance level requirements are those required by the Graduated Rad/Nuc Detector Evaluation and Reporting (GRaDER<sup>SM</sup>) program. For this program, information can be found in the “Compliance Level for GRaDER Instrument Performance” document located at <http://www.dhs.gov/GRaDER>.

### 4. Test and evaluation steps

It is recommended that testing laboratories perform the tests listed in this protocol in the following order:

- Check all items listed in the general requirements
- Perform the radiological tests
- Perform the temperature and humidity tests
- Perform the entire electrical and electromagnetic test except the Electrostatic Discharge (ESD) test
- Perform the impact and the vibration tests
- Perform the moisture and dust test
- Perform the ESD test
- Perform the drop test, as required

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 2 of 94

Excel template sheets are provided by NIST to the testing laboratory to guarantee that all data required is being provided in the test report.

## 5. Recording test results

This Test and Evaluation protocol contains data sheets that shall be used to record and report all test results. Each data sheet is associated with a specific section(s) of the referenced ANSI standard, N42.48. An electronic version of the data sheets is provided in the form of spreadsheets that may be used to record and report the results of the tests. These spreadsheets were verified and validated (V&V) using Microsoft Excel 2007 (compatibility mode).


Instrument status shall be recorded on the “Test Summary” sheet as testing is performed. The comment section in each data sheet shall be used to record changes to the test requirements and methods listed in the ANSI standard. The comment section shall also include the rationale of the changes.

## 6. Test report

A test report summarizing the results of the test shall include the following sections:

- a. Laboratory equipment information:
  1. Identify all participating laboratory facilities. Include points of contact names, mailing address, telephone number, and electronic mail addresses.
  2. Identify the tests performed in the different facilities.
  3. List all supporting equipment name, model number and last day of calibration used for each test.
- b. Test equipment information :
  1. Include manufacturer name, instrument model, instrument serial number, software and firmware version identification, and last day of calibration.
  2. List the operating modes and parameter setting of the instrument and accessory kit(s) used in each test.
- c. Data sheets:
  1. The data sheets listed in this document shall be completed and provided as part of the report.
  2. Include changes to the ANSI standard test requirements or methods and rationale to the changes.

## 7. Guidance for testing ANSI N42.42 data format requirements

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 3 of 94

The standard associated with this Test and Evaluation Protocol requires verification that an output data file is created that complies with ANSI/IEEE N42.42 standard requirements. The range of complexity of the N42.42 compliant instrument output file is extremely broad. Data output files from these instruments are simple files that can be checked manually using a text editor such as Notepad or WordPad. These files can also be verified using additional tools. In principle, all data output files that meet ANSI N42.42 can be verified manually using a text editor as these files are XML files. File reading software, such as Altova XMLSpy® 2009 Standard Edition can also be used for manual viewing and validating of structure and content.


N42.42 schemas can be used to validate the file format as specified in the ANSI/IEEE N42.42 standard. These schemas are available at the NIST web site  
<http://physics.nist.gov/Divisions/Div846/Gp4/ANSIN4242/xml.html>.

There are several XML validators that can be used to verify the XML structure of the N42.42 compliant instrument output file. Examples of these validators can be found at  
<http://www.xmlvalidation.com/> or <http://validator.w3.org/>.

## 8. Considerations

The standard establishes exposure rates for test in Roentgen per hour (R/h). When testing instruments that read in rem per hour, the test field shall be in rem/h instead of R/h. Refer to the “Units and Uncertainties” section in the standard for additional information.



	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 5 of 94


## Pre-Test Data Sheet and Report

<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>		<b>Serial Number:</b>	
<b>Date Performed:</b>		<b>Test Location:</b>	
<b>Requirement:</b>	Verify that the manufacturer supplied an operation and maintenance manual containing the information listed below.		
<b>Test Protocol:</b>	Review the information provided and indicate whether the required information has been provided. Also verify that the documentation is complete and understandable. The documentation should not be in draft form with incomplete sections.		
<b>Note:</b>	Comments are required when the requirement is not verified.		

### Test Results

Requirement	Yes	No
Operating instructions and restrictions	<input type="checkbox"/>	<input type="checkbox"/>
Electrical connection schematic	<input type="checkbox"/>	<input type="checkbox"/>
Spare parts list	<input type="checkbox"/>	<input type="checkbox"/>
Troubleshooting guide.	<input type="checkbox"/>	<input type="checkbox"/>
Description and protocol for communication methods of transmitting and receiving data	<input type="checkbox"/>	<input type="checkbox"/>
Contact information for the manufacturer including name, address, telephone #, fax #, email address, etc.	<input type="checkbox"/>	<input type="checkbox"/>
Power supply requirements	<input type="checkbox"/>	<input type="checkbox"/>
Recommended operational parameters such as: detector response and false alarm probability	<input type="checkbox"/>	<input type="checkbox"/>
Complete description of system or unit	<input type="checkbox"/>	<input type="checkbox"/>
Enclosure specification classification	<input type="checkbox"/>	<input type="checkbox"/>
Inclusion of any hazardous material that may require additional regulation	<input type="checkbox"/>	<input type="checkbox"/>
Description of data analysis software and radionuclide identification procedure	<input type="checkbox"/>	<input type="checkbox"/>
Description of operation and performance of the system or unit	<input type="checkbox"/>	<input type="checkbox"/>

<b>Comments:</b>			
<b>Completed by:</b>		<b>Date:</b>	
<b>Reviewed by:</b>		<b>Date:</b>	

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 6 of 94

## Controls Data Sheet Section 5.1


<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>			<b>Serial Number:</b>
<b>Date Performed:</b>			<b>Test Location:</b>
<b>Requirement:</b>	<p>Controls shall be clearly identified, easily operable under conditions of expected use, and adequately protected from accidental operation. The on/off button or any other control that could cause the instrument not to function as expected by the user shall be protected from accidental operation.</p>		
<b>Note:</b>	Comments are required when the requirement is not verified.		

1st Surface	Verify	
	Yes	No
Did the instrument turn off?		
Did the instrument change mode of operation of configuration?		


2nd Surface	Verify	
	Yes	No
Did the instrument turn off?		
Did the instrument change mode of operation of configuration?		

3rd Surface	Verify	
	Yes	No
Did the instrument turn off?		
Did the instrument change mode of operation of configuration?		




	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 7 of 94

	<b>Verify</b>	
<b>4th Surface</b>	<b>Yes</b>	<b>No</b>
Did the instrument turn off?		
Did the instrument change mode of operation of configuration?		
<b>5th Surface</b>	<b>Verify</b>	
	<b>Yes</b>	<b>No</b>
Did the instrument turn off?		
Did the instrument change mode of operation of configuration?		
<b>6th Surface</b>	<b>Verify</b>	
	<b>Yes</b>	<b>No</b>
Did the instrument turn off?		
Did the instrument change mode of operation of configuration?		
<b>Comments:</b>		
<b>Performed by:</b>		<b>Date:</b>
<b>Reviewed by:</b>		<b>Date:</b>

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 8 of 94

<b>Documentation</b>						
<b>Data Sheet Section 5.2</b>						
<b>Manufacturer:</b>						
<b>Instrument:</b>						
<b>Model:</b>				<b>Serial Number:</b>		
<b>Date Performed:</b>				<b>Test Location:</b>		
<b>Requirement:</b> Manufacturers shall provide instructions to verify proper operation of the instrument.						
Requirements are listed in Clause 10 (Documentation) of the ANSI/IEEE N42.48						
<b>Note:</b> Comments are required when the requirement is not verified.						



	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 10 of 94

**Displays**  
**Data Sheet Section 5.3**

<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>		<b>Serial Number:</b>	
<b>Date Performed:</b>		<b>Test Location:</b>	
<b>Requirement:</b>	<p>The instrument shall directly display the measured exposure rate or dose-equivalent rate with the associated radiological unit (e.g., <math>\mu\text{R/h}</math>, <math>\mu\text{Gy/h}</math>, or <math>\mu\text{Sv/h}</math>).</p> <p>Radionuclide identification results shall be displayed on the instrument. If measurement results can be viewed via a wireless or network link on a secondary device, the failure of that or any secondary device shall not affect the operation of the instrument.</p>		
<b>Note:</b> Comments are required when the requirement is not verified.			

	Verify	
	Yes	No
Is the display backlit?		
Is the display continuously lit?		
Is the display readable in low light level (Verified in section 5.13)?		
Is the display readable in high light level (Verified in section 5.13)?		
Can measurement results be viewed by a secondary device?		
Continues to function properly when the secondary device is switched off?		
Are radionuclide identification results displayed on the instrument?		

<b>Display</b>	LED		LCD		Other	
----------------	-----	--	-----	--	-------	--

<b>Display type</b>	Exposure rate		Dose-equivalent rate		Other	
---------------------	---------------	--	----------------------	--	-------	--


<b>Display range</b>	
----------------------	--

<b>Display units</b>	Units	
----------------------	-------	--

<b>Comments:</b>	
------------------	--


<b>Performed by:</b>		<b>Date:</b>	
----------------------	--	--------------	--

<b>Reviewed by:</b>		<b>Date:</b>	
---------------------	--	--------------	--

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 11 of 94


**Effective Range of Measurement or Indication  
Data Sheet Section 5.4**


<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>		<b>Serial Number:</b>	
<b>Date Performed:</b>		<b>Test Location:</b>	
<b>Requirement:</b>	<p>The effective range of measurement or indication shall be specified by the manufacturer and shall be from 5 µR/h to not less than 2 mR/h.</p> <p>The instrument response over the effective range specified by the manufacturer shall be tested. When exposed to radiation fields that are greater than the effective range of measurement, the instrument shall indicate an over-range condition.</p>		
<b>Note:</b>	Comments are required when the requirement is not verified.		
		<b>Verify</b>	
		<b>Yes</b>	<b>No</b>
For gammas; is the effective range at least 5 µR/h to 2 mR/h?		<input type="checkbox"/>	<input type="checkbox"/>
The instrument has an over-range indication?		<input type="checkbox"/>	<input type="checkbox"/>
For gammas; what is the effective range of measurement as stated by the manufacturer? (include units)			
For gammas; what is the display range shown by the instrument? (include units)			
For neutrons (if available); what is the effective range of measurement as stated by the manufacturer? (include units)			
For neutrons (if available); what is the display range shown by the instrument? (include units)			
<b>Describe over-range display:</b>			
<b>Comments:</b>			
<b>Performed by:</b>			<b>Date:</b>
<b>Reviewed by:</b>			<b>Date:</b>

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 12 of 94

## Audible Alarm Data Sheet Section 5.5

<b>Manufacturer:</b>							
<b>Instrument:</b>							
<b>Model:</b>				<b>Serial Number</b>			
<b>Date Performed:</b>				<b>Test Location:</b>			
<b>Requirements:</b>	<p>The instrument shall provide an audible alarm to indicate an increase in the radiation level that is greater than the alarm set point. Different techniques may be used to differentiate between the type of radiation detected (e.g., gamma, neutron, over-range). The frequency of an audible alarm signal shall be from 1000 Hz to at least 4000 Hz. Where an intermittent alarm signal is provided, the interval shall not exceed 2 s. The A-weighted alarm signal volume at a distance of 30 cm from the instrument shall be at least 80 dB(A) and shall not exceed 100 dB(A).</p> <p>It shall not be possible to disable both the vibration and audible alarm indications simultaneously, except through the restricted mode. When both alarm signals are off, an indication shall be provided on the display to inform the user of this condition.</p> <p>An earphone connection should be available to enable use of the audible function in a high-noise environment.</p>						
<b>Ambient Conditions:</b>		°C		%RH		in HG	
<b>Test Equipment:</b>							
<b>Instrument Mode of operation</b>							
<b>Note:</b>	Comments are required when the requirement is not verified.						

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 13 of 94

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 14 of 94

<b>Measurement Results</b>					
<b>Alarm volume</b>			<b>Frequency</b>		
		dB(A)			Hz
1					
2					
3					
4					
5					
6					
7					
8					
9					
10					
<b>Mean</b>	#DIV/0!	dB(A)	<b>Mean</b>	#DIV/0!	Hz
<b>STD</b>	#DIV/0!	dB(A)	<b>STD</b>	#DIV/0!	Hz
<b>COV %</b>	#DIV/0!		<b>COV %</b>	#DIV/0!	

	Verify	
	Yes	No
Is the alarm frequency within 1000 to 4000 Hz?		
Where an intermittent alarm is provided, is the interval less than 2 seconds?		
Is the alarm volume at a distance of 30 cm within 85 dB(A) and 100 dB(A)?		
If the audible alarm can be disabled, does the instrument have a vibration or visual alarm?		
Is an earphone connection available?		
Alarms are distinguishable for different types of radiation (gamma, neutron, over-range)? If yes, describe differences:		
Does the instrument have preventive measures for disabling all the alarms? If yes, describe:		

**Record intermittent alarm interval (seconds):** \_\_\_\_\_

**Comments:** \_\_\_\_\_


\_\_\_\_\_

\_\_\_\_\_

**Performed by:** \_\_\_\_\_ **Date:** \_\_\_\_\_


**Reviewed by:** \_\_\_\_\_ **Date:** \_\_\_\_\_



	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 15 of 94

## Vibration Alarm Data Sheet Section 5.6

<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>		<b>Serial Number:</b>	
<b>Date Performed:</b>		<b>Test Location:</b>	
<b>Requirements:</b>	<p>The instrument shall have a vibration alarm signal capability. The vibration alarm shall have sufficient intensity to inform the user of an alarm condition.</p> <p>The use of carrying pouches is discouraged. If a holder is used, there should be a rigid connection between the holder and the instrument such that there is no loss of vibration intensity to the user.</p> <p>The intensity of the vibration at the surface of the instrument (instrument pouch or holder, when used) shall be greater than 0.8 g. The vibration motor used by the instrument should rotate between 9000 rpm and 11000 rpm.</p>		
	<b>Note:</b> Comments are required when the requirement is not verified.		
<b>Test Equipment:</b>			
<b>Instrument Mode of operation:</b>			
<b><u>Instrument and Motor Verification</u></b>		<b>Verify</b>	
		<b>Yes</b>	<b>No</b>
Verify that new batteries are installed?		<input type="checkbox"/>	<input type="checkbox"/>
Motor rotation between 9000 and 11000 rpm?		<input type="checkbox"/>	<input type="checkbox"/>
What is the motor rpm?			

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 16 of 94

**Measurement Results**

Intensity (g)
1
2
3
4
5
6
7
8
9
10

Mean intensity	#DIV/0!
----------------	---------

Verify	
Yes	No

Has any measured reading greater than 0.8 g?		
Is the vibration signal intermittent?		

**Comments:**

---




---



---


**Performed by:** \_\_\_\_\_ **Date:** \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_ **Date:** \_\_\_\_\_

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 17 of 94

<b>Size, Mass, Reference point marking, and Explosive Atmospheres</b> <b>Sections 5.7-5.10</b>									
<b>Manufacturer:</b>									
<b>Instrument:</b>									
<b>Model:</b>						<b>Serial Number:</b>			
<b>Date Performed:</b>						<b>Test Location:</b>			
<b>Requirement:</b>	<b>5.7 Size</b> The overall dimensions of the instrument should be similar to that of a personal radiation detector (within the rectangular solid defined by 20 cm in length, 10 cm in width, and 5 cm in depth). Means shall be provided to securely fix the instrument to the user (for example, a clip, ring, or lanyard), with attention given to the necessary orientation of the detector and display.								
	<b>5.8 Mass</b> The mass of the complete instrument should not exceed 400 g.								
	<b>5.9 Reference point marking</b> The instrument shall have reference points on both the front, or back, and side indicating the effective center of the detector.  The instrument shall have an additional reference point indicating its orientation with respect to the wearer. The presence of a clip may be used as the reference point to indicate proper orientation.  All reference points shall be described in the instrument manual.								
	<b>5.10 Explosive atmospheres</b> The manufacturer shall state whether the instrument is certified for use in explosive atmospheres. If certification is claimed, documentation shall be provided. Certification should be based on UL-913-2004.								
<b>Note:</b>	Comments are required when the requirement is not verified.								



	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 19 of 94


### Section 5.11 Battery Lifetime Data Sheet and Report


<b>Manufacturer:</b>						
<b>Instrument:</b>						
<b>Model:</b>				<b>Serial Number:</b>		
<b>Date Performed:</b>				<b>Test Location:</b>		
<b>Requirement:</b>	<p>If non-rechargeable batteries are used, they shall be widely available, not unique to the instrument, and be replaceable in the field without the use of special tools. When rechargeable batteries are used, provisions shall be made to permit recharging from ac or dc (12 V) power sources.</p> <p>The batteries shall be capable of powering the instrument in a non-alarm state for a minimum of 16 h in a 50 <math>\mu\text{R/h}</math> (0.5 <math>\mu\text{Gy/h}</math>) field. The batteries shall be capable of powering the audible alarm continuously for 30 min.</p> <p>The instrument shall have a low battery indicator</p>					
<b>Note:</b>	Comments are required when the requirement is not verified.					

#### Test Results

	Rechargeable	Non-Rechargeable
Battery type used	<input type="checkbox"/>	<input type="checkbox"/>
	<b>Yes</b>	<b>No</b>
Is a low battery indications provided?	<input type="checkbox"/>	<input type="checkbox"/>
The batteries powered the instrument for 16 h in a non-alarming condition in a field of 50 $\mu\text{R/h}$ ?	<input type="checkbox"/>	<input type="checkbox"/>
Was the alarm activated after the 16 h of non-alarming operation?	<input type="checkbox"/>	<input type="checkbox"/>
The batteries powered the instrument for 30 min in an alarming condition?	<input type="checkbox"/>	<input type="checkbox"/>
Did the instrument alarm when the low battery indication was displayed?	<input type="checkbox"/>	<input type="checkbox"/>

<b>Comments:</b>						
<b>Completed by:</b>				<b>Date:</b>		
<b>Reviewed by:</b>				<b>Date:</b>		

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 20 of 94

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 21 of 94


## Section 5.12 Data transmission Data Sheet and Report

<b>Manufacturer:</b>							
<b>Instrument:</b>							
<b>Model:</b>				<b>Serial Number:</b>			
<b>Date Performed:</b>				<b>Test Location:</b>			
<b>Requirement:</b>	<p>The instrument shall have the ability to transfer data to an external device, such as a computer. The transfer should be based on a bi-directional port that meets the requirements of Ethernet, USB, wireless, or other electronic means, such as a removable media device. The technique used shall conform to applicable IEEE protocols. The transferred data shall be in the XML format following the format defined in ANSI N42.42. When used, wireless techniques shall have the ability to be encrypted.</p> <p>Communication protocols shall be described in the technical manual. Proprietary communication formats\ shall not be used and any required drivers shall be made available to the user.</p> <p>Proprietary software should not be required for remote data interpretation. If proprietary software is needed, the software shall be provided by the manufacturer..</p>						
<b>Note:</b>	Comments are required when the requirement is not verified.						

### Test Results


	Yes	No
Additional data transfer requirements were verified in 5.13 and 5.14?	<input type="checkbox"/>	<input type="checkbox"/>
Does the instrument transmit data to an external device?	<input type="checkbox"/>	<input type="checkbox"/>
Is the transfer bidirectional?	<input type="checkbox"/>	<input type="checkbox"/>
What type of port is used (e.g. Ethernet, USB, wireless, other)?		
The output file meets ANSI N42.42 requirements?	<input type="checkbox"/>	<input type="checkbox"/>
Is the data format described in the manual?	<input type="checkbox"/>	<input type="checkbox"/>
Is the communication protocol described in the manual?	<input type="checkbox"/>	<input type="checkbox"/>
Are proprietary communication formats used?	<input type="checkbox"/>	<input type="checkbox"/>
Are drivers provided to the users?	<input type="checkbox"/>	<input type="checkbox"/>
Is proprietary software used?	<input type="checkbox"/>	<input type="checkbox"/>
If yes, was the proprietary software provided?	<input type="checkbox"/>	<input type="checkbox"/>

<b>Comments:</b>							
<b>Completed by:</b>				<b>Date:</b>			
<b>Reviewed by:</b>				<b>Date:</b>			


	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 22 of 94

<b>User Interface</b>				
<b>Data Sheet Section 5.13</b>				
<b>Manufacturer:</b>				
<b>Instrument:</b>				
<b>Model:</b>			<b>Serial Number:</b>	
<b>Date Performed:</b>			<b>Test Location:</b>	
<b>Requirement:</b>	<p>The instrument shall include:</p> <ul style="list-style-type: none"> <li>a) A display that is easily readable over the required temperature range and under different lighting conditions.</li> <li>b) Controls that are user-friendly for routine operation.</li> <li>c) Controls and switches that are designed in a way to minimize accidental operation.</li> <li>d) A menu structure that is simple and easy to be followed intuitively.</li> <li>e) Detect, search/localize, and identification functions.</li> <li>f) The capability to operate if the user is wearing gloves.</li> <li>g) A method to inform the user of the expected time required to collect a spectrum and a means to allow the user to extend or reduce the collection time.</li> <li>h) An automated mode of operation that would automatically start spectrum collection and attempt to identify the radionuclide.</li> <li>i) Provide a status indicator, such as a flashing LED or LCD heartbeat, to inform the user that the instrument is functioning properly, including visual indication of an alarm condition.</li> </ul>			
<b>Note:</b>	Comments are required when the requirement is not verified.			




	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 23 of 94


USER 1			
	Yes		No
<b>Without Gloves in a low level light &lt;150 lux</b>			
It is possible to turn on the instrument?			
It is possible to calibrate the instrument?			
It is possible to go over the menu as described in the manual?			
It is possible to make an exposure rate measurement?			
It is possible to make an identification measurement?			
It is possible to save the data?			
Make a 137Cs identification measurement, was the collection time displayed?			
It is possible to extend the collection time?			
Is this extended time displayed?			
It is possible to reduce the collection time?			
Is this reduced time displayed?			
It is possible to transfer the data to an external device following the manufacturer provided information?			
Is wireless communication used?			
If wireless communication is used, is encryption available?			
It is possible to turn off the instrument?			
<b>Without Gloves in a high level light &gt; 10000 lux and &lt; 32000 lux</b>			
It is possible to turn on the instrument?			
It is possible to calibrate the instrument?			
It is possible to go over the menu as described in the manual?			
It is possible to make an exposure rate measurement?			
It is possible to make an identification measurement?			
It is possible to save the data?			
It is possible to turn off the instrument?			
<b>With Gloves</b>			
It is possible to turn on the instrument?			
It is possible to calibrate the instrument?			
It is possible to go over the menu as described in the manual?			
It is possible to make an exposure rate measurement?			
It is possible to make an identification measurement?			
It is possible to save the data?			
It is possible to turn off the instrument?			

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 24 of 94


USER 2			
	Yes		No
<b>Without Gloves in a low level light &lt;150 lux</b>			
It is possible to turn on the instrument?			
It is possible to calibrate the instrument?			
It is possible to go over the menu as described in the manual?			
It is possible to make an exposure rate measurement?			
It is possible to make an identification measurement?			
It is possible to save the data?			
Make a 137Cs identification measurement, was the collection time displayed?			
It is possible to extend the collection time?			
Is this extended time displayed?			
It is possible to reduce the collection time?			
Is this reduced time displayed?			
It is possible to transfer the data to an external device following the manufacturer provided information?			
Is wireless communication used?			
If wireless communication is used, is encryption available?			
It is possible to turn off the instrument?			
<b>Without Gloves in a high level light &gt; 10000 lux and &lt; 32000 lux</b>			
It is possible to turn on the instrument?			
It is possible to calibrate the instrument?			
It is possible to go over the menu as described in the manual?			
It is possible to make an exposure rate measurement?			
It is possible to make an identification measurement?			
It is possible to save the data?			
It is possible to turn off the instrument?			
<b>With Gloves</b>			
It is possible to turn on the instrument?			
It is possible to calibrate the instrument?			
It is possible to go over the menu as described in the manual?			
It is possible to make an exposure rate measurement?			
It is possible to make an identification measurement?			
It is possible to save the data?			
It is possible to turn off the instrument?			

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 25 of 94

USER 3			
	Yes		No
<b>Without Gloves in a low level light &lt;150 lux</b>			
It is possible to turn on the instrument?			
It is possible to calibrate the instrument?			
It is possible to go over the menu as described in the manual?			
It is possible to make an exposure rate measurement?			
It is possible to make an identification measurement?			
It is possible to save the data?			
Make a 137Cs identification measurement, was the collection time displayed?			
It is possible to extend the collection time?			
Is this extended time displayed?			
It is possible to reduce the collection time?			
Is this reduced time displayed?			
It is possible to transfer the data to an external device following the manufacturer provided information?			
Is wireless communication used?			
If wireless communication is used, is encryption available?			
It is possible to turn off the instrument?			
<b>Without Gloves in a high level light &gt; 10000 lux and &lt; 32000 lux</b>			
It is possible to turn on the instrument?			
It is possible to calibrate the instrument?			
It is possible to go over the menu as described in the manual?			
It is possible to make an exposure rate measurement?			
It is possible to make an identification measurement?			
It is possible to save the data?			
It is possible to turn off the instrument?			
<b>With Gloves</b>			
It is possible to turn on the instrument?			
It is possible to calibrate the instrument?			
It is possible to go over the menu as described in the manual?			
It is possible to make an exposure rate measurement?			
It is possible to make an identification measurement?			
It is possible to save the data?			
It is possible to turn off the instrument?			

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 26 of 94


For additional users add more tables as needed.									
Comments: _____									
<b>Performed by:</b> _____								<b>Date:</b> _____	
<b>Reviewed by:</b> _____								<b>Date:</b> _____	

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 27 of 94


### Spectral Identification Data Sheet Section 5.14

<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>			<b>Serial Number:</b>
<b>Date Performed:</b>			<b>Test Location:</b>
<b>Requirement:</b>	<p>a) The instrument shall have the ability to store and transfer at least fifty complete (unprocessed) spectra. Each stored spectrum shall contain collection and identification results information in the ANSI N42.42 format including:</p> <ul style="list-style-type: none"> <li>• Time and date</li> <li>• Instrument type and serial number</li> <li>• Hardware and software version</li> <li>• Identified radionuclides and associated confidence indications</li> <li>• Spectrum collection time interval</li> <li>• Measured gamma-ray exposure rate</li> <li>• Neutron count rate at the time of measurement, if provided</li> </ul> <p>b) An indication shall be displayed or otherwise provided (e.g., "not identified," "unknown radionuclide") if a radionuclide cannot be identified.</p> <p>c) The manufacturer shall describe the meaning of confidence indications.</p> <p>d) The instrument shall indicate if the exposure rate is too high for radionuclide identification.</p>		
<b>Note:</b>	Comments are required when the requirement is not verified.		


	Yes		No
Is the instrument capable of storing 50 complete unprocessed spectra?			
Is the instrument capable of transferring 50 complete unprocessed spectra?			
Is each spectral save in an ANSI N42.42 format?			
The instrument provides indication such as "not identified" or "unknown radionuclide" when a radionuclide cannot be identified? (Note- use data from test in section 6.10.5 to answer this question)			
The manual describes the meaning of the confidence indications?			
The instrument provides an indication when the exposure rate is too high for a radionuclide identification? (Note- use data from test in section 6.10.6 to answer this question)			

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 28 of 94

The Output File Contains the Following:	Yes	No
Time and date information?		
Instrument type and serial number information?		
Hardware and software version information?		
Identified radionuclide?		
Confidence indications for the radionuclides identified?		
Measured gamma-ray exposure rate?		
If provided, neutron count rate at the time of the measurement?		
Spectrum collection time interval?		
Comments:		
<b>Performed by:</b>		<b>Date:</b>
<b>Reviewed by:</b>		<b>Date:</b>

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 29 of 94


<b>Rate of False Alarms</b>						
<b>Data Sheet Section 6.2</b>						
<b>Manufacturer:</b>						
<b>Instrument:</b>						
<b>Model:</b>			<b>Serial Number:</b>			
<b>Date Performed:</b>			<b>Test Location:</b>			
<b>Requirements:</b>	<p>The false alarm rate for gamma and neutron (when applicable) shall be less or equal than 1 alarm per 10 hours when operated in a stable background environment.</p> <p>The alarm threshold shall be the same as that used for the "time to alarm" test</p>					
<b>Note:</b>	Comments are required when the requirement is not verified.					
<b>Ambient Conditions:</b>	_____ °C	_____ %RH	_____ in HG			
<b>Gamma Alarm Threshold:</b>			<b>Instrument Mode of operation</b>			
<b>Neutron Alarm Threshold:</b>						
Gamma Background measurement	_____	μR/h				
Neutron Background measurement	_____	(Add Units)				
			<b>Yes</b>	<b>No</b>		
For gammas; did the instrument alarm more than once over the test period ?						
If applicable, did the neutron alarm more than once over the test period?						
Record the number of gamma alarms during the test:						
Record the number of neutron alarms during the test:						
<b>Comments:</b>						
<b>Performed by:</b>				<b>Date:</b>		
<b>Reviewed by:</b>				<b>Date:</b>		

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 30 of 94

<b>Time to Alarm: Photons</b>						
<b>Data Sheet Section 6.3</b>						
<b>Manufacturer:</b>						
<b>Instrument:</b>						
<b>Model:</b>				<b>Serial Number:</b>		
<b>Date Performed:</b>				<b>Test Location:</b>		
<b>Requirement:</b>	The alarm shall activate within 2 s after exposure to an increase in the ambient radiation level of 50 $\mu\text{R/h}$ (0.5 $\mu\text{Gy/h}$ ) that occurs over a period of not more than 0.5 s.					
<b>Note:</b>	Comments are required when the requirement is not verified.					
<b>Ambient Conditions:</b>		°C		%RH	in HG	
<b>Test Equipment:</b>						
<b>Sources Data:</b>						
<b>Measurement Results</b>						
<b>Background Field (Cs-137)</b>		$\mu\text{R/h}$	<b>Alarm Threshold</b>		$\mu\text{R/h}$	
<b>Background Field (Am-241)</b>		$\mu\text{R/h}$	<b>Instrument Mode of operation</b>			
<b>Background Field (Co-60)</b>		$\mu\text{R/h}$				
<i>If not 50 <math>\mu\text{R/h}</math> enter:</i>						
<b>Testing Field (Cs-137)</b>		$\mu\text{R/h}$	<b>Instrument reading (Cs-137)</b>		$\mu\text{R/h}$	
<b>Testing Field (Am-241)</b>		$\mu\text{R/h}$	<b>Instrument reading (Am-241)</b>		$\mu\text{R/h}$	
<b>Testing Field (Co-60)</b>		$\mu\text{R/h}$	<b>Instrument reading (Co-60)</b>		$\mu\text{R/h}$	





	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 32 of 94

**Time to alarm - Neutron**  
**Data Sheet Section 6.4**

<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>		<b>Serial Number:</b>	
<b>Date Performed:</b>		<b>Test Location:</b>	
<b>Requirement:</b> The neutron alarm shall activate within 5 s after exposure to an unmoderated neutron field that occurs over a period of not more than 2 s.			
<b>Note:</b> Comments are required when the requirement is not verified.			


<b>Ambient Conditions:</b>	°C	%RH	in HG
<b>Test Equipment:</b>			
<b>Source Data:</b>	<b>Instrument Mode of operation</b>		
<b>Background Field</b>	(add units)	<b>Alarm Threshold</b>	(add units)

**Measurement Results**


Record in table if instrument alarmed or not within 5 seconds.

		Cf-252	
		time to alarm ≤5s	
<b>OPERATIONAL NOTE:</b>  If there is no alarm then a "No alarm" message needs to be recorded in the table	1		(Yes/No entry)
	2		
	3		
	4		
	5		
	6		
	7		
	8		
	9		
	10		





	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 34 of 94

Detection of gradually increasing gamma radiation levels and Neutron detection				
Data Sheet Section 6.5				
<b>Manufacturer:</b>				
<b>Instrument:</b>				
<b>Model:</b>			<b>Serial Number:</b>	
<b>Date Performed:</b>			<b>Test Location:</b>	
<b>Requirement:</b>	<p>The instrument's alarm threshold shall not be affected by slowly increasing radiation levels that may be caused when a wearer is slowly approaching or is being approached by a radiation source.</p> <p>The alarm shall activate within 5 seconds after the instrument reaches the test position.</p>			
<b>Note:</b>	Comments are required when the requirement is not verified.			
<b>Ambient Conditions:</b>		°C	%RH	in HG
<b>Test Equipment:</b>				
<b>Source Data:</b>				
<b>Gamma Background Field:</b>		µR/h	<b>Neutron Background Field:</b>	(add units)
<b>Gamma Alarm Threshold:</b>		µR/h	<b>Neutron Alarm Threshold:</b>	(add units)
<b>Instrument Mode of operation</b>				

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 35 of 94


<u>Measurement Results</u>				
Record in table if instrument alarmed or not within 2 seconds for gammas and 5 seconds for neutrons.				
OPERATIONAL NOTE:  If there is no alarm then a "No alarm" message needs to be recorded in the table		<b>Cs-137</b>	<b>Cf-252</b>	
		time to alarm ≤2s	time to alarm ≤5s	
	1			(Yes/No entry)
	2			
	3			
	4			
	5			
	6			
	7			
	8			
	9			
10				
		<b>Yes</b>	<b>No</b>	
Did the instrument alarm within 2 seconds for gamma?				
Did the instrument alarm within 5 seconds for neutron?				
<b>Comments:</b>				
<b>Performed by:</b>		<b>Date:</b>		
<b>Reviewed by:</b>		<b>Date:</b>		

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 36 of 94

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 37 of 94

**Accuracy - Photons**  
**Data Sheet Section 6.6**


<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>			<b>Serial Number:</b>
<b>Date Performed:</b>			<b>Test Location:</b>
<b>Requirement:</b>	Displayed exposure rates, when provided, shall be within $\pm 30\%$ % of the conventionally true value of the applied exposure rate using Am-241, Cs-137, and Co-60.		
<b>Note:</b>	Comments are required when the requirement is not verified.		
<b>Ambient Conditions:</b>	°C	%RH	in HG
<b>Test Equipment:</b>			
<b>Cs-137 Measurements</b>			
<b>Source Data:</b>	<b>Instrument Mode of operation</b>		
<b>Background</b>	$\mu\text{R/h}$	<b>at test location</b>	
<b>Gamma Alarm Threshold:</b>	$\mu\text{R/h}$		
<b>Maximum instrument range display</b>	$\text{mR/h}$		
<b>Radiation fields in mR/h</b>	For 400 $\mu\text{R/h}$	For 1 $\text{mR/h}$	For 80% max rate
<b>Actual Radiation fields used in the test in mR/h</b>	<b>0.40</b>	<b>1.00</b>	<b>0.00</b>
1			(add instrument units)
2			
3			
4			
5			
6			
7			
8			
9			
10			
Mean	#DIV/0!	#DIV/0!	#DIV/0!
Std dev	#DIV/0!	#DIV/0!	#DIV/0!
COV %	#DIV/0!	#DIV/0!	#DIV/0!
low (-30%)	0.28	0.70	0.00
high (+30%)	0.52	1.30	0.00

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 38 of 94


<u>Am-241 Measurements</u>				
<b>Source Data:</b>	<b>Instrument Mode of operation</b>			
<b>Background</b>	µR/h	<b>at test location</b>		
<b>Gamma Alarm Threshold:</b>	µR/h			
<b>Maximum instrument range display</b>	mR/h			
		For 400 µR/h	For 1 mR/h	For 80% max rate
<b>Radiation fields used for test in mR/h</b>	<b>0.40</b>	<b>1.00</b>	<b>0.00</b>	
<b>Actual Radiation fields used in the test in mR/h</b>				
	1			(add instrument units)
	2			
	3			
	4			
	5			
	6			
	7			
	8			
	9			
	10			
	Mean	#DIV/0!	#DIV/0!	#DIV/0!
	Std dev	#DIV/0!	#DIV/0!	#DIV/0!
	COV %	#DIV/0!	#DIV/0!	#DIV/0!
	low (-30%)	0.28	0.70	0.00
	high (+30%)	0.52	1.30	0.00
<u>Co-60 Measurements</u>				
<b>Source Data:</b>	<b>Instrument Mode of operation</b>			
<b>Background</b>	µR/h	<b>at test location</b>		
<b>Gamma Alarm Threshold:</b>	µR/h			
<b>Maximum instrument range display</b>	mR/h			






	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 40 of 94

<b>Personal Radiation Alarm</b>				
<b>Data Sheet Section 6.7</b>				
<b>Manufacturer:</b>				
<b>Instrument:</b>				
<b>Model:</b>			<b>Serial Number:</b>	
<b>Date Performed:</b>			<b>Test Location:</b>	
<b>Requirement:</b>	<p>The instrument shall provide an alarm that will alert the user to the presence of a relatively high radiation field. The alarm shall be audible and visible, and be different than those associated with the radiation indication alarm described in step 6.3.</p> <p>Using Cs-137, expose the instrument to a 10 mR/h (100 µGy/h) radiation field. The personal alarm shall be activated within 2 s of the exposure. Reduce the radiation field and repeat the exposure two additional times for a total of three trials. The alarm shall activate within the time specified for each trial.</p>			
<b>Note:</b>	Comments are required when the requirement is not verified.			
<b>Ambient Conditions:</b>		°C		%RH
<b>Test Equipment:</b>				
<b>Source Data:</b>				

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 41 of 94

Measurement Results											
<b>Background Field</b>		μR/h	<b>Alarm Threshold</b>								
			μR/h								
<b>Instrument Mode of operation</b>											
<i>If not 10 mR/h enter:</i>											
<b>Testing Field (Cs-137)</b>		mR/h	<b>Instrument reading (Cs-137)</b>								
			mR/h								
Record in table if instrument alarmed or not within 2 seconds.											
	<table border="1" style="margin-left: auto; margin-right: auto;"> <tr> <th style="text-align: center;">Cs-137</th> <th style="text-align: center;">time to alarm ≤2s</th> </tr> <tr> <td style="text-align: center;">1</td> <td>(Yes/No entry)</td> </tr> <tr> <td style="text-align: center;">2</td> <td></td> </tr> <tr> <td style="text-align: center;">3</td> <td></td> </tr> </table>	Cs-137	time to alarm ≤2s	1	(Yes/No entry)	2		3		<b>OPERATIONAL NOTE:</b>  If there is no alarm then a "No alarm" message needs to be recorded in the table	
Cs-137	time to alarm ≤2s										
1	(Yes/No entry)										
2											
3											
		<b>Yes</b>	<b>No</b>								
Did the instrument alarm within 2 seconds for Cs-137?											
It has an audible personal radiation alarm?											
Is the alarm audible alarm different to that of the radiation indication alarm?											
If yes, then describe:											
It has an visible personal radiation alarm?											
Is the alarm visible alarm different to that of the radiation indication alarm?											
If yes, then describe:											
<b>For Cs-137:</b>		Record maximum time to alarm:	seconds								
		Record minimum time to alarm:	seconds								
<b>Comments:</b>											
<b>Performed by:</b>		<b>Date:</b>									
<b>Reviewed by:</b>		<b>Date:</b>									

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 42 of 94

### Over-Range Response Data Sheet Section 6.8

<b>Manufacturer:</b>				
<b>Instrument:</b>				
<b>Model:</b>		<b>Serial Number:</b>		
<b>Date Performed:</b>		<b>Test Location:</b>		

**Requirement:** When exposed to an exposure rate that is two times the maximum exposure rate specified by the manufacturer, the indication of the instrument shall remain at the maximum of that range, and an over-range indication shall be displayed for the duration of the exposure. The instrument shall recover within 5 minute when the radiation field is reduced.

**Note:** Comments are required when the requirement is not verified.

<b>Ambient Conditions:</b>		°C		%RH		in HG
<b>Test Equipment:</b>						
<b>Source Data:</b>						
<b>Instrument Mode of operation</b>						
Manufacturer-Stated Max Exposure Rate:		mR/h				
Over-Range Test Exposure Rate:		mR/h				
Over-Range Exposure Duration:		min.				

	Verify	
	Yes	No
Was an over-range indication displayed?		
Did the instrument recover within 5 minutes?		


**Recorded recovery time:** \_\_\_\_\_ (add units)

**Comments:**  
\_\_\_\_\_  
\_\_\_\_\_

**Performed by:** \_\_\_\_\_ **Date:** \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_ **Date:** \_\_\_\_\_



	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 44 of 94


## Section 6.10.1 Radionuclide Categorization Data Sheet and Report

<b>Manufacturer:</b>								
<b>Instrument:</b>								
<b>Model:</b>					<b>Serial Number:</b>			
<b>Date Performed:</b>					<b>Test Location:</b>			
<b>Requirement:</b>	<p>The manufacturer shall state the radionuclides that the instrument can identify radionuclides by category. The categories selected should be based on the following list:</p> <ul style="list-style-type: none"> <li>- Special Nuclear Materials: Uranium (used to indicate 233U, 235U), 237Np, Pu.</li> <li>- Medical radionuclides: 18F, 67Ga, 51Cr, 75Se, 89Sr, 99Mo, 99mTc, 103Pd, 111In, Iodine (123I, 125I, 131I), 153Sm, 201Tl, 133Xe.</li> <li>- Naturally occurring radioactive materials (NORM): 40K, 226Ra, 232Th and daughters, 238U and daughters.</li> <li>- Industrial radionuclides: 57Co, 60Co, 133Ba, 137Cs, 192Ir, 204Tl, 226Ra, and 241Am.</li> </ul>							
<b>Note:</b>	Comments are required when the requirement is not verified.							

### Test Results

	Yes	No
The manufacturer states the radionuclides that the instrument can identify by category.	<input type="checkbox"/>	<input type="checkbox"/>
The instrument can identify (at a minimum) the four different categories of radionuclides listed in the requirement.	<input type="checkbox"/>	<input type="checkbox"/>

<b>Comments:</b>								
<b>Completed by:</b>					<b>Date:</b>			
<b>Reviewed by:</b>					<b>Date:</b>			

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 45 of 94


**Section 6.10.2.2 Single Radionuclide Identification  
Data Sheet and Report**

<b>Manufacturer:</b>							
<b>Instrument:</b>							
<b>Model:</b>				<b>Serial Number:</b>			
<b>Date Performed:</b>				<b>Test Location:</b>			
<b>Requirement:</b>	<p>The instrument shall be able to identify the following radionuclides within the time specified by the manufacturer with a maximum of 5 min. The manufacturer shall provide radionuclide-specific test results.</p> <p>Medical radionuclides: 67Ga, 99mTc, Iodine (123I, 131I), 201Tl</p> <p>NORM: 40K, 226Ra, 232Th</p> <p>Industrial radionuclides: 22Na, 57Co, 60Co, 133Ba, 137Cs, 152Eu, 192Ir, and 241Am</p> <p>Special nuclear materials: HEU (highly enriched uranium, 235U &gt;90%), Pu [Reactor grade plutonium (&gt; 6% 240Pu)]</p>						
<b>Note:</b>	Comments are required when the requirement is not verified.						

<b>NIST</b>	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	TITLE: Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 46 of 94

Single Radionuclide Identification Test Data					
	<b>Date Performed:</b>	<b><sup>22</sup>Na</b>			<b>Date Performed:</b>
		Shielded	Unshielded		
	1				1
	2				2
	3				3
	4				4
	5				5
	6				6
	7				7
	8				8
	9				9
	10				10
	Corr				Corr
	<b>Date Performed:</b>	<b><sup>40</sup>K</b>			<b>Date Performed:</b>
		Shielded	Unshielded		
	1				1
	2				2
	3				3
	4				4
	5				5
	6				6
	7				7
	8				8
	9				9
	10				10
	Corr				Corr
	<b>Date Performed:</b>	<b><sup>51</sup>Co</b>			<b>Date Performed:</b>
		Shielded	Unshielded		
	1				1
	2				2
	3				3
	4				4
	5				5
	6				6
	7				7
	8				8
	9				9
	10				10
	Corr				Corr
	<b>Date Performed:</b>	<b><sup>60</sup>Co</b>			<b>Date Performed:</b>
		Shielded	Unshielded		
	1				1
	2				2
	3				3
	4				4
	5				5
	6				6
	7				7
	8				8
	9				9
	10				10
	Corr				Corr
	<b>Date Performed:</b>	<b><sup>67</sup>Ga</b>			<b>Date Performed:</b>
		Shielded	Unshielded		
	1				1
	2				2
	3				3
	4				4
	5				5
	6				6
	7				7
	8				8
	9				9
	10				10
	Corr				Corr
	<b>Date Performed:</b>	<b><sup>99m</sup>Tc</b>			<b>Date Performed:</b>
		Shielded	Unshielded		
	1				1
	2				2
	3				3
	4				4
	5				5
	6				6
	7				7
	8				8
	9				9
	10				10
	Corr				Corr



	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	TITLE: Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 47 of 94

<b>Date Performed:</b>		<sup>123</sup> I		<b>Date Performed:</b>		<sup>131</sup> I	
		Shielded	Unshielded			Shielded	Unshielded
	1				1		
	2				2		
	3				3		
	4				4		
	5				5		
	6				6		
	7				7		
	8				8		
	9				9		
	10				10		
	Corr				Corr		
<b>Date Performed:</b>		<sup>133</sup> Ba		<b>Date Performed:</b>		<sup>137</sup> Cs	
		Shielded	Unshielded			Shielded	Unshielded
	1				1		
	2				2		
	3				3		
	4				4		
	5				5		
	6				6		
	7				7		
	8				8		
	9				9		
	10				10		
	Corr				Corr		
<b>Date Performed:</b>		<sup>152</sup> Eu		<b>Date Performed:</b>		<sup>192</sup> Ir	
		Shielded	Unshielded			Shielded	Unshielded
	1				1		
	2				2		
	3				3		
	4				4		
	5				5		
	6				6		
	7				7		
	8				8		
	9				9		
	10				10		
	Corr				Corr		

<b>NIST</b>	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	TITLE: Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 48 of 94

	<b>Date Performed:</b>	<b><sup>201</sup>Tl</b>	<b>Date Performed:</b>	<b><sup>226</sup>Ra</b>
		Shielded    Unshielded		Shielded    Unshielded
	1			1
	2			2
	3			3
	4			4
	5			5
	6			6
	7			7
	8			8
	9			9
	10			10
	Corr			Corr
	<b>Date Performed:</b>	<b><sup>232</sup>Th</b>	<b>Date Performed:</b>	<b>HEU</b>
		Shielded    Unshielded		Shielded    Unshielded
	1			1
	2			2
	3			3
	4			4
	5			5
	6			6
	7			7
	8			8
	9			9
	10			10
	Corr			Corr
	<b>Date Performed:</b>	<b>RGPu</b>	<b>Date Performed:</b>	<b><sup>241</sup>Am</b>
		Shielded    Unshielded		Shielded    Unshielded
	1			1
	2			2
	3			3
	4			4
	5			5
	6			6
	7			7
	8			8
	9			9
	10			10
	Corr			Corr



**TEST AND EVALUATION PROTOCOL**

**TEP NO.**  
N42.48

**PREPARED BY:**  
DIV682

**TITLE:** Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.

**EFF. DATE**  
2010-11-09

**REV.**  
1.02

**PAGE**  
49 of 94

Did the instrument ID correctly 8 out of 10 time?			
Unshielded			
Date of Test	Radionuclide	Yes	No
	<sup>22</sup> Na	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>40</sup> K	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>57</sup> Co	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>60</sup> Co	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>67</sup> Ga	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>99m</sup> Tc	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>123</sup> I	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>131</sup> I	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>133</sup> Ba	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>137</sup> Cs	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>152</sup> Eu	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>192</sup> Ir	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>201</sup> Tl	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>226</sup> Ra	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>232</sup> Th	<input type="checkbox"/>	<input type="checkbox"/>
	HEU	<input type="checkbox"/>	<input type="checkbox"/>
	RGPu	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>241</sup> Am	<input type="checkbox"/>	<input type="checkbox"/>

Did the instrument ID correctly 8 out of 10 time?			
Shielded			
Date of Test	Radionuclide	Yes	No
	<sup>22</sup> Na	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>40</sup> K	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>57</sup> Co	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>60</sup> Co	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>67</sup> Ga	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>99m</sup> Tc	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>123</sup> I	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>131</sup> I	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>133</sup> Ba	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>137</sup> Cs	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>152</sup> Eu	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>192</sup> Ir	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>201</sup> Tl	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>226</sup> Ra	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>232</sup> Th	<input type="checkbox"/>	<input type="checkbox"/>
	HEU	<input type="checkbox"/>	<input type="checkbox"/>
	RGPu	<input type="checkbox"/>	<input type="checkbox"/>
	<sup>241</sup> Am	<input type="checkbox"/>	<input type="checkbox"/>

Source Information			
Unshielded			
Date	Radionuclide	Activity	Exposure rate
	<sup>22</sup> Na		
	<sup>40</sup> K		
	<sup>57</sup> Co		
	<sup>60</sup> Co		
	<sup>67</sup> Ga		
	<sup>99m</sup> Tc		
	<sup>123</sup> I		
	<sup>131</sup> I		
	<sup>133</sup> Ba		
	<sup>137</sup> Cs		
	<sup>152</sup> Eu		
	<sup>192</sup> Ir		
	<sup>201</sup> Tl		
	<sup>226</sup> Ra		
	<sup>232</sup> Th		
	HEU		
	RGPu		
	<sup>241</sup> Am		

(add units)

Source Information			
Shielded			
Date	Radionuclide	Activity	Exposure rate
	<sup>22</sup> Na		
	<sup>40</sup> K		
	<sup>57</sup> Co		
	<sup>60</sup> Co		
	<sup>67</sup> Ga		
	<sup>99m</sup> Tc		
	<sup>123</sup> I		
	<sup>131</sup> I		
	<sup>133</sup> Ba		
	<sup>137</sup> Cs		
	<sup>152</sup> Eu		
	<sup>192</sup> Ir		
	<sup>201</sup> Tl		
	<sup>226</sup> Ra		
	<sup>232</sup> Th		
	HEU		
	RGPu		
	<sup>241</sup> Am		

(add units)

Comments: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Completed by: \_\_\_\_\_

Date: \_\_\_\_\_

Reviewed by: \_\_\_\_\_

Date: \_\_\_\_\_


<b>NIST</b>	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 50 of 94

### Section 6.10.3 Simultaneous Radionuclide Identification Data Sheet and Report

<b>Manufacturer:</b>						
<b>Instrument:</b>						
<b>Model:</b>				<b>Serial Number:</b>		
<b>Date Performed:</b>				<b>Test Location:</b>		
<b>Requirement:</b>	The instrument shall be able to identify at least two radionuclides simultaneously. Use <sup>99m</sup> Tc + <sup>137</sup> Cs for the test.					
<b>Note:</b>	Comments are required when the requirement is not verified.					

#### Simultaneous Radionuclide Identification Test Data

<b>Date Performed:</b>		<b><sup>99m</sup>Tc + <sup>137</sup>Cs</b>		<b><sup>99m</sup>Tc Source data:</b>	
		<b>Unshielded</b>			
	1				
	2				
	3				
	4			<b><sup>137</sup>Cs Source data:</b>	
	5				
	6				
	7				
	8				
	9				
	10				
	Corr				

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 51 of 94

<b>Did the instrument categorized correctly 8 out of 10 time?</b>			
<b>Date of Test</b>	<b>Radionuclide</b>	<b>Yes</b>	<b>No</b>
	<sup>99m</sup> Tc + <sup>137</sup> Cs	<input type="checkbox"/>	<input type="checkbox"/>
<b>Did the instrument ID correctly 8 out of 10 time?</b>			
<b>Date of Test</b>	<b>Radionuclide</b>	<b>Yes</b>	<b>No</b>
	<sup>99m</sup> Tc + <sup>137</sup> Cs	<input type="checkbox"/>	<input type="checkbox"/>
<b>Comments:</b>			
<b>Completed by:</b>		<b>Date:</b>	
<b>Reviewed by:</b>		<b>Date:</b>	

<b>NIST</b>	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	TITLE: Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 52 of 94

### Section 6.10.4 Masking Identification Test Data and Report

**Manufacturer:** \_\_\_\_\_  
**Instrument:** \_\_\_\_\_

**Model:** \_\_\_\_\_ **Serial Number:** \_\_\_\_\_  
**Date Performed:** \_\_\_\_\_ **Test Location:** \_\_\_\_\_

**Requirement:** The instrument shall provide an indication (e.g., the correct identification, "unknown," "unable to identify") when exposed to a radionuclide masked by another radionuclide that is not listed in the library or that has a much higher radiation intensity than the unmasked radionuclide.

**Note:** Comments are required when the requirement is not verified.

**Ambient Conditions:** \_\_\_\_\_ °C \_\_\_\_\_ %RH \_\_\_\_\_ in HG

**Test Equipment Used:** \_\_\_\_\_

**Test using: <sup>67</sup>Ga and <sup>54</sup>Mn**

**Date Performed:** \_\_\_\_\_

<sup>67</sup> Ga + <sup>54</sup> Mn
1
2
3
4
5
6
7
8
9
10
<b>Number Correct</b>

**<sup>67</sup>Ga Source data:** \_\_\_\_\_

**<sup>54</sup>Mn Source data:** \_\_\_\_\_


(check that <sup>54</sup>Mn is not in the instrument library, if in library need to choose a different radionuclide for this test)

Did the instrument ID correctly 8 out of 10 time?			
Date of Test	Radionuclide	Yes	No
	<sup>67</sup> Ga + <sup>54</sup> Mn	<input type="checkbox"/>	<input type="checkbox"/>








	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 55 of 94



	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	TITLE: Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 57 of 94

### Section 6.10.6 Over-Range Characteristics for Identification Test Data and Report

<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>		<b>Serial Number:</b>	
<b>Date Performed:</b>		<b>Test Location:</b>	
<b>Requirement:</b>	The manufacturer shall state the maximum exposure rate (relative to <sup>137</sup> Cs) for identification.		
<b>Note:</b>	Comments are required when the requirement is not verified.		

<b>Ambient Conditions:</b>	°C	%RH	in HG
<b>Test Equipment Used:</b>			
<b>Source Data:</b>			
<b>Manufacturer stated maximum exposure rate:</b>		(add units)	

#### Measurement Results

	<b>Cs-137</b>	<b>Exposure rate used in the test:</b>	0 (add units)
1			
2			
3			
4			
5			
6			
7			
8			
9			
10			
Number Correct			


  

	<b>Yes</b>	<b>No</b>	
Did instrument provide indication that exposure rate was too high?			

<b>Did the instrument ID correctly 8 out of 10 time?</b>			
		<b>Yes</b>	<b>No</b>
Date of Test	Radionuclide		
	<sup>137</sup> Cs	<input type="checkbox"/>	<input type="checkbox"/>

<b>Comments:</b>			
<b>Performed by:</b>		<b>Date:</b>	
<b>Reviewed by:</b>		<b>Date:</b>	


	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 58 of 94

<b>Sections 7.2 - Temperature Test Data and Report</b>																										
<b>Manufacturer:</b>																										
<b>Instrument:</b>																										
<b>Model:</b>			<b>Serial Number:</b>																							
<b>Date Performed:</b>			<b>Test Location:</b>																							
<p><b>Requirements:</b> The instrument shall function correctly at temperatures from -20 °C to +50 °C. No alarms shall occur as a result of the temperature conditions alone. If the manufacturer specifies a broader operating temperature range, the instrument shall be tested at the broader temperature range as specified by the manufacturer.</p> <p>Relative humidity shall be within the range specified in Table 1.</p> <p>NOTE: Record units displayed by instrument</p> <hr/> <p><b>Note:</b> Comments are required when the requirement is not verified.</p>																										
<table style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 30%;"><b>Ambient Conditions:</b></td> <td style="width: 15%;"></td> <td style="width: 10%; text-align: center;">°C</td> <td style="width: 15%;"></td> <td style="width: 10%; text-align: center;">%RH</td> <td style="width: 15%;"></td> <td style="width: 15%; text-align: center;">in HG</td> </tr> <tr> <td><b>Test Equipment Used:</b></td> <td colspan="6"></td> </tr> <tr> <td><b>Source Data:</b></td> <td colspan="6"></td> </tr> </table>						<b>Ambient Conditions:</b>		°C		%RH		in HG	<b>Test Equipment Used:</b>							<b>Source Data:</b>						
<b>Ambient Conditions:</b>		°C		%RH		in HG																				
<b>Test Equipment Used:</b>																										
<b>Source Data:</b>																										


<b>NIST</b>	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 59 of 94

**Measurement Results**

Exposure rate <sup>137</sup> Cs	22° C as read	30° C as read	40° C as read	50° C as read	10° C as read	0° C as read	-10° C as read	-20° C as read	
1									(add units)
2									
3									
4									
5									
6									
7									
8									
9									
10									
Mean	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	
STD	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	
COV %	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	
Readings within acceptance range?	Yes								
	No								
Gamma alarm (within 2s)	Yes								
	No								
Neutron alarm (within 5s)	Yes								
	No								
ID Trial 1 <sup>133</sup> Ba+ <sup>137</sup> Cs									Record all Radionuclides Identified
ID Trial 2 <sup>133</sup> Ba+ <sup>137</sup> Cs									
ID Trial 3 <sup>133</sup> Ba+ <sup>137</sup> Cs									
(± 15%) Acceptance Range:				#DIV/0!	to	#DIV/0!			
				-15%		+ 15 %			
<b>Gamma alarm:</b>	Record maximum time to alarm:				seconds				
	Record minimum time to alarm:				seconds				
<b>Neutron alarm:</b> (if applicable)	Record maximum time to alarm:				seconds				
	Record minimum time to alarm:				seconds				
<b>Comments:</b>	_____								
	_____								
<b>Completed by:</b>	_____				<b>Date:</b>	_____			
<b>Reviewed by:</b>	_____				<b>Date:</b>	_____			

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 60 of 94

Section 7.3 Temperature Shock Test Data and Report											
<b>Manufacturer:</b>											
<b>Instrument:</b>											
<b>Model:</b>						<b>Serial Number:</b>					
<b>Date Performed:</b>						<b>Test Location:</b>					
<b>Requirement:</b>	The instrument shall be fully functional within 1 h of exposure to rapid temperature changes from 22 °C to -20 °C, -20 °C to 22 °C, 22 °C to 50 °C, and 50 °C to 22 °C, with each change being made in less than 5 min.										
	No alarms shall occur as a result of the changing temperature conditions alone. Relative humidity shall be within the range specified in Table 1.										
<b>Note:</b>	Comments are required when the requirement is not verified.										
<b>Ambient Conditions:</b>			°C			%RH			in HG		
<b>Test Equipment Used:</b>											
<b>Sources Data:</b>											

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 61 of 94

**Measurement Results**

22° C		Exposure rate <sup>137</sup> Cs		22 to -20° C				- 20 to 22° C				22 to 50° C				50 to 22° C				(add units)
				15	30	45	60	15	30	45	60	15	30	45	60	15	30	45	60	
1	<b>Nominal Mean</b>																			
2	<b>Acceptance Range</b>																			
3	#DIV/0!	to	#DIV/0!																	
4		low	high																	
5		(-15%)	(+15%)																	
6																				
7																				
8																				
9																				
10																				
<b>Mean</b>	#DIV/0!			<b>Mean</b>	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!
<b>STD</b>	#DIV/0!			<b>STD</b>	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!
<b>COV%</b>	#DIV/0!			<b>COV%</b>	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!
				<b>Conf-Int (-)</b>	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!
				<b>Conf-Int (+)</b>	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!
				Readings within acceptance range?	<b>Yes</b>															
					<b>No</b>															
				Gamma alarm (within 2s)	<b>Yes</b>															
					<b>No</b>															
				Neutron alarm (within 5s)	<b>Yes</b>															
					<b>No</b>															
				ID Trial 1 <sup>133</sup> Ba+ <sup>137</sup> Cs																
				ID Trial 2 <sup>133</sup> Ba+ <sup>137</sup> Cs																
				ID Trial 3 <sup>133</sup> Ba+ <sup>137</sup> Cs																

Record all Radionuclides Identified

**ID Acceptance Criteria:**  
 Correct Radionuclides ID: (A is correct)  
 Temperature shock Pass/Fail: Comparison of ID results at test points with the ID results at ambient temperature being the baseline.

- A: Ba-133, Cs-137
- B:
- C:
- D:

**Gamma alarm:** Record maximum time to alarm: \_\_\_\_\_ seconds

Record minimum time to alarm: \_\_\_\_\_ seconds

**Neutron alarm:** (if applicable) Record maximum time to alarm: \_\_\_\_\_ seconds

Record minimum time to alarm: \_\_\_\_\_ seconds

**Comments:**


\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_


**Performed by:** \_\_\_\_\_ **Date:** \_\_\_\_\_

**Reviewed by:** \_\_\_\_\_ **Date:** \_\_\_\_\_


	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 62 of 94

<b>Sections 7.4 - Humidity</b> <b>Test Data and Report</b>				
<b>Manufacturer:</b>				
<b>Instrument:</b>				
<b>Model:</b>			<b>Serial Number:</b>	
<b>Date Performed:</b>			<b>Test Location:</b>	
<b>Requirements:</b>	The instrument shall function correctly over the range of relative humidity up to 93% RH at 35 °C. No alarms shall occur as a result of the humidity conditions alone.			
<b>Note:</b>	Comments are required when the requirement is not verified.			
<b>Ambient Conditions:</b>		°C	%RH	in HG
<b>Test Equipment Used:</b>				
<b>Source Data:</b>				




	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 63 of 94


<u>Measurement Results</u>					
		Nominal 40% RH 22° C	93% RH 35° C	40% RH 35° C	
		Exposure Rate Readings			
1					(add units)
2					
3					
4					
5					
6					
7					
8					
9					
10					
Mean		#DIV/0!	#DIV/0!	#DIV/0!	
STD		#DIV/0!	#DIV/0!	#DIV/0!	
COV %		#DIV/0!	#DIV/0!	#DIV/0!	
Readings within acceptance range?	<b>Yes</b>				
	<b>No</b>				
Gamma alarm (within 2s)	<b>Yes</b>				
	<b>No</b>				
Neutron alarm (within 5s)	<b>Yes</b>				
	<b>No</b>				
ID Trial 1 <sup>133</sup> Ba+ <sup>137</sup> Cs					Record all Radionuclides Identified
ID Trial 2 <sup>133</sup> Ba+ <sup>137</sup> Cs					
ID Trial 3 <sup>133</sup> Ba+ <sup>137</sup> Cs					
<b>Gamma alarm:</b>	Record maximum time to alarm:				seconds
	Record minimum time to alarm:				seconds
<b>Neutron alarm:</b> (if applicable)	Record maximum time to alarm:				seconds
	Record minimum time to alarm:				seconds
<b>Comments:</b>					
<b>Completed by:</b>				<b>Date:</b>	
<b>Reviewed by:</b>				<b>Date:</b>	

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 64 of 94

Sections 7.5.2 - Moisture and Dust Protection - Dust				
Test Data and Report				
<b>Manufacturer:</b>				
<b>Instrument:</b>				
<b>Model:</b>			<b>Serial Number:</b>	
<b>Date Performed:</b>			<b>Test Location:</b>	
<b>Requirements:</b>	The instrument case design shall meet the requirements stated for IP code 53 [R13], which means that the instrument shall be protected from the ingress of dust and spraying water. For IP53, the ingress of dust is not totally prevented, but dust shall not penetrate in a quantity to interfere with satisfactory operation of the instrument or to impair safety, and water sprayed at an angle up to 60° on either side of the vertical shall have no harmful effects.			
<b>Note:</b>	Comments are required when the requirement is not verified.			
<b>Ambient Conditions:</b>		°C	%RH	in HG
<b>Test Equipment Used:</b>				
<b>Source Data:</b>				

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 65 of 94




	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 67 of 94

<b>Sections 7.5.3 - Moisture and Dust Protection, Moisture Test Data and Report</b>						
<b>Manufacturer:</b>						
<b>Instrument:</b>						
<b>Model:</b>				<b>Serial Number:</b>		
<b>Date Performed:</b>				<b>Test Location:</b>		
<b>Requirements:</b>	<p>The test shall be made using a suitable nozzle (see IEC 60068-2-75:1997) with the water pressure adjusted to give flow rate of 10 L/min ±5%, which should be kept constant during the test. The water temperature should not differ by more than 5 °C from the temperature of the instrument under test. The test duration is 1 min/m<sup>2</sup> of the calculated surface area of the instrument with a minimum duration of 5 min.</p>					
<b>Note:</b>	Comments are required when the requirement is not verified.					
<b>Ambient Conditions:</b>		°C		%RH		in HG
<b>Test Equipment Used:</b>						
<b>Source Data:</b>						

<b>NIST</b>	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	TITLE: Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 68 of 94

		Pre-Test Response	Post Test Response			
		<b>Photon Readings</b>		(add units)		
1						
2						
3						
4						
5						
6						
7						
8						
9						
10						
				<b>Acceptance Range</b>		
				#DIV/0!	to	#DIV/0!
				low 15%		high 15%
	Mean	#DIV/0!	#DIV/0!			
	STD	#DIV/0!	#DIV/0!			
	COV %	#DIV/0!	#DIV/0!			
<b>Was the post-test response within the acceptance range?</b>		<b>Yes</b>				
		<b>No</b>				
<b>Gamma alarm (within 2s)</b>		<b>Yes</b>				
		<b>No</b>				
<b>Neutron alarm (within 5s)</b>		<b>Yes</b>				
		<b>No</b>				
				<b>Record all Radionuclides Identified</b>		
ID Trial 1 <sup>133</sup> Ba+ <sup>137</sup> Cs						
ID Trial 2 <sup>133</sup> Ba+ <sup>137</sup> Cs						
ID Trial 3 <sup>133</sup> Ba+ <sup>137</sup> Cs						
<b>Gamma alarm:</b>		Record maximum time to alarm: _____		seconds		
		Record minimum time to alarm: _____		seconds		
<b>Neutron alarm:</b> (if applicable)		Record maximum time to alarm: _____		seconds		
		Record minimum time to alarm: _____		seconds		
		<b>Yes</b>		<b>No</b>		
<b>Did water penetrate the instrument?</b>						
<b>Comments:</b> _____						
<b>Completed by:</b> _____				<b>Date:</b> _____		
<b>Reviewed by:</b> _____				<b>Date:</b> _____		

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 69 of 94


### Sections 7.6 - Cold Temperature Start Up Test Data and Report

<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>			<b>Serial Number:</b>
<b>Date Performed:</b>			<b>Test Location:</b>
<b>Requirements:</b>	The instrument shall be able to operate when switched on at the cold temperature limit (-20 °C).		
	Pre-Test and Post-Test are made at 22 °C		
<b>Note:</b>	Comments are required when the requirement is not verified.		
<b>Ambient Conditions:</b>	°C	%RH	in HG
<b>Manufacturers stated warm-up time:</b>	(add units)		
<b>Test Equipment Used:</b>			
<b>Source Data:</b>			


<b>NIST</b>	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	TITLE: Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 70 of 94


		Pre-Test Readings	Post-Test Readings	Readings at - 20 °C			
		<b>Photon Readings</b>					
	1				(add units)		
	2						
	3						
	4						
	5						
	6						
	7				<b>Acceptance Range</b>		
	8					to	
	9				#DIV/0!		#DIV/0!
	10				low 15%		high 15%
	Mean	#DIV/0!	#DIV/0!	#DIV/0!			
	STD	#DIV/0!	#DIV/0!	#DIV/0!			
	COV %	#DIV/0!	#DIV/0!	#DIV/0!			
<b>Were the results within the acceptance range?</b>		<b>Yes</b>					
		<b>No</b>					
Gamma alarm (within 2s)		<b>Yes</b>					
		<b>No</b>					
Neutron alarm (within 5s)		<b>Yes</b>					
		<b>No</b>					
ID Trial 1 <sup>133</sup> Ba+ <sup>137</sup> Cs					<b>Record all Radionuclides Identified</b>		
ID Trial 2 <sup>133</sup> Ba+ <sup>137</sup> Cs							
ID Trial 3 <sup>133</sup> Ba+ <sup>137</sup> Cs							
<b>Gamma alarm:</b>		Record maximum time to alarm: _____			seconds		
		Record minimum time to alarm: _____			seconds		
<b>Neutron alarm:</b> (if applicable)		Record maximum time to alarm: _____			seconds		
		Record minimum time to alarm: _____			seconds		
<b>Comments:</b>							
<b>Completed by:</b> _____				<b>Date:</b> _____			
<b>Reviewed by:</b> _____				<b>Date:</b> _____			




	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	TITLE: Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 71 of 94

<b>Section 8.2 Electrostatic Discharge</b>				
<b>Test Data and Report</b>				
<b>Manufacturer:</b>				
<b>Instrument:</b>				
<b>Model:</b>			<b>Serial Number:</b>	
<b>Date Performed:</b>			<b>Test Location:</b>	
<b>Requirement:</b>	During exposure to electrostatic discharges at intensities of up to 6 kV using the contact discharge technique, the instrument shall function correctly. No alarms shall occur as a result of the electrostatic discharge alone.			
<b>Note:</b>	Comments are required when the requirement is not verified.			
<b>Ambient Conditions:</b>				
	°C	% RH	in Hg	
<b>Test Equipment Used:</b>				
<b>Source Data:</b>				
<b>Measurement Results</b>				
<b>Test with sources</b>				
	<b>Pre-Test Response</b>			
1		(add units)		
2				
3				
4		<b>Acceptance Range</b>		
5		#DIV/0!	to	#DIV/0!
6		low (-15%)		high (+15%)
7				
8				
9				
10				
<b>Mean</b>	#DIV/0!			
<b>STD</b>	#DIV/0!			
<b>COV%</b>	#DIV/0!			

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 72 of 94

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 73 of 94

		Post-test Readings			
		Gamma Exposure rate			
		2 kV	4 kV	6 kV	
	1				(add units)
	2				
	3				
	4				
	5				
	6				
	7				
	8				
	9				
	10				
	<b>Mean</b>	#DIV/0!	#DIV/0!	#DIV/0!	
<b>Were the results within the acceptance range?</b>	<b>Yes</b>				
	<b>No</b>				
Gamma alarm (within 2s)	<b>Yes</b>				
	<b>No</b>				
Neutron alarm (within 5s)	<b>Yes</b>				
	<b>No</b>				
	ID Trial 1 <sup>133</sup> Ba+ <sup>137</sup> Cs				Record all Radionuclides Identified
	ID Trial 2 <sup>133</sup> Ba+ <sup>137</sup> Cs				
	ID Trial 3 <sup>133</sup> Ba+ <sup>137</sup> Cs				
<b>Gamma alarm:</b>	Record maximum time to alarm:				seconds
	Record minimum time to alarm:				seconds
<b>Neutron alarm:</b> (if applicable)	Record maximum time to alarm:				seconds
	Record minimum time to alarm:				seconds
<hr/>					
<b>Test without sources</b>					
<b>Did the instrument alarm during testing?</b>					
		2 kV	4 kV	6 kV	
	<b>Yes</b>				
	<b>No</b>				
<b>Comments:</b>					
<hr/>					
<hr/>					
<b>Performed by:</b>				<b>Date:</b>	
<hr/>				<hr/>	
<b>Reviewed by:</b>				<b>Date:</b>	
<hr/>				<hr/>	


	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 74 of 94

### Section 8.3 Radio Frequency Test Data and Report


<b>Manufacturer:</b>			
<b>Instrument:</b>			
<b>Model:</b>			<b>Serial Number:</b>
<b>Date Performed:</b>			<b>Test Location:</b>
<b>Requirement:</b>	The instrument shall not be affected by radio frequency (RF) fields over the frequency range of 80 MHz to 2.5 GHz at an intensity of 50 volts per meter (V/m). When exposed to these RF fields, the instrument shall function correctly. No alarms shall occur as a result of the RF field alone.		
<b>Note:</b>	Comments are required when the requirement is not verified.		
<b>Ambient Conditions:</b>	°C	%RH	In. Hg
<b>Test Equipment Used:</b>			
<b>Frequency Scan Observations Without Radioactive Sources</b>			
<b>Were susceptibilities observed?</b>			
Yes		No	
<b>Did the unit alarm during testing?</b>			
Yes		No	

<b>NIST</b>	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	TITLE: Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 75 of 94

<b>With <sup>137</sup>Cs Source</b>				
	<b>Nominal Response (µR/h) - No RF</b>		<b>Source Data:</b>	
1	(add units)			
2				
3		<b>Acceptance Range</b>		
4		<b>#DIV/0!</b>	<b>to</b>	<b>#DIV/0!</b>
5		low (-15%)		high (+15%)
6		<b>Frequency Scan Observations with Radioactive Sources</b>		
7				
8				
9				
10				
Mean	#DIV/0!			
STD	#DIV/0!			
COV%	#DIV/0!			
		<b>Were susceptibilities observed?</b>		
		Yes	No	
	<b>Comments:</b>			
	<b>Completed by:</b>		<b>Date:</b>	
	<b>Reviewed by:</b>		<b>Date:</b>	

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 76 of 94


<b>Sections 8.4 - Magnetic Fields</b>							
<b>Test Data and Report</b>							
<b>Manufacturer:</b>							
<b>Instrument:</b>							
<b>Model:</b>				<b>Serial Number:</b>			
<b>Date Performed:</b>				<b>Test Location:</b>			
<b>Requirements:</b>	When exposed to direct current (dc) magnetic fields in all three mutually orthogonal orientations relative to a 10 gauss (1 mT) magnetic field, the instrument shall function correctly. No alarms shall occur as a result of the magnetic field alone						
<b>Note:</b>	Comments are required when the requirement is not verified.						
<b>Ambient Conditions:</b>		°C		%RH			in HG
<b>Test Equipment Used:</b>							

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 77 of 94

Measurement Results Without Source								
	Initial Orientation		Second Orientation		Third Orientation			
	Nominal Zero Intensity	10 Gauss (DC)	Nominal Zero Intensity	10 Gauss (DC)	Nominal Zero Intensity	10 Gauss (DC)		
1								
2								
3								
4								
5								
6								
7								
8								
9								
10								
<b>Mean</b>	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	(add units)	
<b>STD</b>	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!		
<b>COV%</b>	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!		
							Acceptance Range	
							Initial Orientation:	#DIV/0! to #DIV/0!
							Second Orientation:	#DIV/0! to #DIV/0!
							Third Orientation:	#DIV/0! to #DIV/0!
							low -15%	high +15%
							<b>Yes</b>	Did the instrument alarm during the test?
							<b>No</b>	
							<b>Yes</b>	Were there any functional changes?
							<b>No</b>	
<b>Observations:</b>								







	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 79 of 94

### Section 8.5 Radiated Emissions Test Data and Report


<b>Manufacturer:</b>													
<b>Instrument:</b>													
<b>Model:</b>			<b>Serial Number:</b>										
<b>Date Performed:</b>			<b>Test Location:</b>										
<b>Requirement:</b>	RF emissions from an instrument shall be less than that which can interfere with other equipment located in the area of use. RF emissions when measured at 3 m shall be less than those shown below:												
	<table border="1" style="margin: auto;"> <thead> <tr> <th style="text-align: center;">Emission Frequency Range (MHz)</th> <th style="text-align: center;">Field Strength (micro volts/meter)</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">30 – 88</td> <td style="text-align: center;">100</td> </tr> <tr> <td style="text-align: center;">88 – 216</td> <td style="text-align: center;">150</td> </tr> <tr> <td style="text-align: center;">216 – 960</td> <td style="text-align: center;">200</td> </tr> <tr> <td style="text-align: center;">Above 960</td> <td style="text-align: center;">500</td> </tr> </tbody> </table>			Emission Frequency Range (MHz)	Field Strength (micro volts/meter)	30 – 88	100	88 – 216	150	216 – 960	200	Above 960	500
Emission Frequency Range (MHz)	Field Strength (micro volts/meter)												
30 – 88	100												
88 – 216	150												
216 – 960	200												
Above 960	500												
	<b>Note:</b> Comments are required when the requirement is not verified.												
<b>Ambient Conditions:</b>	°C	%RH	in HG										
<b>Test Equipment Used:</b>													
	<b>Yes</b>	<b>No</b>											
	<b>Were RF emissions above the limits?</b>												
<b>Comments:</b>													
<b>Performed by:</b>			<b>Date:</b>										
<b>Reviewed by:</b>			<b>Date:</b>										

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 80 of 94

<b>Section 9.1 Vibration</b>				
<b>Test Data and Report</b>				
<b>Manufacturer:</b>				
<b>Instrument:</b>				
<b>Model:</b>			<b>Serial Number:</b>	
<b>Date Performed:</b>			<b>Test Location:</b>	
<b>Requirement:</b>	The instrument shall withstand exposure to vibrations associated with the operation of handheld or hand-carried equipment. The physical condition and functionality of the instrument shall not be affected by exposure (e.g.: solder joints shall hold, nuts and bolts shall not come loose).			
<b>Note:</b>	Comments are required when the requirement is not verified.			
<b>Ambient Conditions:</b>	°C	%RH	in HG	
<b>Test Equipment Used:</b>				
<b>Source Data:</b>				


	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 81 of 94

		Pre-test	After Position A	After Position B	After Position C	
Exposure Rate Readings						
1						(add units)
2						
3						
4						
5						
6						
7						
8						
9						
10						
						Acceptance Range
						#DIV/0! to #DIV/0!
	Mean	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	-15% to 15%
	STD	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	
	COV%	#DIV/0!	#DIV/0!	#DIV/0!	#DIV/0!	
Readings within acceptance range?	Yes					
	No					
Gamma alarm (within 2s)	Yes					
	No					
Neutron alarm (within 5s)	Yes					
	No					
	ID Trial 1 <sup>133</sup> Ba+ <sup>137</sup> Cs					Record all Radionuclides Identified
	ID Trial 2 <sup>133</sup> Ba+ <sup>137</sup> Cs					
	ID Trial 3 <sup>133</sup> Ba+ <sup>137</sup> Cs					
Did the instrument controls function properly?	Yes					
	No					
Did the instrument show visible external damage?	Yes					
	No					
<b>Comments:</b>						
<b>Performed by:</b>				<b>Date:</b>		
<b>Reviewed by:</b>				<b>Date:</b>		


	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 82 of 94

<b>Section 9.2 Drop Test</b>									
<b>Test Data and Report</b>									
<b>Manufacturer:</b>									
<b>Instrument:</b>									
<b>Model:</b>					<b>Serial Number:</b>				
<b>Date Performed:</b>					<b>Test Location:</b>				
<b>Requirement:</b>	<p>After being subjected to drops on each of its six surfaces from a height of 1.5 m onto a concrete floor, the instrument shall function correctly and alarm at a change in the radiation field. It is acceptable if a transient alarm occurs at the moment of exposure.</p>								
<b>Note:</b>	Comments are required when the requirement is not verified.								
<b>Ambient Conditions:</b>		°C		%RH				in HG	
<b>Test Equipment Used:</b>									
<b>Source Data:</b>									




	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 84 of 94


Section 9.3 Impact (Microphonics)									
Test Data and Report									
<b>Manufacturer:</b>									
<b>Instrument:</b>									
<b>Model:</b>					<b>Serial Number:</b>				
<b>Date Performed:</b>					<b>Test Location:</b>				
<b>Requirement:</b>	The instrument shall be unaffected by microphonic conditions such as those that may occur from low intensity impacts from sharp contact with hard surfaces. The physical condition and functionality of the instrument shall not be affected by exposure (e.g., solder joints shall hold, nuts and bolts shall not come loose).								
<b>Note:</b>	Comments are required when the requirement is not verified.								
<b>Ambient Conditions:</b>		°C		%RH		in HG			
<b>Test Equipment Used:</b>									

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 85 of 94


Measurement Results - Without sources												
	Side No. 1		Side No. 2		Side No. 3		Side No. 4		Side No. 5		Side No. 6	
	Yes	No	Yes	No	Yes	No	Yes	No	Yes	No	Yes	No
Did the gamma response remain stable during the test?												
Did the neutron reading change?												
Did the instrument alarm during the test?												
Were there any functional changes due to the impacts?												
Did the instrument display spurious indications?												


Measurement Results - With Sources												
Source Data: _____												
		Pretest Response	Post-test Response									
1				(add units)								
2												
3												
4												
5				<b>Acceptance Range</b> #DIV/0! to #DIV/0! low (-15 %) high (+15 %)								
6												
7												
8												
9												
10												
Mean		#DIV/0!	#DIV/0!									
STD		#DIV/0!	#DIV/0!									
COV%		#DIV/0!	#DIV/0!									
Readings within acceptance range?	Yes											
	No											
Gamma alarm (within 2s)	Yes											
	No											
Neutron alarm (within 5s)	Yes											
	No											
ID Trial 1 <sup>133</sup> Ba+ <sup>137</sup> Cs				Record all Radionuclides Identified								
ID Trial 2 <sup>133</sup> Ba+ <sup>137</sup> Cs												
ID Trial 3 <sup>133</sup> Ba+ <sup>137</sup> Cs												
Did the instrument controls function properly?	Yes											
	No											
Did the instrument show visible external damage?	Yes											
	No											

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 86 of 94

<b>Comments:</b>											
<b>Completed by:</b>						<b>Date:</b>					
<b>Reviewed by:</b>						<b>Date:</b>					
											
<b>Note:</b> Side 6 is opposite to Side 1.											




	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 87 of 94

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 88 of 94


## Sections 10 - Documentation

### Test Data and Report


<b>Manufacturer:</b>				
<b>Instrument:</b>				
<b>Model:</b>			<b>Serial Number:</b>	
<b>Date Performed:</b>			<b>Test Location:</b>	
<b>Requirements:</b>	<p>10.1 Type test report The manufacturer shall provide a report covering the type tests performed in accordance with the requirements of this standard.</p> <p>10.2 Certificate The manufacturer shall provide a certificate or other documentation containing at least the following information:</p> <ul style="list-style-type: none"> <li>- Contacts for the manufacturer including, but not limited to, name, address, telephone number, fax number, e-mail address, etc.</li> <li>- Type of instrument, detector, and types of radiation the instrument is designed to measure</li> <li>- Range of exposure rates the instrument is designed to measure</li> <li>- Reference points and reference orientation for radiation source used for calibration</li> <li>- Location and dimensions of the sensitive volumes of the detectors</li> <li>- Response of the instrument to different appropriate radiation energies</li> <li>- Results of tests for accuracy, linearity, and lower limit of detection</li> <li>- Results of calibration tests (isotopes calibration with and date of next calibration due date)</li> <li>- Weight and dimensions of the instrument</li> <li>- Power supply (battery) requirements</li> <li>- Results of tests under environmental conditions</li> <li>- Results of electrical and mechanical tests</li> <li>- List of radionuclides to which the instrument was tested</li> <li>- FWHM and efficiency for 137Cs</li> </ul> <p>10.3 Operation and maintenance manuals The manufacturer shall supply an operation and maintenance manual containing at least the following information for the user:</p> <ul style="list-style-type: none"> <li>- Operating instructions and restrictions. Instructions shall include information regarding alarm threshold adjustments.</li> <li>- Troubleshooting guide.</li> </ul>			
<b>Note:</b>	Only one data sheet per model is required. Comments are required when the requirement is not verified.			

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 89 of 94



	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 91 of 94

<b>Comments:</b>					
<b>Completed by:</b>				<b>Date:</b>	
<b>Reviewed by:</b>				<b>Date:</b>	

	<b>TEST AND EVALUATION PROTOCOL</b>	<b>TEP NO.</b> N42.48	<b>PREPARED BY:</b> DIV682	
	<b>TITLE:</b> Spectroscopic Personal Radiation Detectors (SPRDs) for Homeland Security.	<b>EFF. DATE</b> 2010-11-09	<b>REV.</b> 1.02	<b>PAGE</b> 92 of 94