

Temperature Log for Vaccines (Celsius)

Month/Year: _____ Days 1-16

Completing the temperature log: Check the temperatures in both the freezer and the refrigerator compartments of your vaccine storage units at least twice each working day. Place an "X" in the box that corresponds with the temperature readings, and you initials. Once the month has ended, save each month's form for 3 years, unless state or local jurisdictions require a longer time period.

If the recorded temperature is in the shaded zone: This represents unacceptable temperature range. Follow these steps: 1. Label vaccine as "potentially compromised." 2. Move vaccine to functioning storage area as quickly as possible. 3. Call *USAMMA immediately to determine if potency of vaccines has been affected. 4. Call your Regional Analyst for further assistance.

Day of the month		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	
Staff Initials																		
Room Temp.																		
Exact Time																		
°C Temp		AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	
Too warm*	≥11°																	
	10°	Take immediate action if temperature is in shaded section*																
	9°																	
Refrigerator temperature	8°																	
	7°																	
	6°																	
	5°																	
	4°																	
	3°																	
	2°																	
	1°																	
	0°	Take immediate action if temperature is in shaded section*																
	≤1°																	
Too warm*	≥12°	Take immediate action if temperature is in shaded section*																
	-13°	Take immediate action if temperature is in shaded section*																
	-14°	Take immediate action if temperature is in shaded section*																
	-15°																	
	-16°																	
Freeze temp	-17°																	

*USAMMA Emergency Telephone Numbers: Phone 301-619-4318/1197/4198, DSN (343), After Hours: 301-676-1184/0857, E-mail: USAMMADOC@amedd.army.mil

Adapted by the Military Vaccine (MILVAX) Agency courtesy of the Immunization Action Coalition.

Vaccine Storage Troubleshooting Record

Date	Time	Storage Unit Temp	Room Temp	Problem	Action Taken	Results	Initials

Temperature Log for Vaccines (Celsius)

Month/Year: _____ Days 17-31

Completing the temperature log: Check the temperatures in both the freezer and the refrigerator compartments of your vaccine storage units at least twice each working day. Place an "X" in the box that corresponds with the temperature readings, and your initials. Once the month has ended, save each month's form for 3 years, unless state or local jurisdictions require a longer time period.

If the recorded temperature is in the shaded zone: This represents unacceptable temperature range. Follow these steps: 1. Label vaccine as "potentially compromised." 2. Move vaccine to functioning storage area as quickly as possible. 3. Call *USAMMA immediately to determine if potency of vaccines has been affected. 4. Call your Regional Analyst for further assistance.

Day of the month	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Staff Initials																
Room Temp.																
Exact Time																
°C Temp	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM	AM	PM
Refrigerator temperature Aim for 4°	≥11°	Take immediate action if temperature is in shaded section*														
	10°															
	9°															
	8°															
	7°															
	6°															
	5°															
	4°															
	3°															
	2°															
Too cold*	1°	Take immediate action if temperature is in shaded section*														
	0°															
	≤1°															
Freezer temp	≥12°	Take immediate action if temperature is in shaded section*														
	-13°															
	-14°															
	-15°															
	-16°															
	≤-17°															

*USAMMA Emergency Telephone Numbers: Phone 301-619-4318/1197/4198, DSN (343), After Hours: 301-676-1184/0857, E-mail: USAMMADOC@amedd.army.mil

Adapted by the Military Vaccine (MILVAX) Agency courtesy of the Immunization Action Coalition.

Vaccine Storage Troubleshooting Record

Date	Time	Storage Unit Temp	Room Temp	Problem	Action Taken	Results	Initials