

Standard Reference Materials:

Summary of the Environmental Research, Analysis, and Control Standards Issued by the National Bureau of Standards

R. Mavrodineanu¹


Center for Analytical Chemistry
National Measurement Laboratory
National Bureau of Standards
Gaithersburg, MD 20899

and

S. D. Rasberry

Office of Standard Reference Materials
National Measurement Laboratory
National Bureau of Standards
Gaithersburg, MD 20899

¹Present Address: 227 Almeria Road
West Palm Beach, FL 33405



U.S. DEPARTMENT OF COMMERCE, Malcolm Baldrige, Secretary
NATIONAL BUREAU OF STANDARDS, Ernest Ambler, Director

Issued March 1986

Library of Congress Catalog Card Number: 86-600504

National Bureau of Standards Special Publication 260-105
Natl. Bur. Stand. (U.S.), Spec. Publ. 260-105, 93 pages (Mar. 1986)
CODEN: XNBSAV

U.S. GOVERNMENT PRINTING OFFICE
WASHINGTON: 1986

For sale by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402

Preface

Standard Reference Materials (SRM's) as defined by the National Bureau of Standards (NBS) are well-characterized materials, produced in quantity and certified for one or more physical or chemical properties. They are used to assure the accuracy and compatibility of measurements throughout the Nation. SRM's are widely used as primary standards in many diverse fields in science, industry, and technology, both within the United States and throughout the world. They are also used extensively in the fields of environmental and clinical analysis. In many applications, traceability of quality control and measurement processes to the national measurement system is carried out through the mechanism and use of SRM's. For many of the Nation's scientists and technologists it is therefore of more than passing interest to know the details of the measurements made at NBS in arriving at the certified values of the SRM's produced. An NBS series of papers, of which this publication is a member, called the NBS Special Publication - 260 Series, is reserved for this purpose.

The 260 Series is dedicated to the dissemination of information on different phases of the preparation, measurement, certification and use of NBS SRM's. In general, much more detail will be found in these papers than is generally allowed, or desirable, in scientific journal articles. This enables the user to assess the validity and accuracy of the measurement processes employed, to judge the statistical analysis, and to learn details of techniques and methods utilized for work entailing the greatest care and accuracy. These papers also should provide sufficient additional information not found on the certificate so that new applications in diverse fields not foreseen at the time the SRM was originally issued will be sought and found.

Inquiries concerning the technical content of this paper should be directed to the author(s). Other questions concerned with the availability, delivery, price, and so forth, will receive prompt attention from:

Office of Standard Reference Materials
National Bureau of Standards
Gaithersburg, MD 20899

Stanley D. Rasberry, Chief
Office of Standard Reference Materials

OTHER NBS PUBLICATIONS IN THIS SERIES

- Catalog of NBS Standard Reference Materials (1984-85 edition), Catherine H. Hudson, ed., NBS Spec. Publ. 260 (February 1984). \$5.50* SN003-003-02558-5.
- Michaelis, R. E., and Wyman, L. L. Standard Reference Materials: Preparation of White Cast Iron Spectrochemical Standards. NBS Misc. Publ. 260-1 (June 1964). COM74-11061**
- Michaelis, R. E., Wyman, L. L., and Flitsch, R., Standard Reference Materials: Preparation of NBS Copper-Base Spectrochemical Standards. NBS Misc. Publ. 260-2 (October 1964). COM74-11063**
- Michaelis, R. E., Yakowitz, H., and Moore, G. A., Standard Reference Materials: Metallographic Characterization of an NBS Spectrometric Low-Alloy Steel Standard. NBS Misc. Publ. 260-3 (October 1964). COM74-11060**
- Hague, J. L. Mcars, T. W., and Michaelis, R. E., Standard Reference Materials: Sources of Information, NBS Misc. Publ. 260-4 (February 1965). COM74-11059
- Alvarez, R., and Flitsch R., Standard Reference Materials: Accuracy of Solution X-Ray Spectrometric Analysis of Copper-Base Alloys. NBS Misc. Publ. 260-5 (March 1965). PB168068**
- Shultz, J. I., Standard Reference Materials: Methods for the Chemical Analysis of White Cast Iron Standards, NBS Misc. Publ. 260-6 (July 1965). COM74-11068**
- Bell, R. K., Standard Reference Materials: Methods for the Chemical Analysis of NBS Copper-Base Spectrochemical Standards. NBS Misc. Publ. 260-7 (October 1965). COM74-11067**
- Richmond, M.S., Standard Reference Materials: Analysis of Uranium Concentrates at the National Bureau of Standards. NBS Misc. Publ. 260-8 (December 1965). COM74-11066**
- Anspach, S. C., Cavallo, L. M. Garfinkel, S. B. Hutchinson, J. M. R., and Smith, C. N., Standard Reference Materials: Half Lives of Materials Used in the Preparation of Standard Reference Materials of Nineteen Radioactive Nuclides Issued by the National Bureau of Standards NBS Misc. Publ. 260-9 (November 1965). COM74-11065**
- Yakowitz, H., Vieth, D. L., Heinrich, K. F. J., and Michaelis, R. E., Standard Reference Materials: Homogeneity Characterization of NBS Spectrometric Standards II: Cartridge Brass and Low-Alloy Steel, NBS Misc. Publ. 260-10 (December 1965). COM74-11064**
- Napolitano, A., and Hawkins, E. G., Standard Reference Materials: Viscosity of Standard Lead-Silica Glass, NBS Misc. Publ. 260-11 (November 1966). NBS Misc. Publ. 260-11**
- Yakowitz, H., Vieth, D. L., and Michaelis, R. E., Standard Reference Materials: Homogeneity Characterization of NBS Spectrometric Standards III: White Cast Iron and Stainless Steel Powder Compact, NBS Misc. Publ. 260-12 (September 1966). NBS Misc. Publ. 260-12**
- Spijkerman, J. L., Snediker, D. K., Ruegg, F. C., and DeVoe, J. R., Standard Reference Materials: Mossbauer Spectroscopy Standard for the Chemical Shift of Iron Compounds, NBS Misc. Publ. 260-13 (July 1967). NBS Misc. Publ. 260-13**
- Menis, O., and Sterling, J. T., Standard Reference Materials: Determination of Oxygen in Ferrous Materials - SRM 1090, 1091, and 1092, NBS Misc. Publ. 260-14 (September 1966). NBS Misc. Publ. 260-14**
- Passaglia, E., and Shouse, P. J., Standard Reference Materials: Recommended Method of Use of Standard Light-Sensitive Paper for Calibrating Carbon Arcs Used in Testing Textiles for Colorfastness to Light, NBS Misc. Publ. 260-15 (June 1967). (Replaced by NBS Spec. Publ. 260-41.)
- Yakowitz, H., Michaelis, R. E., and Vieth, D. L., Standard Reference Materials: Homogeneity Characterization of NBS Spectrometric Standards IV: Preparation and Microprobe Characterization of W-20% MO Alloy Fabricated by Powder Metallurgical Methods, NBS Spec. Publ. 260-16 (January 1969). COM74-11062**
- Catanzaro, E. J., Champion, C. E., Garner, E. L., Marinenko, G., Sappenfield, K. M., and Shields, W. R., Standard Reference Materials: Boric Acid; Isotopic and Assay Standard Reference Materials, NBS Spec. Publ. 260-17 (February 1970). Out of Print.
- Geller, S. B., Mantek, P.A., and Cleveland, N. G., Standard Reference Materials: Calibration of NBS Secondary Standard Magnetic Tape (Computer Amplitude Reference) Using the Reference Tape Amplitude Measurement "Process A," NBS Spec. Publ. 260-18 (November 1969). (See NBS Spec. Publ. 260-29.)
- Paule, R. C., and Mandel, J., Standard Reference Materials: Analysis of Interlaboratory Measurements on the Vapor Pressure of Gold (Certification of Standard Reference Material 745). NBS Spec. Publ. 260-19 (January 1970). PB190071**
- Paule, R. C., and Mandel, J., Standard Reference Materials: Analysis of Interlaboratory Measurements on the Vapor Pressures of Cadmium and Silver, NBS Spec. Publ. 260-21 (January 1971). COM74-11359**
- Yakowitz, H., Fiori, C. E., and Michaelis, R. E., Standard Reference Materials: Homogeneity Characterization of Fe-3 Si Alloy, NBS Spec. Publ. 260-22 (February 1971). COM74-11357**
- Napolitano, A., and Hawkins, E. G., Standard Reference Materials: Viscosity of a Standard Borosilicate Glass, NBS Spec. Publ. 260-23 (December 1970). COM71-00157**
- Sappenfield, K. M., Marinenko, G., and Hague, J. L., Standard Reference Materials: Comparison of Redox Standards, NBS Spec. Publ. 260-24 (January 1972). COM72-50058**

- Hicho, G. E., Yakowitz, H., Rasberry, S. D., and Michaelis, R. E., Standard Reference Materials: A Standard Reference Material Containing Nominally Four Percent Austenite, NBS Spec. Publ. 260-25 (February 1971). COM74-11356**
- Martin, J. F., Standard Reference Materials: National Bureau of Standards-US Steel Corporation Joint Program for Determining Oxygen and Nitrogen in Steel, NBS Spec. Publ. 260-26 (February 1971). 85 cents* PB 81176620
- Garner, E. L., Machlan, L. A., and Shields, W. R., Standard Reference Materials: Uranium Isotopic Standard Reference Materials, NBS Spec. Publ. 260-27 (April 1971). COM74-11358**
- Heinrich, K. F. J., Myklebust, R. L., Rasberry, S. D., and Michaelis, R. E., Standard Reference Materials: Preparation and Evaluation of SRM's 481 and 482 Gold-Silver and Gold-Copper Alloys for Microanalysis, NBS Spec. Publ. 260-28 (August 1971). COM71-50365**
- Geller, S. B., Standard Reference Materials: Calibration of NBS Secondary Standard Magnetic Tape (Computer Amplitude Reference) Using the Reference Tape Amplitude Measurement "Process A-Model 2," NBS Spec. Publ. 260-29 (June 1971). COM71-50282
- Gorozhanina, R. S., Freedman, A. Y., and Shaievitch, A. B. (translated by M. C. Selby), Standard Reference Materials: Standard Samples Issued in the USSR (A Translation from the Russian), NBS Spec. Publ. 260-30 (June 1971). COM71-50283**
- Hust, J. G., and Sparks, L. L., Standard Reference Materials: Thermal Conductivity of Electrolytic Iron SRM 734 from 4 to 300 K, NBS Spec. Publ. 260-31 (November 1971). COM71-50563**
- Mavrodineanu, R., and Lazar, J. W., Standard Reference Materials: Standard Quartz Cuvettes, for High Accuracy Spectrophotometry, NBS Spec. Publ. 260-32 (December 1973). 55 cents* SN003-003-01213-1
- Wagner, H. L., Standard Reference Materials: Comparison of Original and Supplemental SRM 705, Narrow Molecular Weight Distribution Polystyrene, NBS Spec. Publ. 260-33 (May 1972). COM72-50526**
- Sparks, L. L., and Hust, J. G., Standard Reference Materials: Thermoelectric Voltage, NBS Spec. Publ. 260-34, (April 1972). COM72-50371**
- Sparks, L. L., and Hust, J. G., Standard Reference Materials: Thermal Conductivity of Austenitic Stainless Steel, SRM 735 from 5 to 280 K, NBS Spec. Publ. 260-35 (April 1972). 35 cents* COM72-50368**
- Cali, J. P., Mandel, J., Moore, L. J., and Young, D. S., Standard Reference Materials: A Referee Method for the Determination of Calcium in Serum, NBS SRM 915, NBS Spec. Publ. 260-36 (May 1972). COM72-50527**
- Shultz, J. I. Bell., R. K. Rains, T. C., and Menis, O., Standard Reference Materials: Methods of Analysis of NBS Clay Standards, NBS Spec. Publ. 260-37 (June 1972). COM72-50692**
- Richmond, J. C., and Hsia, J. J., Standard Reference Materials: Preparation and Calibration of Standards of Spectral Specular Reflectance, NBS Spec. Publ. 260-38 (May 1972). COM72-50528**
- Clark, A. F., Denson, V.A., Hust, J. G., and Powell, R. L., Standard Reference Materials: The Eddy Current Decay Method for Resistivity Characterization of High-Purity Metals, NBS Spec. Publ. 260-39 (May 1972). COM72-50529**
- McAdie, H. G., Garn, P.D., and Menis, O., Standard Reference Materials: Selection of Thermal Analysis Temperature Standards Through a Cooperative Study (SRM 758, 759, 760), NBS Spec. Publ. 260-40 (August 1972.) COM72-50776**
- Wood, L. A., and Shouse, P. J., Standard Reference Materials: Use of Standard Light-Sensitive Paper for Calibrating Carbon Arcs Used in Testing Textiles for Colorfastness to Light, NBS Spec. Publ. 260-41 (August 1972). COM72-50775**
- Wagner, H. L. and Verdier, P. H., eds., Standard Reference Materials: The Characterization of Linear Polyethylene, SRM 1475, NBS Spec. Publ. 260-42 (September 1972). COM72-50944**
- Yakowitz, H., Ruff, A. W., and Michaelis, R. E., Standard Reference Materials: Preparation and Homogeneity Characterization of an Austenitic Iron-Chromium-Nickel Alloy, NBS Spec. Publ. 260-43 (November 1972). COM73-50760**
- Schooley, J. F., Soulen, R. J., Jr., and Evans, G. A., Jr., Standard Reference Materials: Preparation and Use of Superconductive Fixed Point Devices, SRM 767, NBS Spec. Publ. 260-44 (December 1972). COM73-50037**
- Greifer, B., Maienthal, E. J., Rains, T. C., and Rasberry, S. D., Standard Reference Materials: Powdered Lead-Based Paint, SRM 1579, NBS Spec. Publ. 260-45 (March 1973). COM73-50226**
- Hust, J. G., and Giarratano, P. J., Standard Reference Materials: Thermal Conductivity and Electrical Resistivity Standard Reference Materials: Austenitic Stainless Steel, SRM's 735 and 798, from 4 to 1200 K, NBS Spec. Publ. 260-46 (March 1975). SN003-003-01278-5*
- Hust, J. G., Standard Reference Materials: Electrical Resistivity of Electrolytic Iron, SRM 797, and Austenitic Stainless Steel, SRM 798, from 5 to 280 K, NBS Spec. Publ. 260-47 (February 1974). COM74-50176**
- Mangum, B. W., and Wise, J. A., Standard Reference Materials: Description and Use of Precision Thermometers for the Clinical Laboratory, SRM 933 and SRM 934, NBS Spec. Publ. 260-48 (May 1974). 60 cents* SN003-003-01278-5
- Carpenter, B. S., and Reimer, G. M., Standard Reference Materials: Calibrated Glass Standards for Fission Track Use, NBS Spec. Publ. 260-49 (November 1974). COM74-51185

- Hust, J. G., and Giarratano, P. J., Standard Reference Materials: Thermal Conductivity and Electrical Resistivity Standard Reference Materials: Electrolytic Iron, SRM's 734 and 797 from 4 to 1000 K, NBS Spec. Publ. 260-50 (June 1975). \$1.00* SN003-003-01425-7
- Mavrodineanu, R., and Baldwin, J. R., Standard Reference Materials: Glass Filters As a Standard Reference Material for Spectrophotometry; Selection; Preparation; Certification; Use-SRM 930, NBS Spec. Publ. 260-51 (November 1975). \$1.90* SN003-003-01481-8
- Hust, J. G., and Giarratano, P. J., Standard Reference Materials: Thermal Conductivity and Electrical Resistivity Standard Reference Materials 730 and 799, from 4 to 3000 K, NBS Spec. Publ. 260-52 (September 1975). \$1.05* SN003-003-01464-8
- Durst, R. A., Standard Reference Materials: Standardization of pH Measurements, NBS Spec. Publ. 260-53 (December 1975, Revised). \$1.05 SN003-003-01551-2
- Burke, R. W., and Mavrodineanu, R., Standard Reference Materials: Certification and Use of Acidic Potassium Dichromate Solutions as an Ultraviolet Absorbance Standard, NBS Spec. Publ. 260-54 (August 1977). \$3.00* SN003-003-01828-7
- Ditmars, D. A., Cezairliyan, A., Ishihara, S., and Douglas, T. B., Standard Reference Materials: Enthalpy and Heat Capacity; Molybdenum SRM 781, from 273 to 2800 K, NBS Spec. Publ. 260-55 (September 1977). \$2.20* SN003-003-01836-8
- Powell, R. L., Sparks, L. L., and Hust, J. G., Standard Reference Materials: Standard Thermocouple Materials, Pt.67: SRM 1967, NBS Spec. Publ. 260-56 (February 1978). \$2.20* SN003-003-018864
- Cali, J. P. and Plebanski, T., Guide to United States Reference Materials, NBS Spec. Publ. 260-57 (February 1978). \$2.20* PB 277173
- Barnes, J. D., and Martin, G. M., Standard Reference Materials: Polyester Film for Oxygen Gas Transmission Measurements SRM 1470, NBS Spec. Publ. 260-58 (June 1979) \$2.00* SN003-003-02077
- Chang, T., and Kahn, A. H. Standard Reference Materials: Electron Paramagnetic Resonance Intensity Standard; SRM 2601, NBS Spec. Publ. 260-59 (August 1978) \$2.30* SN003-003-01975-5
- Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., and Moody, J. R., Standard Reference Materials: A Reference Method for the Determination of Sodium in Serum, NBS Spec. Publ. 260-60 (August 1978). \$3.00* SN003-003-01978-0
- Verdier, P. H., and Wagner, H. L., Standard Reference Materials: The Characterization of Linear Polyethylene (SRM 1482, 1483, 1484), NBS Spec. Publ. 260-61 (December 1978). \$1.70* SN003-003-02006-1
- Soulen, R. J., and Dove, R. B., Standard Reference Materials: Temperature Reference Standard for Use Below 0.5 K (SRM 768). NBS Spec. Publ. 260-62 (April 1979). \$2.30* SN003-003-02047-8
- Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., Machlan, L. A., and Gramlich, J. W., Standard Reference Materials: A Reference Method for the Determination of Potassium in Serum. NBS Spec. Publ. 260-63 (May 1979). \$3.75* SN003-003-02068
- Velapoldi, R. A., and Mielenz, K. D., Standard Reference Materials: A Fluorescence Standard Reference Material Quinine Sulfate Dihydrate (SRM 936), NBS Spec. Publ. 260-64 (January 1980). \$4.25* SN003-003-02148-2
- Marinenko, R. B., Heinrich, K. F. J., and Ruegg, F. C., Standard Reference Materials: Micro-Homogeneity Studies of NBS Standard Reference Materials, NBS Research Materials, and Other Related Samples. NBS Spec. Publ. 260-65 (September 1979). \$3.50* SN003-003-02114-1
- Venable, W. H., Jr., and Eckerle, K. L., Standard Reference Materials: Didymium Glass Filters for Calibrating the Wavelength Scale of Spectrophotometers (SRM 2009, 2010, 2013). NBS Spec. Publ. 260-66 (October 1979). \$3.50* SN003-003-02127-0
- Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., Murphy, T. J., and Gramlich, J. W., Standard Reference Materials: A Reference Method for the Determination of Chloride in Serum, NBS Spec. Publ. 260-67 (November 1979). \$3.75* SN003-003-02136-9
- Mavrodineanu, R. and Baldwin, J. R., Standard Reference Materials: Metal-On-Quartz Filters as a Standard Reference Material for Spectrophotometry-SRM 2031, NBS Spec. Publ. 260-68 (April 1980). \$4.25* SN003-003-02167-9
- Velapoldi, R. A., Paule, R. C., Schaffer, R., Mandel, J., Machlan, L. A., Garner, E. L., and Rains, T. C., Standard Reference Materials: A Reference Method for the Determination of Lithium in Serum, NBS Spec. Publ. 260-69 (July 1980). \$4.25* SN003-003-02214-4
- Marinenko, R. B., Biancanello, F., Boyer, P. A., Ruff, A. W., DeRobertis, L., Standard Reference Materials: Preparation and Characterization of an Iron-Chromium-Nickel Alloy for Micro-analysis, NBS Spec. Publ. 260-70 (May 1981). \$2.50* SN003-003-02328-1
- Seward, R. W., and Mavrodineanu, R., Standard Reference Materials: Summary of the Clinical Laboratory Standards Issued by the National Bureau of Standards, NBS Spec. Publ. 260-71 (November 1981). \$6.50* SN003-003-02381-7
- Reeder, D.J., Coxon, B., Enagonio, D., Christensen, R. G., Schaffer, R., Howell, B. F., Paule, R. C., Mandel, J., Standard Reference Materials: SRM 900, Antiepilepsy Drug Level Assay Standard, NBS Spec. Publ. 260-72 (June 1981). \$4.25* SN003-003-02329-9

- Interrante, C. G., and Hicho, G. E., Standard Reference Materials: A Standard Reference Material Containing Nominally Fifteen Percent Austenite (SRM 486), NBS Spec. Publ. 260-73 (January 1982). \$2.75* SN003-003-02386-8
- Marinenko, R. B., Standard Reference Materials: Preparation and Characterization of K-411 and K-414 Mineral Glasses for Microanalysis: SRM 470. NBS Spec. Publ. 260-74 (April 1982). \$3.50 SN003-003-023-95-7
- Weidner, V. R., and Hsia, J. J., Standard Reference Materials: Preparation and Calibration of First Surface Aluminum Mirror Specular Reflectance Standards (SRM 2003a), NBS Spec. Publ. 260-75 (May 1982). \$3.75 SN003-003-023-99-0
- Hicho, G. E. and Eaton, E. E., Standard Reference Materials: A Standard Reference Material Containing Nominally Five Percent Austenite (SRM 485a), NBS Spec. Publ. 260-76 (August 1982). \$3.50 SN003-003-024-33-3
- Furukawa, G. T., Riddle, J. L., Bigge, W. G., and Pfeiffer, E. R., Standard Reference Materials: Application of Some Metal SRM's as Thermometric Fixed Points, NBS Spec. Publ. 260-77 (August 1982). \$6.00 SN003-003-024-34-1
- Hicho, G. E. and Eaton, E. E., Standard Reference Materials: Standard Reference Material Containing Nominally Thirty Percent Austenite (SRM 487), NBS Spec. Publ. 260-78 (September 1982). \$3.75* SN003-003-024-35-0
- Richmond, J. C., Hsia, J. J., Weidner, V. R., and Wilmering, D. B., Standard Reference Materials: Second Surface Mirror Standards of Specular Spectral Reflectance (SRM's 2023, 2024, 2025), NBS Spec. Publ. 260-79 (October 1982). \$4.50* SN003-003-024-47-3
- Schaffer, R., Mandel, J., Sun, T., Cohen, A., and Hertz, H. S., Standard Reference Materials: Evaluation by an ID/MS Method of the AACC Reference Method for Serum Glucose, NBS Spec. Publ. 260-80 (October 1982). \$4.75* SN003-003-024-43-1
- Burke, R. W., and Mavrodineanu, R. (NBS retired), Standard Reference Materials: Accuracy in Analytical Spectrophotometry, NBS Spec. Publ. 260-81 (April 1983). \$6.00* SN003-003-024-8
- Weidner, V. R., Standard Reference Materials: White Opal Glass Diffuse Spectral Reflectance Standards for the Visible Spectrum (SRM's 2015 and 2016). NBS Spec. Publ. 260-82 (April 1983). \$3.75* SN003-003-024-89-9**
- Bowers, G. N., Jr., Alvarez, R., Cali, J. P. (NBS retired), Eberhardt, K. R., Reeder, D. J., Schaffer, R., Uriano, G. A., Standard Reference Materials: The Measurement of the Catalytic (Activity) Concentration of Seven Enzymes in NBS Human Serum SRM 909, NBS Spec. Publ. 260-83 (June 1983). \$4.50* SN003-003-024-99-6
- Gills, T. E., Seward, R. W., Collins, R. J., and Webster, W. C., Standard Reference Materials: Sampling, Materials Handling, Processing, and Packaging of NBS Sulfur in Coal Standard Reference Materials, 2682, 2683, 2684, and 2685, NBS Spec. Publ. 260-84 (August 1983). \$4.50* SN003-003-025-20-8
- Swyt, D. A., Standard Reference Materials: A Look at Techniques for the Dimensional Calibration of Standard Microscopic Particles, NBS Spec. Publ. 260-85 (September 1983). \$5.50* SN003-003-025-21-6
- Hicho, G. E. and Eaton, E. E., Standard Reference Materials: A Standard Reference Material Containing Two and One-Half Percent Austenite, SRM 488, NBS Spec. Publ. 260-86 (December 1983). \$1.75* SN003-003-025-41-1
- Mangum, B. W., Standard Reference Materials: SRM 1969: Rubidium Triple-Point - A Temperature Reference Standard Near 39.30 °C, NBS Spec. Publ. 260-87 (December 1983). \$2.25* SN003-003-025-44-5
- Gladney, E. S., Burns, C. E., Perrin, D. R., Roelandts, I., and Gills, T. E., Standard Reference Materials: 1982 Compilation of Elemental Concentration Data for NBS Biological, Geological, and Environmental Standard Reference Materials. Spec. Publ. 260-88 (March 1984). \$7.00* SN003-003-02565-8
- Hust, J. G., Standard Reference Materials: A Fine-Grained, Isotropic Graphite for Use as NBS Thermophysical Property RM's from 5 to 2500 K, NBS Spec. Publ. 260-89 (September 1984). \$4.50* SN003-003-02608-5
- Hust, J. G., and Lankford, A. B., Standard Reference Materials: Update of Thermal Conductivity and Electrical Resistivity of Electrolytic Iron, Tungsten, and Stainless Steel, NBS Spec. Publ. 260-90 (September 1984). \$3.00* SN003-003-02609-3
- Goodrich, L. F., Vecchia, D. F., Pittman, E. S., Ekin, J. W., and Clark, A. F., Standard Reference Materials: Critical Current Measurements on an NbTi Superconducting Wire Standard Reference Material, NBS Spec. Publ. 260-91 (September 1984). \$2.75* SN003-003-02614-0
- Carpenter, B. S., Standard Reference Materials: Calibrated Glass Standards for Fission Track Use (Supplement to NBS Spec. 260-49). NBS Spec. Publ. 260-92 (September 1984). \$1.50* SN003-003-02610-7
- Ehrstein, J., Preparation and Certification of Standard Reference Materials for Calibration of Spreading Resistance Probes, NBS Spec. Publ. 260-93 (January 1985). \$2.25* SN003-003-02633-6
- Gills, T. E., Koch, W. F., Stolz, J. W., Kelly, W. R., Paulsen, P. J., Colbert, J. C., Kirklind, D. R., Pei, P.T.S., Weeks, S., Lindstrom, R. M., Fleming, R. F., Greenberg, R. R., and Paule, R. C., Methods and Procedures Used at the National Bureau of Standards to Certify Sulfur in Coal SRM's for Sulfur Content, Calorific Value, Ash Content, NBS Spec. Publ. 260-94 (December 1984). \$2.25* SN003-003-02629-8
- Mulholland, G. W., Hartman, A. W., Hembree, G. G., Marx, E., and Lettieri, T. R., Standard Reference Materials: Development of a 1 µm Diameter Particle Size Standard, SRM 1690, NBS Spec. Publ. 260-95 (May 1985). SN003-003-02665-4

Carpenter, B. S., Gramlich, J. W., Greenberg, R. R., and Machlan, L. A., Standard Reference Materials: Uranium-235 Isotopic Abundance Standard Reference Materials for Gamma Spectrometry Measurements, NBS Spec. Publ. 260-96 (In Preparation).

Mavrodineanu, R. and Gills, T. E., Standard Reference Materials: Summary of the Coal, Ore, Mineral, Rock, and Refractory Standards Issued by the National Bureau of Standards, NBS Spec. Publ. 260-97 (September 1985).

Hust, J. G., Standard Reference Materials: Glass Fiberboard SRM for Thermal Resistance, NBS Spec. Publ. 260-98 (August 1985). SN003-003-02674-3

Callanan, J. E., Sullivan, S. A., and Vecchia, D. F., Standard Reference Materials: Feasibility Study for the Development of Standards Using Differential Scanning Calorimetry, NBS Spec. Publ. 260-99 (August 1985). SN003-003-02675-1

Taylor, J. K., Standard Reference Materials: Handbook for SRM Users, NBS Spec. Publ. 260-100 (In Press).

Mangum, B. W., Standard Reference Materials: SRM 1970, Succinonitrile Triple-Point Standard: A Temperature Reference Standard Near 58.08 °C, NBS Spec. Publ. 260-101 (In Preparation).

Weidner, V. R., Mavrodineanu, R., Mielenz, K. D., Velapoldi, R. A., Eckerle, K. L., and Adams, B., Standard Reference Materials: Holmium Oxide Solution Wavelength Standard from 240-650 nm, SRM 2034, NBS Spec. Publ. 260-102 (In Press).

Hust, J. G., Standard Reference Materials: Glass Fiberblanket SRM for Thermal Resistance, NBS Spec. Publ. 260-103 (September 1985). SN003-003-02687-5

Mavrodineanu, R. and Alvarez, R., Standard Reference Materials: Summary of the Biological and Botanical Standards Issued by the National Bureau of Standards, NBS Spec. Publ. 260-104 (November 1985).

Mavrodineanu, R. and Rasberry, S.D., Standard Reference Materials: Summary of the Environmental Research, Analysis, and Control Standards Issued by the National Bureau of Standards, NBS Spec. Publ. 260-105.

* Send order with remittance to Superintendent of Documents, US Government Printing Office Washington, DC 20402. Remittance from foreign countries should include an additional one-fourth of the purchase price for postage.

** May be ordered from: National Technical Information Services (NTIS). Springfield Virginia 22161.

TABLE OF CONTENTS

	<u>Page</u>
Preface	iii
Abstract	x
Introduction	1
Table 1. Categories of Standard Reference Materials available from the National Bureau of Standards	2
Environmental Research, Analysis, and Control Standards	4
Table 2. Summary of the Environmental Research, Analysis, and Control Standards	6
Table 3. Comparative List of Chemical Elements for the SRM's Described in Table 2	12
Appendix I. Alphabetical Index by Standard Reference Material Name	15
Appendix II. Certificates for the Environmental Research, Analysis, and Control Standards (listed in numerical order)	28
Appendix III. Guide for Requesting Development of Standard Reference Materials	81
Appendix IV. Guide to Ordering Standard Reference Materials	82

Abstract

This publication is a summary of the environmental research, analysis, and control standards issued by NBS as Standard Reference Materials (SRM's). The material, composition, certification, use, and remarks concerning each of the SRM's described are presented in tabular form. Copies of the certificates of these SRM's are contained in the appendix for more detailed information.

Key Words: chemical composition; environmental standards; Standard Reference Materials.

Introduction

Since its inauguration in 1901, the National Bureau of Standards (NBS) has issued nearly 2000 different Standard Reference Materials (SRM's). Many of these have been renewed several times, many have been replaced or discontinued as technology changed. Today, over 900 SRM's are available, together with a large number of scientific publications related to the fundamental and applied characteristics of these materials. Each material is certified for chemical composition, chemical properties, or its physical or mechanical characteristics. Each SRM is provided with a Certificate or Certificate of Analysis that contains the essential data concerning its properties or characteristics. The SRM's currently available cover a wide range of chemical, physical, and mechanical properties, and a corresponding wide range of measurement interests in practically all aspects of fundamental and applied science. These SRM's constitute a unique and invaluable means of transferring to the user accurate data obtained at NBS, and provide essential tools that can be used to improve accuracy in practically all areas where measurements are performed.

In addition to SRM's, the National Bureau of Standards issues a variety of Research Materials (RM's) having various properties described in individual "Reports of Investigation." They are intended primarily to further the scientific or technical research on that particular material. Other materials, called Special Reference Materials (GM's), are also available from NBS. These are materials produced and certified by other Government agencies, standard organizations, or other nonprofit organizations, that are considered useful to the public and for which no alternate method of national distribution exists.

The various categories of materials available from NBS are given in table 1. This table lists these materials according to their chemical composition, physical properties, or engineering characteristics. A more detailed alphabetic enumeration of these materials is given in appendix I. Table 1 and appendix I were taken from NBS Special Publication 260, NBS Standard Reference Materials Catalog, 1984-85 Edition¹. This publication lists every material available from the NBS Office of Standard Reference Materials.

Further information on the reference materials available from NBS may be obtained from the Office of Standard Reference Materials, National Bureau of Standards, Gaithersburg, MD 20899. Information on other NBS services may be obtained from the Technical Information and Publications Division, National Bureau of Standards, Gaithersburg, MD 20899.

In addition to reference materials, NBS provides many additional services. These include: Measurement Assurance Programs, Calibration and Related Measurement Services, Proficiency Sample Programs, a National Voluntary Laboratory Accreditation Program, Standards Information Services, Standard Reference Data, and Technical Information and Publications.

¹For sale by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402, under Stock No. 003-003-02558-5 (Price \$5.50, add 25 percent for foreign orders.)

Table 1. Categories of Standard Reference Materials available from the National Bureau of Standards.

CERTIFIED CHEMICAL COMPOSITION STANDARDS

Steels (chip form)	Gases in Metals
Plain carbon	High-Purity Metals
Low alloy	Electron Probe Microanalytical Standards
High alloy	Primary, Working, and Secondary
Stainless	Standard Chemicals
Tool	Microchemical Standards
Steels (granular form)	Clinical Laboratory Standards
Steels (solid form)	Biological Standards
Ingot iron and low alloy	Environmental Standards
Special ingot irons and low alloy	Analyzed gases
Stainless	Analyzed liquids and solids
Specialty	Permeation tubes
High-temperature alloys	Industrial Hygiene Standards
Tool	Spectrochemical Standards
Steelmaking Alloys	Hydrocarbon Blends
Cast irons (chip form)	Metallo-Organic Compounds
Cast Steels, White Cast Irons, Ductile	Fertilizers
Irons, and Blast Furnace Irons	Ores
(solid form)	Minerals, Refractories, Glasses, and
Nonferrous Alloys (chip form)	Carbides
Aluminum "Benchmarks"	Cement
Cobalt	Trace Element Standards
Copper	Nuclear Materials
Copper "Benchmarks"	Special nuclear materials
Lead	Plutonium assay
Magnesium	Plutonium isotopic
Nickel	Uranium assay
Nickel Superalloy, Trace Elements	Uranium isotopic
Nickel oxide	Neutron density standards
Selenium	Fission track glass standards
Tin	Isotopic Reference Standards
Titanium	
Zinc	
Zirconium	
Nonferrous Alloys (solid form)	
Aluminum "Benchmarks"	
Copper	
Copper "Benchmarks"	
Lead	
Nickel	
Titanium	
Zinc	
Zirconium	

Table 1. continued.

CERTIFIED PHYSICAL PROPERTY STANDARDS

Ion Activity Standards

pH standards
pD standards
Ion selective electrodes

Optical Standards

Spectrophotometric
Thermal emittance
Refractive index

Mechanical and Metrology Standards

Magnification
Coating thickness
Glass
Elasticity
Density
Polymer
Rheology

Radioactivity Standards

Alpha-particle standards
Beta-particle and gamma-ray gas standards
Alpha-particle, beta-particle, gamma-ray, and electron-capture solution standards
Contemporary standard for carbon-14 dating laboratories
Environmental standards
Low energy photon sources
Gamma-ray "point-source" standards
Radium gamma-ray solution standards
Radium solution standards for random analysis
Radioactivity standard reference materials currently not in stock

Heat Standards

Superconductive thermometric fixed point devices
Freezing Points
Defining fixed points
Determined reference points

Metallurgical

Melting points
Calorimetric

Mossbauer

Combustion
Solution
Heat source
Enthalpy and heat capacity

X-ray Diffraction

Vapor pressure
Thermal expansion
Thermocouple materials
Thermal resistance

Gas Transmission

Permittivity

Reference Fuels

Resistivity

Magnetic Standards

Magnetic susceptibility
Magnetic moment
Paramagnetic resonance

ENGINEERING TYPE STANDARDS

Standard Rubber and Rubber-Compounding Materials

X-ray and Photographic Standards

Reference Magnetic Tapes

Surface Flammability Standards

Lubricant Standards

Semiconductor Production Standards

Sizing Standards

Water Vapor Permeance

Glass spheres for particle size
Turbidimetric and fineness (cement)

Tape Adhesion Testing Standards

Color Standards

RESEARCH MATERIALS

SPECIAL REFERENCE MATERIALS

Environmental Research, Analysis, and Control Standards

This work is the fourth in a series of NBS Special Publications dedicated to the description of the Standard Reference Materials (SRM's) issued by the National Bureau of Standards (NBS).

The first volume, NBS SP 260-71, "Summary of the Clinical Laboratory Standards Issued by the National Bureau of Standards" (176 pp., November 1981), describes in a tabular form 41 SRM's available in that field and includes copies of the corresponding Certificates of Analysis for further information.

The second volume, NBS SP 260-97, "Summary of the Coal, Ore, Mineral, Rock, and Refractory Standards Issued by the National Bureau of Standards" (134 pp., September 1985), describes in the same manner 39 SRM's available in that field.

The third volume, NBS SP 260-104, "Summary of the Biological and Botanical Standards Issued by the National Bureau of Standards" (68 pp., October 1985), presents in a similar manner the essential data concerning 9 SRM's issued in that field.

The present volume, NBS SP 260-105, "Summary of the Environmental Research, Analysis, and Control Standards Issued by the National Bureau of Standards" (97 pp., March 1986), describes in a tabular form the characteristic properties of 22 SRM's issued in the field of environmental analytical instrumentation and methodology. Copies of the Certificates of Analysis are reproduced in this work also, for further information.

Tables 2 and 3 contain the essential information concerning the material composition, the certification parameters, and use. Under "Remarks," additional data such as storage conditions and stability is provided.¹ All the data and information contained in these tables were extracted from the Certificates or Certificates of Analysis issued for the SRM's included in the table. An examination of these tables gives the reader a general view of these SRM's. For more detailed information, the individual Certificates reproduced in appendix II should be consulted as well as the references cited in each Certificate. The SRM's listed in the tables include all of the environmental standards that were issued or were in preparation by the end of 1984. These SRM's are the result of the concerted efforts of a number of scientists from the NBS National Measurement Laboratory. Each Certificate lists the individuals who contributed to development of the SRM.

Appendix III provides a guide to the reader to assist in requesting NBS to develop new SRM's. A final appendix, appendix IV, is a guide to ordering SRM's.

In addition to the SRM's and their Certificates, NBS issues a series of Special Publications, called the "260 Series," that relate directly to Standard Reference Materials as stated in the preface. The list of available publications in the "260 Series" is given at the beginning of this publication.²

¹NOTE: The use of proprietary designations in table 2 is for information only, and should not be construed as an endorsement of the product by either the Department of Commerce or the National Bureau of Standards.

²For complete bibliographic reference and ordering information, see "Other NBS Publications in This Series," pp. iv.

**TABLE 2. SUMMARY OF THE ENVIRONMENTAL
RESEARCH, ANALYSIS, AND
CONTROL STANDARDS.**

SRM	Material	COMPOSITION
1579 Powdered Lead Based Paint	Collected by the Philadelphia Dept. Public Health, and sieved to a powder passing through a 325 mesh sieve.	Pb: 11.87 \pm 0.04 % by wt., based on samples \geq 0.1 g of as-received material.
1580 Organics in Shale Oil	The shale oil came from the Laramie Energy Technology Center, Laramie, Wyoming, and was collected from the Mahogany Zone of the Colorado Green River Formation.	Fluoranthene: 54; pyrene: 104; benzo[a]pyrene: 21; benzo[e]pyrene: 18; perylene: 3.4; phenol: 407; o-cresol: 385; 2,6-dimethylphenol: 175; benzol[f]quinoline: 16, μ g/g.
1620a Sulfur in Residual Fuel Oil	The material is a commercial "No. 5 Heavy" residual fuel oil as defined by the American Society for Testing and Materials.	S: 4.504 \pm 0.010 wt. %.
1621b Sulfur in Residual Fuel Oil	The material is a commercial "No. 6" residual fuel oil as defined by the American Society for Testing and Materials.	S: 0.950 \pm 0.005 wt. %.
1622b Sulfur in Residual Fuel Oil	Same as for SRM 1621b.	S: 1.982 \pm 0.018 wt. %.
1623a Sulfur in Residual Fuel Oil	Same as for SRM 1620a.	S: 0.240 \pm 0.003 wt. %.
1624a Sulfur in Distillate (Diesel) Fuel Oil	The material is a commercial "No. 2-D" distillate fuel oil as defined by the American Society for Testing and Materials.	S: 0.141 \pm 0.002 wt. %.
1630 Trace Mercury in Coal	Commercial coal crushed to 210-500 micrometer particle size.	Hg: 0.13 μ g/g \pm 1 in the last significant figure.

Certification	Use	Remarks
By x-ray fluorescence, atomic absorption spectrometry, polarography.	In the calibration of apparatus and methods for determining Pb in paint removed from the interior surfaces of old housing.	
By gas chromatography, gas chromatography/mass spectrometry, and high performance liquid chromatography.	For evaluating the reliability of analytical methods used for trace organic compounds in oil materials.	Seven additional organic compounds are included for information only and are not certified.
By gravimetry, ion chromatography, and x-ray fluorescence.	As analytical standard in the determination of total S in fuel oils and similar materials.	Four additional physical properties are indicated but are not certified; similarly, 16 trace elements are mentioned semi-quantitatively only.
As for SRM 1620a. Is valid for 3 years from date of purchase.	As for SRM 1620a.	
As for SRM 1621b.	As for SRM 1621b.	
As for SRM 1620a.	As for SRM 1620a.	
By gravimetry and ion chromatography. Is valid for 3 years from date of purchase.	As for SRM 1620a.	
By neutron activation and flameless atomic absorption spectrometry (0.14 µg/g). The neutron activation procedure is described in the certificate.	As analytical standard for the determination of trace mercury in coal.	Selenium is also given for information, but is not certified (2.1 µg/g).

Table 2. Continued.

SRM	Material	Composition
1632b Trace Elements in Coal (Bituminous)	Obtained from an underground mine that recovers coal from the Pittsburgh seam, crushed and sieved through a -60 mesh at the Humphrey No. 7 mine and coal preparation plant of the Consolidation Coal Co., Christopher Coal Co. Div., Osage, W. Va.	Material should be vacuum dried at ambient temperature for 24 hours prior to use. Values based on a minimum sample size of 250 mg. C (total) 78.11; H 5.07; N 1.56; S 1.89; Volatile matter 35.4; Al 0.855; Ca .204; Fe .759; Mg .0383; K .0748; Na .0515; Ti .0454; values % by wt. As 3.72; Ba 67.5; Cd 0.0573; Co 2.29; Cu 6.28; Pb 3.67; Mn 12.4; Ni 6.10; Rb 5.05; Se 1.29; Th 1.342; U 0.436; Zn 11.89; values µg/g. Not certified: 17 additional constituents (see certificate).
1633a Trace Elements in Coal Fly Ash	Obtained from a coal fired power plant using Pennsylvania and West Virginia Coal. The ash was sieved through a No. 170 sieve.	Determined on 250 mg or more sample dried to constant weight. Al 14.3; Ca 1.11; Fe 9.40; K 1.88; Mg 0.455; Na 0.17; Si 22.8; wt. %. Sb 6.8; As 145; Cd 1.0; Cr 196; Cu 118; Hg 0.16; Ni 127; Pb 72.4; Rb 131; Se 10.3; Sr 830; Th 24.7; Tl 5.7; U 10.2; V 2.97; Zn 220; µg/g. Additional 15 elements determined but not certified.
1634a Trace Elements in Fuel Oil	Commercial No. 6 residual fuel oil as defined by ASTM.	Determined on at least 1 g sample. Pb 2.80; Mn 0.19; Ni 29; Se 0.15; Na 87; V 56; Zn 2.7; µg/g. S 2.85 wt. %. Values for additional 11 elements and 4 physical properties are given but not certified.
1635 Trace Elements in Coal (Subbituminous)	Subbituminous coal from Eagle Mine of the Imperial Coal Co. of Erie, Colorado; sieved through a No. 65 sieve.	Determined on at least 250 mg of dried sample. As 0.42; Cd 0.03; Cr 2.5; Cu 3.6; Pb 1.9; Mn 21.4; Ni 1.74; Se 0.9; Th 0.62; U 0.24; V 5.2; Zn 4.7; µg/g; and Fe 0.239; S 0.33 wt. %. Additional 10 elements determined but not certified.
1636a, 1637a, 1638a Lead in Reference Fuel	Supplied by Phillips Petroleum Co., Bartlesville, Okla. Lead was added as tetraethyl lead motor mix.	Determined on at least 1 g sample. Vial I-11.2; Vial II-18.8; Vial III-25.1; Vial IV-764 µg/g.

Certification	Use	Remarks
Analyses performed in the NBS Center for Analytical Chemistry. Estimated uncertainty 0.006 depending on constituent.	For the calibration of apparatus and the evaluation of techniques employed in the analysis of coal or similar materials.	Should be kept in its original bottle. Should not be exposed to intense sources of radiation, including ultraviolet lamps or sunlight.
Using two to four different analytical procedures for each element as shown in the Certificate, on the dried sample.	For calibration of apparatus and methods used in the analysis of coal ash and similar materials.	The materials should be dried as indicated in the Certificate. Stability (>3 years) not yet established.
Using two to three independent analytical procedures for each element.	For calibration of apparatus and methods used in the analysis of fuel oils and similar materials for trace elements.	Store in tightly sealed bottle. Certificate valid for 3 years from date of purchase.
Using two to three independent analytical procedures for each element. For sample drying, see the Certificate.	For calibration of apparatus and techniques used in trace element analysis of coal and similar materials.	Store in tightly sealed bottle in a cool, dark place. Long term (>1 year) stability not yet established.
By isotope dilution mass spectrometry. SRM 1636a is made from Vials I, II, III, and IV. SRM 1637a is made from Vials I, II, and III. SRM 1638a is made from Vial IV.	For calibration of instruments and techniques used for the analysis of Pb in gasoline.	The vials should be stored at 10-30 °C in darkness, and not reused after first opening.

Table 2. - Continued.

SRM	Material	Composition
1641b Mercury in Water - $\mu\text{g/mL}$	This SRM was prepared at NBS and is delivered in 6 ampoules of 20 mL each.	Hg: $1.52 \pm 0.04 \mu\text{mL}$.
1642b Mercury in Water - ng/mL	Same as SRM 1641b.	Hg: $1.49 \pm 0.06 \text{ng/mL}$.
1643b Trace Elements in Water	Prepared at the U.S. Geological Survey, National Water Quality Laboratory, Arvada, Colo., by using high-purity reagents and sterilization.	Ba 44; Be 19; Cd 20; Cr 18.6; Co 26; Cu 21.9; Fe 99; Pb 23.7; Mn 28; Mo 85; Ni 49; Se 9.7; Ag 9.8; Sr 227; Tl 8; V 45.2; Zn 66 ng/g . Additional 3 elements determined but not certified.
1644 Generator Columns for Polynuclear Aromatic Hydrocarbons	Three 50 cm x 0.6 cm coiled stainless steel tubes each packed with fine quintus quartz coated with 0.5 wt. % of polynuclear aromatic hydrocarbon (PAH).	At 20 °C: anthracene 30.7; benzo(a)anthracene 6.45; benzo(a)pyrene 1.13, all in $\mu\text{g/Kg}$.
1645 River Sediment	Deposit dredged from the bottom of Indiana Harbor Canal, Gary, Ind., freeze-dried, sieved (No. 80 screen) and radiation-sterilized.	Al 2.26; Cr 2.96; Fe 11.3; K 1.26; Mg 0.74; Na 0.54; Zn 0.172, wt. %. Cd 10.2; Cu 109; Co 10.1; Pb 714; Mn 785; Hg 1.1; Ni 45.8; Tl 1.44; Th 1.62; U 1.11; V 23.5; $\mu\text{g/g}$. Additional 13 values determined but not certified.
1646 Estuarine Sediment	Dredged from the Chesapeake Bay, freeze-dried, radiation-sterilized and sieved (No. 100 sieve).	Al 6.25; Ca 0.83; Fe 3.35; Mg 1.09; P 0.054, wt. %. As 11.6; Cd 0.36; Cr 76; Co 10.5; Cu 18; Pb 28.2; Mn 375; Hg 0.063; Ni 32; V 94; Zn 138, $\mu\text{g/g}$.
1648 Urban Particulate Matter	Urban particulate matter was collected in St. Louis, Mo., over a period of 12 months.	Al 3.42; Fe 3.91; K 1.05; Pb 0.655; Na 0.425; Zn 0.476, wt. %. As 115; Cd 75; Cr 403; Cu 609; Ni 82; Se 27; U 5.5; V 140 $\mu\text{g/g}$. Non-certified values are given for 26 elements.

Certification	Use	Remarks
By atomic absorption spectrometry and neutron activation. Use of blank samples is necessary.	For primary calibration of instruments and techniques and as a spike sample in method-of-addition procedures for the determination of Hg in natural waters.	Stability limited to 1 year from date of purchase.
Same as SRM 1641b.	Same as SRM 1641b.	Should be used without dilution. For precautions in use see Certificate.
At least two from the nine analytical procedures employed were used for the determination of each element.	For evaluating the accuracy of trace element determination in fresh water and for instrument calibration.	The certification is valid for two years from the shipping date. For precautions in use see Certificate.
Performed by two independent analytical procedures. Concentrations for the three PAH at other temperatures (10-30 °C) are given in the Certificate.	This SRM is intended to provide accurate concentrations of the three PAH.	For further properties, use, and stability, see Certificate.
Six independent analytical procedures and 100 mg to 1 g of sample were used for certification.	For calibration of apparatus and methods used in the analysis of river sediments or similar materials.	For details on stability, use, and homogeneity, see Certificate. Data valid for 5 years from date of purchase.
By 6 independent analytical procedures on 500 mg or more dried sample, using two to four different procedures for each element.	For calibration of instrumentation and evaluation of analytical methods in sediments and similar matrices.	Certified data valid for 5 years after date of shipping.
On 100 mg or more of dried sample using 9 independent analytical procedures.	Calibration of apparatus and evaluation of methods used in the analysis of atmospheric particulate matter and similar matrices.	Certification valid for 5 years from date of purchase, for samples kept in the original bottle at 10-30 °C in a desiccator and in dark.

Table 3. Comparative List of Chemical Elements for the SRM's Described in Table 2.

Analyzed Liquids and Solids

MULTI-ELEMENT

Concentrations:

Values expressed as microgram per gram, except:

Weight Percent -- %

Nanogram per gram -- *italics*

Parenthesis indicates elements not certified and given for information only.

SRM	Type	Unit Size	Al	Sb	As	Ba
1632b	Trace Elements in Coal (Bituminous)	75 g	0.855%	(0.24)	3.72	67.5
1633a	Trace Elements in Coal Fly Ash	75 g	14.3%	6.8	145	(0.15%)
1634a	Trace Elements in Fuel Oil	100 mL	--	--	(0.12)	--
1635	Trace Elements in Coal (Subbituminous)	75 g	(0.32%)	(0.14)	0.42	--
1643b	Trace Elements in Water (ng/g)	950 mL	--	--	(49)	44
1645	River Sediment	70 g	2.26%	(51)	(66)	--
1646	Estuarine Sediment	75 g	6.25%	(0.4)	11.6	--
1648	Urban Particulate	2 g	3.42%	(45)	115	(737)

SRM	Be	Bi	B	Br	Cd	Ca	C	Ce
1632b	--	--	--	(0.17)	0.0573	0.204%	78.11%	(9)
1633a	(12)	--	--	--	1.00	1.11%	--	(180)
1634a	(0.006)	--	--	(<1)	(0.002)	(16)	--	--
1635	--	--	--	--	0.03	--	--	(3.6)
1643b	19	(11)	(94)	--	20	--	--	--
1645	--	--	--	--	10.2	(2.9%)	--	--
1646	(1.5)	--	--	--	0.36	0.83%	--	(80)
1648	--	--	--	(500)	75	--	--	(55)

SRM	Cs	Cl	Cr	Co	Cu	Eu	F
1632b	(0.44)	(1260)	(11)	2.29	6.28	(0.17)	--
1633a	(11)	--	196	(46)	118	(4)	--
1634a	--	(31)	(0.7)	(0.3)	--	--	--
1635	--	--	2.5	(0.65)	3.6	(0.06)	--
1643b	--	--	18.6	28	21.9	--	--
1645	--	--	2.96%	10.1	109	--	(0.09%)
1646	(3.7)	--	76	10.5	18	(1.5)	--
1648	(3)	(0.45%)	403	(18)	609	(0.8)	--

SRM	Ga	Ge	Hf	H	In	I	Fe	La
1632b	--	--	(0.43)	5.07%	--	--	0.759%	(5.1)
1633a	(58)	--	(8)	--	--	--	9.4%	--
1634a	--	--	--	--	--	--	(31)	--
1635	(1.05)	--	(0.29)	--	--	--	0.239%	--
1643b	--	--	--	--	--	--	99	--
1645	--	--	--	--	--	--	11.3%	(9)
1646	--	(1.4)	--	--	--	--	3.35%	--
1648	--	--	(4.4)	--	(1.0)	(20)	3.91%	(42)

Table 3. Continued.

SRM	Pb	Li	Mg	Mn	Hg	Mo	Ni	N
1632b	3.67	(10)	0.0383%	12.4	--	(0.9)	6.10	1.56
1633a	72.4	--	0.455%	179	0.16	(29)	127	--
1634a	2.80	--	--	0.19	(<0.002)	(0.12)	29	--
1635	1.9	--	--	21.4	--	--	1.74	--
1643b	23.7	--	--	28	--	85	49	--
1645	714	--	0.74%	785	1.1	--	45.8	--
1646	28.2	(49)	1.09%	375	0.063	(2.0)	32	--
1648	0.655%	--	(0.8%)	(860)	--	--	82	--

SRM	P	K	Rb	Sm	Sc	Se	Si	Ag
1632b	--	0.0748%	5.05	(0.87)	(1.9)	1.29	(1.4%)	--
1633a	--	1.88%	131	--	(40)	10.3	22.8%	--
1634a	--	--	--	--	--	0.15	--	--
1635	--	--	--	--	(0.63)	0.9	--	--
1643b	--	--	--	--	--	9.7	--	--
1645	--	1.26%	--	--	(2)	(1.5)	--	--
1646	0.054%	(1.4%)	(87)	--	(10.8)	(0.6)	(31%)	--
1648	--	1.05%	(52)	(4.4)	(7)	27	--	(6)

SRM	Na	Sr	S	Te	Tl	Th	Ti	W
1632b	0.0515%	(102)	1.89%	--	--	1.342	0.0454%	(0.48)
1633a	0.17%	830	(0.18%)	--	5.7	24.7	(0.8%)	--
1634a	87	--	2.85%	--	--	--	--	--
1635	(0.24%)	--	0.33%	--	--	0.62	(0.02%)	--
1643b	--	227	--	--	8.0	--	--	--
1645	0.54%	--	(1.1%)	--	1.44	1.62	--	--
1646	(2.0%)	--	(0.96%)	(0.5)	(0.5)	(10)	(0.51%)	--
1648	0.425%	--	(5.0%)	--	--	(7.4)	(0.40%)	(4.8)

SRM	U	V	Zn
1632b	0.436	(14)	11.89
1633a	10.2	297	220
1634a	--	56	2.7
1635	0.24	5.2	4.7
1643b	--	45.2	66
1645	1.11	23.5	0.172%
1646	--	94	138
1648	5.5	140	0.476%

Table 3. Continued.

Analyzed Liquids and Solids

SINGLE ELEMENT

SRM	Type	Unit Size	Lead	Sulfur	Mercury
1579	Powdered Lead Base Paint	35 g	11.87%	--	--
1620a	Sulfur in Residual Fuel Oil	100 mL	--	4.504%	--
1621b	Sulfur in Residual Fuel Oil	100 mL	--	0.950%	--
1622b	Sulfur in Residual Fuel Oil	100 mL	--	1.982%	--
1623a	Sulfur in Residual Fuel Oil	100 mL	--	0.240%	--
1624a	Sulfur in Residual Fuel Oil	100 mL	--	0.141%	--
1630	Trace Mercury in Coal	50 mg	--	--	0.13 µg/g
1641b	Mercury in Water - µg/mL	120 mL	--	--	1.52 µg/mL
1642b	Mercury in Water - ng/mL	950 mL	--	--	1.49 ng/mL

SRM	Type	Element Certified	Nominal Concentration	No. Units
1636a	Lead in Reference Fuel	Pb	0.03, 0.05, 0.07, and 2.0 g/gal	3 vials each
1637a	Lead in Reference Fuel	Pb	0.03, 0.05, 0.07 g/gal	4 vials each
1638a	Lead in Reference Fuel	Pb	2.0 g/gal	12 vials

Organic Constituents

SRM	Type	Unit of Issue
1580	Shale Oil	Set of 5-2 mL ampoules
1644	Polyaromatic Hydrocarbon Generator Columns	Set of 3 columns

Constituents	SRM 1580 (µg/g)	SRM 1644 (µg/kg)
Anthracene	--	30.7
Benzo[a]anthracene	--	6.45
Benzo[a]pyrene	21	1.13
Benzo[e]pyrene	18	--
Fluoranthene	54	--
o-Cresol	385	--
Phenol	407	--
Perylene	3.4	--
Pyrene	104	--
2,6-Dimethylphenol	175	--
Benzo[f]quinoline	16	--

Appendix I.

Alphabetical Index by Standard Reference Material Name

Name	SRM	Name	SRM
Acetanilide	141c	Aluminum, Freezing Point Standard	44f
Acid Open-Hearth Steel, 0.2% Carbon	19G	Aluminum, Magnetic Gram Susceptibility	763
Acid Potassium Phthalate	84j	Aluminum Oxide, Melting Point	742
AISI 1045 Steel	20g	Aluminum Rod Ultra Purity	RM 1R
AISI 4340 Steel	36l	Aluminum-26 Radioactivity Standard	4229
AISI 4340 Steel	1261a	Americium-241 Alpha-Particle Standard	4904F
AISI 94B17 Steel (Modified)	362	Americium-241 Gamma-ray Standard	4213
AISI 94B17 Steel (Modified)	1262a	Ammonium Dihydrogen Phosphate	194
Albacore Tuna	RM 50	Angiotensin I (Human)	998
Alkali Lead Silicate Glass	712	Anisic Acid	142
Alpha Quartz	1878	Anticonvulsant Drug Level Assay Standard	1599
Alumina (Reduction Grade)	699	Anticpilepsy Drug Level Assay Standard	900
Alumina Silicate Glass	714	Antimony-125-Tellurium-125m, Europium-154, Europium-155 Mixed-Radionuclide Point-Source Standard	4275B
Aluminosilicate Glass	715	Antimony-125-Tellurium-125m, Europium-154, Europium-155 Mixed-Radionuclide Solution Standard	4276B
Aluminum Alloy	85B	A.O.H., 0.4C Spectrographic Steel Standard	413
Aluminum Alloy 6011 (Modified)	858	Argillaceous Limestone	1C
Aluminum Alloy 6011 (Modified)	1258	Arsenic Trioxide Reductometric Standard	83d
Aluminum Alloy 7075	859	Assay-Isotopic Standard for Potassium	985
Aluminum Alloy 7075	1259	Assay-Isotopic Standard for Rhenium	989
Aluminum Block, Eddy Current Conductivity	1860	Assay-Isotopic Standard for Silicon	990
Aluminum Block, Eddy Current Conductivity	1861	Assay-Isotopic Standard for Strontium	987
Aluminum Block, Eddy Current Conductivity	1862	2% Austenite in Ferrite	488
Aluminum Block, Eddy Current Conductivity	1863	5% Austenite in Ferrite	485a
Aluminum Brass Standard for Optical Emission and X-ray Spectroscopic Analysis	1118	15% Austenite in Ferrite	486
Aluminum Brass Standard for Optical Emission and X-ray Spectroscopic Analysis	C1118	30% Austenite in Ferrite	487
Aluminum Brass Standard for Optical Emission and X-ray Spectroscopic Analysis	1119		
Aluminum Brass Standard for Optical Emission and X-ray Spectroscopic Analysis	C1119		
Aluminum Casting Alloy 356	855		
Aluminum Casting Alloy 380	856		
Aluminum Cube Ultra Purity	RM 1C		
Aluminum 2-Ethylhexanoate	1075a		

Name	SRM	Name	SRM
Austenitic Stainless Steel, Thermal Conductivity and Electrical Resistivity	1460	Beryllium on Filter Media	2675
Austenitic Stainless Steel, Thermal Conductivity and Electrical Resistivity	1461	Bessemer Steel (Simulated) 0.1% Carbon	8j
Austenitic Stainless Steel, Thermal Conductivity and Electrical Resistivity	1462	Bilirubin	916
Barium Crown Glass	713	Bis(1-phenyl-1, 3-butanediono) copper (II)	1080a
Barium Cyclohexanebutyrate	1051b	Bis(1-phenyl-1, 3-butanediono) oxovanadium (IV)	1052b
Barium-133 Radioactivity Point-Source Standard	4241B	Black Porcelain Enamel for Directional Hemispherical Reflectance	2021
Barium-133 Radioactivity Standard	4251B	Black Porcelain Enamel for Directional Hemispherical Reflectance	2022
Basalt Rock	688	Blast Furnace Iron Standard (Chill Cast White)	1143a
Base Oil	1083	Blast Furnace Iron Standard (Chill Cast White)	1144a
Basic Electric Spectrographic Steel Standard	404a	B.O.H., 0.4C Spectrographic Steel Standard	417a
Basic Open-Hearth Steel, 0.1% Carbon	15g	Boric Acid	951
Basic Open-Hearth Steel, 0.1% Carbon	335	Boron-Doped Silicon Slices for Resistivity Measurements	1521
Basic Open-Hearth Steel, 0.1% Carbon	1228	Borosilicate Glass	93a
Basic Open-Hearth Steel, 0.2% Carbon	11h	Borosilicate Glass	623
Basic Open-Hearth Steel, 0.4% Carbon	12H	Borosilicate Glass	717
Basic Open-Hearth Steel, 0.5% Carbon	152A	Borosilicate Glass	1825
Basic Open-Hearth Steel, 0.8% Carbon	14f	Borosilicate Glass, Thermal Expansion	731
Basic Open-Hearth Steel, 1% Carbon (Disk)	1227	Bovine Liver	1577a
Basic Open-Hearth Steel, 1.1% Carbon	16f	Bovine Serum Albumin	926
Basic Open-Hearth Steel, 1.1% Carbon	337	Bovine Serum Albumin (7% Solution)	927
0.4C Basic Oxygen Furnace Steel	178	Branched Polyethylene	1476
Bauxite (Arkansas)	69b	Brewers Yeast	1569
Bauxite (Dominican)	697	Bright Copper Microhardness Standard	1894
Bauxite (Jamaican)	698	Bright Nickel Microhardness Standard	1895
Bauxite (Surinam)	696	Bromobenzoic Acid	2142
Benzene in Nitrogen	1805	Burnt Refractory	76a
Benzene in Nitrogen	1806	Burnt Refractory	77a
Benzene Permeation Device	1911	Burnt Refractory	78a
Benzoic Acid	140b	Cadmium Cyclohexanebutyrate	1053a
Benzoic Acid	350a	Cadmium, Vapor Pressure	746
Benzoic Acid Calorimetric Standard	39i	Calcium Carbonate	915
Benzothiazyl Disulfide Rubber Compound	373f	Calcium 2-Ethylhexanoate	1074a
Beryllium-Copper Standard	1122	Calcium in Low-Alloy (Silicon) Steel	1254
Beryllium-Copper Standard	C1122	Calcium Molybdate	71
Beryllium-Copper Standard	C1123	Calibrated Glass Beads	1004
		Calibrated Glass Beads	1017a
		Calibrated Glass Beads	1018a
		Calibrated Glass Spheres	1003a
		Carbon Dioxide in Air	1670
		Carbon Dioxide in Air	1671
		Carbon Dioxide in Air	1672
		Carbon Dioxide in Nitrogen	1674b
		Carbon Dioxide in Nitrogen	1675b
		Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2619a
		Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2620a

Name	SRM	Name	SRM
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2621a	Catalyst Package for Lubricant Oxidation	1817
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2622a	Centerline Drawings for Optical Character Recognition, B	1901
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2623a	Characters	
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2624a	Centroid Color Chart	2106
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2625a	Centroid Color Kit	2107
Carbon Dioxide in Nitrogen (Combustion Efficiency Gas Standard)	2626a	Cesium-137, Barium-137m Point-Source Radioactivity Standard	4200B
Carbon Dioxide in Nitrogen (Mobile Source Emission Gas Standard)	2632	Cesium-137, Barium-137m Point-Source Radioactivity Standard	4207
Carbon Dioxide in Nitrogen (Mobile Source Emission Gas Standard)	2633	Cesium-137 Burn-Up Standard	4233B
Carbon Monoxide in Air (Ambient Air Quality Gas Standard)	2612a	Cesium-134 Radioactivity Standard	4250B
Carbon Monoxide in Air (Ambient Air Quality Gas Standard)	2613a	Channel Black Rubber Compound	375g
Carbon Monoxide in Air (Ambient Air Quality Gas Standard)	2614a	Chlorine-36 Beta-ray Standard	4943
Carbon Monoxide in Nitrogen	1677c	Chlorine-36 Radioactivity Standard	4422L
Carbon Monoxide in Nitrogen	1678c	Chlorobenzoic Acid	2144
Carbon Monoxide in Nitrogen	1679c	Chrome Refractory	103a
Carbon Monoxide in Nitrogen	1680b	Chromium-Molybdenum-Aluminum Steel	106B
Carbon Monoxide in Nitrogen	1681b	Chromium-Molybdenum Steel	36b
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2635	Chromium-Molybdenum Steel	133B
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2636	Chromium-Nickel-Molybdenum Steel	139b
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2637	Chromium-Nickel-Molybdenum Steel	1222
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2638	17Chromium-9 Nickel-0.2 Selenium Steel	339
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2639	Chromium-Nickel Spectrographic Steel Standard	408a
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2640	15Chromium-7 Nickel Steel	344
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2641	16Chromium-4 Nickel Steel	345
Carbon Monoxide in Nitrogen (Mobile Source Emission Gas Standard)	2642	Chromium-51 Radioactivity Standard	4400L-F
Carbon-14 Radioactivity Standard	4245	Chromium Steel	163
Carbon-14 Radioactivity Standard	4246	Chromium-Tungsten Steel	155
Carbon Steel	1224	Chromium-Vanadium Spectrographic Steel Standard	407a
Carbon Steel, 0.6%	13g	Cholesterol	911a
Cast Iron	4k	Chrysotile Asbestos Fibers	1876
Cast Iron	5L	Citrus Leaves	1572
Cast Iron	6g	Clinical Laboratory Thermometer	934
Cast Iron	7G	Cobalt Cyclohexanebutyrate	1055b
Cast Iron Car Wheel	122h	Cobalt-Molybdenum-Tungsten Steel	153A
Cast Steel 3	C1173	Cobalt-57 Radioactivity Standard	4408L-C
Cast Steel Standard	1138a	Cobalt-60 Radioactivity Standard	4915D
Cast Steel Standard	1139a	Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	1115
		Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	C1115
		Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	1116

Name	SRM	Name	SRM
Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	C1116	Cupro-Nickel, 10% (CDA 706) High Purity	874
Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	1117	Cystine	143c
Commercial Bronze Standard for Optical Emission and X-ray Spectroscopic Analysis	C1117	Dextrose	41b
Common Lead Isotopic Standard	981	D-Glucose	917
Copper Concentrate	332	Dibutyltin Bis(2-ethylhexanoate)	1057b
Copper Heat Capacity Test Specimen	RM5	Didymium Glass Filter for Checking the Wavelength Scale of Spectrophotometers	2009
Copper-Nickel-Chromium Cast Iron	115A	Didymium Glass Filter for Checking the Wavelength Scale of Spectrophotometers	2010
Copper Ore, Mill Heads	330	Disodium Hydrogen Phosphate	186IIc
Copper Ore, Mill Tails	331	Disodium Hydrogen Phosphate	2186II
Copper-Thermal Expansion	736a	D-Mannitol	920
Copper, Secondary Freezing Point Standard	45d	Dolomitic Limestone	88a
Cortisol (Hydrocortisone)	921	Doped Platinum	681L1
Creatinine	914	Doped Platinum	681L2
Cr-Mo Low Alloy Steel	1270	Ductile Cast Iron	341
Cr-Mo Steel (ASTM A-213)	291	Electrical Residual Resistivity Ratio Standard	769
Cr-Mo (SAE 4140) Spectrographic Steel Standard	414	Electrolytic Iron	365
Cr-Mo (SAE 4150) Spectrographic Steel Standard	427	Electrolytic Iron	1265a
Cr-Mo (SAE X4130) Spectrographic Steel Standard	418a	Electrolytic Iron, Thermal Conductivity and Electrical Resistivity	1463
Cr-Ni-Mo Steel (AISI 8620)	293	Electrolytic Iron, Thermal Conductivity and Electrical Resistivity	1464
18Cr-10Ni Steel (AISI 304L)	101f	Electronic and Magnetic Alloy Standard	1159
Cr-V Steel (Modified)	363	Electronic and Magnetic Alloy Standard	1160
Cr-V Steel (Modified)	1263a	Enriched Boric Acid	952
Cr-V Steel (SAE 6150)	30f	Equal-Atom Lead Isotopic Standard	982
Crystalline Potassium Dichromate	935	Estuarine Sediment	1646
Crystalline Potassium Iodide, Heterochromatic Stray Radiant Energy Standard	2032	Europium-152 Point-Source Standard	4218E
Crystalline (Ruby) Electron Paramagnetic Resonance Absorption Intensity Standard	2601	Europium-152 Radioactivity Standard	4370B
Cupro-Nickel (CDA 706)	1275	Extra Dense Lead Glass	709
Cupro-Nickel (CDA 715)	1276	Fe-Cr-Ni Alloy Microprobe Standard	479a
Cupro-Nickel, 10% (CDA 706) Doped	875	Fe-3Si Alloy Microprobe Standard	483
		Feldspar	70a
		Feldspar	99a
		Ferrochromium (Low Carbon)	196
		Ferrochromium Silicon	689
		Ferroniobium	340
		Ferrophosphorus	90
		Ferrosilicon	58a
		Ferrosilicon	59a
		Ferrosilicon (75% Si)	195
		First Surface Aluminum Mirror for Specular Reflectance	2003a
		First Surface Mirror, Gold on Glass	2008a

Name	SRM	Name	SRM
Fission Track Glass Standard	961	Gold-198 Radioactivity Standard	4405L-B
Fission Track Glass Standard	962a	Gold-Silver Wires for Microprobe	481
Fission Track Glass Standard	963a	Analysis	
Fission Track Glass Standard	964	Gold, Vapor Pressure	745
Flint Clay	97a	Gray Cast Iron	334
Fluorobenzoic Acid	2143	Halocarbons (in methanol) for Water	1639
Fluorspar	79a	Analysis	
Free-Cutting Brass	1103	High-Alloy Steel (A-743)	C1288
Free-Cutting Brass	C1104	High-Alloy Steel (AISI 310 Mod.)	C1287
Freeze-Dried Urine	2670	High-Alloy Steel, (AISI 414 Mod.)	C1289
Freeze-Dried Urine Certified	2671a	High-Alloy White Cast	892
for Fluoride		High-Alloy White Cast Iron	890
Freeze-Dried Urine Certified	2672a	High-Alloy White Cast Iron	891
for Mercury		High-Carbon Ferrochromium	64c
Fused-Silica Thermal Expansion	739	High-Carbon Ferromanganese	68c
Gadolinium-148 Alpha-Particle	4907	High-Carbon Steel (Modified)	364
Standard		High-Carbon Steel (Modified)	1264a
Gallium Melting-Point Standard	1968	High-Grade Fluorspar	180
Gallium-67 Radioactivity Standard	4416L-D	High-Nickel Steel	126c
Gas Furnace Black Rubber Compound	382a	High-Nickel Steel	1158
Gasometric Set (1095-1099)	1089	High-Purity Gold	685
Gasometric Standard for Unalloyed	357	High-Purity Platinum	680L1A
Zirconium		High-Purity Platinum	680L2A
Gasometric Standard for Unalloyed	358	High-Purity Platinum Thermoelement	1967
Zirconium		High-Purity Zinc	682
Generator Columns for Polynuclear	1644	High-Silicon Steel	179
Aromatic Hydrocarbons		High-Silicon Steel	1134
Gilding Metal	1112	High-Silicon Steel	1135
Gilding Metal	C1112	High-Silicon Steel (Calcium Bearing)	125b
Gilding Metal	1113	High-Sulfur Steel	105
Gilding Metal	C1113	High-Sulfur Steel	129c
Gilding Metal	1114	High-Sulfur Steel	1136
Gilding Metal	C1114	High Temperature Alloy A286	348
Glasses for Microchemical Analysis	1871	High Temperature Alloy M308	1197
Glasses for Microchemical Analysis	1872	High Temperature Alloy L605 and	S1199
Glasses for Microchemical Analysis	1873	S816	
Glasses for Microchemical Analysis	1874	High-Temperature Alloy	1206-2
Glasses for Microchemical Analysis	1875	High-Temperature Alloy	1207-1
Glass Fibers for Microanalysis	RM 31	High-Temperature Alloy	1207-2
Glass Filter for Transmittance	2030	High-Temperature Alloy	1208-1
Measurement		High-Temperature Alloy	1208-2
Glass Filters for Spectrophotometry	930D	Homogeneous River Sediment for	RM 45B
Glass Fluorescence Source	477	Radioactivity Measurements	
Glass Sand	81a	Human Liver, Environmental	4352
Glass Sand	165a	Radioactivity	
Glass Spheres	1019a	Human Lung, Environmental	4351
Gold Coating on Glass Sealing Alloy	1398a	Radioactivity	
Gold Coating on Nickel	1379	Human Serum	909
Gold Coating on Nickel	1380		
Gold Coating on Nickel	1399b		
Gold-Copper Wires for Microprobe	482		
Analysis			
Gold-195 Radioactivity Standard	4421L		

Name	SRM	Name	SRM
Hydrogen in Unalloyed Titanium	352b	Iron Ore (Sibley)	27f
Hydrogen in Unalloyed Titanium	1086	Iron Ore Concentrate (Canada)	690
Hydrogen in Unalloyed Titanium	1087	Iron-59 Radioactivity Standard	4411L-B
Hydrogen in Unalloyed Titanium	1088	Isobutylene-Isoprene (Butyl) Rubber	1495
Hydrogen-3 Radioactivity Standard	4361	Isobutylene-Isoprene (Butyl) Rubber	388L
Hydrogen-3 Radioactivity Standard	4926C	Isotopic Standard for Bromine	977
Hydrogen-3 Toluene Radioactivity Standard	4947	Isotopic Standard for Chlorine	975
4-Hydroxy-3 methoxy-DL-mandelic Acid (VMA)	925	Isotopic Standard for Chromium	979
ICTA High Temperature Set Differential Thermal Analysis	GM 760	Isotopic Standard for Copper	976
ICTA Low Temperature Set Differential Thermal Analysis	GM 757	Isotopic Standard for Magnesium	980
ICTA Mod Temperature Set Differential Thermal Analysis	GM 759	Isotopic Standard for Silver	978
ICTA Mid Temperature Set Differential Thermal Analysis	GM 758	Krypton-85 Gaseous Radioactivity Standard	4308C
ICTA Polystyrene Differential Thermal Analysis	GM 754	Krypton-85 Radioactivity Standard	4235
ICTA Thermogravimetry Set	GM 761	Krypton-85 Radioactivity Standard	4935C
Incoloy, 901 and Hastelloy X	S1198	Lead-Barium Glass	89
Inconels, Alloy 600 (Chips)	864	Lead-Base Bearing Metal	53e
Inconels, Alloy 600 (Solid)	1244	Lead-Base Bearing Metal	1132
Inconels, Alloy 625 (Chips)	865	Lead Cyclohexanebutyrate	1059c
Inconels, Alloy 625 (Solid)	1245	Lead in Reference Fuel	1636a
Incoloy, Alloy 800 (Chips)	866	Lead in Reference Fuel	1637a
Incoloy, Alloy 800 (Solid)	1246	Lead in Reference Fuel	1638a
Incoloy, Alloy 825 (Chips)	867	Lead Nitrate	928
Incoloy, Alloy 825 (Solid)	1247	Lead on Filter Media	2674
Indium-111 Radioactivity Standard	4417L-C	Lead-203 Radioactivity Standard	4420L
Ingot Iron Spectrographic Steel Standard	420a	Lead, Secondary Freezing Point Standard	49e
Intermediate Purity Selenium	726	Lead-Silica Glass	1827
Intermediate-Purity Zinc	728	Lead-Silica Glass (Viscosity)	711
Iodine-123 Radioactivity Standard	4414L-C	Lead-Silica Glass for dc Volume Resistivity	624
Iodine-125 Radioactivity Standard	4407L-H	Lead-Silica Glass for Dielectric Constant	774
Iodine-129 Radioactivity Standard	4949B	Lead 206 Spike Assay and Isotopic Solution Standard	991
Iodine-131 Radioactivity Standard	4401L-I	Leaded-Tin Bronze Alloy	1035
Iron Foil Mössbauer Standard	1541	Light-Sensitive Paper	700d
Iron-55 Low-Energy Photon Standard	4260C	Light-Sensitive Paper	701d
Iron Metal (Clinical Standard)	937	Light-Sensitive Plastic Chip	703
Iron Ore (Labrador)	692	Linear Polyethylene	1475
Iron Ore (Nimba)	693	Linear Polyethylene	1482
		Linear Polyethylene	1483
		Linear Polyethylene	1484
		Linerboard, Standard for Tape Adhesion Testing	1810
		Liquid Absorbance Standard for Ultraviolet and Visible Spectrophotometry	931c
		Lithium Carbonate	924
		Lithium Ore	181
		Lithium Ore	182
		Lithium Ore	183
		Low-Alloy Steel, (AISI 4130)	1225
		Low Alloy Steel	1226
		Low Alloy Steel (A242 Mod.)	C1285
		Low-Alloy Steel, AISI 4130	72g
		Low Alloy Steel (AISI 1526, Modified)	1269
		Low-Alloy Steel (Hy 80)	1286

Name	SRM	Name	SRM
Low-Alloy Steel Set (661-665)	S668	Naval Brass Standards for Optical Emission and Spectroscopic Analysis	1108
Low-Carbon Silicon Steel	131c	Naval Brass Standards for Optical Emission and Spectroscopic Analysis	C1108
Low-Carbon Silicon Steel	1036	Neutral Glass	716
Low-Carbon Stainless Steel (AISI 316L)	166c	Neutron Density Monitor Wire	953
Magnesium-base Alloy	171	Nickel-Chromium Cast Iron	82b
Magnesium Cyclohexanebutyrate	1061c	Nickel-Chromium-Molybdenum Cast Iron	107C
Magnesium Gluconate Dihydrate	929	Nickel-Chromium Steel	32E
Magnetic Coating on Magnetic Substrate (Nickel on Steel)	1365a	Nickel-Copper Alloy	882
Magnetic Coating on Magnetic Substrate (Nickel on Steel)	1366a	Nickel Cyclohexanebutyrate	1065b
Magnetic Coating on Non-Magnetic Substrate (Nickel and Chromium on brass)	1367a	Nickel Oxide, No. 1	671
Magnetic Tape, High Density	6250	Nickel Oxide, No. 2	672
Manganese Fluoride, Magnetic Gram Susceptibility	766	Nickel Oxide, No. 3	673
Manganese Ore	25d	Nickel-63 Radioactivity Standard	4226
Manganese-54 Point-Source Radioactivity Standard	4997E	Nickel Silver (CDA 762)	879
Manganese-54 Radioactivity Standard	4257	Nickel Silver (CDA 770)	880
Manganese Steel	100B	Nickel Spectrographic Steel Standard	409b
Manganous Cyclohexanebutyrate	1062b	Nickel Sphere, Magnetic Moment	772
Maraging Steel	1156	Nickel Steel	33d
Metal on Quartz Filters for Spectrophotometry	2031	Ni-Cr-Mo-V Steel	1173
Metals on Filter Media	2676b	Nicotinic Acid	148
Methane in Air	1658a	Niobium-94 Gamma-ray Standard	4201B
Methane in Air	1659a	Nitric Oxide in Nitrogen	1683b
Methane in Air	1660a	Nitric Oxide in Nitrogen	1684b
Medium Manganese Spectrographic Steel Standard	405a	Nitric Oxide in Nitrogen	1685b
Mercaptobenzothiazole	383a	Nitric Oxide in Nitrogen	1686b
Mercury, Freezing Point	743	Nitric Oxide in Nitrogen	1687b
Mercury-203 Radioactivity Standard	4418L	Nitric Oxide in Nitrogen (Mobile Source Emission Gas Standard)	2627
Mercury in Water, µg/mL	1641b	Nitric Oxide in Nitrogen (Mobile Source Emission Gas Standard)	2628
Mercury in Water, ng/mL	1642b	Nitric Oxide in Nitrogen (Mobile Source Emission Gas Standard)	2629
Microcopy Resolution Test Chart	1010a	Nitric Oxide in Nitrogen (Mobile Source Emission Gas Standard)	2630
Microprobe Standard - Cartridge Brass	478	Nitric Oxide in Nitrogen (Mobile Source Emission Gas Standard)	2631
Mineral Glasses for Microanalysis	470	Nitrogen Dioxide in Air (Stationary Source Emission Gas Standard)	2653
Molybdenum Concentrate	333	Nitrogen Dioxide in Air (Stationary Source Emission Gas Standard)	2654
Molybdenum, Heat Capacity	781	Nitrogen Dioxide in Air (Stationary Source Emission Gas Standard)	2655
Molybdenum-99 Radioactivity Standard	4412L-H	Nitrogen Dioxide in Air (Stationary Source Emission Gas Standard)	2656
Molybdenum-Tungsten-Chromium-Vanadium Steel	134A	Nitrogen Dioxide Permeation Device	1629a
Naval Brass Standards for Optical Emission and Spectroscopic Analysis	1106	4-Nitrophenol	938
Naval Brass Standards for Optical Emission and Spectroscopic Analysis	C1106		
Naval Brass Standards for Optical Emission and Spectroscopic Analysis	1107		
Naval Brass Standards for Optical Emission and Spectroscopic Analysis	C1107		

Name	SRM	Name	SRM
Nodular Cast Iron	342a	Organics in Shale Oil	1580
Nominal One Micrometer Polystyrene Spheres	1690	Oxalic Acid	4990C
Non-Fat Powdered Milk	1549	Oxygen in Ferrous Materials	1090
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1359	Ingot Iron	
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1360	Oxygen in Ferrous Materials (Stainless Steel AISI 431)	1091
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1361b	Oxygen in Ferrous Materials Vacuum Melted Steel	1092
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1362a	Oxygen in Maraging Steel	1094
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1363a	Oxygen in Nitrogen (Gas Standard)	2657
Nonmagnetic Coating on Magnetic Substrate (Copper and Chromium on Steel)	1364a	Oxygen in Nitrogen (Gas Standard)	2658
NPL GM Alpha Alumina	8005	Oxygen in Nitrogen (Gas Standard)	2659
NPL GM Alpha Alumina	8006	Oxygen in Titanium-Base Materials	355
NPL GM Alpha Alumina	8007	Oxygen in Valve Steel	1093
NPL GM Alpha Alumina	8008	Oyster Tissue	1566
NPL GM Graphitized Carbon Black	8001	Palladium, Magnetic Gram	765
NPL GM Graphitized Carbon Black	8002	Susceptibility	
NPL GM Melting Point Set	8000	Penetrant Test Block	1850
NPL GM Non-porous Silica	8003	Peruvian Soil, Environmental	4355
NPL GM Non-porous Silica	8004	Radioactivity	
N-tertiary-Butyl-2-benzothiazolesulfenamide Rubber Compound	384d	Petroleum Crude Oil	1582
Obsidian Rock	278	Phosphate Rock (Florida)	120b
Octaphenylcyclotetrasiloxane	1066a	Phosphor Bronze (CDA 521)	871
Oil Furnace Black Rubber Compound	378b	Phosphor Bronze (CDA 544)	872
Opal Glass Powder	91	Phosphorized Copper, Cu VIII	C1251
Optical Emission and X-ray Spectroscopic Analysis	1102	Phosphorized Copper, Cu IX	C1252
Optical Microscope Linewidth Measurement Standard	474	Phosphorized Copper, Cu X	C1253
Optical Microscope Linewidth Measurement Standard	475	Phosphorus-32 Radioactivity Standard	4406L-G
Optical Microscope Linewidth Measurement Standard	476	Photographic Step Tablet	1008
		Pine Needles	1575
		Plastic Clay	98a
		Platinum, Magnetic Gram	764
		Susceptibility	
		Plutonium-238 Alpha-Particle Standard	4906B
		Plutonium-240 Alpha-Particle Emission-Rate Solution Standard	4338
		Plutonium-239 Alpha-Particle Solution Standard	4331
		Plutonium-242 Alpha-Particle Solution Standard	4334B
		Plutonium Isotopic Standard	946
		Plutonium Isotopic Standard	947
		Plutonium Isotopic Standard	948
		Plutonium Metal	949f
		Plutonium Metal (Standard Matrix Material)	945
		Plutonium-244 Spike Assay and Isotopic Standard	996
		Polychlorinated Biphenyls in Oil	1581
		Polycrystalline Alumina Elasticity Standard	718
		Polyester Plastic Film for Oxygen Gas Transmission	1470
		Polyisobutylene Solution in Cetane	1490
		Polystyrene	1478
		Polystyrene	1479
		Polystyrene (Broad Molecular Weight)	706
		Polystyrene (Narrow Molecular Weight)	705
		Polystyrene Spheres	1691
		Portland Cement (Black)	1880

Name	SRM	Name	SRM
Portland Cement (Blue)	635	Quartz on Filter Media	2679a
Portland Cement (Clear)	639	Quinine Sulfate Dihydrate	936
Portland Cement (Gold)	634	Radiogenic Lead Isotopic Standard	983
Portland Cement (Green)	638	Radium-226 Gamma-ray Standard	4956
Portland Cement (Pink)	637	Radium-226 Gamma-ray Standard	4957
Portland Cement (Red)	633	Radium-226 Gamma-ray Standard	4958
Portland Cement (White)	1881	Radium-226 Gamma-ray Standard	4959
Portland Cement (Yellow)	636	Radium-226 Gamma-ray Standard	4960
Portland Cement Fineness Standard	114n	Radium-226 Gamma-ray Standard	4961
Potassium Chloride	2202	Radium-226 Gamma-ray Standard	4962
Potassium Chloride (Clinical Standard)	918	Radium-226 Gamma-ray Standard	4963
Potassium Chloride (Primary Chemical)	999	Radium-226 Gamma-ray Standard	4964B
Potassium Chloride for Solution Calorimetry	1655	Radium Standard (Blank Solution)	4952B
Potassium Dichromate	136d	Radon-226 for Radon Analysis	4953C
Potassium Dihydrogen Phosphate	200	Red Brass	1109
Potassium Dihydrogen Phosphate	1861c	Red Brass	C1109
Potassium Dihydrogen Phosphate	2186I	Red Brass	1110
Potassium Erucate	1076	Red Brass	C1110
Potassium Feldspar	607	Red Brass	1111
Potassium Fluoride	2203	Red Brass	C1111
Potassium Hydrogen Phthalate	185e	Reduced Iron Oxide	691
Potassium Hydrogen Tartrate	188	Reference Fuel Isooctane	1816a
Potassium Iodide with Attenuator	2033	Reference Fuel n-Heptane	1815a
Potassium Nitrate	193	Reflection Step Tablet	2061
Potassium Tetroxalate	189	Refractive Index Glass	1820
Powdered Lead Based Paint	1579	Refractive Index Silicone Liquids	1823
Priority Pollutant Polynuclear Aromatic Hydrocarbons (in Acetonitrile)	1647	Refractive Index, Soda-Lime Glass	1822
Propane in Air	1665b	Relative Stress-Optical Coefficient Glass	708
Propane in Air	1666b	Resulfurized-Rephosphorized Steel	C1221
Propane in Air	1667b	Rice Flour	1568
Propane in Air	1668b	River Sediment	1645
Propane in Air	1669b	River Sediment, Environmental Radioactivity	4350B
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2643	Rocky Flats Soil Number 1, Environmental Radioactivity	4353
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2644	Rubidium Melting Point	1969
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2645	Rutile Ore	670
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2646	Scanning Electron Microscope Magnification Standard	484c
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2647	Scanning Electron Microscope Performance Standard	2069
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2648	Secondary Standard Flexible Disk Cartridge (Computer Amplitude Reference)	3210
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2649	Secondary Standard Magnetic Tape	3200
Propane in Nitrogen (Mobile Source Emission Gas Standard)	2650	Secondary Standard Magnetic Tape Cassette	1600
Propane in Nitrogen and Oxygen (Mobile Source Emission Gas Standard)	2651	Secondary Standard Magnetic Tape Cartridge (Computer Amplitude Reference)	3216
Propane in Nitrogen and Oxygen (Mobile Source Emission Gas Standard)	2652	Second Surface Aluminum Mirror for Specular Reflectance	2023
Quartz Cuvette for Spectrophotometry	932		
Quartz for Hydrofluoric Acid Solution Calorimetry	1654		

Name	SRM	Name	SRM
Second Surface Aluminum Mirror for Specular Reflectance	2024	Soda-Lime Sheet Glass	1831
Second Surface Aluminum Mirror with Wedge for Specular Reflectance	2025	Soda-Lime Silica Glass	622
Selenium-Bearing Steel	1170b	Soda-Lime Silica Glass	710
Selenium-75 Radioactivity Standard	4409L-D	Soda-Lime Silica Glass for Liquidus Temperature	773
Sheet Brass	37E	Sodium Bicarbonate	191a
Silica Brick	198	Sodium Bicarbonate	2191
Silica Brick	199	Sodium Carbonate	192a
Silicon-Aluminum Alloy	87a	Sodium Carbonate	2192
Silicon Bronze	158A	Sodium Chloride	2201
Silicon Density Standard	1840	Sodium Chloride (Clinical Standard)	919
Silicon Density Standard	1841	Sodium Cyclohexanebutyrate	1069b
Silicon Metal	57a	Sodium Oxalate Reductometric Standard	40h
Silicon Powder, Spacing Standard for X-ray Diffraction	640a	Sodium Pyruvate	910
Silicon Power Device Level Resistivity Standard	1522	Sodium Tetraborate Decahydrate (Borax)	187b
Silicon Resistivity Standard for Eddy Current Testers	1523	Solder	127b
Silver 2-Ethylhexanoate	1077a	Solder	1131
Silver-Gold Thermocouple Wire	733	Special Nuclear Container DOT 6M, 15 gal.	9940
Silver, Vapor Pressure	748	Special Nuclear Container, 55 gal.	9941
Sintered and Arc-Cast Tungsten, Thermal Conductivity and Electrical Resistivity	1465	Special Nuclear Container Type A, 10 gal.	9942
Sintered and Arc-Cast Tungsten, Thermal Conductivity and Electrical Resistivity	1466	Special Nuclear Container, Type A, 55 gal.	9943
Sintered and Arc-Cast Tungsten, Thermal Conductivity and Electrical Resistivity	1467	Special Nuclear Material Package	9910
Sintered and Arc-Cast Tungsten, Thermal Conductivity and Electrical Resistivity	1468	Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	461
Smoke Density Chamber Standard (Flaming Exposure Condition)	1007a	Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	462
Smoke Density Chamber Standard (Non-flaming Exposure Condition)	1006b	Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	463
Soda-Lime Container Glass	621	Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	464
Soda-Lime Flat Glass	620	Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	465
Soda-Lime Float Glass	1830	Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	466
Soda-Lime Glass	1826	Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	467
Soda-Lime Glass Powder	92	Spectrographic Ingot Iron and Low-Alloy Steel Standard (Rod)	468
		Spectrographic Ingot Iron and Low-Alloy Steel Standard	1166
		Spectrographic Stainless Steel Standard	442
		Spectrographic Stainless Steel Standard	443
		Spectrographic Stainless Steel Standard	444
		Spectrographic Stainless Steel Standard (Disc)	D849
		Spectrographic Stainless Steel Standard (Disc)	D850
		Spectrographic Stainless Steel Standard (Group II)	445

Name	SRM	Name	SRM
Spectrographic Stainless Steel Standard (Group II)	446	Spectroscopic Titanium-Base Standard	644
Spectrographic Stainless Steel Standard (Group II)	447	Spectroscopic Titanium-Base Standard	645
Spectrographic Stainless Steel Standard (Group II)	448	Spectroscopic Titanium-Base Standard	646
Spectrographic Stainless Steel Standard (Group II)	449	Spheroidized Iron Carbide in Ferrite	493
Spectrographic Stainless Steel Standard (Group II)	450	Spreading Resistance Calibration (100) n-Type Silicon	2529
Spectrographic Stainless Steel Standard (Rod)	849	Spreading Resistance Calibration (100) p-Type Silicon	2528
Spectrographic Stainless Steel Standard (Rod)	850	Spreading Resistance Calibration (111) n-Type Silicon	2527
Spectrographic Steel Standard (Disc)	D803a	Spreading Resistance Calibration (111) p-Type Silicon	2526
Spectrographic Steel Standard (Disc)	D807a	Stabilized Wine	1590
Spectrographic Steel Standard (Rod)	803a	Stainless Steel	121d
Spectrographic Steel Standard (Rod)	804a	Stainless Steel	123c
Spectrographic Steel Standard (Rod)	805a	Stainless Steel	160b
Spectrographic Steel Standard (Rod)	807a	Stainless Steel (AISI 446)	367
Spectrographic Steel Standard (Rod)	808a	Stainless Steel (AISI 446)	1267
Spectrographic Steel Standard (Rod)	809a	Stainless Steel, 13% Chromium	73c
Spectrographic Steel Standard (Rod)	817b	Stainless Steel, Cr-Ni	C1151
Spectrographic Steel Standard (Rod)	820a	Stainless Steel, Cr-Ni	1151a
Spectrographic Steel Standard (Rod)	821	Stainless Steel, Cr-Ni	C1152
Spectrographic Steel Standard (Rod)	827	Stainless Steel, Cr-Ni	1152a
Spectrographic Tool Steel Standard	436	Stainless Steel, Cr-Ni	C1153
Spectrographic Tool Steel Standard	437	Stainless Steel, Cr-Ni	1153a
Spectrographic Tool Steel Standard	438	Stainless Steel, Cr-Ni	C1154
Spectrographic Tool Steel Standard	439	Stainless Steel, Cr-Ni	1154a
Spectrographic Tool Steel Standard	440	Stainless Steel, Cr-Ni-Mo	1155
Spectrographic Tool Steel Standard	441	Stainless Steel, Cr-Ni-Nb	1172
Spectrographic Tool Steel Standard	837	Stainless Steel, Cr-Ni-Ti	1171
Spectrographic Tool Steel Standard	840	Stainless Steel for Pitting or Crevice Corrosion	1890
Spectrographic Tool Steel Standard (Disc)	D837	Stainless Steel Thermal Expansion	738
Spectrographic Tool Steel Standard (Disc)	D840	Stearic Acid Rubber Compound	372h
Spectrographic Tool Steel Standard (Disc)	D841	Steel (AISI 1211)	368
Spectrographic Zinc-Base Die-Casting Alloy A	625	Steel (Lead-Bearing)	1169b
Spectrographic Zinc-Base Die-Casting Alloy B	626	Strontium Cyclohexanebutyrate	1070a
Spectrographic Zinc-Base Die-Casting Alloy C	627	Strontium-85 Radioactivity Standard	4403L-B
Spectrographic Zinc-Base Die-Casting Alloy D	628	Strontium-89 Radioactivity Standard	4945D
Spectrographic Zinc-Base Die-Casting Alloy E	629	Styrene-butadiene Rubber (Type 1500)	386h
Spectrographic Zinc-Base Die-Casting Alloy F	630	Succinonitrile Freezing Point	1970
Spectroscopic Titanium-Base Standard	641	Sucrose	17c
Spectroscopic Titanium-Base Standard	642	Sulfate and Nitrate on Filter Media	2673
Spectroscopic Titanium-Base Standard	643	Sulfur Dioxide in Nitrogen	1661a
		Sulfur Dioxide in Nitrogen	1662a
		Sulfur Dioxide in Nitrogen	1663a
		Sulfur Dioxide in Nitrogen	1664a
		Sulfur Dioxide in Nitrogen	1693
		Sulfur Dioxide in Nitrogen	1694

Name	SRM	Name	SRM
Sulfur Dioxide in Nitrogen	1696	Titanium-Base Alloy (Unalloyed)	650
Sulfur Dioxide Permeation Tube (2 cm tube)	1627	Titanium-Base Alloy (Unalloyed)	651
Sulfur Dioxide Permeation Tube (5 cm tube)	1626	Titanium-Base Alloy (Unalloyed)	652
Sulfur Dioxide Permeation Tube (10 cm tube)	1625	Titanium Dioxide	154b
Sulfur in Coal	2682	Toluene	211c
Sulfur in Coal	2683	Tomato Leaves	1573
Sulfur in Coal	2684	Tool Steel (AISI M2)	132b
Sulfur in Coal	2685	Tool Steel (AISI M2)	1157
Sulfur in Residual Fuel Oil	1619	Tool Steel Abrasive Wear Standard	1857
Sulfur in Residual Fuel Oil	1620a	Tracealloy (Nickel-Base High-Temperature Alloy)	897
Sulfur in Residual Fuel Oil	1621b	Tracealloy (Nickel-Base High-Temperature Alloy)	898
Sulfur in Residual Fuel Oil	1622b	Tracealloy (Nickel-Base High-Temperature Alloy)	899
Sulfur in Residual Fuel Oil	1623a	Trace Elements in a Glass Matrix	610
Sulfur in Residual Fuel Oil	1624a	Trace Elements in a Glass Matrix	611
Sulfur Rubber Compound	371g	Trace Elements in a Glass Matrix	612
Superconductive Thermometric Fixed Point Device	767a	Trace Elements in a Glass Matrix	613
Superconductive Thermometric Fixed Point Device	768	Trace Elements in a Glass Matrix	614
Surface Flammability Standard	1002c	Trace Elements in a Glass Matrix	615
Synthetic Sapphire	720	Trace Elements in a Glass Matrix	616
Technetium-99 Radioactivity Standard	4288	Trace Elements in a Glass Matrix	617
Technetium-99m Radioactivity Standard	4410H-I	Trace Elements in Coal (Bituminous)	1632a
Tetrachloroethylene in Nitrogen	1808	Trace Elements in Coal (Sub- bituminous)	1635
Thallium-201 Radioactivity Standard	4404L-F	Trace Elements in Coal Fly Ash	1633a
Thermal Resistance, Fibrous Glass Batt	1451	Trace Elements in Fuel Oil	1634a
Thermal Resistance, Fibrous Glass Board	1450b	Trace Elements in Water	1643a
Thorium-228, Thallium-208 Gamma-ray Point-Source Standard	4206C	Trace Mercury in Coal	1630
Tin-Base Bearing Metal	54D	2,2,4-Trimethylpentane	217c
Tin, Freezing Point	741	Tripalmitin	1595
Tin-113-Indium-113m Radioactivity Standard	4402L-C	Tris, Basimetric	723a
Tin-121m Point-Source Gamma-ray Emission-Rate Standard	4264B	Tris, for Solution Calorimetry	724a
Tin, Secondary Freezing Point Standard	42g	Tris(hydroxymethyl)aminomethane	922
Titanium Alloy	654a	Tris(hydroxymethyl)aminomethane hydrochloride	923
Titanium-Base Alloy	173b	Tris(1-phenyl-1, 3-butanediono) Chromium (III)	1078b
Titanium-Base Alloy	176	Tris(1-phenyl-1, 3-butanediono) Iron (III)	1079b
		Triphenyl Phosphate	1071b
		Tungsten Carbide	276a
		Tungsten-Chromium-Vanadium Steel	50c
		Tungsten Concentrate	277
		Tungsten, Heat Capacity	782
		Tungsten-20% Molybdenum Alloy Electron Microprobe Standard	480
		Tungsten Thermal Expansion	737
		Unalloyed Copper	1034
		Unalloyed Copper, Cu "O"	393
		Unalloyed Copper, Cu IV	457
		Unalloyed Copper, Cu XI	454
		Unalloyed Copper, Cu I (Chip)	394
		Unalloyed Copper, Cu II (Chip)	395
		Unalloyed Copper, Cu III (Chip)	396
		Unalloyed Copper, Cu V (Chip)	398
		Unalloyed Copper, Cu VI (Chip)	399
		Unalloyed Copper, Cu VII (Chip)	400
		Unalloyed Copper, Cu I (Rod)	494

Name	SRM	Name	SRM
Unalloyed Copper, Cu II (Rod)	495	Wear-Metals in Lubricating Oil (300 ppm)	1085
Unalloyed Copper, Cu III (Rod)	496	Wheat Flour	1567
Unalloyed Copper, Cu V (Rod)	498	White Cast Iron	338
Unalloyed Copper, Cu VI (Rod)	499	White Cast Iron (Disc)	1145
Unalloyed Copper, Cu VII (Rod)	500	White Cast Iron (Disc)	1146
Unalloyed Titanium	354	White Cast Iron (Disc)	1150
Uranium Isotopic Standard (Nominally depleted to 0.02%)	U-0002	White Ceramic Tile for Directional Hemispherical Reflectance	2019b
Uranium Isotopic Standard	U-005a	White Ceramic Tile for Directional Hemispherical Reflectance	2020
Uranium Isotopic Standard (Nominally 1% Enriched)	U-010	White Iron	3d
Uranium Isotopic Standard (Nominally 1.5% Enriched)	U-015	White Opan Glass Diffuse Spectral Reflectance Standard for the Visible Spectrum	2015
Uranium Isotopic Standard	U-020	Xenon-127 Gaseous Radioactivity Standard	4309G
Uranium Isotopic Standard	U-030a	Xenon-133 Gaseous Radioactivity Standard	4307I
Uranium Isotopic Standard (Nominally 5% Enriched)	U-050	Xenon-133 Gaseous Radioactivity Standard	4415L-I
Uranium Isotopic Standard (Nominally 10% Enriched)	U-100	Xenon-133, Xenon-137, Krypton-85 Mixed Gaseous Radioactivity Standard	4310B
Uranium Isotopic Standard (Nominally 15% Enriched)	U-150	X-ray Film Step Tablet	1001
Uranium Isotopic Standard (Nominally 20% Enriched)	U-200	X-ray Powder Diffraction Intensity Standard	674
Uranium Isotopic Standard (Nominally 35% Enriched)	U-350	X-ray Powder Diffraction (Mica) Low 2 Theta	675
Uranium Isotopic Standard (Nominally 50% Enriched)	U-500	Ytterbium-169 Radioactivity Standard	4419L-B
Uranium Isotopic Standard (Nominally 75% Enriched)	U-750	Zinc-Base Alloy (Die Casting)	94c
Uranium Isotopic Standard (Nominally 80% Enriched)	U-800	Zinc Concentrates	113a
Uranium Isotopic Standard (Nominally 85% Enriched)	U-850	Zinc Concentrates	329
Uranium Isotopic Standard (Nominally 90% Enriched)	U-900	Zinc Cyclohexanebutyrate	1073b
Uranium Isotopic Standard (Nominally 93% Enriched)	U-930	Zinc, Freezing Point	740
Uranium Isotopic Standard (Nominally 97% Enriched)	U-970	Zinc, Freezing Point Standard	43h
Uranium Metal	960	Zinc Metal	683
Uranium Oxide	950b	Zinc Oxide Rubber Compound	370e
Uranium Oxide	969	Zircaloy-2	360a
Uranium-233 Spike Assay and Isotopic Solution Standard	995	Zircaloy-4 Metal	1237
Uranium-235 Spike Assay and Isotopic Solution Standard	993	Zircaloy-4 Metal	1238
Urban Dust/Organics	1649	Zircaloy-4 Metal	1239
Urban Particulate Matter	1648	Zirconium-Barium Chromate Formulation for Heat-Source Powder Calorimetry	1651
Urea	912a	Zirconium-Barium Chromate Formulation for Heat-Source Powder Calorimetry	1652
Urea	2141	Zirconium-Barium Chromate Formulation for Heat-Source Powder Calorimetry	1653
Urea	2152	Zirconium Metal	1234
Uric Acid	913	Zirconium Metal	1235
Vanadium and Nickel in Residual Fuel Oil	1618	Zirconium Metal	1236
Vanadium in Curde Oil	8505		
Vanadium-49 Low-Energy Photon Standard	4266		
Waspaloy	349		
Wear-Metals in Lubricating Oil (100 ppm)	1084		

U. S. Department of Commerce
Frederick B. Dent
Secretary
National Bureau of Standards
Richard W. Roberts, Director

Appendix II. Certificates for the Environmental Research,
Analysis, and Control Standards (listed in
numerical order).

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1579

Powdered Lead Based Paint

This Standard Reference Material is intended for use in the calibration of apparatus and methods used in the determination of lead in paint removed from the interior surfaces of old housing. The certified value is based on at least a 100 milligram sample of the as-received, total material.

Lead Content 11.87 ± 0.04 Weight Percent

The certified value of 11.87 percent lead is the weighted average value determined by a statistical analysis of the results of 32 determinations by atomic absorption spectrometry (average 11.84 percent lead, $s = 0.13$ percent lead), and 16 determinations by polarography (average 11.93 percent lead, $s = 0.13$ percent lead). The standard error of the weighted average is 0.02 percent lead, and the half-width of the 95 percent confidence interval is taken to include ± 0.04 percent lead by weight.

X-ray fluorescence spectrometry showed the bottle-to-bottle inhomogeneity of the material with respect to lead content to be no greater than 0.02 percent lead; no within-bottle inhomogeneity was detected.

Analyses for lead and determinations of homogeneity were carried out in the NBS Analytical Chemistry Division by the following persons:

X-ray Fluorescence: S. D. Rasberry
Atomic Absorption Spectrometry: T. C. Rains and T. A. Rush
Polarography: E. J. Maienthal

Statistical calculations were carried out by J. Mandel of the NBS Institute for Materials Research.

The overall direction and coordination of the technical measurements leading to this certificate were performed under the chairmanship of B. Greifer.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T. W. Mears.

Washington, D. C. 20234
January 23, 1973

J. Paul Cali, Chief
Office of Standard Reference Materials

(over)

Preparation, Testing, and Analysis

Collection

The paint for this Standard Reference Material was collected by the staff of the Philadelphia Department of Public Health from the interior surfaces of dwellings undergoing renovation. The paint was softened with a hand torch, scraped from the plaster and wood substrates, and collected in plastic bags as a heterogeneous mixture of many different kinds of paints. In the laboratory, non-paint matter such as bits of metal, plastic, glass, and wood were removed and the paint mixture was ground in a disk mill to produce a material suitable for feeding into a jet mill. The paint was comminuted in a jet mill operating at 100 psig air pressure, then sieved through a 100-mesh vibrating screen to remove the coarse, non-grindable fraction. Two additional passes through the jet mill at 97 to 107 psig gave a fine powder with 99.31 weight percent passing through a 325 mesh sieve.

Homogeneity


Sample homogeneity was ascertained by x-ray fluorescence analysis for lead content on 17 samples chosen at random from the total lot. A statistical analysis of the data from 136 observations showed the bottle-to-bottle variability among the samples to be no greater than 0.02 percent lead. No within-bottle variation with respect to lead was detected.

Dissolution

A procedure used to dissolve the sample is summarized briefly: dry ash the weighed paint for 2 hours at 450 ° C, digest with 2:5 HCl - HNO₃ containing HF, evaporate to dryness; treat with HNO₃, evaporate to dryness; treat twice with HCl and evaporate to dryness each time. Extract the solids twice with portions of acetic acid - ammonium acetate solution, heating for several hours just below boiling. Combine the extracts and heat the mixture (including solids) for one hour, just below boiling. Cool the mixture and determine lead in solution. (The solids need not be removed for polarographic analysis.)

An alternate procedure for sample dissolution is: dry ash the weighed paint for 6 hours at 500 ° C, cool, then digest for 2 hours in 1:1 HCl - HNO₃. Separate the insoluble solids from the solution by centrifuging, and wash 3 times with 1:10 HNO₃ combining the rinsings with the principal solution. Determine lead in solution.

Details of the dissolution procedures, the analytical procedures, and results will be published in the 260 series of NBS Special Publications.



National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1580

Organics in Shale Oil

This SRM is intended primarily for evaluating the reliability of analytical methods for the determination of trace level organic compounds in an oil matrix, i.e., shale oil, petroleum crude oil, or coal-derived liquids.

Certified Values of Constituent Organic Compounds: *The certified values for selected organic constituents are shown in Table 1. These values are based on results obtained by two independent, analytical methods (see Table 2). Non-certified values, which are given for information only, are listed in Table 3.*

NOTICE AND WARNINGS TO USER

Expiration of Certification: This certification is valid, within the limits certified, for 3 years from the date of purchase. In the event that the certification should become invalid before then, purchasers will be notified by NBS.

Storage: Sealed ampoules, as received, should be stored in the dark at temperatures between 10-30 °C.

Use: Samples for analysis should be withdrawn from ampoules immediately after opening and processed without delay for any certified value in Table 1 to be valid within the stated uncertainty. Certified values are not applicable to ampoules stored after opening, even if resealed.

PREPARATION AND ANALYSIS

The shale oil for this SRM came from a 150-ton retort for *in-situ* simulated combustion of oil shale, operated by the Laramie Energy Technology Center, Laramie, Wyoming. The shale was from the Mahogany Zone of the Colorado Green River Formation. The shale oil had been supplied in November 1975 to the Oak Ridge National Laboratory (ORNL) where it underwent centrifugation to separate the oil from water and sludge. The shale oil was provided to NBS by Bruce R. Clark, ORNL, Oak Ridge, Tennessee.

At NBS, the centrifuged sample was filtered through fine filter paper and mixed in a 20-liter, Teflon-stoppered, glass bottle by rolling for 40 hours. Samples were aliquoted into 2-mL amber glass ampoules. Although not intended to be representative of all shale oils, SRM 1580 provides a typical specimen of this matrix for use in developing analytical methods.

Randomly selected ampoules were analyzed. Each analyst examined at least six ampoules, sometimes measuring replicates from one ampoule. No trend was found in measured values with the ampouling sequence.

Two independent techniques were employed for the determination of the certified values for the organic constituents. Three different methods of sample preparation were used prior to analysis: simple dilution of the shale oil with methylene chloride (or other suitable solvent), acid/basic extraction to isolate acidic, basic, and neutral components; and a high performance liquid chromatographic fractionation. The following techniques were employed for the final quantitative analysis: gas chromatography (GC), gas chromatography/mass spectrometry (GC/MS) with single ion monitoring for selective detection, and high performance liquid chromatography (HPLC) with selective fluorescence detection. All GC/MS analyses used the standard addition method for quantitation. The GC and HPLC analyses employed either internal standard, external standard, or standard addition methods. The analytical methods and the corresponding values are summarized in Table 2.

Consultation on the statistical design of the experimental work was provided by K. R. Eberhardt of the Statistical Engineering Division.

The coordination of the technical measurements leading to certification were performed under the direction of H. S. Hertz, S. N. Chesler, L. R. Hilpert, W. E. May, and S. A. Wise.

The technical and support aspects involved in preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Washington, D.C. 20234
November 24, 1980
(Revision of Certificate
dated 3-10-80)

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

The following members of the staff of the Center for Analytical Chemistry, Organic Analytical Research Division, performed the analytical determinations.

- | | |
|-----------------------|-----------------|
| 1. J. M. Brown-Thomas | 5. P. L. Konash |
| 2. S. N. Chesler | 6. W. E. May |
| 3. F. R. Guenther | 7. R. M. Parris |
| 4. L. R. Hilpert | 8. K. L. Richie |

TABLE 1. Certified Values of Organic Constituents

Compound	Concentration ($\mu\text{g/g}^a$)
Fluoranthene	54 \pm 10
Pyrene	104 \pm 18
Benzo[a]pyrene	21 \pm 6
Benzo[e]pyrene	18 \pm 8
Perylene	3.4 \pm 2.2
Phenol	407 \pm 50
<i>o</i> -Cresol	385 \pm 50
2,6-Dimethylphenol	175 \pm 30
Benzo[f]quinoline (5,6-Benzoquinoline)	16 \pm 4

^aThe estimated uncertainty listed for a constituent is based on judgment and represents an evaluation of the combined effects of method imprecision, possible systematic errors among methods, and material inhomogeneity. The estimated uncertainty is intended to correspond to approximately 95% confidence limits.

TABLE 2. Summary of Results by the Analytical Methods Used in Certification

Compound	Concentration ($\mu\text{g/g}^a$)	Number of Ampoules Analyzed	Sample Preparation Technique	Analytical Technique
Fluoranthene	55 \pm 5	6	Direct Injection	GC/MS
	53 \pm 2	9	HPLC	HPLC
Pyrene	101 \pm 5	6	Direct Injection	GC/MS
	107 \pm 8	10	HPLC	HPLC
Benzo[a]pyrene	20 \pm 1	6	Direct Injection	GC/MS
	23 \pm 1	8	HPLC	HPLC
Benzo[e]pyrene	17 \pm 1	6	Direct Injection	GC/MS
	20 \pm 3	8	HPLC	HPLC
Perylene	2.8 \pm 0.6	5	Direct Injection	GC/MS
	3.9 \pm 0.6	11	HPLC	HPLC
Phenol	412 \pm 35	8	HPLC	GC/MS
	402 \pm 4	8	Acid/Base Extraction	GC
<i>o</i> -Cresol	386 \pm 42	8	HPLC	GC/MS
	384 \pm 9	8	Acid/Base Extraction	GC
2,6-Dimethylphenol	183 \pm 23	9	HPLC	GC/MS
	168 \pm 8	8	Acid/Base Extraction	GC
Benzo[f]quinoline (5,6-Benzoquinoline)	16 \pm 1	7	HPLC	HPLC
	15 \pm 1	8	Acid/Base Extraction	Multi-dimensional GC

^aUncertainty is the standard deviation of a single measurement.

TABLE 3. Non-Certified Values of Organic Compounds in Shale Oil

NOTE: The values shown in this table are not certified because they are not based on the results of two independent methods. These values are included for information only.

<u>Compound</u>	<u>Concentration ($\mu\text{g/g}$)</u>
<i>p</i> -Cresol	(270) ^a
<i>m</i> -Cresol	(330) ^a
2,5-Dimethylphenol	(320) ^a
2,4-Dimethylphenol	(380) ^a
2,5,6-Trimethylphenol	(360) ^a
2,4,6-Trimethylphenol	(120) ^a
Phenanthridine	(45) ^b

^aAcid/base extraction - GC analysis
^bHPLC extraction - HPLC analysis

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1620a

Sulfur in Residual Fuel Oil

Sulfur Concentration 4.504 ± 0.010 weight percent

This Standard Reference Material is intended for use as an analytical standard in the determination of total sulfur in fuel oils or materials of similar matrices. SRM 1620a is a commercial "No. 5 Heavy" residual fuel oil as defined by American Society for Testing and Materials, ASTM.

Sulfur was certified using three independent methods of analysis: gravimetry, ion chromatography, and x-ray fluorescence.

The standard error of the certified value includes observed variability within and between measurement methods and any observed material heterogeneity.

NOTICE AND RECOMMENDED USE: Due to the high sulfur content of SRM 1620a, it is recommended that the bottle be shaken vigorously before sampling. Homogeneity and stability testing at NBS indicates that the best results are achieved when the material is shaken before use.

Analyses for certification were performed by W. F. Koch and E. R. Deardorff of the Inorganic Analytical Research Division and P. A. Pella of the Gas and Particulate Science Division.

The statistical analysis of the certification data was performed by R. C. Paule of the National Measurement Laboratory.

The overall direction and coordination of the technical measurements leading to certification were performed under the chairmanship of E. L. Garner, Chief of the Inorganic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T. E. Gills.

Washington, D.C. 20234
December 22, 1981

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

SRM 1620a was also tested and found to exhibit the physical properties shown in Table 1. In addition, semi-quantitative values obtained by emission spectrometry are given in Table 2. These values are not certified, but supplied for information only.

Table 1
Physical Properties for SRM 1620a

Flash Point ^a °C	Kinematic Viscosity ^b 50 °C (cSt)	Pour Point ^c °C	Density @ 20 °C ^d g/cm ³
70	47.75	2	1.096

These measurements were performed by S. Weeks, Materials Chemistry Division, Center for Material Science.

Methods Used for Physical Tests

- a. ASTM D-93-80 Flash Point by Pensky-Martens Closed Tester
- b. ASTM D445-79 Kinematic Viscosity of Transparent and Opaque Liquids
- c. ASTM D97-66 (1978) Pour Point of Petroleum Oils
- d. ASTM D4052-81 Density and Relative Density of Liquids by Digital Density Meter (modified)

Table 2
Semi-Quantitative Emission Spectrometry
Analysis for SRM 1620a

Element	μg/mL	Element	μg, mL
Al	20	Mo	<1
B	<1	Na	31
Ca	9	Ni	<1
Cr	<1	Si	13
Cu	<1	Sn	<1
Fe	<5	Ti	<1
Mg	<1	V	<1
Mn	<1	Zn	23

Note: SRM 1620a was analyzed using the rotating disc method. This method is based on absolute amounts of sample since no internal standard is used to correct for the amount of sample actually analyzed. Differences in actual values may range from factors of 1-3.

These measurements were performed by J. A. Norris, Inorganic Analytical Research Division, Center for Analytical Chemistry.

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1621b

Sulfur in Residual Fuel Oil

Sulfur Concentration 0.950 ± 0.005 weight percent

This Standard Reference Material is intended for use as an analytical standard in the determination of total sulfur in fuel oils or materials of similar matrices. SRM 1621b is a commercial "No. 6" residual fuel oil as defined by American Society for Testing and Materials, ASTM.

Sulfur was certified using three independent methods of analysis: gravimetry, ion chromatography, and x-ray fluorescence.

The standard error of the certified value includes observed variability within and between measurement methods and any observed material heterogeneity.

NOTICE: The certification of SRM 1621b is valid for 3 years from date of purchase.

Analyses for certification were performed by W. F. Koch and E. R. Deardorff of the Inorganic Analytical Research Division and P. A. Pella of the Gas and Particulate Science Division.

The statistical analysis of the certification data was performed by R. C. Paule of the National Measurement Laboratory.

The overall direction and coordination of the technical measurements leading to certification were performed under the chairmanship of E. L. Garner, Chief of the Inorganic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T. E. Gills.

Washington, D.C. 20234
December 22, 1981

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

SRM 1621b was also tested and found to exhibit the physical properties shown in Table 1. In addition, semi-quantitative values obtained by emission spectrometry are given in Table 2. These values are not certified, but supplied for information only.

Table 1
Physical Properties for SRM 1621b

Flash Point ^a °C	Kinematic Viscosity ^b 50 °C (cSt)	Pour Point ^c °C	Density @ 20 °C ^d g/cm ³
111	89.2	11	0.929

These measurements were performed by S. Weeks, Materials Chemistry Division, Center for Material Science.

Methods Used for Physical Tests


- a. ASTM D-93-80 Flash Point by Pensky-Martens Closed Tester
- b. ASTM D445-79 Kinematic Viscosity of Transparent and Opaque Liquids
- c. ASTM D97-66 (1978) Pour Point of Petroleum Oils
- d. ASTM D4052-81 Density and Relative Density of Liquids by Digital Density Meter (modified)

Table 2
Semi-Quantitative Emission Spectrometry
Analysis for SRM 1621b

Element	µg/mL	Element	µg/mL
Al	6	Mo	<1
B	<1	Na	8
Ca	9	Ni	6
Cr	3	Si	6
Cu	<1	Sn	<1
Fe	<5	Ti	<1
Mg	<1	V	15
Mn	1	Zn	15

Note: SRM 1621b was analyzed using the rotating disc method. This method is based on absolute amounts of sample since no internal standard is used to correct for the amount of sample actually analyzed. Differences in actual values may range from factors of 1-3.

These measurements were performed by J. A. Norris, Inorganic Analytical Research Division, Center for Analytical Chemistry.



National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1622b

Sulfur in Residual Fuel Oil

Sulfur Concentration 1.982 ± 0.018 weight percent

This Standard Reference Material is intended for use as an analytical standard in the determination of total sulfur in fuel oils or materials of similar matrices. SRM 1622b is a commercial "No. 6" residual fuel oil as defined by American Society for Testing and Materials, ASTM.

Sulfur was certified using three independent methods of analysis: gravimetry, ion chromatography, and x-ray fluorescence.

The standard error of the certified value includes observed variability within and between measurement methods and any observed material heterogeneity.

NOTICE: The certification of SRM 1622b is valid for 3 years from date of purchase.

Analyses for certification were performed by W. F. Koch and E. R. Dcardorff of the Inorganic Analytical Research Division and P. A. Pella of the Gas and Particulate Science Division.

The statistical analysis of the certification data was performed by R. C. Paule of the National Measurement Laboratory.

The overall direction and coordination of the technical measurements leading to certification were performed under the chairmanship of E. L. Garner, Chief of the Inorganic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T. E. Gills.

Washington, D.C. 20234
December 22, 1981

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

SRM 1622b was also tested and found to exhibit the physical properties shown in Table 1. In addition, semi-quantitative values obtained by emission spectrometry are given in Table 2. These values are not certified, but supplied for information only.

Table 1
Physical Properties for SRM 1622b

Flash Point ^a °C	Kinematic Viscosity ^b 50 °C (cSt)	Pour Point ^c °C	Density @ 20 °C ^d g/cm ³
65	377.34	-7	0.984

These measurements were performed by S. Weeks, Materials Chemistry Division, Center for Material Science.

Methods Used for Physical Tests

- a. ASTM D-93-80 Flash Point by Pensky-Martens Closed Tester
- b. ASTM D445-79 Kinematic Viscosity of Transparent and Opaque Liquids
- c. ASTM D97-66 (1978) Pour Point of Petroleum Oils
- d. ASTM D4052-81 Density and Relative Density of Liquids by Digital Density Meter (modified)

Table 2
Semi-Quantitative Emission Spectrometry
Analysis for SRM 1622b

Element µg/mL		Element µg/mL	
Al	8	Mo	<1
B	<1	Na	25
Ca	24	Ni	15
Cr	1	Si	13
Cu	<1	Sn	<1
Fe	<5	Ti	<1
Mg	2	V	50
Mn	1	Zn	11

Note: SRM 1622b was analyzed using the rotating disc method. This method is based on absolute amounts of sample since no internal standard is used to correct for the amount of sample actually analyzed. Differences in actual values may range from factors of 1-3.

These measurements were performed J. A. Norris, Inorganic Analytical Research Division, Center for Analytical Chemistry

U. S. Department of Commerce
Malcolm Baldrige
Secretary

National Bureau of Standards
Ernest Amble, Director

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1623a

Sulfur in Residual Fuel Oil

Sulfur Concentration.0.240 ± 0.003 weight percent

This Standard Reference Material is intended for use as an analytical standard in the determination of total sulfur in fuel oils or materials of similar matrices. SRM 1623a is a commercial "No. 5 Heavy" residual fuel oil as defined by American Society for Testing and Materials, ASTM.

Sulfur was certified using three independent methods of analysis: gravimetry, ion chromatography, and x-ray fluorescence.

The standard error of the certified value includes observed variability within and between measurement methods and any observed material heterogeneity.

NOTICE: The certification of SRM 1623a is valid for 3 years from date of purchase.

Analyses for certification were performed by W. F. Koch and E. R. Deardorff of the Inorganic Analytical Research Division and P. A. Pella of the Gas and Particulate Science Division.

The statistical analysis of the certification data was performed by R. C. Paule of the National Measurement Laboratory.

The overall direction and coordination of the technical measurements leading to certification were performed under the chairmanship of E. L. Garner, Chief of the Inorganic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T. E. Gills.

Washington, D.C. 20234
December 22, 1981

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

SRM 1623a was also tested and found to exhibit the physical properties shown in Table 1. In addition, semi-quantitative values obtained by emission spectrometry are given in Table 2. These values are not certified, but supplied for information only.

Table 1
Physical Properties for SRM 1623a

Flash Point ^a °C	Kinematic Viscosity ^b 50 °C (cSt)	Pour Point ^c °C	Density @ 20 °C ^d g/cm ³
140	53.82	17	0.918

These measurements were performed by S. Weeks, Materials Chemistry Division, Center for Material Science.

Methods Used for Physical Tests

- a. ASTM D-93-80 Flash Point by Pensky-Martens Closed Tester
- b. ASTM D445-79 Kinematic Viscosity of Transparent and Opaque Liquids
- c. ASTM D97-66 (1978) Pour Point of Petroleum Oils
- d. ASTM D4052-81 Density and Relative Density of Liquids by Digital Density Meter (modified)

Table 2
Semi-Quantitative Emission Spectrometry
Analysis for SRM 1623a

Element	µg/mL	Element	µg/mL
Al	5	Mo	<1
B	<1	Na	9
Ca	9	Ni	1
Cr	1	Si	<1
Cu	<1	Sn	<1
Fe	<5	Ti	<1
Mg	<1	V	3
Mn	<1	Zn	15

Note: SRM 1623a was analyzed using the rotating disc method. This method is based on absolute amounts of sample since no internal standard is used to correct for the amount of sample actually analyzed. Differences in actual values may range from factors of 1-3.

These measurements were performed by J. A. Norris, Inorganic Analytical Research Division, Center for Analytical Chemistry.

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1624a

Sulfur in Distillate (Diesel) Fuel Oil

Sulfur Concentration 0.141 ± 0.002 weight percent

This Standard Reference Material is intended for use as an analytical standard in the determination of total sulfur in fuel oils or materials of similar matrices. SRM 1624a is a commercial "No. 2-D" distillate fuel oil as defined by American Society for Testing and Materials, ASTM.

Sulfur was certified using two independent methods of analysis: gravimetry and ion chromatography.

The standard error of the certified value includes observed variability within and between measurement methods and any observed material heterogeneity.

NOTICE: The certification of SRM 1624a is valid for 3 years from date of purchase.

Analyses for certification were performed by W. F. Koch and E. R. Deardorff of the Inorganic Analytical Research Division and P. A. Pella of the Gas and Particulate Science Division.

The statistical analysis of the certification data was performed by R. C. Paule of the National Measurement Laboratory.

The overall direction and coordination of the technical measurements leading to certification were performed under the chairmanship of E. L. Garner, Chief of the Inorganic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T. E. Gills.

Washington, D.C. 20234
December 22, 1981

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

SRM 1624a was also tested and found to exhibit the physical properties shown in Table 1. In addition, semi-quantitative values obtained by emission spectrometry are given in Table 2. These values are not certified, but supplied for information only.

Table 1
Physical Properties for SRM 1624a

Flash Point ^a °C	Kinematic Viscosity ^b 40 °C (cSt)	Cloud Point ^c °C	Density @ 20 °C ^d g/cm ³
53	2.57	-14	0.848

These measurements were performed by S. Weeks, Materials Chemistry Division, Center for Material Science.

Methods Used for Physical Tests

- a. ASTM D-93-80 Flash Point by Pensky-Martens Closed Tester
- b. ASTM D445-79 Kinematic Viscosity of Transparent and Opaque Liquids
- c. ASTM D2500-66 (1976) Cloud Point of Petroleum Oils
- d. ASTM D4052-81 Density and Relative Density of Liquids by Digital Density Meter (modified)

Table 2
Semi-Quantitative Emission Spectrometry
Analysis for SRM 1624a

Element µg/mL		Element µg/mL	
Al	1	Mo	<1
B	<1	Na	<1
Ca	7	Ni	<1
Cr	<1	Si	<1
Cu	<1	Sn	<1
Fe	<5	Ti	<1
Mg	<1	V	<1
Mn	<1	Zn	<1

Note: SRM 1624a was analyzed using the rotating disc method. This method is based on absolute amounts of sample since no internal standard is used to correct for the amount of sample actually analyzed. Differences in actual values may range from factors of 1-3.

These measurements were performed by J. A. Norris, Inorganic Analytical Research Division, Center for Analytical Chemistry.

U.S. Department of Commerce
Juanita M. Kreps
Secretary

National Bureau of Standards
Ernest Ambler, Director

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1630

Trace Mercury in Coal

This Standard Reference Material is intended as an analytical standard for the determination of trace mercury in coal. The material is a commercially available coal that was crushed to a size of 210 to 500 micrometers with a roll crusher. From a total of 500 packaged bottles, 30 were randomly selected for analysis. Duplicate determinations were made on 0.5 g portions of 25 of these bottles, and single determinations were made on the other five. The mercury content of this material was obtained by destructive neutron activation analysis.

The recommended value is the average of these 55 determinations on 30 bottles, which was found to be:

$$\text{Mercury content} = 0.13 \mu\text{g/g}$$

The recommended value is not expected to change by more than ± 1 in the last significant figure.

A study of homogeneity showed no variability among bottles that could not be accounted for by analytical error. Duplicate samples from the same bottle indicated a homogeneity for mercury of $\pm 5\%$ (relative).

The mercury content was also determined by flameless atomic absorption spectrometry, yielding an average value of $0.14 \mu\text{g/g}$.

Selenium was also determined using destructive neutron activation analysis. The value obtained, which is not certified but included for information only, was found to be $2.1 \mu\text{g/g}$.

The homogeneity testing and analyses for certification were performed in the NBS Analytical Chemistry Division by T. E. Gills and H. Rook under the direction of P. D. LaFleur.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by C. L. Stanley.

Washington, D.C. 20234
August 1, 1979
(Revision of Certificate dated 11-2-71
Editorial Revision only.)

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

ANALYTICAL PROCEDURE

The bottles containing the samples were allowed to remain open at room temperature (about 25 °C) for twenty-four hours.

The coal samples, along with solution standards of mercury and NBS Standard Reference Material 1571 (Orchard Leaves) used as a control, were encapsulated in cleaned quartz vials. The geometry of both the samples and the standards were optimized so that flux monitors were not needed. The samples were irradiated for four hours at a thermal flux of $6 \times 10^{13} \text{ n} \cdot \text{cm}^{-2} \text{ sec}^{-1}$. The samples were allowed to decay for three days to minimize the personnel dose rate. The samples were postweighed into porcelain boats and burned in a combustion tube. The volatile mercury compounds and other volatile products liberated during burning were trapped in a liquid nitrogen cold trap. The cold trap was allowed to warm to room temperature. The mercury compounds were then transferred to polyethylene bottles by washing the cold trap with concentrated nitric acid and water. For this analysis, ^{197}Hg produced by $^{196}\text{Hg}(n,\gamma)^{197}\text{Hg}$ was used as the measuring activity.

Bromine-82, an interfering isotope, was separated from the sample by using the classical silver bromide precipitation.

The samples were counted on a 22 cm³ Ge(Li) detector connected to a 2048-multichannel analyzer. The accumulated data was processed by computer for peak identification and integration. The concentrations were determined by using a Standard Comparator Method.


NOTE TO USER

It is suggested that persons using SRM 1630 to check their analytical technique should adopt the following criteria. If the average, \bar{X} , of N replicate measurements on this SRM is found to lie in the interval—

$$0.127 - \frac{0.013}{\sqrt{N}} < \bar{X} < 0.127 + \frac{0.013}{\sqrt{N}}$$

then the analytical technique used gives a result compatible with that found at NBS. However, if the value \bar{X} lies outside this interval, then the technique should be examined for possible bias or miscalibration.

NOTE: The above expression is not rigorously correct. It does not include a possible component for between laboratory variability nor sources of systematic error.



National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1632b

Trace Elements in Coal

(Bituminous)

This Standard Reference Material (SRM) is intended for use in the calibration of apparatus and the evaluation of techniques employed in the analysis of coal or similar materials. SRM 1632b is a bituminous coal with a nominal sulfur content of 1.9%. It is in the form of a fine powder (-60 mesh).

Certified Values of Constituent Elements: The certified values for the constituent elements are given in Table 1. The certified values are based on measurements using proven techniques and methods. Noncertified values are given in Table 2 and are provided for information only. These values are based on measurements made using a single technique or method. While no reason exists to suspect systematic bias in the information values, no attempt was made to determine if such a bias exists that is attributable to the technique and/or method used. A list of analytical techniques and methods used for the different analyses is given in Table 3. As part of its update certification program, NBS will periodically update many of these values to certification status.

Expiration of Certification: The certification of SRM 1632b will be valid up to 5 years from the purchase date. Should any of the certified constituents become invalid prior to that date, purchasers will be notified by NBS.

Use: This material should be vacuum dried at ambient temperature for 24 hours prior to use. The certified concentrations are reported on a "dry-weight" basis, thus the concentration determined on undried samples should be adjusted for the moisture content of the sample. Typical moisture loss using the drying procedure stated above is 1.3%.

A minimum sample size of 250 mg of the dried material is required for the certified values to be valid.

This SRM should be kept in its original bottle. It should not be exposed to intense source of radiation, including ultra-violet lamps or sunlight.

The statistical analysis of the certification data was performed by R.C. Paule of the National Measurement Laboratory.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T.E. Gills.

Gaithersburg, MD 20899
June 20, 1985

Stanley D. Rasberry, Chief
Office of Standard Reference Materials

(over)

Source and Preparation of Material: The coal for this SRM was obtained from the Humphrey No. 7 mine and coal preparation plant of the Consolidation Coal Company, Christopher Coal Company Division, Osage, West Virginia. This mine produces bituminous coal with a sulfur content of 1.8 - 1.9 percent (dry basis). This coal was obtained from an underground mine that recovers coal from the Pittsburgh seam, which is considered the single most valuable and extensive coal seam in the United States.

Approximately 900 kg of the coal for SRM 1632b was oven dried prior to processing, in accordance with procedures outlined in ASTM D2013. The coal was reduced in size to -60 mesh and sieved prior to blending. The coal was then blended in a stainless steel cone blender (approximate capacity 0.85 cubic meter). After blending the coal was packaged in polyethylene-lined aluminum cans and was subsequently repackaged in fifty gram units.

Analysis

Major, Minor, and Trace Constituents: In general, the major, minor, and trace constituents were certified using two or more independent methods of analysis or two or more different laboratories. For those constituents that were determined using a single method, technique, or laboratory, the values are given for information only. (See Table 3).

Calorific Value: The calorific value was determined using measurements made in an isoperibol calorimeter, an isothermal calorimeter, and an adiabatic calorimeter at two different laboratories.

Moisture, Ash, and Volatile Matter: The moisture, ash, and volatile matter values were determined on measurements made using the standard ASTM methods, D3173, D3174, and D3175, respectively. In addition, commercial instruments commonly used for the determination of the parameters provided additional values.

Table 1. Certified Values of Constituent Elements

<u>Major Constituents</u>		<u>Minor Constituents</u>	
<u>Elements</u>	<u>Content</u> <u>Wt. Percent</u>	<u>Elements</u>	<u>Content</u> <u>Wt. Percent</u>
Carbon (Total)	78.11 ± 0.37 ^a	Aluminum	0.855 ± 0.019
Hydrogen	5.07 ± 0.06	Calcium	0.204 ± 0.006
Nitrogen	1.56 ± 0.07	Iron	0.759 ± 0.045
Sulfur	1.89 ± 0.06	Magnesium	0.0383 ± 0.0008
Volatile matter	35.4 ± 1.1	Potassium	0.0748 ± 0.0028
		Sodium	0.0515 ± 0.0011
		Titanium	0.0454 ± 0.0017

Trace Constituents

<u>Element</u>	<u>Content</u> <u>µg/g</u>	<u>Element</u>	<u>Content</u> <u>µg/g</u>
Arsenic	3.72 ± 0.09	Manganese	12.4 ± 1.0
Barium	67.5 ± 2.1	Nickel	6.10 ± 0.27
Cadmium	0.0573 ± 0.0027	Rubidium	5.05 ± 0.11
Cobalt	2.29 ± 0.17	Selenium	1.29 ± 0.11
Copper	6.28 ± 0.30	Thorium	1.342 ± 0.036
Lead	3.67 ± 0.26	Uranium	0.436 ± 0.012
		Zinc	11.89 ± 0.78

Calorific Value^{b,c}

14005 ± 35 Btu/lb (32.57 ± 0.08 MJ kg⁻¹)

Ash, wt. %

6.79 ± 0.16

^aThe listed ± uncertainties for carbon, hydrogen, volatile matter, and calorific value are two standard deviations of the certified value. The listed ± uncertainties for all other constituents are two standard deviations for the certified values and include an allowance for minor sample heterogeneity. The observed sample variability was generally less than two percent of the constituent value.

^bThe calorific value (MJ kg⁻¹) may decrease upon aging or normal oxidation of the coals. NBS will continue to monitor this value and report any substantive change in the certified calorific value to the purchaser. The reference date for the calorific value is May 1985.

^cThe calorific value is determined as HHV2 (Higher Heating Value-Moisture Free).

Table 2. Noncertified Values for Constituent Elements

<u>Trace Constituents</u>			
<u>Element</u>	<u>Content</u> <u>µg/g</u>	<u>Element</u>	<u>Content</u> <u>µg/g</u>
Antimony	(0.24)	Lithium	(10)
Bromine	(17)	Molybdenum	(0.9)
Cerium	(9)	Samarium	(0.87)
Cesium	(0.44)	Scandium	(1.9)
Chlorine	(1260)	Silicon, wt %	(1.4)
Chromium	(11)	Strontium	(102)
Europium	(0.17)	Tungsten	(0.48)
Hafnium	(0.43)	Vanadium	(14)
Lanthanum	(5.1)		

Table 3. Analytical Techniques and Methods Used for the Characterization of SRM 1632b

Method/ Element	A	B	C	D	E	F	G	H	I	J	K	L	M
Al			•	•								•	
As			•		•								
Ash Content							• 2		•				
Ba			•										
Br			•										
C (Total)							• 5	•	•		•		
Ca		•	•	•								•	
Cal Val									•	•			
Cd	•	•											
Ce			•										
Cl			•										
Co			•	•									
Cr			•										•
Cs			•										
Cu					•								•
Eu			•										
Fe	•		•									•	
H							• 5	•	•				
Hf			•										
K		•	•	•								•	
La			•										
Li				•									
Mg	•	•	•										
Mn			•	•									
Mo			•										
N							• 6						
Na			•	•									
Ni	•												•
Pb	•	•											
Rb		•	•	•									
S						•	• 4		•			•	
Sb			•										
Sc			•										
Se			•		•								
Si			•									•	
Sm			•										
Sr			•										
Th		•	•										
Ti			•	•								•	•
U		•	•										
V			•										•
Volatile Matter							• 3		•				
W			•										
Zn	•	•	•										

Analytical Methods

- A. Atomic absorption spectrometry
- B. Isotope dilution mass spectrometry
- C. Instrumental neutron activation analysis
- D. Flame emission spectrometry
- E. Flameless atomic absorption spectrometry
- F. Ion chromatography
- G. ASTM Methods: (1)D3173, (2)D3174, (3)D3175, (4)D3177, (5)D3178, (6)D3179
- H. Combustion coulometry
- I. Commercial coal analyzers: moisture, ash, sulfur, Btu, volatile matter, carbon, hydrogen, nitrogen
- J. Commercial calorimeter
- K. Gas chromatography
- L. X-ray fluorescence
- M. Inductively coupled plasma emission spectrometry

Analysts

NBS

E.S. Beary
W.A. Bowman
K.A. Brletic
T.A. Butler
J.D. Fassett
J.W. Gramlich
R.R. Greenberg
W.F. Koch
L.A. Machlan
A. Marlow
D.M. Mo
T.J. Murphy
P.J. Paulsen
P. Pella
T.C. Rains
T.A. Rush
T.A. Sleater
S.F. Stone
R. Zeisler

Laboratories

E. Huffman
Huffman Laboratories
Wheat Ridge, Colorado 80034

R. Peck
Dickerson Laboratory
El Paso, Texas 79912

J.B. Bodkin
College Earth & Mineral Science
The Pennsylvania State University
University Park, Pa. 16802

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1633a

Trace Elements in Coal Fly Ash

This Standard Reference Material (SRM) is intended for use in the evaluation of analytical methods for the determination of constituent elements in coal fly ash or materials with a similar matrix.

SRM 1633a is a fly ash that was sieved through a No. 170 sieve with a nominal sieve opening of 90 μm .

Certified Values of Constituent Elements: The certified values for the constituent elements are shown in Table 1. The analytical techniques used and the analysts are given in Table 3. The certified values are based on results obtained by reference methods of known accuracy or from two or more independent, reliable analytical methods. Noncertified values are given for information only in Table 2.

Notice and Warnings to Users: This certification is invalid 5 years from date of purchase of the SRM. The constituents certified or analyzed are reviewed periodically and may be updated to reflect improved measurement. Updated certificates will be made available upon request.

Use: This material should be dried to a constant weight before using. Recommended procedures for drying are: (1) Vacuum drying for 24 hours at ambient temperature using a cold trap at or below -50°C and a pressure not greater than 30 Pa (0.2 mm Hg); (2) drying for 2 hours in an oven at 105°C ; (3) drying in a desiccator over P_2O_5 or Mg_2ClO_4 . Samples of the dried material weighing at least 250-mg should be used for analysis. When not in use the material should be kept in a tightly sealed bottle.

Source and Preparation of Material: The fly ash material was supplied by a coal fired power plant and is a product of Pennsylvania and West Virginia coals. It was selected as a typical fly ash and is not intended as a fly ash from a specific coal or combustion process. The material was sieved and blended for 2 hours in a Vee blender. The material was then removed and placed in a series of bulk containers from which specific samples were taken for homogeneity testing and certification analysis. Twelve bottles were selected for the homogeneity test. Samples from each bottle were analyzed for cobalt, chromium, europium, iron, scandium, and thorium using nondestructive neutron activation analysis. The observed standard deviations for both 50 and 250 mg sample sizes were consistent with counting statistics, indicating that the fly ash is homogeneous within $\pm 5\%$ (relative) based on these elements. The homogeneity testing and certification analyses were performed in the NBS Center for Analytical Chemistry.

The overall direction and coordination of the analytical measurements leading to the initial certification were performed in the Center for Analytical Chemistry under the chairmanship of L.A. Machlan.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by W.P. Reed and T.E. Gills.

Gaithersburg, MD 20899
January 5, 1985
(Revision of certificate
dated April 18, 1979)

Stanley D. Rasberry, Chief
Office of Standard Reference Materials

(over)

Table 1. Certified Values of Constituent Elements

<u>Major Constituents</u>	<u>Content Wt. Percent</u>	<u>Minor Constituents</u>	<u>Content Wt. Percent</u>
Aluminum	14.3 ± 1.0 ^a	Magnesium	0.455 ± 0.010
Iron	9.4 ± 0.1	Sodium	0.17 ± 0.01
Potassium	1.88 ± 0.06		
Silicon	22.8 ± 0.8		
Calcium	1.11 ± 0.01		

Trace Constituents

<u>Element</u>	<u>Content µg/g</u>	<u>Element</u>	<u>Content µg/g</u>
Antimony	6.8 ± 0.4	Rubidium	131 ± 2
Arsenic	145 ± 15	Selenium	10.3 ± 0.6
Cadmium	1.00 ± 0.15	Strontium	830 ± 30
Chromium	196 ± 6	Thorium	24.7 ± 0.3
Copper	118 ± 3	Thallium	5.7 ± 0.2
Manganese	179 ± 8	Uranium	10.2 ± 0.1
Mercury	0.16 ± 0.01	Vanadium	297 ± 6
Nickel	127 ± 4	Zinc	220 ± 10
Lead	72.4 ± 0.4		

^aThe uncertainties of the certified values are based on judgment and represent an evaluation of the combined effects of method imprecision, possible systematic errors among methods, and material variability for samples of 250-mg or more. (No attempt was made to derive exact statistical measures of imprecision because several methods were involved in the determination of most constituents).

Supplemental Information

Note: The following values are not certified because they are not based on the results of either a reference method or of two or more independent methods. These values are included for information only.

Table 2. Noncertified Values for Constituent Elements

<u>Element</u>	<u>Content Wt. Percent</u>	<u>Element</u>	<u>Content µg/g</u>
Barium	0.15	Beryllium	12
Titanium	0.8	Cerium	180
Sulfur	0.18	Cobalt	46
		Cesium	11
		Europium	4
		Gallium	58
		Hafnium	8
		Molybdenum	29
		Scandium	40

Table 3. Analytical Methods Used for Certified Constituent Elements

Method/ Element	A	B	C	D	E	F	G	H	I
Aluminum	.		.						.
Antimony			.				.		
Arsenic	.		.						
Cadmium			
Calcium	.	.			.				
Chromium	.	.	.						
Copper	.	.	.						
Iron	.	.	.						
Lead		.		.	.				
Magnesium	.	.							
Manganese	.		.						.
Mercury	.		.						
Nickel				
Potassium	.	.			.				
Rubidium				
Selenium	.		.				.		
Silicon					.			.	
Sodium	.		.						
Strontium	.				.	.			
Thallium		.					.		
Thorium		.	.						
Uranium		.							
Vanadium	.	.	.						
Zinc			

Analytical Methods

- A. Atomic Absorption Spectrometry or Flame Emission Spectrometry
- B. Isotope Dilution Mass Spectrometry
- C. Neutron Activation Analysis
- D. Polarography
- E. X-ray Fluorescence Spectrometry
- F. Inductively-Coupled Plasma Emission Spectrometry
- G. Isotope Dilution Spark Source Mass Spectrometry
- H. Gravimetry
- I. Direct Coupled Plasma Emission Spectrometry

Analysts

NBS Center for Analytical Chemistry

- | | |
|----------------------|------------------------|
| 1. J. B. Baldwin | 16. W. R. Kelly |
| 2. T. J. Brady | 17. H. M. Kingston |
| 3. E. R. Deardorff | 18. E. C. Kuehner |
| 4. M. G. Dias | 19. R. M. Lindstrom |
| 5. L. J. Powell | 20. L. A. Machlan |
| 6. M. S. Epstein | 21. E. J. Maienthal |
| 7. R. F. Fleming | 22. J. S. Maples |
| 8. E. L. Garner | 23. J. D. Messman |
| 9. T. E. Gills | 24. L. J. Moore |
| 10. C. A. Grabnegger | 25. P. J. Paulsen |
| 11. J. W. Gramlich | 26. P. A. Pella |
| 12. R. R. Greenberg | 27. T. C. Rains |
| 13. S. Hanamura | 28. K. J. R. Rosman |
| 14. S. H. Harrison | 29. T. A. Rush |
| 15. E. G. Heald | 30. P. A. Sleeth |
| | 31. R. L. Watters, Jr. |

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1634a

Trace Elements in Fuel Oil

This Standard Reference Material is intended for use in the evaluation of methods and the calibration of apparatus used in the analysis of fuel oils and other materials with similar matrices for trace elements. SRM 1634a is a commercial "No. 6" residual fuel oil as defined by the American Society for Testing and Materials (ASTM). This SRM was certified using two or more independent methods of analysis and a single method that has been carefully evaluated with respect to its accuracy and precision. Methods were selected to include those that are commonly used in the field and in laboratories.

The certified values are given in table 1 and are based on at least a 1.0 g sample of the material which is the minimum amount that should be used for analysis.

Table 1

Element ¹	Content ² $\mu\text{g/g}$	Element ¹	Content ² , Wt %
Lead ^a	2.80 \pm 0.08	Sulfur ^{c,g,h}	2.85 \pm 0.05
Manganese ^{b,c}	0.19 \pm 0.02		
Nickel ^{d,e}	29 \pm 1		
Selenium ^{b,c}	0.15 \pm 0.02		
Sodium ^{b,c}	87 \pm 4		
Vanadium ^{a,b,d}	56 \pm 2		
Zinc ^{b,d}	2.7 \pm 0.2		

1. Method of Analysis

- | | |
|---------------------------------------|--|
| a. Isotope Dilution Mass Spectrometry | e. Inductive Coupled Plasma Spectrometry |
| b. Neutron Activation Analysis | f. Gravimetry |
| c. Atomic Absorption Spectrometry | g. Ion Chromatography |
| d. Spark Source Mass Spectrometry | h. X-ray Fluorescence |

2. The uncertainties shown are expressed at the 95% confidence level and include any observed material heterogeneity, possible method differences, and errors of measurement.

NOTICE: The certification of SRM 1634a is valid for 3 years from date of purchase.

The statistical analysis of the certification data was performed by K.R. Eberhardt of the Statistical Engineering Division.

The overall direction and coordination of the analytical measurements leading to certification were performed in the Inorganic Analytical Research Division, E.L. Garner, Chief.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T.E. Gills.

Washington, D.C. 20234
 February 19, 1982

George A. Uriano, Chief
 Office of Standard Reference Materials

(over)

PREPARATION, TESTING, AND ANALYSIS

A random scheme for sample selection was used in assessing the homogeneity of this material. The elements calcium and vanadium were measured by x-ray fluorescence as indicators of homogeneity. Based on these elements, the material variability for this lot of 1634a is within $\pm 2\%$ relative.

Long-term stability of this SRM has not been rigorously established. When not in use, the material should be stored in the tightly sealed bottle. NBS will continue to monitor this material and any substantive change in its certification will be reported to the purchasers.

Analyses for the various elements were performed in the Center for Analytical Chemistry, Inorganic Analytical Research Division, by I. L. Barnes, T.A. Butler, E.R. Deardorff, J.W. Gramlich, S. Hanamura, H.M. Kingston, W.F. Koch, G.M. Lambert, R.M. Lindstrom, L.A. Machlan, J.R. Moody, P.J. Paulson, T.C. Rains, T.A. Rush, and R. Zeisler.

The homogeneity studies were performed in the Gas and Particulate Science Division by P.A. Pella and M. Watson.

The physical properties were measured by S. Weeks, Materials Chemistry Division, Center for Materials Science.

The values and physical properties data in table 2 are *not certified* because they are based on a non-reference method or were not determined by two or more independent methods. The values are included for information only.

Table 2

Supplemental Information

Element	Content, $\mu\text{g/g}$	Physical Properties	
Arsenic	(0.12)		
Beryllium	(0.006)	Flash Point ^a	64 °C
Bromine	(<1)		
Cadmium	(0.002)	Kinematic Viscosity ^b	321.66
Calcium	(16)	at 50 °C	
Chlorine	(31)	Pour Point ^c	-10 °C
Chromium	(0.7)		
Cobalt	(0.3)	Density at 20 °C ^d	0.995 g/cm ³
Iron	(31)		
Mercury	(<0.002)		
Molybdenum	(0.12)		

Methods Used for Physical Tests

- ASTM D-93-80 Flash Point by Pensky-Martens Closed Tester
- ASTM D445-79 Kinematic Viscosity of Transparent and Opaque Liquids
- ASTM D97-66 (1978) Pour Point of Petroleum Oils
- ASTM D4052-81 Density and Relative Density of Liquids by Digital Density Meter (modified)

National Bureau of Standards Certificate of Analysis

Standard Reference Material 1635

Trace Elements in Coal (Subbituminous)

This Standard Reference Material is intended for use in the calibration of apparatus and the evaluation of techniques employed in the trace element analysis of coal and similar materials. The material should be dried without heat to constant weight before use.

The recommended procedures for drying are either vacuum drying at ambient temperature for 24 hours, or freeze drying in which the drying chamber is kept at room temperature. The moisture content of this material is approximately 20%. Because of this moisture level, it is recommended that small individual samples be dried immediately before use. Drying of large samples may result in a violent discharge of water vapor and resultant loss of sample. When not in use, the material should be kept in a tightly sealed bottle and stored in a cool, dark place. Long-term (>1 year) stability of this SRM has not been rigorously established. NBS will continue to monitor this material and any substantive change will be reported to purchasers.

The certified values given below are based on at least a 250-mg sample of the dried material, the minimum amount that should be used for analysis.

Element ¹	Content, $\mu\text{g}/\text{g}^2$	Element ¹	Content, $\mu\text{g}/\text{g}^2$
Arsenic ^{a,b}	0.42 ± 0.15	Thorium ^{c,e}	0.62 ± 0.04
Cadmium ^{c,d,e}	0.03 ± 0.01	Uranium ^c	0.24 ± 0.02
Chromium ^{c,e}	2.5 ± 0.3	Vanadium ^{c,g}	5.2 ± 0.5
Copper ^{a,c,e}	3.6 ± 0.3	Zinc ^{c,d}	4.7 ± 0.5
Lead ^{c,d}	1.9 ± 0.2		
Manganese ^{a,e}	21.4 ± 1.5	Element ¹	Wt. % ²
Nickel ^{c,d}	1.74 ± 0.10	Iron ^{c,d,e,f}	0.239 ± 0.005
Selenium ^{a,e}	0.9 ± 0.3	Sulfur ^{f,h}	0.33 ± 0.03

1. Methods of Analysis:

- | | |
|---------------------------------------|--------------------------------|
| a. Atomic Absorption Spectrometry | e. Neutron Activation |
| b. Photon Activation | f. Spectrophotometry |
| c. Isotope Dilution Mass Spectrometry | g. Flame Emission Spectrometry |
| d. Polarography | h. Gravimetry |

2. The estimated uncertainty is based on judgment and represents an evaluation of the combined effects of method imprecision, possible systematic errors among methods, and material variability for samples of 250-mg or more. (No attempt was made to derive exact statistical measures of imprecision because several methods were involved in the determination of most constituents.)

The overall direction and coordination of the analytical measurements leading to this certificate were performed in the Analytical Chemistry Division under the chairmanship of L. J. Moore.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by W. P. Reed.

Washington, D.C. 20234
August 22, 1979
(Revision of Certificate
dated 1-23-78)

George A. Uriano
Office of Standard Reference Materials

(over)

PREPARATION, TESTING, and ANALYSIS

This material was prepared from one lot of subbituminous coal from the Eagle Mine of The Imperial Coal Company, Erie, Colorado. The material was ground and sieved thru a No. 65 (230 μm) sieve by the Colorado School of Mines Research Institute. The material was then blended in a V-type blender.

Samples for homogeneity testing were taken from the top, middle, and bottom of three bulk containers of coal, and analyzed by neutron activation analysis for sodium, scandium, chromium, iron, cobalt, lanthanum, cerium, and thorium. Replicate analyses of 250-mg samples indicated a homogeneity for these elements of $\pm 2.5\%$ (relative) except for chromium, which was homogeneous within counting statistics of $\pm 6\%$. The homogeneity measurements were performed in the NBS Analytical Chemistry Division by R. R. Greenberg. Certification analyses for the various elements were made in the NBS Analytical Chemistry Division by T. J. Brady, B. I. Diamondstone, L. P. Dunstan, M. S. Epstein, M. Gallorini, E. L. Garner, T. E. Gills, J. W. Gramlich, R. R. Greenberg, S. H. Harrison, G. M. Hyde, G. J. Lutz, L. A. Machlan, E. J. Maienthal, J. D. Messman, T. J. Murphy, and T. C. Rains.

The following values are *not certified* because they were based on a non-reference method, or were not determined by two or more independent methods. They are included for information only.

<u>Element</u>	<u>Content</u> <u>($\mu\text{g/g}$)</u>
Antimony	(0.14)
Cerium	(3.6)
Cobalt	(0.65)
Europium	(0.06)
Gallium	(1.05)
Hafnium	(0.29)
Scandium	(0.63)
	<u>(wt. %)</u>
Aluminum	(0.32)
Sodium	(0.24)
Titanium	(0.02)

U. S. Department of Commerce
Philip M. Klutznick
Secretary

National Bureau of Standards
Ernest Ambler, Director

National Bureau of Standards

Certificate of Analysis

Standard Reference Materials 1636a, 1637a, 1638a

Lead in Reference Fuel

This Standard Reference Material is intended for use in the calibration of instruments and the evaluation of techniques used for the analysis of lead in gasoline. Samples of the reference fuel are supplied at four lead concentrations, nominally 0.03, 0.05, 0.07, and 2.0 g/gal. These Standard Reference Materials are made up of various combinations of the four concentrations, see Table 1 on the back of this certificate.

Certified Values: The certified values for the lead content, expressed in units of $\mu\text{g/g}$, are shown below. These certified values are based on results obtained by isotope dilution mass spectrometry, a definitive method of known accuracy.

<u>Vial</u> <u>Identification</u>	<u>Nominal</u> <u>Lead Concentration</u> <u>g/gal</u>	<u>Certified</u> <u>Lead Concentration</u> <u>$\mu\text{g/g}$</u>
I	0.03	11.2 ± 0.2^a
II	0.05	18.8 ± 0.1
III	0.07	25.1 ± 0.2
IV	2.0	764 ± 4

^aThe uncertainties shown are the 95 percent confidence intervals for a single determination plus allowance for known sources of possible error.

Use: The certification of these materials is based on a minimum sample size of 1.0 gram and only samples equal to or greater than 1 gram should be used for any analytical determination to be related to the certified values of this certificate.

Stability: The ampoules should be stored at temperatures between 10-30 °C. They should not be exposed to intense sources of radiation, including ultraviolet lamps or sunlight. The ampoules should be opened only at time of use. No attempt should be made to keep the material in opened ampoules for future use.

Source and Preparation of Material: The reference fuel containing lead at the four concentration levels were supplied by Phillips Petroleum Company of Bartlesville, Oklahoma. The 91-octane number (Research Octane Number) reference fuel is a mixture of 91 percent by volume (0.899 mole-fraction) 2,2,4-trimethylpentane and 9 percent by volume (0.101 mole-fraction) n-heptane. Lead was added in the form of tetraethyl lead motor mix.

Analyses leading to certification were performed in the Inorganic Analytical Research Division by T. J. Murphy and I. L. Barnes.

The overall direction and coordination of the technical measurements leading to this certificate were performed by E. L. Garner, Chief of the Inorganic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of these Standard Reference Materials were coordinated through the Office of Standard Reference Materials by T. E. Gills.

Washington, D.C. 20234
February 5, 1980

George A. Uriano, Chief
Office of Standard Reference Material

(over)

Table 1
Composition of SRM's 1636a, 1637a, 1638a

SRM	Nominal Concentration	No. Units.
1636a	0.03,0.05,0.07, 2.0 g/gal	3 vials each
1637a	0.03,0.05,0.07 g/gal	4 vials each
1638a	2.0 g/gal	12 vials each

Additional Information: Because the volume of the reference material varies with temperature, the various concentrations of lead are certified by weight, i.e., micrograms of lead per gram of fuel. For convenience to the user, information is given for the concentration in the customary units, grams per gallon and grams per liter, at 23 °C. These data are shown in Table 2.

Table 2


Vial Identification	Nominal Concentration g/gal	Density ^a at 23 °C g/mL	Lead Concentration ^b at 23 °C	
			g/gal	g/L
I	0.03	0.6888	0.0292	0.0077
II	0.05	0.6888	0.0490	0.0129
III	0.07	0.6888	0.0654	0.0173
IV	2.0	0.6895	1.994	0.527

^aThe density (ρ) of each concentration was measured at 23 °C using a modification of ASTM Method D1217. The temperature coefficient of these materials is $0.0008 \text{ g} \cdot (\text{mL})^{-1} \cdot (\text{°C})^{-1}$

^bThe conversion of the certified values ($\mu\text{g/g}$) to C(g/gal) and C(g/L) was done using equations 1 and 2 respectively.

$$\text{Eq. 1} \quad C_{\text{g/gal}} = \frac{3.785 \rho C \mu\text{g/g}}{10^3}$$

$$\text{Eq. 2} \quad C_{\text{g/L}} = \frac{\rho C \mu\text{g/g}}{10^3}$$



National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1641b

Mercury in Water - $\mu\text{g/mL}$

This Standard Reference Material is intended for use in the primary calibration of instruments and techniques used for the determination of mercury in natural waters. It is designed for the preparation of calibration solutions and for use as a "spike" sample in a "method-of-additions" type analytical procedure.

Mercury concentration $1.52 \pm 0.04 \mu\text{g/mL}$

The estimated uncertainty, $0.04 \mu\text{g/mL}$, includes the effects of the observed random variability and an upper bound estimate of possible systematic errors. The random variability expressed as two standard deviations of the certified value is $\pm 0.02 \mu\text{g/mL}$ and reflects both internal and between-method variabilities for the NBS atomic absorption and neutron activation measurements. The upper bound estimate of possible systematic errors is $\pm 0.02 \mu\text{g/mL}$.

Stability: The long-term stability of trace mercury solutions has been a constant problem. At or below the $\mu\text{g/mL}$ level, mineral acid stabilization is not sufficient. However, the addition of trace gold to a nitric acid solution of mercury was found to stabilize the concentration of mercury in the two previous issues of this Mercury in Water SRM. Although the mercury concentration of SRM 1641b has not changed significantly in eight months, the stability will continue to be monitored. However, SRM 1641b should *not* be used after ONE YEAR FROM date of purchase.

Precautions: Traces of mercury vapor are present in most laboratory situations. Therefore, contamination of reagents, equipment, and common laboratory materials may cause a severe problem. Apparatus for analysis at this level must be scrupulously cleaned immediately before use, and only the purest-grade reagents, with respect to mercury, should be used.

SRM 1641b was prepared by J.R. Moody. Atomic absorption analyses were performed by T.C. Rains and T.A. Butler; and neutron activation analyses by R. Zeisler, Inorganic Analytical Research Division.

The overall direction and coordination of technical measurements leading to certification were performed under the chairmanship of E.L. Garner, Inorganic Analytical Research Division. The statistical evaluation was done by R.C. Paule.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Analytical: Two independent techniques were used in the certification of this Standard Reference Material: atomic absorption spectrometry and instrumental neutron activation analysis.

Use: This SRM consists of six ampoules, each containing approximately 20 mL of solution. Dilutions may be made by the addition of accurately measured aliquots, withdrawn from an ampoule, to known volumes of pure or natural water (spiking mode) using conventional techniques. Blank determinations should be made of the water and other reagents used.

The reliability of the dilution process will depend on the care exercised and the reliability of the calibration of the volumetric apparatus, which should have an uncertainty no greater than one percent. The volumetric apparatus should be scrupulously cleaned. Diluted solutions should be used without delay, as their stability cannot be guaranteed. SRM 1642b, which is certified for mercury at the ng/mL level, should be used to validate methodology for these concentrations. The long-term retention of unused portions of this Standard Reference Material in opened ampoules is not recommended.

Washington, D.C. 20234
April 13, 1983

George A. Uriano, Chief
Office of Standard Reference Materials

National Bureau of Standards

Certificate

Standard Reference Material 1642b

Mercury in Water – ng/mL

This Standard Reference Material is intended for use in the primary standardization of instruments and techniques used for the determination of mercury in water. It is intended for use as received, without dilution or other alteration. The concentration of mercury in this Standard Reference Material is at, or near, the detection limit of most commercial instruments used for the determination of mercury in water. It is to be used for the primary standardization of these instruments near these detection limits where many analytical problems occur.

Mercury Concentration 1.49 ± 0.06 ng/mL

The estimated uncertainty, 0.06, includes the effects of the observed random variability and an upper bound estimate of possible systematic errors. The random variability expressed as two standard deviations of the certified value is ± 0.04 and reflects both internal and between-method variabilities for the NBS atomic absorption and neutron activation measurements. The upper bound estimate of possible systematic errors is ± 0.02 .

Stability: Trace mercury solutions have been a constant problem when long-term storage is required. Below the $\mu\text{g/mL}$ level, mineral acid stabilization is not sufficient. A stabilizing technique has been applied to this Standard Reference Material that allows for prolonged storage. Gold, as the tetrachloride, has been added in a concentration 10 times that of the mercury. The gold ion, in conjunction with the normal mineral acid, has proven to be an effective stabilizer. It is recommended that this Standard Reference Material not be used after ONE YEAR FROM DATE OF PURCHASE.

Precautions: Traces of mercury vapor are present in most laboratory situations. Therefore, contamination of reagents, equipment, and common laboratory materials is a severe problem. Apparatus for analyses at this level must be scrupulously cleaned immediately before use, and only the purest-grade reagents should be employed. After use, the bottle should be capped tightly and placed inside the aluminized bag, which should be folded and sealed with a sealing tape. This safeguard will assist in maintaining the integrity of the sample.

Analytical: Two independent techniques were used in the certification of this Standard Reference Material: atomic absorption spectroscopy and neutron activation analysis.

Use: This Standard Reference Material should be used, as received, without dilution. It may be carried through the chemical manipulations required for the analytical procedure normally used for the analysis of natural waters.


This Standard Reference Material was prepared by J.R. Moody. Atomic absorption analyses were performed by I.C. Rains and T.A. Butler and neutron-activation analyses were performed by R. Zeisler.

The overall direction and coordination of the technical measurements leading to the certification were performed under the chairmanship of E.L. Garner. The statistical evaluation was done by R.C. Paule.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Washington, D.C. 20234
July 15, 1982

George A. Uriano, Chief
Office of Standard Reference Materials



National Bureau of Standards

Certificate

Standard Reference Material 1643b

Trace Elements in Water

This Standard Reference Material (SRM) is intended primarily for use in evaluating the accuracy of trace element determinations in filtered and acidified fresh water and for calibrating instrumentation used in these determinations. SRM 1643b consists of approximately 950 mL of water in a polyethylene bottle, which is sealed in an aluminized bag to maintain stability. SRM 1643b simulates the elemental composition of fresh water. Nitric acid is present at a concentration of 0.5 moles per liter to stabilize the trace elements.

Concentrations of Constituent Elements: The concentrations of the trace elements that were determined are shown in Table 1. The certified values are based on results obtained either by reference methods of known accuracy or by two or more independent, reliable analytical methods. Noncertified values, which are given for information only, appear in parentheses.

Notice and Warnings to Users:

Expiration of Certification: This certification is invalid two years after the shipping date.

Precautions: The bottle should be shaken before use because of possible water vapor condensation. To prevent possible contamination of the SRM, do not insert pipets into the bottle. After use, the bottle should be capped tightly and placed inside the aluminized bag, which should be folded and sealed with sealing tape. This safeguard will protect the SRM from possible environmental contamination and long-term loss of water.

Elemental determinations of ng/g levels are limited by contamination. Apparatus should be scrupulously cleaned and only the purest grade reagents employed. Sampling and manipulations, such as evaporations, should be done in a clean environment, for example, a Class 100 clean hood.

The overall direction and coordination of the technical measurements leading to this certification were performed under the direction of E. Garner, Chief of the Inorganic Analytical Research Division.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Washington, DC 20234
May 18, 1984

Stanley D. Rasberry, Chief
Office of Standard Reference Materials

(over)

(Table 1)
Concentrations of Constituent Elements

Element	Concentration,* ng/g	Element	Concentration,* ng/g
Arsenic ^{1,5}	(49)**	Lead ^{3,4b}	23.7 ± 0.7
Barium ^{2a,2b,5}	44 ± 2	Manganese ^{1,2a,3}	28 ± 2
Beryllium ^{1,2a}	19 ± 2	Molybdenum ^{2a,5}	85 ± 3
Bismuth ¹	(11)	Nickel ^{2a,3}	49 ± 3
Boron ^{2a}	(94)	Selenium ^{1,5}	9.7 ± 0.5
Cadmium ^{2b,3,5}	20 ± 1	Silver ^{1,5}	9.8 ± 0.8
Chromium ^{4b}	18.6 ± 0.4	Strontium ^{2a,5}	227 ± 6
Cobalt ^{1,5}	26 ± 1	Thallium ^{4b}	8.0 ± 0.2
Copper ^{3,4b}	21.9 ± 0.4	Vanadium ^{4b}	45.2 ± 0.4
Iron ^{2a,4a,5}	99 ± 8	Zinc ^{2a,5}	66 ± 2

* The estimated uncertainty is based on judgment and represents an evaluation of the combined effects of method imprecision and possible systematic errors among methods. To convert to nanograms per milliliter, multiply by the density of the SRM. The density at 23 °C is 1.017 grams per milliliter.

** Values in parentheses are not certified.

- | | |
|---|---|
| 1. Atomic absorption spectrometry, electrothermal | 4. Isotopic dilution mass spectrometry, |
| 2. Atomic emission spectrometry, | a. resonance ionization |
| a. dc plasma | b. thermal ionization |
| b. flame | 5. Neutron activation, |
| 3. Laser enhanced ionization flame spectrometry | instrumental |

Source and Preparation of Material: SRM 1643b was prepared at the U.S. Geological Survey, National Water Quality Laboratory, Arvada, Colorado, under the direction of V.J. Janzer of that laboratory and J.R. Moody of the NBS Center for Analytical Chemistry. Only high-purity reagents were used and the containers were acid-cleaned and sterilized before use. In the preparation, a polyethylene cylindrical tank was filled with distilled water and sufficient nitric acid to make the solution approximately 0.5 moles HNO₃ per liter. Solutions containing known amounts of calcium, sodium, magnesium, potassium, and the elements to be determined were added to the acidified water solution with constant stirring. After thoroughly mixing, the solution was filtered, sterilized, and then transferred to one-liter polyethylene bottles. The approximate concentrations, in µg/mL, of Ca, Na, Mg, and K are respectively 35, 8, 15, and 3.

Analysts:

Center for Analytical Chemistry, National Bureau of Standards

- | | |
|--------------------|------------------------|
| 1. K. A. Brletic | 10. J. R. Moody |
| 2. T. A. Butler | 11. L. J. Powell |
| 3. E. C. Deal | 12. T. C. Rains |
| 4. M. S. Epstein | 13. T. A. Rush |
| 5. J. D. Fassett | 14. S. F. Stone |
| 6. K. Fitzpatrick | 15. G. C. Turk |
| 7. H. M. Kingston | 16. R. L. Watters, Jr. |
| 8. R. M. Lindstrom | 17. R. Zeisler |
| 9. L. A. Machlan | |

National Bureau of Standards

Certificate

Standard Reference Material 1644

Generator Columns for Polynuclear Aromatic Hydrocarbons

SRM 1644 is intended to provide accurate concentrations of anthracene, benzo(a)anthracene (1,2-benzanthracene), and benzo(a)pyrene (3,4-benzopyrene) in water. The SRM consists of three 50 cm x 0.6 cm (coiled) stainless steel tubes, each packed with fine quintus quartz (sea sand) coated with approximately 0.5 percent by weight of the polynuclear aromatic hydrocarbon (PAH) of interest.

Principle of Operation: A saturated aqueous solution of the PAH of interest is generated by flowing high-purity water slowly through the column. Because the aqueous solubility of a compound is a well-defined thermodynamic quantity, a saturated solution has a fixed concentration (1, 2, 3).

Equilibration and Use of Generator Columns: To equilibrate a new column before initial use, purge with high-purity water, such as commercial HPLC grade water. The volume required for equilibration of each column is: 500 mL for anthracene, 1000 mL for benzo(a)anthracene, and 500 mL for benzo(a)pyrene. After equilibration, pump the high-purity water at a constant temperature (± 0.1 °C) through the column at a flow rate between 0.1 and 5 mL/min to produce a saturated solution. Record the temperature. The solution should be used immediately after generation to avoid sorption losses.

If either the temperature is changed by as much as 1 °C or the flow is interrupted for less than one hour, pump 25 mL water through the column under the new conditions to restore equilibrium prior to sample collection. However, if the flow is interrupted for more than one hour, pump 50 mL water prior to sample collection. During periods of frequent use, a column can be kept equilibrated by maintaining a steady, but low, flow rate of approximately 0.1 mL/min through the column. The flow rate can be increased to collect a large volume of sample and then decreased again. Columns should be purged with 10 liters of oil-free nitrogen prior to storage periods of more than one month.

Certified Concentrations: When used as directed, these columns generate saturated solutions. The concentrations of the compounds in these solutions at temperatures between 10 and 30 °C were determined by two independent analytical methods. The data obtained were combined by fitting an empirical expression of the form $\ln[\text{Conc}] = A + B(1/T) + C(1/T^2)$ by least squares. In this equation, $\ln[\text{Conc}]$ is the natural logarithm of the concentration, T is the absolute temperature, and A, B, and C are constants for each compound. The derived equations for anthracene, benzo(a)anthracene, and benzo(a)pyrene were used to calculate the certified concentrations at one degree intervals between 10 and 30 °C. These certified concentrations are given in Tables I-III.

Service Life of Columns: Generator columns for anthracene, benzo(a)anthracene, and benzo(a)pyrene are certified for either two years or for total aqueous purge volumes of 7.5×10^2 , 3×10^3 , and 1.5×10^4 liters, respectively, whichever comes first.

Consultation on the statistical design of the experimental work and statistical analysis of the data was provided by K. R. Eberhardt of the Statistical Engineering Division. Coordination of the technical measurements leading to certification was performed by W. E. May, R. A. Velapoldi, and H. S. Hertz. Technical measurements leading to the development and certification of SRM 1644 were performed by the following members of the Center for Analytical Chemistry: W. E. May, J. M. Brown-Thomas, W. J. Sonnefeld, R. A. Velapoldi, and P. A. White.

The technical and support aspects concerning the certification and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Washington, D.C. 20234
April 27, 1981

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

The anthracene generator column, after equilibration, produces aqueous solutions that contain very small amounts of phenanthrene. Because of its relatively high solubility (1000 $\mu\text{g}/\text{kg}$), residual amounts of phenanthrene are depleted from the column at a rapid rate.

The benzo(a)anthracene generator column, after equilibration, produces aqueous solutions that contain as much as 0.3 $\mu\text{g}/\text{kg}$ anthracene in addition to the certified concentrations of benzo(a)anthracene. The concentration of anthracene in the generator column effluent varies with time (volume) and, therefore, is not certified.

The benzo(a)pyrene generator column, after equilibration, produces aqueous solutions that contain as much as 0.1 $\mu\text{g}/\text{kg}$ benzo(a)anthracene and 1 $\mu\text{g}/\text{kg}$ chrysene in addition to the certified concentrations of benzo(a)pyrene. The concentration of these non-analyte components vary with time and, therefore, are not certified.

Analyses of the Saturated Aqueous Solutions: The concentrations of the saturated aqueous solutions of anthracene, benzo(a)anthracene, and benzo(a)pyrene in the effluent from the respective generator columns were determined by two independent analytical techniques. The first technique was high-performance liquid chromatography (HPLC). It involved quantitative extraction of the PAH of interest from the aqueous effluent by an "extractor column" packed with an octadecylsilane (C_{18}) bonded phase; use of an acetonitrile-water eluant to transfer components from the extractor column to an analytical C_{18} column for separation of the analyte from non-analyte components; and detection of the analyte by measuring its absorbance at 254 nm. The second technique involved the use of a "standard addition" spectrofluorimetric technique for "on stream" analysis. The aqueous effluent from the generator column was mixed with PAH standards dissolved in acetonitrile and the PAH concentration of the resultant mixture was determined by fluorescence at the following excitation (λ_{ex}) and emission (λ_{em}) wavelengths: anthracene, $\lambda_{\text{ex}} = 254 \text{ nm}$, $\lambda_{\text{em}} = 384$ and 404 nm ; benzo(a)anthracene, $\lambda_{\text{ex}} = 290 \text{ nm}$, $\lambda_{\text{em}} = 395 \text{ nm}$; benzo(a)pyrene, $\lambda_{\text{ex}} = 296 \text{ nm}$, $\lambda_{\text{em}} = 414 \text{ nm}$. The PAH effluent concentration was determined mathematically. Where necessary, corrections for inner filter effects or the emission-absorbance contributions by impurities were made to obtain the values for determining the certified PAH concentrations listed in Tables I-III.

References:

1. May, W. E., The Solubility Behavior of Polycyclic Aromatic Hydrocarbons in Aqueous Systems, American Chemical Society Advances in Chemistry Series 185 (7), 143-192 (1980).
2. May, W. E., Wasik, S. P., and Freeman, D. H., Determination of the Aqueous Solubility of Polynuclear Aromatic Hydrocarbons by a Coupled-Column Liquid Chromatographic Technique, Anal. Chem. 50, 1 (1978).
3. Schwarz, F. P. and Miller, J. M., Determination of the Aqueous Solubilities of Organic Liquids at 10°C, 20°C, and 30°C by Elution Chromatography, Anal. Chem. 52, 2162-2164 (1980).

Table 1. Certified Aqueous Concentrations of Anthracene and Their Uncertainties, in Micrograms/Kilogram and Nanomoles/Liter, as a Function of Temperature

Temperature °C	Concentration and Its Uncertainty ¹	
	µg/kg	nmol/L
10	16.6 ± 0.7	93.1 ± 4.0
11	17.6 ± 0.6	98.7 ± 3.6
12	18.7 ± 0.6	105 ± 3.3
13	19.8 ± 0.5	111 ± 3.1
14	21.1 ± 0.5	118 ± 2.9
15	22.4 ± 0.5	126 ± 2.9
16	23.8 ± 0.5	134 ± 2.9
17	25.4 ± 0.5	142 ± 2.9
18	27.0 ± 0.5	151 ± 3.0
19	28.8 ± 0.5	161 ± 3.1
20	30.7 ± 0.6	172 ± 3.2
21	32.8 ± 0.6	184 ± 3.2
22	35.0 ± 0.6	196 ± 3.2
23	37.4 ± 0.6	210 ± 3.2
24	39.9 ± 0.6	224 ± 3.3
25	42.7 ± 0.6	239 ± 3.7
26	45.7 ± 0.8	256 ± 4.4
27	48.9 ± 1.0	273 ± 5.5
28	52.4 ± 1.3	293 ± 7.2
29	56.1 ± 1.7	313 ± 9.5
30	60.1 ± 2.2	336 ± 12

¹The uncertainties are 99 percent (Working-Hotelling) confidence bands for the entire regression curve. The difference between the true and certified concentrations should be less than the stated uncertainty at the 99 percent confidence level.

Table II. Certified Aqueous Concentrations of Benzo(a)anthracene and Their Uncertainties, in Micrograms/Kilogram and Nanomoles/Liter, as a Function of Temperature

Temperature °C	Concentration and Its Uncertainty ¹	
	µg/kg	nmol/L
10	3.38 ± 1.2	14.8 ± 5.3
11	3.60 ± 1.1	15.8 ± 4.6
12	3.83 ± 0.91	16.8 ± 4.0
13	4.09 ± 0.79	17.9 ± 3.4
14	4.36 ± 0.68	19.1 ± 3.0
15	4.65 ± 0.59	20.4 ± 2.6
16	4.96 ± 0.54	21.7 ± 2.4
17	5.29 ± 0.55	23.2 ± 2.4
18	5.65 ± 0.60	24.8 ± 2.6
19	6.04 ± 0.68	26.4 ± 3.0
20	6.45 ± 0.77	28.2 ± 3.4
21	6.90 ± 0.87	30.2 ± 3.8
22	7.38 ± 0.94	32.3 ± 4.1
23	7.90 ± 1.0	34.5 ± 4.4
24	8.45 ± 1.0	36.9 ± 4.5
25	9.05 ± 1.0	39.5 ± 4.6
26	9.69 ± 1.0	42.3 ± 4.6
27	10.4 ± 1.1	45.4 ± 4.7
28	11.1 ± 1.2	48.4 ± 5.0
29	11.9 ± 1.3	51.9 ± 5.9
30	12.8 ± 1.7	55.8 ± 7.3

¹The uncertainties are 99 percent (Working-Hotelling) confidence bands for the entire regression curve. The difference between the true and certified concentrations should be less than the stated uncertainty at the 99 percent confidence level.

Table III. Certified Aqueous Concentrations of Benzo(a)pyrene and Their Uncertainties, in Micrograms/Kilogram and Nanomoles/Liter, as a Function of Temperature

Temperature °C	Concentration and Its Uncertainty ¹	
	µg/kg	nmol/L
10	0.59 ± 0.06	2.34 ± 0.23
11	0.63 ± 0.05	2.50 ± 0.20
12	0.67 ± 0.04	2.65 ± 0.17
13	0.71 ± 0.04	2.81 ± 0.15
14	0.76 ± 0.03	3.01 ± 0.12
15	0.81 ± 0.03	3.21 ± 0.11
16	0.87 ± 0.03	3.44 ± 0.10
17	0.93 ± 0.03	3.68 ± 0.10
18	0.99 ± 0.03	3.92 ± 0.11
19	1.06 ± 0.03	4.19 ± 0.12
20	1.13 ± 0.03	4.47 ± 0.13
21	1.21 ± 0.04	4.79 ± 0.15
22	1.30 ± 0.04	5.14 ± 0.16
23	1.39 ± 0.04	5.50 ± 0.16
24	1.49 ± 0.04	5.89 ± 0.17
25	1.59 ± 0.04	6.28 ± 0.17
26	1.71 ± 0.04	6.76 ± 0.18
27	1.83 ± 0.05	7.23 ± 0.19
28	1.96 ± 0.05	7.74 ± 0.22
29	2.11 ± 0.07	8.33 ± 0.27
30	2.26 ± 0.09	8.92 ± 0.35

¹The uncertainties are 99 percent (Working-Hotelling) confidence bands for the entire regression curve. The difference between the true and certified concentrations should be less than the stated uncertainty at the 99 percent confidence level.

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1645

River Sediment

This Standard Reference Material (SRM) is intended for use for the calibration of apparatus and the verification of methods used in the analysis of river sediments and material with a similar matrix.

Certified Values of Constituent Elements: The certified values for the constituent elements are shown in Table 1. The analytical techniques used and the names and affiliations of the analysts are shown in Table 3. Certified values are based on results obtained by reference methods of known accuracy and analyses performed by two or more analysts; or alternatively, from results obtained by two or more independent, reliable analytical methods. Noncertified values are given for information only in Table 2. All values are based on measurements made on a dried sample of at least 100 mg for all constituents except iron and chromium for which a 1-g sample was used.

Notice and Warnings to Users:

Expiration of Certification: This certification is invalid 5 years from the date of purchase.

Stability: This material has been freeze-dried and is essentially free of moisture. However, its stability has not been rigorously assessed. NBS will continue to monitor this material and if substantive changes in certification occur the purchasers will be notified. The material should be kept in its original bottle and stored at temperatures between 10-30 °C. The material should be dried without heat to a constant weight before using. Recommended procedures for drying are: (1) drying for 24 hours using a cold trap at or below -50 °C and a pressure not greater than 30 Pa (0.2 mm Hg); (2) drying for 24 hours in a desiccator over P₂O₅ or Mg (ClO₄)₂.

Use: Material of this kind is intrinsically heterogeneous. Consequently, the analyst should endeavor to minimize any segregation by thoroughly mixing the contents of the bottle by shaking and/or rolling before each use. In addition, when taking a portion for analysis, the analyst should strive to remove as representative a sample as possible.

Source and Preparation of Material: The material for this SRM was prepared from material dredged from the bottom of the Indiana Harbor Canal near Gary, Indiana. The material was screened to remove foreign objects, freeze-dried, and sieved to pass a No. 80 (180 μm) screen. The material was thoroughly mixed in a V-blender and bottled. The bulk material was radiation-sterilized to minimize alteration due to biological activity.

The collection, freeze-drying and homogenization of this SRM were performed under the supervision and direction of H.L. Rook, Gas and Particulate Science Division.

The overall direction and coordination of the technical measurements leading to certification were performed under the chairmanship of J.K. Taylor, Center for Analytical Chemistry.

The technical and support aspects involved in the preparation, current and previous certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T.E. Gills and W.P. Reed.

Washington, D.C. 20234
May 5, 1982
(Revision of Certificate
dated 11-16-78)

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

Homogeneity Assessment and Certification: The homogeneity of this material was established using a minimum sample size of 100 milligrams for all constituents except iron and chromium for which the sample size was 1.0 gram.

Randomly selected bottles were used for the analytical measurements. Each analyst examined at least 6 different bottles, some of them measuring replicate samples from each bottle. Accordingly, it is believed that all bottles of this SRM have substantially the same composition. Measurements and calibrations were made to reduce random and systematic errors to no more than one percent, relative.

Table 1. Certified Values of Constituent Elements

<u>Major Constituents</u>		<u>Minor Constituents</u>	
<u>Element</u>	<u>Content</u> <u>wt. percent^a</u>	<u>Element</u>	<u>Content</u> <u>wt. percent^a</u>
Aluminum ^b	2.26 ± 0.04	Magnesium ^b	0.74 ± 0.02
Chromium	2.96 ± 0.28	Sodium ^b	0.54 ± 0.01
Iron	11.3 ± 1.2	Zinc	0.172 ± 0.017
Potassium ^b	1.26 ± 0.05		

<u>Trace Constituents</u>			
<u>Element</u>	<u>Content</u> <u>µg/g^a</u>	<u>Element</u>	<u>Content</u> <u>µg/g^a</u>
Cadmium	10.2 ± 1.5	Nickel	45.8 ± 2.9
Copper	109 ± 19	Thallium	1.44 ± 0.07
Cobalt ^b	10.1 ± 0.6	Thorium	1.62 ± 0.22
Lead	714 ± 28	Uranium	1.11 ± 0.05
Manganese	785 ± 97	Vanadium	23.5 ± 6.9
Mercury	1.1 ± 0.5		

^aThe uncertainties of the certified values for the elements, except those noted by superscript "b," include those errors associated with both measurement and material variability. They represent the 95 percent tolerance limits for an individual sub-sample, i.e., 95 percent of the sub-samples from a unit of this SRM would be expected to have a composition within the indicated range of values 95 percent of the time.

^bThese elements are certified as a part of the NBS update certification program. For each element a "best value" is given based on all methods of measurement that were used as well as a *standard error* of this value. Both are based on considerations of variability both within and between analytical methods.

Supplemental Information

Note: The following values are not certified because they are not based on the results of either a reference method or of two or more independent methods. These values are included for information only.

Table 2. Noncertified Values for Constituent Elements

<u>Element</u>	<u>Content wt. Percent</u>	<u>Element</u>	<u>Content µg/g</u>
Calcium	(2.9)	Antimony	(51)
Fluorine	(0.09)	Arsenic	(66)
Sulfur	(1.1)	Lanthanum	(9)
		Scandium	(2)
		Selenium	(1.5)

Additional Information: The values listed below are based on measurements made in one laboratory and while no reason exists to suspect systematic bias in these numbers, no attempt was made to evaluate such bias attributable to either the method or the laboratory. The method used for each set of measurements is also listed. The indicated uncertainties are two times the standard deviation of the mean. These values are included for information only.

Table 3

<u>Constituent</u>	<u>Content wt. percent</u>
Kjeldahl Nitrogen	(0.0797% ± 0.0048)
Total Phosphorus	(0.051% ± 0.001) [1]
Loss on Ignition (800 °C)	(10.72% ± 0.28)
Oil and Grease (Freon)	(1.71% ± 0.26) [3]
Chemical Oxygen Demand (Dichromate)	(149,400 mg/kg ± 9,000)[2]

References

1. ASTM Method E-350
2. Standards Methods for the Examination of Water and Waste Water, 14th Edition (1975), Section 508, pp 550.
3. Ibid., Section 502, pp 518.

Table 3A Methods and Analysts

Method/ Element	A	B	C	D	E	F
Aluminum	•		•		•	
Arsenic			•			
Antimony	•					
Cadmium			•	•		
Calcium			•			
Chromium		•	•			
Cobalt			•		•	
Copper		•	•			
Fluorine						•
Iron	•		•			
Lanthanum			•			
Lead		•		•		
Magnesium	•				•	
Manganese		•	•			
Mercury	•		•			
Nickel		•		•		
Potassium	•		•			
Scandium			•			
Selenium	•					
Sodium	•		•			
Sulfur						•
Thallium		•				
Thorium		•				
Uranium		•				
Vanadium	•		•			
Zinc	•			•		

Analytical Methods

- A. Atomic Absorption Spectrometry
- B. Isotope Dilution Mass Spectrometry
- C. Neutron Activation Analysis
- D. Polarography
- E. D. C. Plasmas Atomic Emission Spectrometry
- F. Ion Chromatography

Analysts

NBS Center for Analytical Chemistry

- | | |
|--------------------|---------------------|
| 1. T. J. Brady | 10. R. R. Greenberg |
| 2. E. R. Deardorff | 11. S. H. Harrison |
| 3. L. P. Powell | 12. G. J. Lutz |
| 4. M. S. Epstein | 13. L. A. Machlan |
| 5. R. Filby | 14. E. J. Maienthal |
| 6. M. Gallorini | 15. T. C. Rains |
| 7. E. L. Garner | 16. H. L. Rook |
| 8. T. E. Gills | 17. T. A. Rush |
| 9. J. W. Gramlich | 18. W. P. Schmidt |

Cooperators

19. L. Kosta (Nuclear Chemistry Section, Josef Stefan Institute, Ljubljana, Yugoslavia)

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1646

Estuarine Sediment

This Standard Reference Material is intended primarily for calibrating instrumentation and evaluating the reliability of analytical methods for the determination of major, minor, and trace elements in sediments, and similar matrices.

Values of Constituent Elements: The *certified* values for the constituent elements are shown in Table 1. They are based on results obtained either by definitive methods or by two or more independent, reliable analytical methods. *Non-certified values*, which are given for information only, appear in Table 2. All values are based on a minimum sample size of 500 mg of the material dried as indicated under "Instructions for Drying".

Notice to Users:

Expiration of Certification: The certification of this SRM will be invalid 5 years after date of shipping.

Use: The material should be kept in its original bottle and shaken well before each use. A minimum sample of 500 mg of the dried material (see Instructions for Drying) should be used for any analytical determination to be related to a certified value of this certificate.

Statistical consultation was provided by K. R. Eberhardt of the Statistical Engineering Division.

The overall direction and coordination of the technical measurements leading to certification were performed in the Inorganic Analytical Research Division, E. L. Garner, Chief.

The technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by R. Alvarez.

Washington, D.C. 20234
June 7, 1982
(Revision of Certificate
dated 1-6-82)

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

Table 1. Certified Concentration of Constituent Elements

Element	Concentration, weight %	Element	Concentration, weight %
Aluminum ^{2b,c;6}	6.25 ± 0.20	Magnesium ^{1c;2c}	1.09 ± 0.08
Calcium ^{2b,c;6}	0.83 ± 0.03	Phosphorus ^{2a;6}	0.054 ± 0.005
Iron ^{2c;4a;6}	3.35 ± 0.10		
Element	Concentration, µg/g	Element	Concentration, µg/g
Arsenic ^{1d;4b}	11.6 ± 1.3	Manganese ^{1c;2c}	375 ± 20
Cadmium ^{1b;3a,b;4b}	0.36 ± 0.07	Mercury ^{1a;4b}	0.063 ± 0.012
Chromium ^{1c;3b;4a}	76 ± 3	Nickel ^{1b;2c;5}	32 ± 3
Cobalt ^{1b;4a}	10.5 ± 1.3	Vanadium ^{2a;3a}	94 ± 1
Copper ^{1c;2c;4b}	18 ± 3	Zinc ^{1b,c;2c;3b;5}	138 ± 6
Lead ^{1b;3a;5}	28.2 ± 1.8		

1. Atomic absorption spectrometry
 - a. cold vapor
 - b. graphite furnace
 - c. flame
 - d. hydride generation
2. Atomic emission spectrometry
 - a. dc plasma
 - b. flame
 - c. inductively coupled plasma

3. Isotope dilution mass spectrometry
 - a. thermal ionization
 - b. spark source
4. Neutron activation
 - a. instrumental
 - b. radiochemical
5. Polarography
6. X-ray fluorescence spectrometry

Notes: (1.) Analytical values are based on the "dry-weight" of material (see Instructions for Drying). Mercury should be determined on samples without drying and the results adjusted to a "dry-weight" basis by determining the moisture content of separate samples.

(2.) The estimated uncertainty for an element is based on judgment and represents an evaluation of the combined effects of method imprecision, possible systematic errors among methods, and material variability for samples 500 mg or more.

Table 2. Non-certified Concentrations of Constituent Elements

Note: The values shown in this table are not certified because they are not based on the results of either a definitive method or two or more independent analytical methods. These values are included, for information only, to provide additional information on the composition.

<u>Element</u>	<u>Concentration, Weight %</u>	<u>Element</u>	<u>Concentration, Weight %</u>
Potassium	(1.4)	Sulfur	(0.96)
Silicon	(31)	Titanium	(0.51)
Sodium	(2.0)		

<u>Element</u>	<u>Concentration, µg/g</u>	<u>Element</u>	<u>Concentration, µg/g</u>
Antimony	(0.4)	Molybdenum	(2.0)
Beryllium	(1.5)	Rubidium	(87)
Cerium	(80)	Scandium	(10.8)
Cesium	(3.7)	Selenium	(0.6)
Europium	(1.5)	Tellurium	(0.5)
Germanium	(1.4)	Thallium	(0.5)
Lithium	(49)	Thorium	(10)

Analysts:

Inorganic Analytical Research Division, National Bureau of Standards. I. L. Barnes, M. B. Blackburn, C. G. Blundell, T. A. Butler, M. S. Epstein, T. E. Gills, J. W. Gramlich, R. R. Greenberg, S. Hanamura, W. R. Kelly, H. M. Kingston, L. Machlan, E. J. Maienthal, J. D. Messman, T. J. Murphy, T. C. Rains, T. A. Rush, R. Sedivy, and R. L. Waters, Jr.

Cooperating Analysts:

University of Tokyo, Tokyo, Japan; present address: Meteorological Research Institute; Tsukuba, Ibaraki, Japan; Y. Dokiya (NBS Guest Worker).

Division of Chemistry, National Research Council of Canada, Ottawa, Canada; S. Berman, A. Desaulniers, R. Sturgeon, A. Mykytuik, J. McLaren, V. Boyko, and P. Semeniuk.

Instructions for Drying: Except for mercury, elements should be determined on samples that have been dried at 110°C for 2 hours.

Mercury should be determined on undried samples. However, because the certified concentration is reported on a "dry-weight" basis, the concentration determined on undried samples should be adjusted for the moisture content of the samples.

Source and Preparation of Material: The material for this SRM was supplied by R. Huggett, Virginia Institute of Marine Sciences, Gloucester Point, Va. It had been dredged from the Chesapeake Bay at a location: 37° 11.1' N, 76° 17.1' W. The material was freeze-dried at Eastern Freeze-Dry Corporation, Lancaster, Pa., and radiation sterilized at Neutron Products Inc., Dickerson, Md. At NBS, the sediment was sieved through a screen with openings of 1.00 mm (No. 18) to remove coarse contaminants; ball-milled to pass a sieve with openings of 150 µm (No. 100); thoroughly mixed in a V-blender, placed in polyethylene bags, and bottled.

Homogeneity Assessment: A preliminary evaluation of homogeneity was made by instrumental neutron activation using samples of approximately 250 mg taken from various locations of the bulk materials. The samples were irradiated and the activities from radionuclides of Ce, Co, Cr, Cs, Eu, Fe, Rb, Sc and Th were counted. Except for Ce and Th, the observed sample-to-sample variations for the elements were approximately the same as the counting statistics indicating satisfactory homogeneity for these elements within approximately 2%. The homogeneity of the material for As, Cd, Hg, N, and Zn was evaluated by various analytical techniques using samples weighing 250 to 300 mg and found to be satisfactory. The homogeneity of the remaining certified elements was determined using sample weights not exceeding one gram.

The uncertainties of the elemental concentrations in Table 1 take into account possible material inhomogeneity for samples weighing 500 mg.

National Bureau of Standards

Certificate of Analysis

Standard Reference Material 1648

Urban Particulate Matter

This Standard Reference Material (SRM) is intended for use in the calibration of apparatus and evaluation of methods used in the analysis of atmospheric particulate matter and materials with a similar matrix.

Certified Values of Constituent Elements: The certified values for the constituent elements are shown in Table 1. The analytical techniques used in the certification are shown in Table 3. The certified values are based on measurements of 6 to 30 samples by each of the analytical techniques indicated. Noncertified values are given for information only in Table 2.

Notice and Warnings to Users: This material is a naturally occurring urban dust to be used for analytical purposes only. It may contain a number of chemicals of unknown toxicities, therefore, the utmost caution and care must be exercised in its use.

Expiration of Certification: This certification is invalid after 5 years from date of purchase. Should it be shown to be invalid prior to that time, users will be notified by NBS.

Stability: This material should be kept in its original bottle and stored at temperatures between 10-30 °C. It should not be exposed to intense source of radiation, including ultraviolet lamps or sunlight. Ideally, the bottle should be kept in a desiccator at the recommended temperature.

Use: A minimum of 100 mg of the dried material (See Drying Instructions) should be used for any analytical determination to be related to the certified values of this certificate.

Source and Preparation of Material: This SRM was prepared from urban particulate matter collected in the St. Louis, Missouri, area in a baghouse specially designed for this purpose. The material was removed from the filter bags and combined in a single lot. This product was screened through a fine-mesh sieve to remove extraneous materials and thoroughly blended in a V-blender. The material was then bottled and sequentially numbered. The material was collected over a period in excess of 12 months and, therefore, is a time-integrated sample. While not represented to be typical of the area in which it was collected, its use should typify the analytical problems of atmospheric samples obtained from industrialized urban areas.

Homogeneity Assessment: Randomly selected bottles were used for the analytical measurements. Each analyst examined at least 6 bottles, some of them measuring replicates from each bottle. No correlation was found between measured values and the bottling sequence. Also, the results of measurements of samples from different bottles were not significantly different than the measurements of replicate samples from single bottles. Accordingly, all bottles of this SRM have been assigned the same certified values of constituent elements.

Instructions for Drying: This material should be dried at 105 °C for 8 hours before use. Because the certified concentrations are reported on a "dry-weight" basis, the concentrations determined on undried samples should be adjusted for the moisture content of the samples.

The overall direction and coordination of the technical measurements leading to certification were performed under the chairmanship of J.K. Taylor.

The technical and support aspects involved in the preparation, current and previous certification, and issuance of this Standard Reference Material were coordinated through the Office of Standard Reference Materials by T.E. Gills and W.P. Reed.

Washington, D.C. 20234
May 11, 1982
(Revision of Certificate
dated 11-16-78)

George A. Uriano, Chief
Office of Standard Reference Materials

(over)

Table 1. Certified Values of Constituent Elements

<u>Major Constituents</u>		<u>Minor Constituents</u>	
<u>Element</u>	<u>Content^a</u> <u>Wt. Percent</u>	<u>Element</u>	<u>Content^a</u> <u>Wt. Percent</u>
Aluminum ^b	3.42 ± 0.11	Lead	0.655 ± 0.008
Iron	3.91 ± 0.10	Sodium ^b	0.425 ± 0.002
Potassium ^b	1.05 ± 0.01	Zinc	0.476 ± 0.014

Trace Constituents

<u>Element</u>	<u>Content^a</u> <u>µg/g</u>	<u>Element</u>	<u>Content^a</u> <u>µg/g</u>
Arsenic	115 ± 10	Nickel	82 ± 3
Cadmium	75 ± 7	Selenium ^b	27 ± 1
Chromium	403 ± 12	Uranium	5.5 ± 0.1
Copper	609 ± 27	Vanadium ^b	140 ± 3

^aThe uncertainties shown for the elements except those noted by superscripts include errors associated with both measurement and material variability. They represent the 95 percent tolerance limits for individual subsamples, i.e., 95 percent of the subsamples from a single unit of this SRM would be expected to have a composition within the indicated range of values 95 percent of the time.

^bThese elements were recently certified as a part of the NBS update certification program. The values for the indicated constituent are the "best value" based on all measurement methods used and the associated uncertainty is expressed as the standard error considering variability within and between analytical methods.

Table 2. Noncertified Values for Constituent Elements

Note: The following values are not certified because they are not based on the results of either a reference method or two or more independent methods. These values are included for information only.

<u>Major Constituents</u>		<u>Minor Constituents</u>	
<u>Element</u>	<u>Content</u> <u>Wt. Percent</u>	<u>Element</u>	<u>Content</u> <u>Wt. Percent</u>
Sulfur	(5.0)	Chlorine	(0.45)
Magnesium	(0.8)	Titanium	(0.40)

<u>Trace Constituents</u>			
<u>Element</u>	<u>Content</u> <u>µg/g</u>	<u>Element</u>	<u>Content</u> <u>µg/g</u>
Antimony	(45)	Lanthanum	(42)
Barium	(737)	Rubidium	(52)
Bromine	(500)	Manganese	(860)
Cerium	(55)	Samarium	(4.4)
Cesium	(3)	Scandium	(7)
Cobalt	(18)	Silver	(6)
Europium	(0.8)	Thorium	(7.4)
Hafnium	(4.4)	Tungsten	(4.8)
Indium	(1.0)		
Iodine	(20)		

Supplemental Information

The values listed below are based on measurements made in a single laboratory and are given for information only. While no reason exists to suspect systematic bias in these numbers, no attempt was made to evaluate such bias attributable to either the method or the laboratory. The method used for each set of measurements is also listed. The uncertainties indicated are two times the standard deviation of the means.

<u>Constituent</u>	<u>Content Wt. Percent</u>
Nitrogen (NO ₃)	(1.07 ± 0.06)
Nitrogen (NH ₄)	(2.01 ± 0.08)
Sulfate	(15.42 ± 0.14)
SiO ₂	(26.8 ± 0.38)
Freon Soluble	(1.19 ± 0.47)

Methods Used:

Nitrate - Extraction with water and measurement by ASTM Method D992

Ammonia - NaOH addition followed by steam distillation and titration

Sulfate - Extraction with water and measurement by ASTM D516

SiO₂ - Solution and measurement by ASTM Method E350

Freon Soluble - Extraction with Freon 113, using the Method described in "Standard Methods in Examination of Water and Waste Water," 14th Edition, p. 518, American Public Health Association, Washington, D.C.

Analysts

Inorganic Analytical Chemistry Division

- | | |
|-----------------------|---------------------|
| 1. C. G. Blundell | 9. J. W. Gramlich |
| 2. R. W. Burke | 10. R. R. Greenberg |
| 3. T. A. Butler | 11. L. A. Machlan |
| 4. E. R. Deardorff | 12. E. J. Maienthal |
| 5. B. I. Diamondstone | 13. T. J. Murphy |
| 6. M. S. Epstein | 14. L. P. Powell |
| 7. M. Gallorini | 15. T. C. Rains |
| 8. E. L. Garner | 16. T. A. Rush |

Table 3 Methods and Analysis

Method/ Element	A	B	C	D	E	F	G	H	I
Ag			•						
Al			•					•	
As			•		•				
Ba			•						
Br			•						
Cd	•	•	•	•					
Ce			•						
Cl							•		
Co			•						
Cr		•	•						
Cs			•						
Cu	•	•			•				
Eu			•						
Fe	•	•	•		•				
Hf			•						
I			•			•			
In			•						
K	•		•						
La			•						
Mg			•						
Mn	•		•						•
Na	•		•						•
Ni	•	•		•					
Pb	•	•		•					
Rb			•						
S							•		
Sb			•						
Sc			•						
Se	•		•						•
Sm			•						
Th			•						
Ti			•						
U		•							
V	•		•						
W			•						
Zn	•	•	•	•					

Analytical Methods

- A. Atomic Absorption Spectrometry
- B. Isotope Dilution Mass Spectrometry
- C. Neutron Activation Analysis
- D. Polarography
- E. Spectrophotometry
- F. Photon Activation Analysis
- G. Ion Chromatography
- H. D.C. Plasma Atomic Emission Spectrometry
- I. Flame Emission Spectrometry

Guide for Requesting Development of Standard Reference Materials

The National Bureau of Standards develops Standard Reference Materials (SRM's) to provide a basis for comparison of measurements on materials and to aid in the control of production processes. The Office of Standard Reference Materials evaluates the requirements of science, industry, and government for carefully characterized reference materials, then directs the production and distribution of these materials.

NBS currently has over 1000 SRM's available, about 100 new ones in preparation, and requests for the development of many more. The demand for new SRM's greatly exceeds the Bureau's capacity to produce and certify these materials. Consequently, requests for new SRM's of limited use are deferred in favor of those that serve a substantial area of interest. In determining which requests receive top priority, NBS relies heavily upon information supplied by industry and interested organizations.

The Bureau welcomes all requests for SRM's. Both the Bureau and potential users would be helped if these requests included as much of the information below as possible.

1. Short title of the proposed Standard Reference Material.
2. Purpose for which this SRM is intended.
3. Reason why the SRM would be useful.
4. Special characteristics and/or requirements of the material. Include necessary additional information, if more than one SRM is needed for standardization in an area.
5. An estimate of the possible present and future (6-10 years) demand for such an SRM in your operations and elsewhere. (National and international estimates are very useful).

6. Facts about whether such an SRM (or a similar one) could be produced by, or obtained from a source other than NBS. If so, justify preparation by NBS.

7. Other pertinent information to justify the SRM, such as: (a) an estimate of the range of application, economic significance of the measurement affected, and scientific and/or technological significance, including estimates of the impact upon industrial productivity or growth, and (b) supporting letters from industry leaders, trade organizations, interested committees, and others.

In developing an NBS-SRM, the candidate material must meet one or more of the criteria listed below:

1. The SRM must permit users to attain more accurate measurements.
2. The production of the SRM must not be economically or technically feasible elsewhere.
3. The SRM must serve as an industry-wide standard for commerce, provided by a unique neutral source.
4. NBS production of the SRM would provide readily available, highly characterized material useful to science, industry, or government.

NBS has recognized the need to enlarge the scope of the SRM program to include all types of well characterized materials that can be used to calibrate a measurement system, or to produce scientific data that can be widely used. Input from science, industry, and government assists NBS in continuing to provide Standard Reference Materials that will be valuable in many areas.

GUIDE TO ORDERING STANDARD REFERENCE MATERIALS

ORDERING

Orders should be addressed to:
Office of Standard Reference Materials
Room B311, Chemistry Building
National Bureau of Standards
Gaithersburg, MD 20899
Telephone: (301) 921-2045

Orders should give number of units, catalog number, and name of the material requested. For example, 1 each, No. 11h, Basic-Open-Hearth Steel, 0.2 percent C. The materials described in this Catalog are distributed only in the units listed or in multiples thereof.

Acceptance of an order does not imply acceptance of any provision set forth in the order contrary to the policy, practice, or regulations of the National Bureau of Standards or the U.S. Government.

Orders received for "out-of-stock" materials are cancelled if only out-of-stock items are ordered. On other orders, shipment is made of available materials and out-of-stock items are cancelled. Back-orders are not accepted for out-of-stock materials; if a renewal lot of material is available, it will be furnished automatically.

TERMS

Prices quoted are in U.S. dollars, and are published in the SRM Price List. When SRM Price Lists are issued they are sent to persons or organizations who have requested them. These prices are subject to revision without notice and orders will be billed for the prices in effect at the time of shipment. No discounts are given on purchases of SRM's, RM's, or GM's.

Remittances of the purchase price need not accompany purchase orders. Payment of invoices is expected within 30 days of receipt of an invoice. Payment on foreign orders may be made by any of the following:

- a. banker's draft against U.S.A. bank
- b. bank to bank transfer to U.S.A. bank
- c. cash against documents
- d. sight draft
- e. International Money Order
- f. UNESCO coupons

Letters of Credit cannot be accepted. If a Letter of Credit or any method of payment other than those listed above is to be used, you must secure the services of an agent in the United States to act in your behalf. Your agent would purchase the material and our invoice would indicate that he is the purchaser. The material would be shipped to your agent, who would transship in accordance with your instructions.

NBS cannot "prepay and add" shipping charges to the invoice. Restricted categories such as hydrocarbons, organic sulfur compounds, compressed gasses, rubber compounding materials, radioactive standards, and similar materials are shipped FOB Gaithersburg, MD.

LATE CHARGES

Unless otherwise notified, payment for SRM's is due within 30 days of shipment of the order to the customer. For non-Federal customers, the U.S. Treasury regulations require late charges, based on the current value of funds to Treasury, be assessed for each 30-day period or portion thereof that the payment is overdue.

PROFORMA INVOICE (PRICE QUOTATION)

Proforma invoice service will frequently require three to four weeks to process, and will be furnished only to those requiring such service.

DOMESTIC SHIPMENTS

Shipments of material (except for certain restricted categories) intended for the United States and Canada are normally shipped prepaid (providing that the parcel does not exceed the weight limitations as prescribed by postal laws and regulations).

FOREIGN SHIPMENTS

The regulations of various nations covering the importation of SRM's, GM's, and RM's differ widely; any attempt to list all possible variations would be impractical. Therefore, where the shipping practices outlined below do not apply, purchasers will be informed of the best method of shipment for their countries.

Most orders will be shipped by prepaid International Air Parcel Post. Exceptions are items in restricted categories and those shipments that exceed parcel post weight limitations. These exceptions will be shipped FOB Gaithersburg, MD, unless an agent (shipping or brokerage firm) located in the United States is required. Where an agent is required, the purchaser will be so notified and will be requested to designate an agent of his/her choice. In this case, the material will be packaged for overseas shipment and will be forwarded to the agent FOB Gaithersburg, MD.

DOCUMENTATION

Listed below are the only documents that we will furnish. All documents are printed in English.

- a. six commercial invoices
- b. two sight drafts
- c. two packing slips
- d. customs invoices for Canada, New Zealand, Australia, and South Africa
- e. Certificate of Origin
- f. parcel post receipts for parcel post shipments
- g. air waybill for air shipments

If documents other than those listed above are required, the services of an agent in the United States will be needed to purchase and ship the materials.

Note: Orders and inquiries submitted in English will be processed more rapidly than those requiring translations.