

Laboratory Safety Centrifuges


Centrifuges, which operate at high speed, have great potential for injuring users if not operated properly. Unbalanced centrifuge rotors can result in injury or death. Sample container breakage can release aerosols that are harmful if inhaled.

The majority of all centrifuge accidents result from user error. To avoid injury, workers should follow the manufacturer's operating instructions for each make and model of centrifuge that they use.

Follow these steps for the safe operation of centrifuges:

- Ensure that centrifuge bowls and tubes are dry.
- Ensure that the spindle is clean.
- Use matched sets of tubes, buckets and other equipment.
- Always use safety centrifuge cups to contain potential spills and prevent aerosols.
- Inspect tubes or containers for cracks or flaws before using them.
- Avoid overfilling tubes or other containers (e.g., in fixed angle rotors, centrifugal force may drive the solution up the side of the tube or container wall).
- Ensure that the rotor is properly seated on the drive shaft.
- Make sure that tubes or containers are properly balanced in the rotor.
- Only check O-rings on the rotor if you are properly trained.
- Apply vacuum grease in accord with the manufacturer's guidelines.
- Do not exceed the rotor's maximum run speed.
- Close the centrifuge lid during operation.

continued on page 2



The majority of all centrifuge accidents result from user error.

For assistance, contact us. We can help. It's confidential.



**Occupational Safety
and Health Administration**
www.osha.gov 1-800-321-6742

Laboratory Safety Centrifuges

continued from page 1

- Make sure that the centrifuge is operating normally before leaving the area.
- Make sure that the rotor has come to a complete stop before opening the lid.

When centrifuging infectious materials, wait 10 minutes after the rotor comes to a complete stop before opening the lid. If a spill occurs, use appropriate decontamination and cleanup procedures for the spilled materials. Report all accidents to your supervisor immediately.

When centrifuging infectious materials, wait 10 minutes after the rotor comes to a complete stop before opening the lid.

For assistance, contact us. We can help. It's confidential.



**Occupational Safety
and Health Administration**
www.osha.gov 1-800-321-6742