

INFECTION CONTROL SURVEYOR WORKSHEET

(Rev. 68 Issued: 11-24-10, Effective: 11-24-10, Implementation: 11-24-10)

Name of State Agency or AO (please print at right)

Instructions: The following is a list of items that must be assessed during the on-site survey, in order to determine compliance with the infection control Condition for Coverage. Items are to be assessed primarily by surveyor observation, with interviews used to provide additional confirming evidence of observations. In some cases information gained from interviews may provide sufficient evidence to support a deficiency citation.

The interviews and observations should be performed with the most appropriate staff person(s) for the items of interest (e.g., the staff person responsible for sterilization should answer the sterilization questions).

A minimum of one surgical procedure must be observed during the site visit, unless the ASC is a low volume ASC with no procedures scheduled during the site visit. The surveyor(s) must identify at least one patient and follow that case from registration to discharge to observe pertinent practices. For facilities that perform brief procedures, e.g., colonoscopies, it is preferable to follow at least two cases.

When performing interviews and observations, any single instance of a breach in infection control would constitute a breach for that practice.

Citation instructions are provided throughout this instrument, indicating the applicable regulatory provision to be cited on the Form CMS-2567 when deficient practices are observed.

PART 1 - ASC CHARACTERISTICS

1. ASC Name (please print)
2. Address, State and Zip Code (please print)
3. 10-digit CMS Certification Number
4. What year did the ASC open for operation?
5. Please list date(s) of site visit:
6. What was the date of the most recent previous federal (CMS) survey:

PLEASE COMPLETELY FILL IN EACH BUBBLE USING A DARK PEN.

7. Does the ASC participate in Medicare via accredited "deemed" status? YES NO

- 7a. If YES, by which CMS-recognized accreditation organization? (Check only ONE):
- Accreditation Association for Ambulatory Health Care (AAAHC)
 - American Associate for Accred. of Ambulatory Surgery Facilities (AAAASF)
 - American Osteopathic Association (AOA)
 - The Joint Commission (TJC)

7b. If YES, according to the ASC, what was the date of the most recent accreditation

/ /
 m m d d y y y y

8. What is the ownership of the facility?
- Physician-owned
 - Hospital-owned
 - National corporation (including joint ventures with physicians)
 - Other (*please print*):

9. What is the primary procedure performed at the ASC (i.e., what procedure type reflects the majority of procedures performed at the ASC)? (**Fill in only ONE bubble**)

- Dental
- Endoscopy
- Ear/Nose/Throat
- OB/Gyn
- Ophthalmologic
- Orthopedic
- Pain
- Plastic/reconstructive
- Podiatry
- Other (*please print*):

10. What additional procedures are performed at the ASC? (**Fill in all that apply**)
Do not include the procedure type indicated in question 9.

- Dental
- Endoscopy
- Ear/Nose/Throat
- OB/Gyn
- Ophthalmologic
- Orthopedic
- Pain
- Plastic/reconstructive
- Podiatry
- Other (*please print*):

11. Who does the ASC perform procedures on? (**Fill in only ONE bubble**)
- Pediatric patients only
 - Adult patients only
 - Both pediatric and adult patients

12. What is the average number of procedures performed at the ASC per month?

per month

13. How many Operating Rooms (including procedure rooms) does the ASC have? Number actively maintained:
- | | | | | | | | | | |
|--|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|-----------------------|
| | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> | <input type="radio"/> |
| | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9+ |

14. Please indicate how the following services are provided: (*fill in all that apply*)

	Contract	Employee	Other	If Other, Please <i>print</i> :
Anesthesia	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	

Environmental Cleaning	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
Linen	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
Nursing	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
Pharmacy	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
Sterilization/Reprocessing	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	
Waste Management	<input type="radio"/>	<input type="radio"/>	<input type="radio"/>	

INFECTION CONTROL PROGRAM

15. Does the ASC have an explicit infection control program? YES NO

NOTE! If the ASC does not have an explicit infection control program, a condition-level deficiency related to 42 CFR 416.51 **must** be cited.

16. Does the ASC’s infection control program follow nationally recognized infection control guidelines? YES NO

NOTE! If the ASC does not follow nationally recognized infection control guidelines, a deficiency related to 42 CFR 416.51(b) **must** be cited. Depending on the scope of the lack of compliance with national guidelines, a condition-level citation may also be appropriate.

16a. Is there documentation that the ASC considered and selected nationally-recognized infection control guidelines for its program? YES NO

- 16b. Which nationally-recognized infection control guidelines has the ASC selected for its program? **(Fill in all that apply)**
- CDC/HICPAC Guidelines:
 - Guideline for Isolation Precautions (CDC/HICPAC)
 - Hand hygiene (CDC/HICPAC)
 - Disinfection and Sterilization in Healthcare Facilities (CDC/HICPAC)
 - Environmental Infection Control in Healthcare Facilities (CDC/HICPAC)
 - Perioperative Standards and Recommended Practices (AORN)
 - Guidelines issued by a specialty surgical society / organization (List)

Please specify **(please print and limit to the space provided)**:

Others

(Please specify (please print and limit to the space provided):

NOTE! If the ASC cannot document that it considered and selected specific guidelines for use in its infection control program, a deficiency related to 42 CFR 416.51(b) **must** be cited. This is the case even if the ASC's infection control practices comply with generally accepted standards of practice/national guidelines. If the ASC neither selected any nationally recognized guidelines nor complies with generally accepted infection control standards of practice, then the ASC should be cited for a condition-level deficiency related to 42 CFR 416.51.

17. Does the ASC have a licensed health care professional qualified through training in infection control and designated to direct the ASC's infection control program? YES NO

NOTE! If the ASC cannot document that it has designated a qualified professional with training (not necessarily certification) in infection control to direct its infection control program, a deficiency related to 42 CFR 416.51(b)(1) **must** be cited. Lack of a designated professional responsible for infection control should be considered for citation of a condition-level deficiency related to 42 CFR 416.51.

17a. If YES, Is this person an: ASC employee
(Fill in only ONE bubble) ASC contractor

17b. Is this person certified in infection control (i.e., CIC) (Note: §416.50(b)(1) does **not** require that the individual be certified in infection control.) YES NO

17c. If this person is **NOT** certified in infection control, what type of infection control training has this person received?

17d. On average, how many hours per week does this person spend in the ASC directing the infection control program? *hours per week*

(Note: §416.51(b)(1) does **not** specify the amount of time the person must spend in the ASC directing the infection control program, but it is expected that the designated individual spends sufficient time on-site directing the program, taking into consideration the size of the ASC and the volume of its surgical activity.)

18. Does the ASC have a system to actively identify infections that may have been related to procedures performed at the ASC? YES NO

18a. If YES, how does the ASC obtain this information? The ASC sends e-mails to patients after discharge
(Fill in ALL that apply) The ASC follows-up with their patients' primary care providers after discharge
 The ASC relies on the physician performing the procedure to obtain this information at a follow-up visit after discharge, and report it to the ASC
 Other (please *print*):

18b. Is there supporting documentation confirming this tracking activity? YES NO

NOTE! If the ASC does not have an identification system, a deficiency related to 42 CFR 416.44(a)(3) and 42 CFR 416.51(b)(3) **must** be cited.

18c. Does the ASC have a policy/procedure in place to comply with State notifiable disease reporting requirements? YES
 NO

NOTE! If the ASC does not have a reporting system, a deficiency **must** be cited related to 42 CFR 416.44(a)(3). CMS does not specify the means for reporting; generally this would be done by the State health agency.

19. Do staff members receive infection control training? YES
 NO

19a. If YES, how do they receive infection control training? In-service
 Computer-based training
 Other (*please print*):

(Fill in all that apply)

19b. Which staff members receive infection control training? Medical staff
 Nursing staff
 Other staff providing direct patient care
 Staff responsible for on-site sterilization/high-level disinfection
 Cleaning staff
 Other (*please print*):

(Fill in all that apply)

19c. Is training: the same for all categories of staff
 different for different categories of staff

19d. Indicate frequency of staff infection control training Upon hire
 Annually
 Periodically / as needed
 Other (please print):

(Fill in all that apply)

19e. Is there documentation confirming that training is provided to all categories of staff listed above? YES
 NO

NOTE! If training is not provided to appropriate staff upon hire/granting of privileges, with some refresher training thereafter, a deficiency **must** be cited in relation to 42 CFR 416.51(b) and (b)(3). If training is completely absent, then consideration should be given to condition-level citation in relation to 42 CFR 416.51, particularly when the ASC's practices fail to comply with infection control standards of practice.

20. How many procedures were observed during the site visit? 1 2 3 4 Other

If other, please *print* the number:

--	--

procedures

PART 2 – INFECTION CONTROL & RELATED PRACTICES

INSTRUCTIONS:

- Please **completely fill in ONE bubble** for each “Was Practice Performed?” and “Manner of Confirmation” question, unless otherwise noted.
- Please use a dark pen to fully fill in each bubble.
- Unless otherwise indicated, a “No” response to any question below **must** be cited as a deficient practice in relation to 42 CFR 416.51(a).
- If N/A is response, please explain why there is no associated observation, or why the question is not applicable, *in the COMMENTS box at the end of each section.*

I. Hand Hygiene

Observations are to focus on staff directly involved in patient care (e.g., physicians, nurses, CRNAs, etc.).

Hand hygiene should be observed not only during the case being followed, but also while making other observations in the ASC throughout the survey. Interviews are used primarily to provide additional evidence for what the surveyor has observed, but may in some cases substitute for direct observation to support a citation of deficient practice.

Practices to be Assessed	Was Practice Performed?	Manner of Confirmation
A. All patient care areas have:		
Note: 42 CFR 416.51(a) should be cited only if the answer to both a and b is “No.”		
a. Soap and water available	<input type="radio"/> Yes <input type="radio"/> No	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
b. Alcohol-based hand rubs available	<input type="radio"/> Yes <input type="radio"/> No	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
I. If alcohol-based hand rub is available in patient care areas, it is installed as required. (There are LSC requirements at 42 CFR 416.44(b)(5) for installation of alcohol-based hand rubs)	<input type="radio"/> Yes <input type="radio"/> No	
B. Staff perform hand hygiene:		
a. After removing gloves	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
b. After direct patient contact	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
c. Before performing invasive procedures (e.g., placing an IV)	<input type="radio"/> Yes <input type="radio"/> No	<input type="radio"/> Observation <input type="radio"/> Interview

N/A Both

Practices to be Assessed	Was Practice Performed?	Manner of Confirmation
d. After contact with blood, body fluids, or contaminated surfaces (even if gloves are worn)	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
C. Regarding gloves, staff:		
a. Wear gloves for procedures that might involve contact with blood or body fluids	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
b. Wear gloves when handling potentially contaminated patient equipment	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
c. Remove gloves before moving to the next tasks and/or patient	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
D. Additional breaches in hand hygiene, not captured by the questions above, were identified (If YES, please specify further in comments)	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both

Comments: <i>(please print and limit comments to the space provided)</i>	
---	--

II. Injection Practices (injectable medications, saline, other infusates)
Observations are to be made of staff who prepare and administer medications and perform injections (e.g., anesthesiologists, certified registered nurse anesthetists, nurses).

Practices to be Assessed	Was Practice Performed?	Manner of Confirmation
A. Needles are used for only one patient	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
B. Syringes are used for only one patient	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both

Practices to be Assessed	Was Practice Performed?	Manner of Confirmation
C. Medication vials are always entered with a new needle	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
D. Medication vials are always entered with a new syringe	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
E. Medications that are pre-drawn are labeled with the time of draw, initials of the person drawing, medication name, strength and expiration date or time	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
<p>Note: A “No” answer should result in citation as a deficient practice in relation to 42 CFR 416.48(a), Administration of Drugs</p>		
F. a. Single dose (single-use) medication vials are used for only one patient (A “No” response must be cited in relation to 42 CFR 416.48(a).)	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
b. Manufactured prefilled syringes are used for only one patient	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
c. Bags of IV solutions are used for only one patient	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
d. Medication administration tubing and connectors are used for only one patient	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both

G. *Please print* all injectable medications/infusates that are in a vial/container used for **more than one patient**:

Name of Medication	Average number of patients per vial/container

Practices to be Assessed	Was Practice Performed?	Manner of Confirmation
H. Multi-dose injectable medications are used for only one patient	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both

(Note: a “No” answer here is not necessarily a breach in infection control and does not result in a citation. However, a “No” response to the related questions I - K should be cited).

(Fill in N/A if no multi-dose medications/infusates are used).

If YES, please skip to “L”

If NO, please answer “I-K”:

I. The rubber septum on a multi-dose vial used for more than one patient is disinfected with alcohol prior to each entry	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
J. Multi-dose medications used for more than one patient are dated when they are first opened and are discarded within 28 days of opening or according to manufacturer’s recommendations, whichever comes first	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
K. Multi-dose medications, used for more than one patient , are not stored or accessed in the immediate areas where direct patient contact occurs	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
L. All sharps are disposed of in a puncture-resistant sharps container	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
M. Sharps containers are replaced when the fill line is reached	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
N. Additional breaches in injection practices, not captured by the questions above were identified (If YES, please specify further in comments)	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both

Comments:
(please print and limit comments to the space provided)

III. Single Use Devices, Sterilization, and High Level Disinfection

Pre-cleaning must always be performed prior to sterilization and high-level disinfection

Sterilization must be performed for critical equipment (i.e., instruments and equipment that enter normally sterile tissue or the vascular system, such as surgical instruments)

High-level disinfection must be performed for semi-critical equipment (i.e., items that come into contact with non-intact skin or mucous membranes such as reusable flexible endoscopes, laryngoscope blades)

Observations are to be made of staff who perform equipment reprocessing (e.g., surgical techs), unless these activities are performed under contract or arrangement off-site from the ASC.

SINGLE-USE DEVICES

(Choose N/A if single-use devices are never reprocessed and used again)

(Surveyor to confirm there is a contract or other documentation of an arrangement with a reprocessing facility by viewing it)

Practices to be Assessed	Was Practice Performed?	Manner of Confirmation
A. a. If single-use devices are reprocessed, they are devices that are approved by the FDA for reprocessing	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
b. If single-use devices are reprocessed, they are reprocessed by an FDA-approved reprocessor.	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both

STERILIZATION

A. Critical equipment is sterilized	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
B. Are sterilization procedures performed on-site? (If NO, skip to “F”)	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both

(A “No” answer does not result in a citation, since ASCs are permitted to provide for sterilization off-site, under a contractual arrangement.)

(Surveyor to confirm there is a contract or other documentation of an arrangement for off-site sterilization by viewing it)

a. If YES to B , please indicate method of sterilization:	<input type="radio"/> Steam autoclave <input type="radio"/> Peracetic acid <input type="radio"/> Other <i>(please</i>	
--	---	--

C. Items are pre-cleaned according to manufacturer’s instructions or evidence-based guidelines prior to sterilization	<input type="radio"/> Yes <input type="radio"/> No	<input type="radio"/> Observation <input type="radio"/> Interview
---	---	--

N/A Both

Practices to be Assessed	Was Practice Performed?	Manner of Confirmation
D. a. Medical devices and instruments are visually inspected for residual soil and re-cleaned as needed before packaging and sterilization	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
b. A chemical indicator is placed in each load	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
c. A biologic indicator is performed at least weekly and with all implantable loads	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
d. Each load is monitored with mechanical indicators (e.g. time, temperature, pressure)	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
e. Documentation for each piece of sterilization equipment is maintained and up to date and includes results from each load	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
E. Items are appropriately contained and handled during the sterilization process to assure that sterility is not compromised prior to use	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
F. After sterilization, medical devices and instruments are stored in a designated clean area so that sterility is not compromised	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
G. Sterile packages are inspected for integrity and compromised packages are reprocessed	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
H. Additional breaches in sterilization practices not captured by the questions above were identified (If YES, please specify further in comments)	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both

Comments:
(please print and limit comments to the space provided)

HIGH-LEVEL DISINFECTION

Practices to be Assessed	Was Practice Performed?	Manner of Confirmation
A. Semi-critical equipment is high-level disinfected or sterilized	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
B. Is high-level disinfection performed on site? (If NO, Skip to "F")	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
(A "No" answer does not result in a citation, since ASCs are permitted to provide for high-level disinfection off-site, under a contractual arrangement.)		
(Surveyor to confirm there is a contract or other documentation of an arrangement for off-site sterilization by viewing it)		
a. <i>If answer to B was YES, please indicate method of high-level disinfection:</i>	<input type="radio"/> Manual <input type="radio"/> Automated <input type="radio"/> Other <i>(please print):</i>	
C. Items are pre-cleaned according to manufacturer's instructions or evidence-based guidelines prior to high-level disinfection	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
D. a. Medical devices and instruments are visually inspected for residual soil and re-cleaned as needed before high-level disinfection	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
b. High-level disinfection equipment is maintained according to manufacturer instructions	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
c. Chemicals used for high-level disinfection are:		
I. Prepared according to manufacturer instructions	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
II. Tested for appropriate concentration according to	<input type="radio"/> Yes	<input type="radio"/> Observation

manufacturer's instructions	<input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Interview <input type="radio"/> Both
III. Replaced according to manufacturer's instructions	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
Practices to be Assessed	Was Practice Performed?	Manner of Confirmation
IV. Documented to have been prepared and replaced according to manufacturer's instructions	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
d. Instruments requiring high-level disinfection are:		
I. Disinfected for the appropriate length of time as specified by manufacturer's instructions or evidence-based guidelines	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
II. Disinfected at the appropriate temperature as specified by manufacturer's instructions on evidence-based guidelines	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
E. Items that undergo high-level disinfection are allowed to dry before use	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
F. Following high-level disinfection, items are stored in a designated clean area in a manner to prevent contamination	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
G. Additional breaches in high-level disinfection practices, not captured by the questions above were identified (If YES, please specify further in comments)	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
Comments: <i>(please print and limit comments to the space provided)</i>		

IV. Environmental Infection Control

Observations are to be made of staff who perform environmental cleaning (e.g., surgical technicians, cleaning staff, etc.)

Practices to be Assessed	Was Practice Performed?	Manner of Confirmation
A. Operating rooms are cleaned and disinfected after each surgical or invasive procedure with an EPA-registered disinfectant	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
B. Operating rooms are terminally cleaned daily	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
C. High-touch surfaces in patient care areas are cleaned and disinfected with an EPA-registered disinfectant	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
D. The ASC has a procedure in place to decontaminate gross spills of blood	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
E. Additional breaches in environmental cleaning not captured by the questions above were identified (If YES, please specify further in comments)	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both

Comments: <i>(please print and limit comments to the space provided)</i>	
---	--

V. Point of Care Devices (e.g., blood glucose meter)

Observations are to be made of staff who perform fingerstick testing (e.g., nurses)

If N/A is *filled in*, please clarify why *in the comments box below* why it was not applicable or not observed.

Practices to be Assessed	Was Practice Performed?	Manner of Confirmation

1. Does the ASC have a blood glucose meter? If NO, STOP HERE.	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
A. A new single-use, auto-disabling lancing device is used for each patient	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
B. The glucose meter is not used on more than one patient unless the manufacturer's instructions indicate this is permissible	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
C. The glucose meter is cleaned and disinfected after every use.	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both
D. Additional breaches in appropriate use of point of care devices (like glucose meters) not captured by the questions above were identified (If YES, please specify further in comments)	<input type="radio"/> Yes <input type="radio"/> No <input type="radio"/> N/A	<input type="radio"/> Observation <input type="radio"/> Interview <input type="radio"/> Both

Comments: <i>(please print and limit comments to the space provided)</i>	
---	--