



National Institute of Standards & Technology

Certificate of Analysis

Standard Reference Material[®] 1622e

Sulfur in Residual Fuel Oil

This Standard Reference Material (SRM) is intended for use in the calibration of instruments and the evaluation of methods used in the determination of total sulfur in fuel oils or materials of similar matrix. A unit of SRM 1622e consists of 100 mL of commercial No. 6 residual fuel oil as defined by ASTM D396-95 Standard Specification for Fuel Oils [1].

Certified Value of Sulfur (mass fraction): 2.1468 % \pm 0.0041 %

The certified sulfur content is based on analyses by isotope dilution thermal ionization mass spectrometry (ID-TIMS) [2]. Homogeneity testing was performed using X-ray fluorescence spectrometry. No significant bottle to bottle variability was observed.

The uncertainty in the certified value is calculated as $U = ku_c$, where u_c is the *combined standard uncertainty* calculated according to the *ISO Guide* [3] and k is a coverage factor. The value of u_c is intended to represent, at the level of one standard deviation, the combined effect of uncertainty components associated with material inhomogeneity and ID-TIMS measurement uncertainty. In the absence of Type B uncertainties, (which are negligible here in comparison with Type A), the expanded uncertainty (U) is given for a 95 % confidence interval. The coverage factor, $k = 2.36$, is the Student's t -value for a 95 % confidence interval with seven degrees of freedom.

Expiration of Certification: The certification of this SRM is valid until **01 July 2015**, within the uncertainty specified, provided the SRM is handled and stored in accordance with the instructions given in the certificate (see *Use and Handling*). However, the certification will be nullified if the SRM is damaged, contaminated, or modified.

Stability: This material is considered to be stable during the period of certification. NIST will monitor this material and will report any significant changes in certification to the purchaser. Return of the attached registration card will facilitate notification.

The overall direction and coordination of the technical measurements leading to certification of this SRM were performed by G.C. Turk of the NIST Analytical Chemistry Division.

Analytical measurements were performed by W.R. Kelly, R.D. Vocke, A.F. Marlow, P.A. Pella, and T.L. Quinn of the NIST Analytical Chemistry Division.

Statistical consultation for this SRM was provided by K.R. Eberhardt of the NIST Statistical Engineering Division.

The support aspects involved in the preparation, certification, and issuance of this SRM was coordinated through the Standard Reference Materials Program by B.S. MacDonald of the NIST Measurement Services Division.

Willie E. May, Chief
Analytical Chemistry Division

John Rumble, Jr., Chief
Measurement Services Division

Gaithersburg, MD 20899
Certificate Issue Date: 02 April 2003
See Certificate Revision History on Last Page

INSTRUCTIONS FOR USE

The SRM bottle should only be opened for the minimum time required to dispense the material. To relate analytical determinations to the certified value in this Certificate of Analysis, a minimum sample mass of 140 mg should be used. After use, the bottle should be tightly recapped and stored under normal laboratory conditions away from direct sunlight.

SUPPLEMENTAL INFORMATION

The additional properties of SRM 1622e are listed below. These properties were determined by a commercial firm under contract to NIST using ASTM methods. The results are **NOT** certified and are provided as additional information on the matrix.

Physical Property Test	ASTM Standard Used	Result
Density @ 15 °C	D 1250-80 (1990)	1005.1 kg/m ³
@ 60 ° F	D 287-92 (1995)	9.2 API
Flash Point, PMCC D 93-94	77 °C	
Pour Point	D 97-93	14 °C
Heat of Combustion, Gross	D 240-92 ^{e1}	42.49 MJ/kg (18 268 Btu/lb)
Kinematic Viscosity @ 40 °C	D 445-94 ^{e1}	1763 •10 ⁻⁶ m ² /s (1763 cSt)
@ 50 °C	D 445-94 ^{e1}	776.9 •10 ⁻⁶ m ² /s (776.9 cSt)
@ 100 °C	D 445-94 ^{e1}	54.7 •10 ⁻⁶ m ² /s (54.74 cSt)
Carbon	D 5291-92	86.5 %
Hydrogen	D 5291-92	10.1 %

ASTM Standards

D 93-94	Standard Test Methods for Flash Point by Pensky-Martens Closed Tester
D 97-93	Standard Test Methods for Pour Point of Petroleum Products
D 240-92 ^{e1}	Standard Test Method for Heat of Combustion of Liquid Hydrocarbon Fuels by Bomb Calorimeter
D 287-92 (1995)	Standard Test Method for API Gravity of Crude Petroleum and Petroleum Products (Hydrometer Method)
D 445-94 ^{e1}	Standard Test Method for Kinematic Viscosity of Transparent and Opaque Liquids (and the Calculation of Dynamic Viscosity)
D 1250-80 (1990) ^{e1}	Standard Guide for Petroleum Measurement Tables
D 5291-92	Standard Test Methods for Instrumental Determination of Carbon, Hydrogen, and Nitrogen in Petroleum Products and Lubricants

REFERENCES

- [1] ASTM D 396-95, *Standard Specification for Fuel Oils*; Annu. Book ASTM Stand.; Vol. 05.01, West Conshohocken, PA.
- [2] Kelly, W.R.; Paulsen, P.J.; Murphy, K.E.; Vocke, R.D., Jr.; Chen, L.-T.; *Determination of Sulfur in Fossil Fuels by Isotope Dilution Thermal Ionization Mass Spectrometry*; Anal. Chem., Vol. 66, pp. 2505-2513 (1994).
- [3] *Guide to the Expression of Uncertainty in Measurement*; ISBN 92-67-10188-9, 1st Ed. ISO, Geneva, Switzerland (1993); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC (1994).

Certificate Revision History: 02 April 2003 (This revision reflects an extension in the expiration date); 10 April 1997 (Original certificate date).

Users of this SRM should ensure that the certificate in their possession is current. This can be accomplished by contacting the SRM Program at: Telephone (301) 975-6776 Fax (301) 926-4751, e-mail srminfo@nist.gov, or via the Internet <http://www.nist.gov/srm>.