



# National Institute of Standards & Technology

## Certificate of Analysis

### Standard Reference Material<sup>®</sup> 2658a

#### Oxygen in Nitrogen

(Nominal Amount-of-Substance Fraction – 10 % mol/mol)

*This certificate reports the certified value for Lot 72-C-XX.*

This Standard Reference Material (SRM) is a primary gas mixture that, the amount-of-substance fraction expressed as concentration [1], may be related to secondary working standards. The SRM is intended for the calibration of instruments used for oxygen determinations and for other applications.

This SRM mixture is supplied in a DOT 3AL specification aluminum (6061 alloy) cylinder with a water volume of 6 L. Mixtures are shipped with a nominal pressure exceeding 12.4 MPa (1800 psi), which provides the user with 0.73 m<sup>3</sup> (25.8 ft<sup>3</sup>) of useable mixture. The cylinder is the property of the purchaser and is equipped with a CGA-590 brass valve, which is the recommended outlet for this oxygen mixture. NIST recommends that this cylinder **NOT** be used below 0.7 MPa (100 psi).

**Certified Value:** This SRM mixture has been certified for oxygen concentration. The certified value, given below, applies to the identified cylinder and NIST sample number.

Oxygen Concentration: 9.397 % mol/mol ± 0.049 % mol/mol

Cylinder Number:

NIST Sample Number:

The uncertainty of each certified value includes the estimated uncertainties in the NIST standards, the analytical comparisons to the lot standard (LS), and the uncertainty of comparing the LS with each of the mixtures comprising this lot. This uncertainty is expressed as an expanded uncertainty  $U = ku_c$ , with  $u_c$  determined by experiment and a coverage factor  $k = 2$ . The true value for the oxygen amount-of-substance fraction is asserted to lie in the interval defined by the certified value ±  $U$  with a level of confidence of approximately 95 % [2].

**Expiration of Certification:** This certification is valid until **01 July 2011**, within the measurement uncertainty specified, provided the SRM is handled and stored in accordance with the instructions given in this certificate. However, the certification will be nullified if the SRM is contaminated or modified. NIST will monitor this SRM over the period of its certification. If changes occur that affect the certification before the expiration of this certificate, NIST will notify the purchaser. Registration (see attached sheet) will facilitate notification.

**Hydrotest Date:** 07/97

**Blend Date:** 08/97

**Cylinder and Gas Handling Information:** NIST recommends the use of a high-purity, two-stage pressure regulator with a stainless steel diaphragm and CGA-590 outlet to safely reduce the pressure and to deliver this SRM mixture to the instrument. The regulator should be purged several times to prevent accidental contamination of the sample.

The analytical measurements leading to the certification of this current SRM lot were performed by G.C. Rhoderick of the NIST Analytical Chemistry Division.

Stephen A. Wise, Chief  
Analytical Chemistry Division

Gaithersburg, MD 20899  
Certificate Issue Date: 19 September 2005  
*See Certificate Revision History on Last Page*

Robert L. Watters, Jr., Chief  
Measurement Services Division

The overall direction and coordination of the technical work required for this SRM's certification were performed by F.R. Guenther of the NIST Analytical Chemistry Division.

The support aspects involved in the issuance of this SRM were coordinated through the NIST Measurement Services Division.

**Mixture Preparation:** The gas mixtures comprising this SRM lot were prepared in accordance with NIST technical specifications by a commercial specialty gas vendor under contract to NIST. The specifications stipulate that each SRM mixture be identical in oxygen concentration and stable with time.

**Analytical Methods:** Analyses of the oxygen concentration for this lot of cylinders were conducted by comparing each cylinder mixture to a representative cylinder chosen from the lot, the lot standard (LS), using a research gas chromatograph (GC) equipped with a thermal conductivity detector (TCD). Assignment of the oxygen concentration to the LS was accomplished by comparison to primary gravimetric standards using GC/TCD.

**Homogeneity Analysis:** Each of the oxygen mixtures that comprise this SRM lot was compared to the LS using GC/TCD. An analysis of variance indicated that sample-to-sample oxygen concentration differences were statistically insignificant. This indicates that within the precision of the NIST measurements, all of the cylinders comprising this SRM lot have identical oxygen concentrations. Therefore, one concentration has been assigned to the entire SRM lot.

**Oxygen Concentration Value Assignment:** The certified oxygen concentration for this SRM lot was computed from the assigned concentration for the lot standard and the homogeneity analysis.

**Stability:** Periodic analyses of SRM units from this lot are performed at NIST to monitor stability. If significant changes in the oxygen concentration are observed, the purchaser will be notified. Refer to the "Cylinder and Gas Handling Information" section for proper handling of this SRM. Registration (see attached sheet) will facilitate notification.

**Other Analyses:** Additional analyses performed during the certification procedure are given below. The concentrations reported below are **NOT** certified values. These values are given for informational purposes only.

Constituent	Concentration	Method
Argon	20 $\mu\text{mol/mol}$	Mass Spectrometry
Carbon Dioxide	3.8 $\mu\text{mol/mol}$	Mass Spectrometry
Water	3 $\mu\text{mol/mol}$	Moisture Analyzer
Total Other Hydrocarbons (expressed as methane)	< 0.0014 $\mu\text{mol/mol}$	GC/FID

#### REFERENCES

- [1] Taylor, B.N.; *Guide for the Use of the International System of Units (SI)*; NIST Special Publication 811; U.S. Government Printing Office: Washington, DC (1995).
- [2] ISO; *Guide to the Expression of Uncertainty in Measurement*; ISBN 92-67-10188-9, 1st ed.; International Organization for Standardization: Geneva, Switzerland (1993); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at <http://physics.nist.gov/Pubs/>.

**Certificate Revision History:** 14 September 2005 (This technical revision reports an extension of the certification period); 06 March 2003 (This technical revision reports a change in the expiration date); 30 December 1999 (Original certificate date).

*Users of this SRM should ensure that the certificate in their possession is current. This can be accomplished by contacting the SRM Program at: telephone (301) 975-6776; fax (301) 926-4751; e-mail [srminfo@nist.gov](mailto:srminfo@nist.gov); or via the Internet at <http://www.nist.gov/srm>.*