

X17B3 SAFETY CHECKLIST			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
Rev: C	Effective: 5/24/2010	Page 1 of 1	Number: LS-SCL-0039
Prepared by: Philip Marino		Reviewed by: P. Marino	Approved by: B. Chmiel

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

- 1. X17B1 Safety Checklist complete and in place.
- 2. X17B2 Safety Checklist complete and in place.
- 3. X17B3 hutch – threshold door piece (painted yellow) bolted in place under each of 2 doors going into hutch.
- 4. X17B3 hutch – lead shielding mats against threshold door piece in each of 2 doors going into hutch.
- 5. X17B3 hutch interior – water cooled backstop/ Bremsstrahlung shield #5 (see photo) in place and banded at downstream end of hutch.

OPCO / Beamline Staff: _____ Date: _____