

X20A SAFETY CHECKLIST			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
Rev: D	Effective: 5/24/2010	Page 1 of 1	Number: LS-SCL-0044
Prepared by: Philip Marino		Reviewed by: P. Marino	Approved by: B. Chmiel

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

After completing this form call Operation Coordinator for Safety Approval Form (SAF)

For locations of beamline checkpoints see Beamline Schematic.

- 1. Valid Padlock Index and log, no locks open
 - A. Check to make sure all safety shutter interlocks are satisfied (white panel at top of equipment rack)

Check points 2-19 are along beamline X20A, beginning at upstream end.

Best access to 2-17 is between X19&X20.

- 2. Bremsstrahlung shield 1A, B, C (BS #1) (built into ring sawtooth wall), as in photo 1.
- 3. Lead scatter shielding A, B, C around beamline pipes and BE window, as in photo 2.
 - B. Water flow for Be window, located overhead above valve 1.
 - C. Gate valve 1 open (indicator on the top of valve).
- 4. Exclusion Zone 1A, B, C (EZ #1). Scatter Shield under EZ #1 as in photo 3.
- 5. Bremsstrahlung shield 2A, B, C (BS # 2) as in photo 4.
 - D. Check pressure for X20A at Ion pump controller located below Bremsstrahlung shield 2, should be less than 1×10^{-5} torr.
- 6. Exclusion Zone 2A, B, C (EZ #2) as in photo 5.
- 7. Bremsstrahlung shield 3A, B, C (BS # 3) as in photo 6.
- 8. Bremsstrahlung shield (BS # 4) as in photo 7.
- 9. Lead shielding slit #2 and bellows #2 as in photo 8.
 - E. Water flow for mirror, located near floor.
- 10. Lead shielding on observation cross #1 and spacer cross, as in photo 8.
- 11. Lead shielding on mirror tank, as in photo 9.
- 12. Bremsstrahlung shield (BS # 5) as in photo 10.
- 13. Lead shielding on bellows #3 and downstream spool piece, as in photo 10.
- 14. Bremsstrahlung shield (BS # 6) and lead covering upstream pipe as in photo 11.
- 15. Bremsstrahlung shield (BS # 7) as in photo 12.
 - F. Water flow for monochromator, located near floor behind roughing pump.
- 16. Lead shielding on bellows #4 and on top of observation cross #2 as in photo 13.
- 17. Leaded Plexiglas on monochromator viewports (on top 2 viewports).

Checkpoint 18 is inside the x20C hutch and can be viewed through its door window.

- 18. Lead shielding on bellows #5 as in photo on hutch door.

For numbered Safety items:

OPCO / Beamline Staff: _____ Date: _____

For lettered User system checks:

Beamline Rep.: _____ Date: _____