

X26A SAFETY CHECKLIST			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
Rev: C	Effective:5/24/2010	Page 1 of 1	Number: LS-SCL-0057
Prepared by: Philip Marino		Reviewed by: P. Marino	Approved by: B. Chmiel

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

1. Valid Padlock Index and log, no padlocks open

ITEMS 2-3 REFER TO PHOTO #1

2. Scatter Shielding over Bellows, Be Window, and Slit box as per photo.
 3. Lead Bremsstrahlung shield (BS#1) in place and banded.

ITEMS 4-7 REFER TO PHOTO #2

4. Monochromator wrapped in lead and banded.
 5. Viewport on monochromator covered in lead.
 6. Vacuum Bellows #2 wrapped in lead.
 7. Lead Bremsstrahlung shield (BS#2) in place and banded.

ITEMS 8-11 REFER TO PHOTO #3

8. Vacuum Bellows #3 wrapped in lead.
 9. Mirror tank wrapped in lead and banded.
 10. Slits A2 wrapped in lead and banded.
 11. Ion chamber pipe wrapped in lead and banded at X26A hutch wall

INSIDE OF X26A HUTCH

12. Bremsstrahlung shield beam-stop, in place and banded.

. OPCO / Beamline Staff: _____ Date: _____