

## Beamline Hazard Analysis - Beamline X16C

LS-BHA-X16C

Revision: 02

Effective: 10/17/11

Page 1 of 1

The following tasks are included in the beamline safety review. Any operations not listed here, or performed by others than those listed, need additional review by ES&H personnel and/or beamline personnel.

Prepared By: P. Stephens

Approved By: Lori Stiegler

Tasks / Equipment	Performed by:		Controls
	Beamline Staff	Users	
Inert compressed gas	Y	Y	<ul style="list-style-type: none"> <li>Compressed gas training – TQ-COMPGAS1 or NSLS User training</li> <li>Transport cylinders using a cylinder cart</li> <li>Properly secure cylinders</li> <li>Wear safety glasses with side shields while connecting lines</li> <li>Leak check gas lines</li> </ul>
Solvents	Y	Y	<ul style="list-style-type: none"> <li>Limit quantities at beamline to &lt; 250 ml</li> <li>Use PPE - safety glasses, nitrile gloves</li> </ul>
High temperature	Y	Y	<ul style="list-style-type: none"> <li>Sample furnace inaccessible when hot</li> <li>Training</li> </ul>
Cutting/razor blades	Y	Y	<ul style="list-style-type: none"> <li>Be aware of hand positioning</li> <li>Use safety knife when possible</li> <li>Consider using cut resistant gloves for frequent cutting tasks</li> </ul>