

Beamline Hazard Analysis - Beamline X24A

LS-BHA-X24A

Revision: 02

Effective: 10/18/11

Page 1 of 2

The following tasks are included in the beamline safety review. Any operations not listed here, or performed by others than those listed, need additional review by ES&H personnel and/or beamline personnel.

Prepared By: B. Karlin, J. Woicik

Approved By: Lori Stiegler

Tasks / Equipment	Performed by:		Controls
	Beamline Staff	Users	
Inert compressed gas	Y	Y	<ul style="list-style-type: none"> Compressed gas training – TQ-COMPGAS1 or NSLS User training Transport cylinders using a cylinder cart Properly secure cylinders Wear safety glasses with side shields while connecting lines Leak check gas lines
Solvents	Y	Y	<ul style="list-style-type: none"> Limit quantities at beamline to < 250 ml Use PPE - safety glasses, nitrile gloves
Pouring cryogenics ≤ 5L	Y	N	<ul style="list-style-type: none"> Use goggles Long pants and sleeves - avoid clothing which can trap cryogenics (pockets, cuffs) Cryogen or heavy gloves Enclosed shoes Never pour from above chest level
Using the LN2 Fill Station or pressurized transfer	Y	Y	<ul style="list-style-type: none"> LS-PROC-LN2FILL-MANL training for fill station Use full face shield along with goggles or safety glasses Long pants and sleeves - avoid clothing which can trap cryogenics (pockets, cuffs) Cryogen or heavy gloves Enclosed shoes
Class 2/3a laser	Y	N	<ul style="list-style-type: none"> Laser registration, warning signs
Ergonomics concerns – awkward postures – adjusting beamline equipment	Y	N	<ul style="list-style-type: none"> Be aware of body posture Take frequent breaks
Ergonomics concerns – pinch points/sharp edges – valves	Y	N	<ul style="list-style-type: none"> Warning signs Training
Soldering	Y	N	<ul style="list-style-type: none"> Use a designated area, periodically clean surfaces
High temperature – bakeout chamber	Y	N	<ul style="list-style-type: none"> Warning signs Training
Working at heights	Y	N	<ul style="list-style-type: none"> Be aware of body posture, training
Material Handling/Crane	Y	N	<ul style="list-style-type: none"> Training
Cutting/razor blades	Y	Y	<ul style="list-style-type: none"> Be aware of hand positioning Use safety knife when possible

Beamline Hazard Analysis - Beamline X24A

LS-BHA-X24A

Revision: 02

Effective: 10/18/11

Page 2 of 2

The following tasks are included in the beamline safety review. Any operations not listed here, or performed by others than those listed, need additional review by ES&H personnel and/or beamline personnel.

- | | | | |
|--|--|--|--|
| | | | <ul style="list-style-type: none">• Consider using cut resistant gloves for frequent cutting tasks |
|--|--|--|--|