Brookhaven National Laboratory – Photon Sciences - National Synchrotron Light Source						
Beamline Hazard Analysis - Beamline X26C						
LS-BHA-X26C	Revision: 02	Effective: 10/18/11	Page 1 of 1			
The following tasks are included in the beamline safety review. Any operations not listed here, or performed by others than						
those listed, need additional review by ES&H personnel and/or beamline personnel.						

Prepared By:	A. Orville, S. Myers	Approved By:	Lori Stiegler

^{*} Approval signatures on file with master copy

Tasks / Equipment	Performed by:		Controls	
,	Beamline Staff	Users		
Solvents	Υ	Υ	 Limit quantities at beamline to < 250 ml Use PPE - safety glasses, nitrile gloves 	
Pouring cryogens <u><</u> 5L	Υ	Y	 Use goggles Long pants and sleeves - avoid clothing which can trap cryogens (pockets, cuffs) Cryogen or heavy gloves Enclosed shoes Never pour from above chest level 	
Using the LN2 Fill Station or pressurized transfer	Y	Y	 LS-PROC-LN2FILL-MANL training for fill station Use full face shield along with goggles or safety glasses Long pants and sleeves - avoid clothing which can trap cryogens (pockets, cuffs) Cryogen or heavy gloves Enclosed shoes 	
Manipulating cryogenic samples	Y	Y	 Use goggles Long pants Impervious disposable gloves over cotton liners Enclosed shoes Use tongs or tools whenever possible 	
Cryogenic autofill system	Υ	Υ	Ensure hutch ventilation system is operating	
Class 2/3a laser	Υ	Υ	Laser registration, warning signs	
Ergonomics concerns – pinch points/sharp edges – motion of detector			BLOSA training	
Working at heights	Υ	N	Be aware of body posture, training	
Cutting/razor blades	Y	Y	 Be aware of hand positioning Use safety knife when possible Consider using cut resistant gloves for frequent cutting tasks 	