

## Progress Toward SRM 2372: Human DNA Quantitation Standard

Margaret Kline, Amy Decker, Janette Redman,  
Peter Vallone, David Duewer and John Butler  
National Institute of Standards & Technology  
Chemical Science & Technology Laboratory, Gaithersburg MD

National Institute of Justice 7<sup>th</sup> Annual DNA Grantees Workshop,  
June 26-28, 2006, at the Marriott Crystal Gateway Hotel in Arlington, VA.

### Disclaimers

**Funding:** Interagency Agreement 2003-IJ-R-029 between  
the **National Institute of Justice** and NIST Office of Law  
Enforcement Standards

Points of view are those of the authors and do not necessarily represent the official position or policies of the US Department of Justice. Certain commercial equipment, instruments and materials are identified in order to specify experimental procedures as completely as possible. In no case does such identification imply a recommendation or endorsement by the National Institute of Standards and Technology nor does it imply that any of the materials, instruments or equipment identified are necessarily the best available for the purpose.

**Our publications and presentations are made available at:**  
<http://www.cstl.nist.gov/biotech/strbase/NISTpub.htm>



### National Institute of Justice

The Research, Development, and Evaluation Agency of the U.S. Department of Justice

#### Current Areas of NIST Effort with Forensic DNA

##### • Standards

- Standard Reference Materials
- Standard Information Resources (STRBase website)
- Interlaboratory Studies

##### • Technology

- Research programs in SNPs, miniSTRs, Y-STRs, mtDNA, qPCR
- Assay and software development

##### • Training Materials

- Review articles and workshops on STRs, CE, validation
- PowerPoint and pdf files available for download

### The issue:

How can we reproducibly measure the concentration of extracted DNA?




### Interlaboratory Studies for DNA Quantitation

- NIST Mixed Stain Study #2 (1999)
  - [DNA] range 0.5 ng/μL to 5 ng/μL; **x1.8 variability**
  - *J Forensic Sci* 2001;46:1199-1210
- NIST Mixed Stain Study #3 (2001)
  - [DNA] range 1 ng/μL to 4 ng/μL; **x1.7 variability**
  - *Anal Chem* 2003;75:2463-2469
- NIST Quantitation Study 2004
  - [DNA] range 0.05 ng/μL to 1.5 ng/μL; **x1.8 variability**
  - *J Forensic Sci* 2005;50:571-578

### Lessons Learned from the NIST QS04

- At low [DNA] 0.05 ng/μL labs recovered 73% more DNA from a Teflon tube than from a polypropylene tube.
- The “new” qPCR methods used in the study consistently provided quantitative results for samples with [DNA] 0.16 ng/μL and lower.
- The qPCR methods appeared to be biased to one another.







### Would a Quantitative DNA Standard Help?

- Yes
  - If the bias seen among the QS04 qPCR methods is from the use different calibrators.
  - If the SRM material produced is fit for purpose.

### Investigation of “bias” between qPCR Methods

- The experimental design:
  - 5 qPCR methods
    - Quantifiler, Quantifiler Y, Alu, CFS, CADOJ
  - 6 calibrators
    - 3 commercial sources
    - 3 NIST extracted materials (1 Proteinase K treatment)
  - 7 dilutions of each calibrator made daily
    - 10.0, 4.0, 1.6, 0.64, 0.26, 0.1, and 0.04 ng/μL
  - Duplicate plates (run on the same day)





## Requirements for NIST SRM 2372 Human DNA Quantitation Standard

### Material must be fit for purpose:

- **Homogeneity**
  - All tubes are the same
- **Stability**
  - Will withstand shipping and normal storage
- **Recoverability**
  - What went in the tubes comes out
- **Traceability**
  - Values assigned are traceable to the designated certification method.

## SRM 2372 Human DNA Quantitation Standard

Anticipated 2006 issue



Component A: Male  
 Component B: Female  
 Component C: Mixture

Planned Amounts: Each component 110 μL of Human Genomic DNA with a concentration targeted @ 50 ng/μL.

The [DNA] for each component will be listed in the Certificate of Analysis.

## Preparation of 2372 Components

- Female and male components isolated from Buffy coats at NIST using “salting out” extraction procedure
  - Miller et al. *Nucleic Acids Res.* 1988, 16(3) 1215
- Mixture material (multiple male and female donors) purchased as purified freeze-dried material.

## Component A: Male

- **Single Male donor (Buffy coat).**
- After initial extraction, EtOH ppt material was solubilized and treated with RNase.
- Following RNase treatment the material was retreated with Proteinase K and EtOH ppt.
- The ppt material was washed in 70% EtOH and allowed to dry.
- Final solubilization was in TE<sup>-4</sup> (10 mM Tris 0.1 mM EDTA, pH 8.0) buffer.


## Component B: Female

- **Multiple Female donors (Buffy coats).**
- After initial extraction, EtOH ppt material was solubilized and retreated with Proteinase K.
- The ppt material was washed in 70% EtOH and allowed to dry.
- Final solubilization was in TE<sup>-4</sup> buffer.

## Component C: Mixture

- **Multiple Male and Female donors**
  - (tissue extract, freeze-dried, Sigma #D7011)
- Solubilized in TE<sup>-4</sup> buffer.
- Allowed to equilibrate 2 weeks prior to measuring absorbance.
- Material has been used for stability testing.

## Gel Image Pilot DNA Components



All materials appear to be intact DNA.

The mobility of the commercially obtained mixture material appears slightly different than that of the NIST extracted material.

## Certification Plan

- Use a NIST National-Reference Spectrophotometer to **certify** (*value and uncertainty*) absorbance values at 230, 260, 270, 280, and 330 nm.
- This provides the traceability path...  
**so that future materials can have the same properties.**

## How do we get Traceability?

- **Photometry**
- Regular transmittance scale is maintained on the National Reference Spectrophotometer in the NIST Analytical Chemistry Division which is validated by the double-aperture method of light addition and benchmarked through **international intercomparisons** using optical filter artifact standards.
- Liquid or powder SRMs are assigned certified values for absorbance per unit pathlength or specific absorptivity at specified wavelengths.
- NIST Special Publication 260-136
  - Definitions of Terms and Modes Used at NIST for Value-Assignment of Reference Materials for Chemical Measurements

## Absorbance at 260 nm Facts:

- DNA, RNA, EDTA, and Phenol all absorb
- Absorption coefficients are affected by:
  - Ionic strength of the solution (*needs to be low*)
  - pH of the solution (*needs to be controlled*)
    - Beaven et al., The nucleic acids, 1955 (1) 493 - 553;
    - Wilfinger et al. BioTechniques 1997 (22) 474 - 481
- Method reliable only in 5 ng/μL to 90 ng/μL range of DNA concentration.
- (you have to have light passing through the solution!)
  - Sambrook and Russell, Molecular Cloning, 3<sup>rd</sup> Ed 2001, A8.20

## Spectrophotometric Determination

- 260 nm & 280 nm readings
- 260 nm allows calculation of the **conventional [DNA]**
- OD = 1 ≈ 50 μg/mL dsDNA  
≈ 40 μg/mL ssDNA  
≈ 33 μg/mL oligos
- 260 / 280 ratio ≈ 1.8 to 2.0  
(Provides an estimate of contaminating protein)


### Additional wavelengths:

- 230 nm Significant absorbance indicates:
  - Phenolate ion
  - Thiocyanates
  - And other organic compounds
- 270 nm
  - Water saturated with phenols absorbs
  - 260:270 ~ 1.2 indicates preparation free of phenol  
Stulnig and Amberger 1994 BioTechniques;16 :403-404
- 330 nm and higher absorbance
  - Caused by light scattering indicating presence of particulate matter

### Certification Plan continued

- Recoverability
- Stability
- Homogeneity
  - Use the Cary 100 Bio instrument (110 µL cell)
  - Made stable and accurate by virtue of the appropriate use of **wavelength** (SRM 2034) and **absorbance standards** (SRM 2031) before and after component measurements.
- Validity and commutability of conventional [DNA] with qPCR and other field methods.
  - qPCR methods performed at NIST
  - Interlaboratory study

### Stability of a DNA Solution



A DNA solution was prepared in a sterile Teflon container Sept 2003. The material was aliquoted into sterile 5 mL Teflon vials and held at lab ambient temperature, 4 °C, -20 °C, and -80 °C.

After 33 months of storage there is no difference in the apparent quality of the material.

**Recommended Stored Temperature 4 °C**  
(will be shipped on ice bricks NOT dry ice)

### Stability of [DNA] in Different Tubes

Can the end user get out what was put in?

Five different tubes were evaluated at :  
 3 different storage temperatures  
 3 different [DNA]

Quantifier used to evaluate the [DNA]  
 Duplicate tubes, duplicate qPCR runs


Duration 7 months : Averaged results for 5 time points

[DNA]	A	B	C	D	E
0.20	1.00	0.74	1.14	0.72	0.69
1.00	1.00	0.88	0.98	0.86	0.88
5.00	1.00	0.99	0.91	0.94	0.72

[DNA] in ng/µL

30 data points / tube type / [DNA]

### 10 weeks of UV measurements



To make sure that **you** can retrieve what **we** put in the **Sarstedt** tubes.

Multiple aliquots of a preliminary material were tested for 10 weeks. The DNA did **not** change in that time period. The instrument required calibration with both wavelength (SRM 2034) and absorbance (SRM 2031) standards before and after measuring the DNA solution.


### Absorbance of the Pilot Lots

Dil	Female		Male		Mix	
	260 ratio	260 ratio	260 ratio	260 ratio	260 ratio	260 ratio
1:2	1.299	2.6	0.580	1.2	1.401	2.8
1:5	0.518	2.6	0.235	1.2	0.556	2.8
1:10	0.263	2.6	0.114	1.1	0.279	2.8
Mean [DNA]	130.3		57.9		139.5 ng/µL	
sd	0.89		0.83		0.55	
rsd	0.7		1.4		0.4 %	
260 : 280	1.97		1.97		1.97	
260 : 270	1.2		1.2		1.2	

### Dilution scheme for qPCR

Female			Male			Mix		
130.3	57.9	139.5	μL	300	μL	300	μL	300
DNA	DNA	DNA	TE	TE	TE	TE	TE	TE
10	10	10	23.0	277.0	10	10	21.5	278.5
5	5	5	10	10	5	5	10	10
2.5	2.5	2.5	10	10	2.5	2.5	10	10
1.25	1.25	1.25	10	10	1.25	1.25	10	10
0.625	0.625	0.625	10	10	0.625	0.625	10	10
0.313	0.313	0.313	10	10	0.313	0.313	10	10
0.156	0.156	0.156	10	10	0.156	0.156	10	10
0.078	0.078	0.078	10	10	0.078	0.078	10	10

### Quantifier Results for Preliminary Components of SRM 2372



### SRM 2372 Human DNA Quantitation Standard



#### Components

- A: Male/single donor/RNased/NIST
- B: Female/multiple donors/NIST
- C: Mixture/male & female/commercial

#### Quantities

110 μL of Human Genomic DNA  
 Absorbance of 1.0 OD  
 Conventional [DNA] 50 ng/μL.

#### Certification

Absorbance by US National Spectrophotometer  
 Homogeneity by Cary 100 Bio  
**Validation of conventional [DNA] by Interlaboratory Study**

### What is Delaying Release?

- **Need to extract more DNA** in order to reach goal of producing >1,500 units (there is a great deal of interest in SRM 2372 outside of the forensic community—e.g., pharmaceutical industry)
  - ~30 units (3 mL) are required by the NIST National-Reference Spectrophotometer for its measurements
- **Additional studies to be performed:** interlaboratory (performed by multiple forensic labs), homogeneity (monitored by NIST statisticians), and continual stability testing for the life of the product

### An Interlaboratory Study Will Be Performed to Demonstrate Commutability of SRM 2372

You will have 3 weeks to return your data once we ship the final packaged material (~August 2006).

**Any Volunteers?**

Contact: [margaret.kline@nist.gov](mailto:margaret.kline@nist.gov)

### Thank you for your Attention!!



#### Acknowledgements

##### Funding:

Interagency Agreement between National Institute of Justice and NIST Office of Law Enforcement Standards

##### NIST Project Team:

- John Butler      Pete Vallone
- Margaret Kline      Jan Redman
- Amy Decker      Becky Hill
- Dave Duewer

Email: [Margaret.Kline@nist.gov](mailto:Margaret.Kline@nist.gov)