Identifier: SOP-5164 (formerly SOP-15.04-R1)

Effective Date: 6/17/08

Revision: 0

Next Review Date: 6/17/13



Waste & Environmental Services

Standard Operating Procedure

for ROUTINE VALIDATION OF HIGH EXPLOSIVE (HE) ANALYTICAL DATA

APPROVAL SIGNATURES:

Subject Matter Expert:	Organization	Signature	Date	
Bill Hardesty	WES-EDA Signature on file		4/21/08	
Quality Assurance Specialist:	Organization	Signature	Date	
Laura Ortega	QA-IQ	Signature on file	5/14/08	
Responsible Line Manager:	Organization	Signature	Date	
Craig Eberhart	WES-EDA	Signature on file	4/21/08	

Title: Routine Validation of High Explosive (HE) Analytical	No.: SOP-5164	Page 2 of 12
Data	Revision: 0	

1.0 **PURPOSE AND SCOPE**

This procedure represents the minimum standards for evaluating routine Explosives analytical data. This procedure is a mandatory document and shall be implemented by all Los Alamos National Laboratory (LANL or Laboratory) personnel and contractors who evaluate routine Explosives analytical data for the specific LANL projects.

2.0 **BACKGROUND AND PRECAUTIONS**

2.1 **Background**

This procedure conforms to the requirements of Environmental Protection Agency (EPA) Methodologies and the EPA document, "U.S. EPA Contract Laboratory Program National Functional Guidelines for Organic Data Review." LANL data validation is performed according to procedures based upon the NNSA Model Data Validation Procedure. Data qualifiers and reason codes are assigned according to the specifications in this method specific procedure.

2.2 **Precautions**

Nothing in this procedure precludes the data validator from going beyond the minimum requirements specified within this procedure. If additional directions are required, the data validator shall reference NNSA Model Data Validation Procedure, EPA method specific guidelines and/or National Functional Guidelines for Organic Data Review. Implementation of this procedure may be followed by a more focused and data use-specific evaluation of the data by the project chemist, especially if the implementation of this procedure indicates the data may contain technical deficiencies.

3.0 **EQUIPMENT AND TOOLS**

None.

4.0 STEP-BY-STEP PROCESS DESCRIPTION

Qualifications for Data Validators

4.1 Data 1. Possess a minimum of a bachelor's degree in chemistry, or one of the physical sciences Validator AND either two (2) years of experience in generating analytical data in an environmental analytical laboratory **AND** two (2) years of data validation experience. 2. Complete Attachment 1, Data Validation Cover Sheet, and Attachment 2, High Explosive (HE) Analytical Data Validation Checklist, during data validation. 3. Refer to Attachment 3, Guidance for the Qualifier and Reason Code Application for additional guidance.

Title: Routine Validation of High Explosive (HE) Analytical	No.: SOP-5164	Page 3 of 12
Data	Revision: 0	

4.2 Records

Data Validator 1. Submit the following records generated by this procedure to the Records Processing Facility:

- Completed Data Validation Cover Sheets; and
- Completed High Explosive (HE) Analytical Data Validation Checklists.

5.0 PROCESS FLOW CHART

For specific validation criteria follow the NNSA Model for Data Validation.

6.0 ATTACHMENTS

Attachment 1 5164-1 Data Validation Cover Sheet (1 page)

Attachment 2 5164-2 High Explosive (HE) Analytical Data Validation Checklist (3 pages)

Attachment 3 5164-3 Guidance for the Qualifier and Reason Code Application (4 pages)

7.0 REVISION HISTORY

Author: Bill Hardesty

Revision No. [Enter current revision number, beginning with Rev.0]	Effective Date [DCC inserts effective date for revision]	Description of Changes [List specific changes made since the previous revision]	Type of Change [Technical (T) or Editorial (E)]
0		New Document	Т

Using a CRYPTOCard, click here to record "self-study" training to this procedure.

If you do not possess a CRYPTOCard or encounter problems, contact the EP training specialist.

Title: Routine Validation of High Explosive (HE) Analytical	No.: SOP-5164	Page 4 of 12
Data	Revision: 0	

ATTACHMENT 1: EXAMPLE OF A DATA VALIDATION COVER SHEET

5164-1

Example of a Data Validation Cover Sheet



	Section I.						
REQUEST NUMBER: VALIDATION DATE:						ι	LAB CODE:
CONT	RACT I	_ABOR/	ATORY NAME:				
VALID	ATOR:		ORGANIZATION	۷:			
ANAL	YTICAL	SUITE	(CHECK ALL THAT APPLY):				
ד 🗆	PH-GR	(O	☐ HIGH EXPLOSIVES		(IN FUF	RANS	☐ LCMSMS PERCHLORATES
ן □ ד	PH-DR	:O	☐ METALS	□ РСВ	CONG	ENERS	ORGANOCHLORINE
	☐ GENERAL CHEMISTRY ☐ RADIOCHEMISTRY ☐ LCMSMS HIGH				GH	PESTICIDES/POLYCHLORINATED BIPHENYLS	
(THER	(DESCI	RIBE):				
	-						
			Section II.	Complete	ness Cl	heck	
YES	NO	N/A	(CHECK ONE)	YES	NO	N/A	(CHECK ONE)
			1. CHAIN-OF-CUSTODY FORM(S)				6. RAW/BSS DATA
			2. CASE NARRATIVE				7. QUALITY CONTROL FORMS
			3. SAMPLE RESULT FORMS				8. QUANTITATION REPORTS
			4. SAMPLE CHROMATOGRAMS				9. TICS FORMS
			5. STANDARD CHROMATOGRAMS				10. TICS MASS SPECTRA
	Comments/problems noted (include information about requests for further information submitted to the contract laboratory and agreed-upon date of resolution and contract laboratory point of contact):						
VALID	VALIDATOR'S SIGNATURE: DATE:						
SOP-5	SOP-5164, Revision 0.0 LOS ALAMOS						
Environmental Restoration Project							

Title: Routine Validation of High Explosive (HE) Analytical No.: SOP-5164 Page 5 of 12 Revision: 0

ATTACHMENT 2: HIGH EXPLOSIVE (HE) ANALYTICAL DATA VALIDATION CHECKLIST

5164-2

High Explosive (HE) Analytical Data Validation Checklist



Yes	No	N/A			Assign Qualifier Listed Below If Criterion = Yes	
(Ch	eck O	ne)			Non-detected Detected Analyte Analyte	
			1.	The holding time was >1 and ≤2 times the applicable holding time requirement.	UJ, H9	J-, H9
			2.	The holding time was >2 times the applicable holding time requirement.	R, H9a	J-, H9a
			3.	The affected analytes are regarded as Rejected because the analytical holding time was exceeded.	R, H9b	R, H9b
			4.	The affected results were not analyzed with a valid 5-point calibration curve and/or a standard at the reporting limit.	UJ, R, H7	J, H7
			5.	The affected analytes were analyzed with an initial calibration curve that exceeded the %RSD criteria and/or the associated multi-point calibration correlation coefficient is <0.995.	UJ, H7a	J, H7a
			6.	The Initial Calibration Verification (ICV) and/or Continuing Calibration Verification (CCV) were recovered outside the method limits.	UJ, H7c	J, H7c
			7.	The ICV and/or CCV were not analyzed at the appropriate method frequency.	UJ, H7d	J, H7d
			8.	Required calibration information is missing or samples were analyzed on an expired calibration. Contact the SMO or external laboratory for information.	R, H7f	R, H7f
			9.	The sample result is ≤5 times the concentration of the related analyte in the method blank.	N/A	U, H4
			10.	The affected analytes are considered estimated and biased high because this analyte was identified in the method blank but was >5x.	N/A	J+, H4a

Title: Routine Validation of High Explosive (HE) Analytical No.: SOP-5164 Page 6 of 12

Revision: 0

Yes	No	N/A		Assign Qualifier Listed Below If Criterion = Yes	
(Ch	eck O	ne)		Non-detected Detected Analyte Analyte	
			11. The sample result is ≤5 times the concentration of the related analyte in the trip blank, rinsate blank, and/or equipment blank.	N/A	U, H4d
			12. Required method blank information is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.	R, H4e	R, H4e
			13. The analyte RT shifted by more than 0.05 minutes from the mid-level standard of the initial calibration.	R, H0	J, H0
			14. Analyte is positively confirmed but outside the IS retention time window; however, spectral matches must be provided (hexp – diode array detector).	N/A	J+, H0a
			15. Required retention time documentation is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.	R, H0b	R, H0b
			16. The surrogate is <10%R. Follow the external laboratory limits located within the associated data package.	R, H3	J-, H3
			17. The surrogate is < the Lower Acceptance Limit (LAL) but ≥10%R. Follow the external laboratory limits located within the associated data package.	UJ, H3a	J-, H3a
			18. The surrogate %R value is > the Upper Acceptance Limit (UAL). Follow the external laboratory limits located within the associated data package.	N/A	J+, H3b
			19. At least one surrogate is > the UAL and one surrogate is < the LAL. Follow the external laboratory limits located within the associated data package.	UJ, H3c	J, H3c
			20. Required surrogate information is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.	R, H3d	R, H3d
			21. The LCS percent recover was <10%. Follow the external laboratory limits located within the associated data package.	R, H12	R, H12

Title: Routine Validation of High Explosive (HE) Analytical No.: SOP-5164 Page 7 of 12

Revision: 0

Yes	No	N/A	A Assign Qualifier Listed Criterion = Ye		
(Ch	eck O	ne)		Non-detected Detected Analyte Analyte	
			22. The LCS percent recovery was < the LAL but >10%. Follow the external laboratory limits located within the associated data package.	UJ, H12a	J-, H12a
			23. The LCS percent recovery was > the UAL. Follow the external laboratory limits located within the associated data package.	N/A	J+, H12b
			24. The LCS documentation is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.	R, H12c	R, H12c
			25. The analyte was not confirmed on a second dissimilar column [R] or diode array spectrum does not match the library (U).	N/A	R, U, H8
			26. The second dissimilar column documentation is missing or diode array spectrums are missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.	R, H8a	R, H8a
			27. Duplicate, dilution, or reanalysis.	UJ, H88	J, H88
			28. The affected analytes have elevated detection limits and may not meet project DQOs because the sample was diluted without any target analytes identified due to matrix interference. Qualify as Reject if the analytical laboratory cannot provide proof for cleanup or matrix interference.	UJ, R, H15	R, H15
			29. Qualification of data via data validation did not occur based on Quality Control requirements in this procedure. Adhere to the external laboratory qualifiers found within the Form I analytical data summary sheets generated by the external laboratory.	U, U_LAB	J, J_LAB, NQ, NQ
			30. The LANL project chemist identified quality deficiencies in the reported data that requires further qualification. This code can only be used and/or under advisement by the LANL project chemist.	UJ, R, H19	J, R, H19

Title: Routine Validation of High Explosive (HE) Analytical Data	No.: SOP-5164	Page 8 of 12
	Revision: 0	

ATTACHMENT 3: GUIDELINES FOR THE QUALIFIER AND REASON CODE APPLICATION

5164-3

Guidelines for the Qualifier and Reason Code Application

Records Use only

Los Alamos NATIONAL LABORATORY EST.1943

No.	Valid Flag Code Nondetect	Valid Flag Code Detect	Valid Reason Code	Valid Reason Description
1	R	J	Н0	The analyte RT shifted by more than 0.05 minutes from the mid-level standard of the initial calibration. Reject nondetects for HPLC.
2	N/A	J	Н0а	Analyte is positively confirmed but outside the retention time window; however, spectral matches must be provided (hexp – diode array detector).
3	R	R	H0b	Required retention time documentation is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.
4	R	R	H12	The LCS percent recovery was <10%. Follow external laboratory limits located within the associated data package.
5	υJ	J-	H12a	The LCS percent recovery was < the Lower Acceptance Limit (LAL) but >10%. Follow external laboratory limits located within the associated data package.
6	N/A	J+	H12b	The LCS percent recovery was > than the Upper Acceptance Limit (UAL). Follow the external laboratory limits located within the associated data package.
7	R	R	H12c	The LCS documentation is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.

Title: Routine Validation of High Explosive (HE) Analytical Data	No.: SOP-5164	Page 9 of 12
	Revision: 0	

No.	Valid Flag Code Nondetect	Valid Flag Code Detect	Valid Reason Code	Valid Reason Description	
8	UJ, R	R	H15	The affected analytes have elevated detection limits and may not meet project DQOs because the sample was diluted without any target analytes identified due to matrix interference. Qualify as Reject if the analytical laboratory cannot provide proof for cleanup or matrix interference.	
9	UJ, R	J, R	H19	The LANL project chemist identified quality deficiencies in the reported data that requires further qualification. This code can ONLY be used and/or under advisement by the project chemist.	
10	R	J-	Н3	The surrogate is <10%R, which indicates the potential for a severely low bias in the results. Follow external laboratory limits located within the associated data package.	
11	UJ	J-	НЗа	The surrogate is < the LAL but ≥10%R, which indicates the potential for a low bias in the results. Follow the external laboratory limits located within the associated data package.	
12	N/A	J+	H3b	The surrogate %R value is > the UAL, which indicates a potential for a high bias in the results and a potential for false positive results. Follow the external laboratory limits located within the associated data package.	
13	ΠΊ	J	Н3с	At least one surrogate is > the UAL and one surrogate is < the LAL, which indicates a greater than normal degree of uncertainty in the result. Follow external laboratory limits located within the associated data package.	
14	R	R	H3d	Required surrogate information is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.	
15	N/A	U	H4	The sample result is ≤5X the concentration of the related analyte in the method blank, which indicates the reported detection is considered indistinguishable from contamination in the blank.	
16	N/A	J	H4a	The affected analytes are considered estimated and biased high because this analyte was identified in the method blank but was >5X.	

CONTROLLED DOCUMENT

Title: Routine Validation of High Explosive (HE) Analytical Data	No.: SOP-5164	Page 10 of 12
	Revision: 0	

No.	Valid Flag Code Nondetect	Valid Flag Code Detect	Valid Reason Code	Valid Reason Description	
17	H4d	U	H4d	The sample result is ≤5X the concentration of the related analyte in the trip blank, rinsate blank, or equipment blank, which indicates the reported detection is considered indistinguishable from contamination in the blank.	
18	R	R	H4e	Required method blank information is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information.	
19	UJ, R	J	H7	The affected results were not analyzed with a valid 5-point calibration curve and/or a standard at the reporting limit.	
20	UJ	J	Н7а	The affected analytes were analyzed with an initial calibration curve that exceeded the %RSD criteria and/or the associated multipoint calibration correlation coefficient is <0.995.	
21	UJ	J	H7c	The Initial Calibration Verification (ICV) and/or Continuing Calibration Verification (CCV) were recovered outside the method-specific limits.	
22	ΠΊ	J	H7d	The ICV and/or CCV were not analyzed at the appropriate method frequency.	
23	R	R	H7f	Required calibration information is missing or Samples were analyzed on an expired calibration. Contact the SMO or external laboratory for information.	
24	N/A	R, U	Н8	The analyte was not confirmed on a second dissimilar column or diode array spectrums do not match library.	
25	ΠΊ	J	H88	Duplicate, dilution, or reanalysis.	
26	R	R	Н8а	The required second dissimilar column or diode array documentation is missing. Data may not be acceptable for use. Contact the SMO or external laboratory for information	
27	UJ	J-	Н9	The extraction/analytical holding time was exceeded by <2X the published method for holding times.	
28	R	J-	Н9а	The extraction/analytical holding time was exceeded by >2X the published method for holding times.	

Title: Routine Validation of High Explosive (HE) Analytical Data	No.: SOP-5164	Page 11 of 12
	Revision: 0	

No.	Valid Flag Code Nondetect	Valid Flag Code Detect	Valid Reason Code	Valid Reason Description
29	R	R	H9b	The affected analytes are regarded as rejected because the analytical holding time was exceeded.
30	U	J, NQ	U_LAB, J_LAB, NQ	Qualification of data via data validation did not occur based on Quality Control requirements in this procedure. Adhere to the external laboratory qualifiers found with the Form 1 analytical data summary sheets generated by the external laboratory.

Title: Routine Validation of LC/MS/MS	No.: SOP-5164	Page 12 of 12
	Revision: 0	

This page left blank intentional.